


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081315\
 Data File : BM002379.D
 Acq On : 13 Aug 2015 02:13
 Operator : UM/IZ
 Sample : DFTPP96
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP96

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Aug 14 03:42:49 2015



AutoFind: Scans 2627, 2628, 2629; Background Corrected with Scan 2622

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.1	41240	PASS
68	69	0.00	2	2.0	823	PASS
69	198	0.00	100	34.7	41962	PASS
70	69	0.00	2	0.8	324	PASS
127	198	10	80	50.1	60488	PASS
197	198	0.00	2	0.6	767	PASS
198	198	100	100	100.0	120794	PASS
199	198	5	9	6.9	8309	PASS
275	198	10	60	26.9	32437	PASS
365	198	1	100	3.1	3793	PASS
441	443	0.01	100	77.5	16707	PASS
442	198	50	100	90.7	109592	PASS
443	442	15	24	19.7	21565	PASS

Scan 2622 (18.510 min): BM002379.D
DFTPP96

m/z	Abundance
40.00	575.0
41.10	471.0
43.00	499.0
44.00	3872.0
55.00	459.0
57.10	320.0
73.00	508.0
207.00	1376.0
208.00	349.0
281.00	418.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

m/z	Abundance
38.00	616.0
39.10	3944.0
40.00	771.0
41.00	600.0
43.00	474.0
44.00	3943.0
50.00	11042.0
51.10	42448.0
52.00	2366.0
55.00	731.0
56.00	1442.0
57.00	3158.0
61.00	540.0
62.00	612.0
63.00	1691.0
65.10	1077.0
68.10	899.0
69.00	43232.0
70.00	307.0
73.00	857.0
74.00	4193.0
75.00	6986.0
76.10	2270.0
77.10	50152.0
78.00	3354.0
79.00	3480.0
80.00	2622.0
81.00	3808.0
82.00	931.0
83.00	1218.0
85.00	774.0
86.00	1026.0
86.90	539.0
91.00	958.0
92.00	958.0
93.00	5271.0
94.00	339.0
95.00	398.0
96.10	378.0
98.00	4326.0
99.00	3218.0
101.00	2024.0
102.90	722.0
104.00	1243.0
105.00	1152.0
106.00	383.0
107.00	14821.0
108.00	2590.0
109.10	586.0
110.00	29120.0
111.00	4409.0
111.90	753.0
115.90	904.0
117.00	10748.0
117.90	748.0
122.00	999.0
123.00	1704.0
123.90	748.0
125.00	755.0
127.00	58448.0
128.00	4896.0
129.00	21224.0
130.00	1851.0
130.90	426.0
132.90	300.0
133.90	624.0
135.00	1988.0
135.90	681.0
137.00	961.0
141.00	2686.0
141.90	900.0
143.00	626.0
146.00	513.0
147.00	1698.0
148.00	3109.0
149.00	830.0
150.00	368.0
151.10	437.0
151.70	330.0
153.00	943.0
154.00	642.0
154.90	1493.0
156.00	2295.0
157.00	673.0
157.90	574.0
159.00	333.0
159.90	960.0
161.00	1191.0
162.00	398.0
165.00	1078.0
165.90	823.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

167.00	4803.0
168.00	2519.0
169.10	395.0
172.00	363.0
173.00	640.0
174.00	1338.0
175.00	2359.0
176.10	583.0
176.90	1068.0
178.00	415.0
179.00	3488.0
180.00	2824.0
181.00	1370.0
184.00	457.0
185.00	1806.0
186.00	15210.0
187.00	4030.0
187.90	441.0
188.90	918.0
191.00	656.0
192.00	1317.0
193.00	1272.0
194.10	406.0
196.00	3614.0
197.10	744.0
197.90	108040.0
198.90	7887.0
200.00	660.0
201.60	434.0
202.90	710.0
204.00	3895.0
205.00	6180.0
206.00	27112.0
207.00	5019.0
208.00	1129.0
208.90	410.0
210.20	557.0
211.00	1045.0
214.90	383.0
215.90	625.0
216.90	6507.0
218.00	1201.0
221.00	4764.0
221.80	1040.0
222.90	1325.0
224.00	14095.0
225.00	3817.0
226.00	459.0
227.00	5472.0
228.00	698.0
229.00	1376.0
231.00	537.0
234.00	326.0
234.90	488.0
235.90	312.0
237.00	460.0
242.00	657.0
243.00	715.0
244.00	11749.0
245.00	1432.0
246.00	2383.0
246.90	525.0
249.00	488.0
254.10	503.0
255.00	58800.0
256.00	8811.0
257.00	774.0
258.00	3027.0
259.00	625.0
265.00	1036.0
273.00	1731.0
274.00	4625.0
275.00	26040.0
276.00	3661.0
277.00	2093.0
277.90	346.0
281.00	423.0
283.00	317.0
284.90	460.0
293.00	457.0
296.00	6745.0
297.00	983.0
303.00	715.0
314.00	328.0
314.90	874.0
315.90	357.0
323.00	2208.0
324.00	430.0
326.90	440.0
334.00	1133.0
334.90	375.0
341.00	305.0
345.90	337.0
352.00	763.0
353.00	512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

354.00	802.0
365.00	2932.0
366.00	366.0
372.00	1109.0
373.00	350.0
401.90	414.0
403.00	611.0
421.00	567.0
421.90	513.0
423.00	4392.0
424.00	827.0
441.00	11778.0
442.00	74192.0
443.00	15128.0
444.00	1421.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

m/z	Abundance
38.00	643.0
39.10	4136.0
40.00	749.0
41.10	533.0
43.00	502.0
44.00	3854.0
49.10	390.0
50.00	12420.0
51.10	47984.0
52.10	2667.0
55.10	734.0
56.00	1515.0
57.00	3533.0
61.00	599.0
62.00	741.0
63.00	2019.0
65.00	1070.0
68.00	988.0
69.00	49176.0
70.00	337.0
73.10	876.0
74.00	4855.0
75.00	7774.0
76.10	2540.0
77.10	58352.0
78.10	4043.0
79.00	3951.0
80.00	2902.0
81.00	4400.0
82.00	1097.0
83.00	1144.0
85.00	753.0
86.00	1028.0
87.00	489.0
91.00	1120.0
92.00	901.0
93.00	6167.0
94.00	490.0
95.00	382.0
96.00	400.0
97.00	306.0
98.00	4915.0
99.00	3570.0
99.90	354.0
101.00	2417.0
102.90	962.0
104.00	1451.0
105.00	1596.0
106.10	376.0
107.00	18368.0
108.00	2952.0
109.10	747.0
110.00	35432.0
111.00	5246.0
112.00	654.0
115.90	922.0
117.00	13000.0
118.00	979.0
122.00	1269.0
123.00	2062.0
123.90	1098.0
124.90	737.0
127.00	71248.0
128.00	5522.0
129.00	26432.0
130.00	2317.0
131.00	504.0
134.00	670.0
135.00	2171.0
136.00	961.0
137.00	1146.0
139.90	310.0
141.00	3395.0
141.90	1263.0
143.00	839.0
145.00	321.0
145.90	583.0
147.00	2031.0
148.00	3562.0
149.00	980.0
151.10	515.0
151.90	380.0
153.00	1140.0
154.00	907.0
155.00	1964.0
156.00	2581.0
156.90	562.0
157.90	687.0
158.90	653.0
160.00	986.0
161.00	1632.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

162.00	421.0
164.90	1256.0
165.90	980.0
167.00	6254.0
167.90	2713.0
169.00	511.0
170.90	315.0
171.90	694.0
173.00	879.0
174.00	1376.0
175.00	2841.0
176.00	836.0
177.00	1359.0
178.00	456.0
179.00	4985.0
180.00	3689.0
181.00	1783.0
182.00	318.0
183.90	371.0
185.00	2524.0
186.00	19800.0
187.00	5650.0
187.90	593.0
189.00	1093.0
190.90	856.0
192.00	1663.0
193.00	1954.0
194.00	441.0
194.80	349.0
196.00	4461.0
197.10	820.0
197.90	142336.0
198.90	10206.0
200.00	825.0
201.40	691.0
202.90	1071.0
204.00	4939.0
205.00	8058.0
206.00	35696.0
207.00	6021.0
208.00	1402.0
209.00	512.0
210.20	609.0
211.00	1469.0
211.80	312.0
214.90	439.0
216.00	809.0
217.00	9020.0
218.00	1298.0
221.00	6044.0
221.80	1326.0
223.00	2099.0
224.00	19848.0
225.00	4888.0
226.00	619.0
227.00	8203.0
228.00	1003.0
229.00	1717.0
231.10	684.0
234.00	544.0
234.90	497.0
235.90	382.0
236.90	674.0
241.00	428.0
242.00	1108.0
243.10	1015.0
244.00	15304.0
245.00	2069.0
246.00	3097.0
247.00	547.0
249.00	597.0
252.90	341.0
254.00	754.0
255.00	80624.0
256.00	11961.0
257.00	1051.0
258.00	3815.0
259.00	652.0
264.90	1758.0
273.00	2560.0
274.00	6725.0
275.00	39096.0
276.00	5070.0
277.00	3340.0
278.10	484.0
281.00	553.0
283.00	398.0
284.00	301.0
285.00	502.0
293.00	649.0
296.00	8892.0
296.90	1274.0
303.00	1206.0
304.10	365.0
314.00	398.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

315.00	1127.0
316.00	596.0
321.00	471.0
323.00	3697.0
324.00	695.0
326.90	647.0
328.00	335.0
332.90	304.0
334.00	2230.0
334.90	641.0
341.00	517.0
345.90	611.0
352.00	1058.0
353.00	869.0
354.00	1046.0
365.00	4581.0
366.00	618.0
372.00	1818.0
373.00	391.0
383.00	535.0
389.90	329.0
402.00	726.0
403.00	1012.0
403.90	373.0
420.90	1096.0
422.00	746.0
423.00	7176.0
424.00	1343.0
441.00	19296.0
442.00	126744.0
443.00	25016.0
444.00	2008.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

m/z	Abundance
38.10	463.0
39.10	2684.0
39.90	696.0
41.00	613.0
43.00	426.0
44.00	3905.0
49.10	313.0
50.00	8477.0
51.00	33288.0
52.00	1690.0
55.00	706.0
56.00	1132.0
57.00	2442.0
60.90	337.0
62.00	495.0
63.00	1281.0
65.00	647.0
68.10	583.0
69.00	33480.0
70.00	330.0
73.00	655.0
74.00	3187.0
75.00	5453.0
76.00	1851.0
77.10	39464.0
78.00	2803.0
79.00	2631.0
80.00	2022.0
81.00	3212.0
82.00	789.0
83.00	839.0
85.00	593.0
86.00	806.0
86.90	439.0
91.10	907.0
92.00	699.0
93.00	4484.0
93.90	358.0
95.00	354.0
96.00	408.0
97.20	333.0
97.90	3383.0
99.00	2738.0
101.00	1887.0
102.90	655.0
104.00	1170.0
105.00	1189.0
106.10	396.0
107.00	13314.0
108.00	2287.0
109.10	612.0
110.00	25408.0
111.00	3696.0
112.00	443.0
116.00	709.0
117.00	9584.0
118.00	790.0
122.00	868.0
123.00	1597.0
124.00	665.0
125.00	642.0
127.00	51768.0
128.00	3873.0
129.00	18960.0
130.00	1553.0
130.90	365.0
134.00	648.0
135.00	1707.0
136.00	619.0
137.00	844.0
141.00	2506.0
141.90	952.0
143.00	502.0
146.00	433.0
147.00	1389.0
148.00	2878.0
149.00	747.0
151.00	331.0
153.00	852.0
154.00	689.0
154.90	1440.0
156.00	2046.0
157.00	459.0
157.90	546.0
159.00	376.0
160.00	789.0
161.00	1199.0
162.00	450.0
165.00	964.0
166.00	751.0
167.00	5120.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

168.00	2189.0
169.00	460.0
172.00	397.0
173.00	743.0
174.00	986.0
175.00	2090.0
176.10	676.0
177.10	938.0
178.00	352.0
178.90	3660.0
180.00	3009.0
180.90	1179.0
181.90	318.0
184.00	334.0
185.00	1782.0
186.00	14625.0
187.00	4465.0
188.00	481.0
189.00	1012.0
191.10	663.0
192.00	1226.0
193.00	1291.0
194.00	354.0
195.00	303.0
196.00	3324.0
197.10	739.0
197.90	112008.0
198.90	6836.0
199.90	690.0
201.40	517.0
203.00	889.0
204.00	3846.0
205.00	6365.0
206.00	28328.0
207.00	5128.0
208.00	1106.0
209.00	513.0
210.20	558.0
211.00	1213.0
216.00	652.0
217.00	7280.0
218.00	1063.0
221.00	5270.0
221.80	1061.0
223.00	1781.0
224.00	15871.0
225.00	4080.0
226.00	491.0
227.00	5984.0
227.90	852.0
228.90	1271.0
231.00	553.0
233.90	560.0
234.90	394.0
237.00	542.0
241.00	417.0
242.00	846.0
243.00	690.0
244.00	13035.0
245.00	1654.0
245.90	2491.0
247.00	446.0
248.90	470.0
252.90	377.0
254.10	610.0
255.00	66688.0
256.00	9639.0
257.00	738.0
258.00	3507.0
259.00	467.0
264.90	1374.0
273.00	2082.0
274.00	5403.0
275.00	32176.0
276.00	4219.0
276.90	2559.0
277.90	410.0
280.90	354.0
283.10	337.0
285.00	482.0
292.90	478.0
296.00	8381.0
296.90	1147.0
303.10	923.0
313.90	405.0
315.00	1044.0
316.00	544.0
323.00	3067.0
323.90	578.0
327.00	516.0
328.00	325.0
333.00	322.0
334.00	1954.0
335.00	483.0
341.00	443.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96

345.90	768.0
352.00	954.0
353.00	742.0
354.00	936.0
365.00	3868.0
366.00	628.0
371.00	319.0
372.00	1738.0
373.00	526.0
383.00	460.0
390.00	350.0
402.00	592.0
403.00	1345.0
404.00	412.0
421.00	858.0
422.00	893.0
423.00	6810.0
424.00	1281.0
441.00	19048.0
442.00	127840.0
443.00	24552.0
444.00	2047.0

Instrument :
BNA_M
ClientSampleId :
DFTPP96