


Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM081316\  
 Data File : BM007060.D  
 Acq On : 13 Aug 2016 09:32  
 Operator : UM/SJ  
 Sample : DFTPP44  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP44

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-EPA-BM071416.M  
 Title : SVOA CALIBRATION  
 Last Update : Wed Jul 20 04:22:37 2016



AutoFind: Scans 2556, 2557, 2558; Background Corrected with Scan 2551

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 32.5      | 84266   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.5       | 1435    | PASS             |
| 69          | 198          | 0.00         | 100          | 35.8      | 92920   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 463     | PASS             |
| 127         | 198          | 10           | 80           | 43.2      | 111965  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 1032    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 259392  | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 17552   | PASS             |
| 275         | 198          | 10           | 60           | 29.4      | 76370   | PASS             |
| 365         | 198          | 1            | 100          | 4.1       | 10729   | PASS             |
| 441         | 443          | 0.01         | 100          | 79.5      | 33320   | PASS             |
| 442         | 198          | 50           | 100          | 83.9      | 217578  | PASS             |
| 443         | 442          | 15           | 24           | 19.3      | 41933   | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 318.0     |
| 37.90  | 784.0     |
| 39.10  | 5188.0    |
| 40.00  | 1047.0    |
| 40.90  | 326.0     |
| 44.00  | 3026.0    |
| 49.00  | 994.0     |
| 50.10  | 20592.0   |
| 51.00  | 75432.0   |
| 52.10  | 3997.0    |
| 55.00  | 519.0     |
| 56.00  | 2125.0    |
| 57.00  | 5163.0    |
| 60.90  | 975.0     |
| 62.00  | 1193.0    |
| 63.00  | 3747.0    |
| 63.90  | 340.0     |
| 64.20  | 367.0     |
| 65.00  | 1283.0    |
| 68.00  | 1354.0    |
| 69.00  | 84192.0   |
| 70.00  | 516.0     |
| 73.00  | 883.0     |
| 74.00  | 9734.0    |
| 75.00  | 14228.0   |
| 76.10  | 4279.0    |
| 77.10  | 95864.0   |
| 78.10  | 6738.0    |
| 79.00  | 7016.0    |
| 80.00  | 5171.0    |
| 81.00  | 7156.0    |
| 82.00  | 1897.0    |
| 83.00  | 1493.0    |
| 83.80  | 481.0     |
| 85.00  | 1289.0    |
| 86.00  | 2062.0    |
| 87.00  | 931.0     |
| 88.00  | 499.0     |
| 88.90  | 330.0     |
| 91.00  | 1808.0    |
| 92.00  | 1942.0    |
| 93.00  | 12846.0   |
| 93.90  | 844.0     |
| 96.00  | 587.0     |
| 97.00  | 366.0     |
| 98.00  | 10097.0   |
| 99.00  | 8576.0    |
| 100.00 | 743.0     |
| 101.00 | 4083.0    |
| 103.00 | 1185.0    |
| 104.00 | 2321.0    |
| 105.00 | 2370.0    |
| 106.10 | 711.0     |
| 107.00 | 31688.0   |
| 108.00 | 4609.0    |
| 109.00 | 838.0     |
| 110.00 | 54120.0   |
| 111.00 | 9004.0    |
| 112.00 | 1234.0    |
| 113.00 | 392.0     |
| 116.00 | 1970.0    |
| 117.00 | 31168.0   |
| 118.00 | 2282.0    |
| 122.00 | 2433.0    |
| 123.00 | 3409.0    |
| 124.00 | 1339.0    |
| 125.00 | 1295.0    |
| 126.10 | 378.0     |
| 127.00 | 94416.0   |
| 128.00 | 6792.0    |
| 129.00 | 39032.0   |
| 130.00 | 3100.0    |
| 131.00 | 750.0     |
| 132.00 | 335.0     |
| 133.90 | 793.0     |
| 135.00 | 3319.0    |
| 135.90 | 1734.0    |
| 137.00 | 1945.0    |
| 138.10 | 560.0     |
| 139.90 | 833.0     |
| 141.00 | 5473.0    |
| 142.00 | 1497.0    |
| 143.00 | 1272.0    |
| 146.10 | 720.0     |
| 147.00 | 3103.0    |
| 148.00 | 6952.0    |
| 149.10 | 1372.0    |
| 150.90 | 621.0     |
| 153.00 | 1675.0    |
| 154.00 | 1347.0    |
| 155.00 | 2752.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

|        |          |
|--------|----------|
| 156.00 | 4346.0   |
| 157.00 | 757.0    |
| 158.00 | 967.0    |
| 159.00 | 675.0    |
| 160.00 | 1929.0   |
| 161.00 | 2828.0   |
| 162.00 | 722.0    |
| 165.00 | 2877.0   |
| 166.10 | 1802.0   |
| 167.00 | 11973.0  |
| 168.00 | 8376.0   |
| 169.00 | 1135.0   |
| 170.90 | 537.0    |
| 172.00 | 1305.0   |
| 172.90 | 1553.0   |
| 174.00 | 2258.0   |
| 175.00 | 4272.0   |
| 176.00 | 1476.0   |
| 176.90 | 1712.0   |
| 178.00 | 736.0    |
| 178.90 | 8950.0   |
| 180.00 | 6706.0   |
| 181.00 | 2446.0   |
| 182.00 | 665.0    |
| 183.90 | 869.0    |
| 185.00 | 3947.0   |
| 186.00 | 31368.0  |
| 187.00 | 8986.0   |
| 188.00 | 943.0    |
| 189.00 | 2436.0   |
| 190.00 | 359.0    |
| 190.90 | 1135.0   |
| 192.00 | 3229.0   |
| 193.00 | 3416.0   |
| 194.00 | 856.0    |
| 195.00 | 332.0    |
| 196.00 | 5539.0   |
| 197.00 | 866.0    |
| 198.00 | 209408.0 |
| 198.90 | 14088.0  |
| 200.00 | 1704.0   |
| 201.30 | 562.0    |
| 201.90 | 648.0    |
| 203.00 | 1887.0   |
| 204.00 | 9772.0   |
| 205.00 | 15301.0  |
| 206.00 | 58744.0  |
| 207.00 | 10501.0  |
| 208.00 | 2154.0   |
| 209.10 | 906.0    |
| 210.00 | 1332.0   |
| 211.00 | 3046.0   |
| 212.00 | 305.0    |
| 215.00 | 960.0    |
| 215.90 | 969.0    |
| 217.00 | 17280.0  |
| 217.90 | 2218.0   |
| 221.00 | 8066.0   |
| 221.90 | 2068.0   |
| 223.00 | 3127.0   |
| 224.00 | 35200.0  |
| 225.00 | 8909.0   |
| 226.00 | 1021.0   |
| 227.00 | 16464.0  |
| 228.00 | 2313.0   |
| 229.00 | 3123.0   |
| 229.90 | 375.0    |
| 231.00 | 1387.0   |
| 233.00 | 310.0    |
| 234.00 | 830.0    |
| 234.90 | 1434.0   |
| 235.90 | 760.0    |
| 237.00 | 1248.0   |
| 239.00 | 800.0    |
| 239.90 | 616.0    |
| 241.00 | 836.0    |
| 242.00 | 1828.0   |
| 243.00 | 1874.0   |
| 244.00 | 25064.0  |
| 245.00 | 3789.0   |
| 246.00 | 5467.0   |
| 247.10 | 1218.0   |
| 249.10 | 901.0    |
| 250.00 | 319.0    |
| 252.90 | 590.0    |
| 253.90 | 847.0    |
| 255.00 | 129280.0 |
| 256.00 | 19264.0  |
| 257.10 | 1606.0   |
| 258.00 | 8415.0   |
| 258.90 | 1472.0   |
| 265.00 | 3130.0   |
| 265.90 | 613.0    |
| 272.00 | 479.0    |
| 273.00 | 3693.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

|        |          |
|--------|----------|
| 274.00 | 10558.0  |
| 275.00 | 57744.0  |
| 276.00 | 8329.0   |
| 277.00 | 5170.0   |
| 278.00 | 883.0    |
| 281.10 | 468.0    |
| 282.00 | 449.0    |
| 283.10 | 471.0    |
| 284.00 | 553.0    |
| 284.90 | 647.0    |
| 291.90 | 451.0    |
| 293.00 | 1271.0   |
| 294.90 | 381.0    |
| 296.00 | 18360.0  |
| 296.90 | 2952.0   |
| 297.90 | 357.0    |
| 303.00 | 2106.0   |
| 303.90 | 702.0    |
| 308.00 | 364.0    |
| 313.90 | 691.0    |
| 315.00 | 2122.0   |
| 316.00 | 1156.0   |
| 321.10 | 848.0    |
| 322.10 | 306.0    |
| 323.00 | 4326.0   |
| 324.00 | 985.0    |
| 326.90 | 1098.0   |
| 328.10 | 517.0    |
| 331.90 | 518.0    |
| 333.10 | 546.0    |
| 334.00 | 4088.0   |
| 335.10 | 759.0    |
| 340.90 | 721.0    |
| 346.00 | 1158.0   |
| 352.00 | 1979.0   |
| 353.00 | 1234.0   |
| 354.10 | 1466.0   |
| 365.00 | 5965.0   |
| 366.00 | 1112.0   |
| 371.00 | 518.0    |
| 372.00 | 2179.0   |
| 373.00 | 631.0    |
| 383.00 | 642.0    |
| 390.10 | 413.0    |
| 391.00 | 351.0    |
| 402.00 | 1213.0   |
| 403.00 | 1319.0   |
| 404.00 | 599.0    |
| 421.10 | 863.0    |
| 422.00 | 878.0    |
| 423.00 | 7021.0   |
| 424.00 | 2030.0   |
| 441.00 | 20496.0  |
| 442.00 | 136128.0 |
| 443.00 | 27232.0  |
| 444.10 | 2967.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

| m/z    | Abundance |
|--------|-----------|
| 38.10  | 1280.0    |
| 39.00  | 6419.0    |
| 40.10  | 1009.0    |
| 41.20  | 371.0     |
| 44.00  | 2876.0    |
| 49.00  | 1479.0    |
| 50.10  | 26104.0   |
| 51.10  | 99024.0   |
| 52.00  | 5043.0    |
| 53.10  | 345.0     |
| 55.00  | 712.0     |
| 56.00  | 3090.0    |
| 57.00  | 7041.0    |
| 57.90  | 365.0     |
| 61.00  | 1596.0    |
| 62.00  | 1787.0    |
| 63.10  | 3931.0    |
| 64.10  | 618.0     |
| 65.10  | 2183.0    |
| 67.90  | 1791.0    |
| 69.00  | 109704.0  |
| 70.00  | 525.0     |
| 73.00  | 1431.0    |
| 74.00  | 12550.0   |
| 75.00  | 17616.0   |
| 76.00  | 5658.0    |
| 77.10  | 123576.0  |
| 78.10  | 8494.0    |
| 79.00  | 9638.0    |
| 80.00  | 6600.0    |
| 81.00  | 10201.0   |
| 82.00  | 2609.0    |
| 83.00  | 2426.0    |
| 84.10  | 490.0     |
| 85.00  | 1808.0    |
| 86.00  | 2811.0    |
| 87.00  | 1028.0    |
| 87.90  | 387.0     |
| 88.90  | 371.0     |
| 90.90  | 1993.0    |
| 92.00  | 2695.0    |
| 93.00  | 16243.0   |
| 94.10  | 1092.0    |
| 95.00  | 411.0     |
| 95.90  | 854.0     |
| 98.00  | 13228.0   |
| 99.00  | 9651.0    |
| 99.90  | 815.0     |
| 101.00 | 6113.0    |
| 102.10 | 314.0     |
| 103.00 | 1917.0    |
| 104.00 | 3263.0    |
| 105.00 | 3624.0    |
| 106.10 | 876.0     |
| 107.00 | 45264.0   |
| 108.00 | 5572.0    |
| 109.10 | 1119.0    |
| 110.00 | 70120.0   |
| 111.00 | 11621.0   |
| 111.90 | 1367.0    |
| 112.90 | 542.0     |
| 115.90 | 2761.0    |
| 117.00 | 43840.0   |
| 118.10 | 2507.0    |
| 118.90 | 562.0     |
| 120.00 | 573.0     |
| 122.00 | 2440.0    |
| 123.00 | 4305.0    |
| 124.00 | 1640.0    |
| 125.00 | 1516.0    |
| 126.10 | 583.0     |
| 127.00 | 129992.0  |
| 128.00 | 10364.0   |
| 129.00 | 55456.0   |
| 130.00 | 4779.0    |
| 131.00 | 806.0     |
| 133.10 | 469.0     |
| 133.90 | 1277.0    |
| 135.00 | 4372.0    |
| 136.00 | 1637.0    |
| 137.00 | 2832.0    |
| 138.10 | 527.0     |
| 138.80 | 323.0     |
| 140.00 | 1024.0    |
| 141.00 | 7373.0    |
| 142.10 | 1940.0    |
| 143.00 | 1764.0    |
| 143.90 | 494.0     |
| 144.90 | 449.0     |
| 145.90 | 1305.0    |
| 147.00 | 3972.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

|        |          |
|--------|----------|
| 148.00 | 10155.0  |
| 149.00 | 2535.0   |
| 150.00 | 552.0    |
| 151.10 | 925.0    |
| 151.90 | 523.0    |
| 153.00 | 2326.0   |
| 154.00 | 2159.0   |
| 155.00 | 4936.0   |
| 156.00 | 6131.0   |
| 157.00 | 968.0    |
| 157.90 | 1484.0   |
| 158.90 | 1088.0   |
| 160.00 | 2792.0   |
| 161.00 | 3891.0   |
| 162.00 | 1052.0   |
| 163.00 | 354.0    |
| 165.00 | 3430.0   |
| 166.00 | 2578.0   |
| 167.00 | 17280.0  |
| 168.00 | 9452.0   |
| 168.90 | 1835.0   |
| 169.90 | 354.0    |
| 171.10 | 837.0    |
| 172.00 | 1580.0   |
| 173.00 | 1771.0   |
| 174.00 | 3697.0   |
| 175.10 | 6623.0   |
| 176.00 | 1698.0   |
| 176.90 | 2346.0   |
| 178.00 | 936.0    |
| 179.00 | 11418.0  |
| 180.00 | 8415.0   |
| 181.00 | 3662.0   |
| 182.10 | 477.0    |
| 184.00 | 1027.0   |
| 185.00 | 5908.0   |
| 186.00 | 44880.0  |
| 187.00 | 13485.0  |
| 188.00 | 1565.0   |
| 189.00 | 2949.0   |
| 189.90 | 726.0    |
| 191.10 | 1530.0   |
| 192.00 | 3582.0   |
| 193.00 | 5153.0   |
| 194.00 | 1391.0   |
| 194.90 | 472.0    |
| 196.00 | 9112.0   |
| 196.90 | 1242.0   |
| 198.00 | 300032.0 |
| 199.00 | 19968.0  |
| 200.00 | 2016.0   |
| 201.60 | 1600.0   |
| 203.00 | 2530.0   |
| 204.00 | 12923.0  |
| 205.00 | 18928.0  |
| 206.00 | 88400.0  |
| 207.10 | 13575.0  |
| 208.00 | 3638.0   |
| 209.00 | 1253.0   |
| 210.30 | 1299.0   |
| 211.10 | 3469.0   |
| 211.90 | 654.0    |
| 214.90 | 1325.0   |
| 216.00 | 2046.0   |
| 217.00 | 24328.0  |
| 218.00 | 3517.0   |
| 218.90 | 359.0    |
| 221.00 | 11425.0  |
| 221.90 | 2940.0   |
| 223.00 | 5581.0   |
| 224.00 | 49336.0  |
| 225.00 | 13305.0  |
| 226.00 | 1379.0   |
| 227.00 | 22712.0  |
| 228.00 | 3264.0   |
| 229.00 | 3964.0   |
| 229.90 | 499.0    |
| 231.00 | 1848.0   |
| 232.90 | 384.0    |
| 233.90 | 1649.0   |
| 235.00 | 1902.0   |
| 236.00 | 1153.0   |
| 237.00 | 2071.0   |
| 237.80 | 361.0    |
| 239.00 | 1004.0   |
| 240.00 | 524.0    |
| 241.00 | 1413.0   |
| 242.00 | 2408.0   |
| 243.00 | 2902.0   |
| 244.00 | 37080.0  |
| 245.00 | 5130.0   |
| 246.00 | 8145.0   |
| 247.00 | 1990.0   |
| 249.10 | 1224.0   |
| 253.00 | 590.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

|        |          |
|--------|----------|
| 254.10 | 1345.0   |
| 255.00 | 204032.0 |
| 256.00 | 30832.0  |
| 257.00 | 2958.0   |
| 258.00 | 13472.0  |
| 258.90 | 1568.0   |
| 259.80 | 330.0    |
| 264.00 | 449.0    |
| 265.00 | 5490.0   |
| 266.00 | 873.0    |
| 270.90 | 453.0    |
| 272.00 | 581.0    |
| 273.00 | 5588.0   |
| 274.00 | 17128.0  |
| 275.00 | 86184.0  |
| 276.00 | 12068.0  |
| 277.00 | 9389.0   |
| 277.90 | 1593.0   |
| 281.00 | 317.0    |
| 283.00 | 833.0    |
| 284.00 | 808.0    |
| 285.10 | 1540.0   |
| 289.00 | 412.0    |
| 292.00 | 540.0    |
| 293.00 | 2178.0   |
| 295.00 | 328.0    |
| 296.00 | 27344.0  |
| 297.00 | 3713.0   |
| 298.10 | 314.0    |
| 301.00 | 395.0    |
| 301.90 | 567.0    |
| 303.10 | 3016.0   |
| 304.00 | 865.0    |
| 308.10 | 490.0    |
| 308.90 | 307.0    |
| 310.00 | 300.0    |
| 312.90 | 315.0    |
| 314.00 | 1318.0   |
| 315.00 | 3560.0   |
| 316.00 | 1694.0   |
| 316.90 | 474.0    |
| 321.00 | 1003.0   |
| 321.80 | 347.0    |
| 323.00 | 8228.0   |
| 324.00 | 1773.0   |
| 327.00 | 1726.0   |
| 327.90 | 911.0    |
| 328.80 | 416.0    |
| 332.00 | 535.0    |
| 332.90 | 932.0    |
| 334.00 | 5852.0   |
| 335.00 | 1353.0   |
| 341.00 | 895.0    |
| 346.00 | 2149.0   |
| 346.90 | 406.0    |
| 352.10 | 2106.0   |
| 353.10 | 1819.0   |
| 354.00 | 2287.0   |
| 355.00 | 660.0    |
| 365.00 | 12772.0  |
| 366.00 | 1616.0   |
| 370.90 | 485.0    |
| 372.00 | 3941.0   |
| 373.00 | 1221.0   |
| 382.90 | 986.0    |
| 390.00 | 647.0    |
| 391.00 | 519.0    |
| 392.00 | 540.0    |
| 402.00 | 1481.0   |
| 403.00 | 2623.0   |
| 404.00 | 1144.0   |
| 421.00 | 1988.0   |
| 422.00 | 1376.0   |
| 423.10 | 13138.0  |
| 424.00 | 2544.0   |
| 425.10 | 322.0    |
| 441.00 | 39104.0  |
| 442.00 | 245504.0 |
| 443.00 | 48976.0  |
| 444.00 | 4641.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

| m/z    | Abundance |
|--------|-----------|
| 38.10  | 800.0     |
| 39.10  | 5606.0    |
| 40.10  | 1132.0    |
| 41.00  | 490.0     |
| 44.00  | 2210.0    |
| 49.00  | 756.0     |
| 50.00  | 20104.0   |
| 51.10  | 78344.0   |
| 52.00  | 3918.0    |
| 55.00  | 541.0     |
| 56.00  | 2613.0    |
| 57.00  | 5037.0    |
| 61.10  | 956.0     |
| 62.00  | 1321.0    |
| 63.00  | 4215.0    |
| 64.10  | 524.0     |
| 65.10  | 1472.0    |
| 68.00  | 1161.0    |
| 69.00  | 84864.0   |
| 70.00  | 349.0     |
| 73.00  | 1090.0    |
| 74.00  | 9488.0    |
| 75.00  | 15012.0   |
| 76.00  | 5266.0    |
| 77.10  | 99216.0   |
| 78.00  | 7685.0    |
| 79.00  | 7648.0    |
| 80.00  | 5810.0    |
| 81.00  | 7426.0    |
| 82.00  | 1716.0    |
| 83.00  | 1713.0    |
| 83.90  | 459.0     |
| 85.00  | 1231.0    |
| 86.00  | 2777.0    |
| 86.90  | 1138.0    |
| 89.00  | 370.0     |
| 91.00  | 1962.0    |
| 92.00  | 2213.0    |
| 93.00  | 12999.0   |
| 94.00  | 880.0     |
| 95.10  | 377.0     |
| 96.00  | 666.0     |
| 97.10  | 357.0     |
| 98.00  | 11593.0   |
| 99.00  | 7897.0    |
| 99.90  | 840.0     |
| 101.00 | 3999.0    |
| 102.00 | 340.0     |
| 103.00 | 1239.0    |
| 104.00 | 3115.0    |
| 105.00 | 2863.0    |
| 106.00 | 731.0     |
| 107.00 | 35392.0   |
| 108.00 | 5551.0    |
| 108.90 | 788.0     |
| 110.00 | 61616.0   |
| 111.00 | 9376.0    |
| 111.90 | 999.0     |
| 113.00 | 511.0     |
| 115.90 | 1819.0    |
| 117.00 | 34776.0   |
| 118.00 | 2235.0    |
| 120.00 | 600.0     |
| 122.00 | 3104.0    |
| 122.90 | 4355.0    |
| 124.00 | 2024.0    |
| 125.00 | 1304.0    |
| 126.10 | 541.0     |
| 127.00 | 111488.0  |
| 128.00 | 8468.0    |
| 129.00 | 46872.0   |
| 130.00 | 3321.0    |
| 131.10 | 649.0     |
| 132.90 | 473.0     |
| 134.00 | 1661.0    |
| 135.00 | 4040.0    |
| 136.00 | 1521.0    |
| 137.10 | 2299.0    |
| 137.80 | 304.0     |
| 140.10 | 682.0     |
| 141.00 | 7025.0    |
| 141.90 | 2045.0    |
| 142.90 | 1241.0    |
| 144.10 | 303.0     |
| 146.00 | 1018.0    |
| 147.00 | 4018.0    |
| 148.00 | 8266.0    |
| 149.00 | 2244.0    |
| 150.00 | 626.0     |
| 151.00 | 703.0     |
| 151.90 | 715.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

|        |          |
|--------|----------|
| 153.00 | 2309.0   |
| 154.00 | 2008.0   |
| 155.00 | 4148.0   |
| 156.00 | 6066.0   |
| 157.10 | 957.0    |
| 157.90 | 1080.0   |
| 158.90 | 1199.0   |
| 160.00 | 1634.0   |
| 161.00 | 3642.0   |
| 162.00 | 694.0    |
| 163.00 | 405.0    |
| 164.00 | 331.0    |
| 165.00 | 2699.0   |
| 166.00 | 2101.0   |
| 167.00 | 14342.0  |
| 168.00 | 8053.0   |
| 169.00 | 959.0    |
| 170.20 | 458.0    |
| 170.90 | 536.0    |
| 172.00 | 1436.0   |
| 173.00 | 1809.0   |
| 174.00 | 3500.0   |
| 175.10 | 5797.0   |
| 175.90 | 1976.0   |
| 177.00 | 2595.0   |
| 177.90 | 823.0    |
| 179.00 | 10252.0  |
| 180.00 | 8176.0   |
| 181.00 | 4009.0   |
| 182.00 | 603.0    |
| 183.90 | 850.0    |
| 185.00 | 5236.0   |
| 186.00 | 41312.0  |
| 187.10 | 12089.0  |
| 188.00 | 1266.0   |
| 189.00 | 2744.0   |
| 190.00 | 693.0    |
| 191.00 | 1570.0   |
| 192.00 | 3325.0   |
| 193.00 | 4767.0   |
| 194.00 | 1092.0   |
| 195.00 | 446.0    |
| 196.00 | 6699.0   |
| 197.10 | 989.0    |
| 198.00 | 268736.0 |
| 199.00 | 18600.0  |
| 200.00 | 1440.0   |
| 201.40 | 1052.0   |
| 202.00 | 341.0    |
| 202.90 | 2378.0   |
| 204.00 | 12147.0  |
| 205.00 | 18840.0  |
| 206.00 | 77728.0  |
| 207.00 | 13980.0  |
| 208.00 | 3774.0   |
| 209.00 | 1157.0   |
| 210.00 | 1111.0   |
| 211.00 | 3483.0   |
| 215.00 | 1463.0   |
| 216.00 | 1940.0   |
| 217.00 | 23352.0  |
| 218.00 | 3529.0   |
| 220.20 | 323.0    |
| 221.00 | 9722.0   |
| 221.90 | 2686.0   |
| 223.00 | 5136.0   |
| 224.00 | 43176.0  |
| 225.00 | 11258.0  |
| 226.10 | 1337.0   |
| 227.00 | 22320.0  |
| 228.00 | 2939.0   |
| 228.90 | 4266.0   |
| 231.00 | 1940.0   |
| 232.00 | 317.0    |
| 233.00 | 403.0    |
| 234.00 | 1510.0   |
| 235.00 | 1337.0   |
| 236.00 | 954.0    |
| 237.00 | 1491.0   |
| 239.00 | 1170.0   |
| 240.00 | 960.0    |
| 241.00 | 972.0    |
| 242.00 | 2879.0   |
| 243.00 | 2236.0   |
| 244.00 | 35720.0  |
| 245.10 | 4388.0   |
| 246.00 | 9086.0   |
| 247.00 | 1710.0   |
| 249.00 | 1256.0   |
| 249.90 | 512.0    |
| 251.10 | 469.0    |
| 252.00 | 391.0    |
| 253.00 | 758.0    |
| 254.10 | 861.0    |
| 255.00 | 194688.0 |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

|        |          |
|--------|----------|
| 256.00 | 27984.0  |
| 257.00 | 2401.0   |
| 258.00 | 11933.0  |
| 259.00 | 1835.0   |
| 261.00 | 346.0    |
| 264.00 | 351.0    |
| 265.00 | 6688.0   |
| 266.00 | 583.0    |
| 271.00 | 330.0    |
| 271.90 | 381.0    |
| 273.00 | 5693.0   |
| 274.00 | 15916.0  |
| 275.00 | 85184.0  |
| 276.00 | 11747.0  |
| 277.00 | 7762.0   |
| 278.00 | 1512.0   |
| 281.00 | 545.0    |
| 281.90 | 319.0    |
| 283.10 | 1250.0   |
| 284.10 | 775.0    |
| 285.00 | 2009.0   |
| 292.10 | 486.0    |
| 292.90 | 1557.0   |
| 293.90 | 467.0    |
| 295.00 | 729.0    |
| 296.00 | 26840.0  |
| 297.00 | 4347.0   |
| 301.00 | 497.0    |
| 301.90 | 353.0    |
| 303.00 | 4052.0   |
| 303.90 | 969.0    |
| 308.90 | 366.0    |
| 309.90 | 396.0    |
| 312.90 | 356.0    |
| 313.90 | 1320.0   |
| 315.00 | 2892.0   |
| 316.00 | 1653.0   |
| 321.00 | 744.0    |
| 323.00 | 8045.0   |
| 324.00 | 1525.0   |
| 326.90 | 1740.0   |
| 328.00 | 985.0    |
| 332.00 | 563.0    |
| 332.90 | 893.0    |
| 334.00 | 5911.0   |
| 335.00 | 1715.0   |
| 336.00 | 424.0    |
| 341.10 | 1012.0   |
| 342.00 | 401.0    |
| 345.90 | 2072.0   |
| 346.90 | 396.0    |
| 350.90 | 312.0    |
| 352.00 | 2147.0   |
| 353.10 | 1745.0   |
| 354.00 | 1894.0   |
| 355.00 | 644.0    |
| 365.00 | 13452.0  |
| 365.90 | 2082.0   |
| 371.10 | 547.0    |
| 372.10 | 4278.0   |
| 373.00 | 1155.0   |
| 382.90 | 1575.0   |
| 383.90 | 383.0    |
| 389.80 | 397.0    |
| 392.00 | 375.0    |
| 401.90 | 1304.0   |
| 403.00 | 2822.0   |
| 404.00 | 1247.0   |
| 421.00 | 2065.0   |
| 422.00 | 2104.0   |
| 423.00 | 13490.0  |
| 424.00 | 3062.0   |
| 425.00 | 568.0    |
| 441.00 | 40360.0  |
| 442.00 | 271104.0 |
| 443.00 | 49592.0  |
| 444.10 | 4839.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44

Scan 2551 (17.592 min): BM007060.D  
DFTPP44

| m/z    | Abundance |
|--------|-----------|
| 39.90  | 766.0     |
| 40.90  | 456.0     |
| 44.00  | 2883.0    |
| 48.90  | 454.0     |
| 73.00  | 574.0     |
| 83.90  | 472.0     |
| 96.00  | 481.0     |
| 191.00 | 397.0     |
| 207.00 | 2261.0    |
| 208.00 | 344.0     |
| 281.10 | 581.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP44