


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM081320\
Data File : BM027054.D
Acq On : 13 Aug 2020 08:20
Operator : JU/CG
Sample : DFTPP25
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP25

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM080720MA.M
Title : SVOA CALIBRATION
Last Update : Fri Aug 14 12:48:04 2020



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.5	54632	PASS
68	69	0.00	2	2.0	1873	PASS
69	198	0.00	100	47.9	91709	PASS
70	69	0.00	2	0.5	485	PASS
127	198	10	80	43.7	83674	PASS
197	198	0.00	2	0.6	1096	PASS
198	198	100	100	100.0	191381	PASS
199	198	5	9	6.8	13022	PASS
275	198	10	60	30.8	59032	PASS
365	198	1	100	5.2	10047	PASS
441	198	0.01	100	13.1	24981	PASS
442	198	50	100	89.4	171114	PASS
443	442	15	24	18.8	32232	PASS

Scan 2459 (17.451 min): BM027054.D
DFTPP25

m/z	Abundance
39.90	317.0
44.10	498.0
73.00	284.0
91.20	164.0
207.00	1295.0
208.00	183.0
209.20	161.0
281.00	282.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

m/z	Abundance
37.00	410.0
38.00	1262.0
39.10	5367.0
39.90	547.0
43.90	941.0
49.00	725.0
50.10	17432.0
51.00	64384.0
52.10	3670.0
55.00	377.0
56.00	2189.0
57.00	5238.0
58.00	154.0
61.10	993.0
62.10	1444.0
63.00	3477.0
64.00	734.0
65.10	1520.0
66.90	243.0
68.00	2332.0
69.00	104360.0
70.00	630.0
73.00	1341.0
74.00	12051.0
75.00	18080.0
76.00	6291.0
77.10	108872.0
78.00	7607.0
79.00	8664.0
80.00	6811.0
81.00	7988.0
82.00	2271.0
83.00	1666.0
83.90	341.0
85.00	1606.0
86.00	2702.0
86.90	1089.0
88.00	343.0
89.10	250.0
91.00	1683.0
92.10	2329.0
93.00	15332.0
94.00	894.0
95.00	167.0
96.00	633.0
98.00	11665.0
99.00	8898.0
100.00	704.0
101.00	3890.0
102.00	415.0
103.00	1499.0
104.00	2627.0
105.00	2885.0
106.10	812.0
107.00	31128.0
108.00	4622.0
109.00	699.0
110.00	51264.0
111.00	8617.0
112.00	1077.0
112.80	294.0
116.00	2298.0
117.00	34752.0
118.00	2091.0
118.80	229.0
119.90	371.0
122.00	2317.0
123.00	3474.0
124.00	1567.0
125.00	1280.0
127.00	92080.0
128.00	8214.0
129.00	40200.0
130.00	3604.0
131.00	904.0
132.20	219.0
133.00	238.0
133.90	1037.0
135.00	3350.0
135.90	1307.0
137.00	2383.0
137.80	246.0
138.10	259.0
139.00	214.0
139.90	665.0
141.00	5450.0
141.90	1768.0
143.00	1556.0
144.00	427.0
145.00	267.0
146.00	1129.0
147.00	2877.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

148.00	8360.0
149.00	1843.0
149.90	404.0
150.90	527.0
151.90	501.0
153.00	1867.0
154.00	1100.0
155.00	3392.0
156.00	4457.0
157.00	741.0
157.90	1030.0
159.00	796.0
160.00	1976.0
161.00	2569.0
162.00	573.0
162.90	232.0
164.00	301.0
165.00	2602.0
166.10	1776.0
167.00	12233.0
168.00	8309.0
169.00	958.0
169.90	429.0
170.90	572.0
172.00	1357.0
173.00	1344.0
174.00	2699.0
175.00	5302.0
176.00	1112.0
176.90	2421.0
178.00	821.0
179.00	8619.0
180.00	6301.0
181.00	3053.0
181.90	414.0
182.90	305.0
183.90	753.0
185.10	4365.0
186.00	32184.0
187.00	9015.0
187.90	945.0
189.00	2305.0
189.90	337.0
191.00	1120.0
192.00	2825.0
193.10	2776.0
193.90	684.0
194.90	392.0
196.00	5970.0
197.10	1130.0
198.00	195200.0
199.00	14020.0
200.10	1083.0
200.80	357.0
201.40	897.0
202.10	166.0
203.00	1894.0
204.00	8888.0
205.00	14633.0
206.00	57080.0
207.00	8649.0
208.00	2530.0
209.00	1023.0
210.10	853.0
211.00	2590.0
211.80	566.0
214.90	656.0
216.00	1445.0
217.00	16680.0
218.00	2054.0
219.00	233.0
221.00	7195.0
221.90	2392.0
222.90	3876.0
224.00	32504.0
225.00	8559.0
226.00	1108.0
227.00	15938.0
228.00	2098.0
229.00	2867.0
229.80	439.0
231.00	1093.0
232.90	292.0
234.00	968.0
235.00	1043.0
236.00	818.0
237.00	930.0
238.00	294.0
239.00	454.0
240.00	732.0
240.90	848.0
242.00	2186.0
243.10	1789.0
244.00	24232.0
245.00	3012.0
246.00	6263.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

247.00	1415.0
247.80	210.0
248.90	775.0
251.00	220.0
252.00	333.0
253.00	453.0
254.10	1112.0
255.00	133312.0
256.00	19776.0
257.00	1588.0
258.00	8804.0
259.00	1423.0
260.00	209.0
260.90	213.0
264.00	188.0
265.00	4073.0
266.00	586.0
267.00	165.0
270.00	213.0
270.90	453.0
272.00	335.0
273.00	3710.0
274.00	9424.0
275.00	53464.0
276.00	7219.0
277.00	5617.0
278.00	1351.0
278.80	207.0
281.00	451.0
281.90	242.0
282.90	758.0
284.10	320.0
284.90	653.0
285.90	206.0
289.00	188.0
289.90	209.0
290.90	162.0
291.90	323.0
292.90	948.0
294.00	292.0
295.00	429.0
296.00	19848.0
297.00	2709.0
298.00	224.0
300.90	308.0
301.90	217.0
303.10	2083.0
304.00	617.0
308.00	260.0
308.90	167.0
309.90	187.0
314.00	723.0
315.00	2397.0
316.10	975.0
321.00	740.0
322.00	327.0
323.00	5583.0
324.00	960.0
324.90	160.0
326.00	217.0
327.00	1471.0
328.00	471.0
332.00	282.0
333.00	298.0
334.00	3997.0
335.00	724.0
335.90	157.0
341.10	689.0
341.90	205.0
345.90	1231.0
346.90	270.0
352.00	1725.0
353.00	1037.0
354.00	1443.0
355.00	497.0
359.10	167.0
365.00	8179.0
365.90	1124.0
369.90	237.0
370.90	447.0
372.00	2353.0
373.00	696.0
383.10	531.0
383.90	192.0
390.10	508.0
391.00	249.0
392.20	267.0
401.00	164.0
402.00	966.0
403.00	1132.0
404.00	698.0
404.90	261.0
421.00	1192.0
422.10	1142.0
423.00	7668.0
424.00	2359.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

425.00	568.0
441.10	18736.0
442.10	133120.0
443.10	25792.0
444.10	2400.0
445.10	175.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

m/z	Abundance
37.00	407.0
38.00	1057.0
39.00	5390.0
40.00	712.0
40.90	216.0
43.90	604.0
49.00	445.0
50.00	17760.0
51.10	63992.0
52.00	3414.0
55.00	382.0
56.00	2300.0
57.00	4941.0
58.00	370.0
61.00	1232.0
62.00	1249.0
63.00	3604.0
64.10	718.0
65.00	1886.0
66.90	172.0
68.10	2115.0
69.00	108776.0
70.00	527.0
73.10	1372.0
74.00	12138.0
75.00	17688.0
76.10	4779.0
77.10	107200.0
78.00	7744.0
79.00	8888.0
80.00	7000.0
81.00	8902.0
82.00	2053.0
83.00	1781.0
83.90	192.0
85.00	1407.0
86.00	2970.0
87.00	1231.0
88.10	382.0
88.90	151.0
91.00	1636.0
91.90	2252.0
93.00	16137.0
94.00	917.0
95.10	185.0
96.00	496.0
97.10	455.0
98.00	11297.0
99.00	9399.0
99.90	830.0
100.90	3745.0
102.00	305.0
103.00	1340.0
104.00	2872.0
105.00	3005.0
106.10	964.0
107.00	33896.0
108.00	5355.0
109.00	882.0
110.00	53536.0
111.00	9071.0
111.90	1276.0
113.00	420.0
116.00	1585.0
117.00	34368.0
118.00	2700.0
119.00	295.0
120.00	379.0
121.00	162.0
122.00	2278.0
123.00	3630.0
124.00	1868.0
125.00	1357.0
126.20	573.0
127.00	99272.0
128.00	7057.0
129.00	42768.0
130.00	3725.0
130.90	791.0
132.00	426.0
133.00	301.0
134.00	1609.0
135.00	3596.0
136.00	1710.0
137.10	2250.0
138.10	380.0
139.10	184.0
140.10	674.0
141.00	5971.0
142.00	1857.0
142.90	1494.0
144.10	343.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

144.90	328.0
146.00	1143.0
147.00	3344.0
148.00	9596.0
148.90	1647.0
150.00	425.0
151.00	683.0
152.10	680.0
152.90	1843.0
153.90	1424.0
155.00	3469.0
156.10	4822.0
157.00	1103.0
157.90	1557.0
159.00	791.0
160.00	1907.0
161.00	2679.0
162.00	843.0
162.90	282.0
163.90	310.0
165.00	2691.0
166.00	2013.0
167.00	14666.0
168.00	7528.0
169.00	1401.0
170.10	394.0
171.00	590.0
171.90	1488.0
173.00	1924.0
174.00	2975.0
175.00	5649.0
176.00	1446.0
177.00	2552.0
177.90	802.0
179.00	11038.0
180.00	7734.0
181.00	3541.0
182.00	547.0
182.80	256.0
183.90	748.0
185.00	4769.0
186.00	36840.0
187.10	10369.0
188.00	1172.0
189.00	2495.0
189.90	567.0
191.00	1348.0
192.00	4097.0
193.00	3847.0
194.10	805.0
195.00	238.0
196.00	6377.0
196.90	1337.0
198.00	231488.0
198.90	15328.0
199.90	1135.0
201.00	353.0
201.60	706.0
203.00	2327.0
204.00	10055.0
205.00	16768.0
206.00	69432.0
207.10	10432.0
208.00	2924.0
208.90	1170.0
210.10	955.0
211.00	2839.0
213.00	173.0
214.90	898.0
216.00	1962.0
217.00	19344.0
217.90	3149.0
218.90	158.0
221.00	9368.0
222.00	2348.0
223.00	4685.0
224.00	40360.0
225.00	9995.0
226.00	1133.0
227.00	19688.0
228.00	2804.0
229.00	3636.0
229.90	527.0
231.00	1774.0
232.00	206.0
232.90	331.0
234.00	1186.0
234.90	1150.0
236.00	844.0
237.00	1606.0
238.20	231.0
239.00	965.0
240.10	420.0
241.00	1045.0
242.00	1681.0
243.10	2283.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

244.10	29552.0
245.00	3957.0
246.00	7465.0
247.00	1405.0
248.00	317.0
248.90	1008.0
250.00	151.0
250.90	231.0
252.00	465.0
253.10	688.0
254.00	907.0
255.00	166080.0
256.00	23264.0
257.10	1872.0
258.00	10790.0
259.00	1621.0
260.00	398.0
261.00	269.0
263.90	299.0
265.00	5692.0
266.00	582.0
271.00	532.0
272.10	738.0
273.00	4911.0
274.00	12127.0
275.00	73048.0
276.00	10243.0
277.00	7486.0
278.00	1420.0
278.90	263.0
280.90	453.0
282.00	241.0
282.90	1002.0
284.00	642.0
285.00	1099.0
285.90	205.0
289.00	383.0
289.90	263.0
290.90	189.0
292.00	415.0
293.10	1365.0
293.90	476.0
295.10	682.0
296.00	27840.0
297.00	3258.0
297.90	204.0
301.00	258.0
302.10	327.0
303.00	2900.0
304.00	708.0
307.90	252.0
308.80	171.0
309.90	260.0
313.10	177.0
314.00	1264.0
315.00	3246.0
316.00	1281.0
321.00	914.0
322.10	435.0
323.10	6330.0
324.00	1405.0
325.90	169.0
326.90	1708.0
328.00	717.0
329.00	179.0
332.00	601.0
333.10	716.0
334.00	5011.0
335.00	1250.0
335.90	173.0
341.00	935.0
341.90	448.0
346.00	1722.0
346.90	371.0
351.10	206.0
352.10	2671.0
353.00	1405.0
354.00	2176.0
355.10	698.0
359.00	245.0
365.00	12982.0
366.10	2080.0
367.00	250.0
370.90	372.0
372.10	3437.0
373.10	1006.0
382.90	906.0
384.00	336.0
389.90	460.0
390.90	185.0
391.20	217.0
391.90	174.0
401.00	214.0
402.00	1520.0
403.00	2315.0
404.00	757.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

405.00	170.0
421.00	1531.0
422.00	1614.0
423.00	11788.0
424.10	2854.0
425.00	543.0
441.10	30992.0
442.10	209280.0
443.10	40296.0
444.00	3686.0
445.00	202.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

m/z	Abundance
36.90	206.0
38.10	655.0
39.10	2759.0
40.00	408.0
40.90	156.0
43.90	649.0
49.10	439.0
50.10	10020.0
51.10	35520.0
52.00	1695.0
55.20	385.0
56.00	1306.0
57.00	2760.0
58.00	158.0
61.00	708.0
62.00	689.0
63.00	2161.0
64.00	441.0
65.00	1002.0
68.00	1174.0
69.00	61992.0
70.00	298.0
73.10	1077.0
74.00	7622.0
75.00	10752.0
76.10	3322.0
77.10	60872.0
78.00	4489.0
79.00	5073.0
80.00	3527.0
81.00	5231.0
82.00	1257.0
83.00	884.0
84.00	223.0
85.00	813.0
86.00	1540.0
86.90	718.0
87.90	204.0
91.00	1070.0
92.00	1152.0
93.00	9516.0
94.10	680.0
95.10	160.0
95.90	334.0
98.00	6884.0
99.00	5052.0
99.90	408.0
101.00	2536.0
102.90	779.0
104.00	1791.0
105.00	1647.0
106.00	595.0
107.00	20416.0
108.00	2945.0
109.10	703.0
110.00	31928.0
111.00	5028.0
111.90	343.0
113.00	344.0
116.00	1211.0
117.00	21264.0
118.00	1506.0
118.80	190.0
119.90	189.0
121.90	1545.0
123.00	2243.0
123.90	1056.0
124.90	796.0
127.00	59672.0
128.00	4718.0
129.00	25400.0
130.00	2401.0
131.00	422.0
131.90	221.0
133.10	366.0
133.80	773.0
134.90	1810.0
136.00	891.0
136.90	1352.0
137.90	340.0
139.80	416.0
141.00	4323.0
142.10	1019.0
143.00	926.0
143.90	367.0
146.00	1112.0
146.90	1952.0
148.00	5562.0
148.90	937.0
150.00	251.0
151.00	648.0
152.00	255.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

152.90	1360.0
154.00	878.0
154.90	2012.0
156.00	2853.0
157.00	607.0
158.00	962.0
159.00	646.0
159.90	1184.0
161.00	2183.0
162.00	538.0
163.00	161.0
165.00	1777.0
166.00	1323.0
167.00	9996.0
168.00	4301.0
169.00	707.0
170.00	257.0
170.90	410.0
172.00	919.0
173.00	917.0
173.90	2200.0
175.00	3447.0
176.10	746.0
177.00	1719.0
178.00	529.0
179.00	6980.0
180.00	5015.0
181.00	2134.0
182.00	541.0
183.90	616.0
185.00	2878.0
186.00	23392.0
187.00	7419.0
188.10	701.0
189.00	1940.0
189.80	258.0
191.00	758.0
192.00	2255.0
193.00	2113.0
193.90	446.0
195.00	178.0
196.00	4284.0
197.10	821.0
198.00	147456.0
199.00	9719.0
199.90	665.0
201.60	600.0
203.00	1273.0
204.00	6570.0
205.00	10876.0
206.00	45264.0
207.00	7382.0
208.00	1647.0
209.00	891.0
209.90	764.0
211.10	1854.0
212.20	252.0
215.00	536.0
216.00	1333.0
217.00	12861.0
218.00	1398.0
219.20	231.0
220.30	236.0
221.00	5986.0
221.80	1397.0
223.00	2587.0
224.10	27376.0
225.00	7650.0
226.00	611.0
227.00	13490.0
228.00	1284.0
229.00	2724.0
230.10	352.0
231.00	1204.0
231.90	175.0
232.90	308.0
233.90	849.0
235.00	870.0
236.00	841.0
237.00	1036.0
238.00	193.0
239.00	434.0
240.00	326.0
241.00	487.0
242.10	1431.0
243.10	1530.0
244.00	20848.0
245.00	2958.0
246.00	5618.0
247.00	1045.0
248.10	191.0
248.80	845.0
249.80	157.0
251.00	289.0
251.90	235.0
253.00	487.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25

254.10	533.0
255.00	115544.0
256.00	16048.0
257.00	1180.0
258.00	7645.0
259.10	1437.0
260.00	250.0
261.00	181.0
263.80	227.0
265.00	3731.0
266.10	432.0
271.10	299.0
272.00	339.0
273.00	3983.0
274.00	8391.0
275.00	50584.0
276.00	6937.0
277.00	5425.0
278.00	775.0
279.10	205.0
281.10	343.0
282.00	293.0
283.10	665.0
284.00	364.0
285.00	911.0
288.90	238.0
290.00	165.0
293.00	1271.0
294.00	394.0
295.00	462.0
296.00	18392.0
297.00	2592.0
297.80	191.0
300.90	208.0
301.90	442.0
303.00	2359.0
304.00	482.0
307.90	332.0
309.00	168.0
309.90	226.0
313.00	249.0
314.00	1164.0
315.00	2284.0
315.90	1339.0
317.00	201.0
321.10	646.0
322.00	265.0
323.10	5199.0
324.00	1027.0
327.00	1483.0
328.10	422.0
329.00	152.0
331.90	405.0
333.00	593.0
334.00	3486.0
334.90	852.0
340.90	950.0
346.00	1276.0
347.10	277.0
350.90	153.0
352.00	1666.0
353.00	1198.0
354.00	1437.0
355.00	319.0
359.10	249.0
365.00	8982.0
366.10	1180.0
370.10	221.0
371.00	285.0
372.00	2764.0
373.10	637.0
383.10	819.0
384.10	218.0
389.90	225.0
391.00	233.0
402.00	1134.0
403.10	1472.0
404.00	833.0
421.10	1150.0
422.00	1791.0
423.10	9219.0
424.10	1933.0
425.10	185.0
440.10	213.0
441.10	25216.0
442.10	170944.0
443.10	30608.0
444.10	2868.0
445.00	244.0

Instrument :
BNA_M
ClientSampleId :
DFTPP25