

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081415\
 Data File : BM002410.D
 Acq On : 15 Aug 2015 06:07
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02038

Quant Time: Aug 17 02:06:53 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 17 02:04:41 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
3	Naphthalene	1.021	1.003	1.8	118	0.00
4	2-Methylnaphthalene	0.765	0.729	4.7	115	0.00
5	1-Methylnaphthalene	0.756	0.721	4.6	114	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
7 S	Acenaphthylene-d8	2.045	2.032	0.6	110	0.00
8	Acenaphthylene	2.140	2.124	0.7	110	0.00
9 C	Acenaphthene	1.410	1.403	0.5	111	0.00
10 S	Fluorene-d10	1.436	1.396	2.8	108	0.00
11	Fluorene	1.615	1.581	2.1	109	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
13	Phenanthrene	1.155	1.143	1.0	107	0.00
14 S	Anthracene-d10	0.965	0.953	1.2	106	0.00
15	Anthracene	1.170	1.163	0.6	106	0.00
16 C	Fluoranthene	1.308	1.248	4.6	104	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
18 S	Pyrene-d10	0.887	0.950	-7.1	104	0.00
19	Pyrene	1.172	1.259	-7.4	103	0.00
20	Benzo(a)anthracene	1.196	1.173	1.9	103	-0.01
21	Chrysene	1.119	1.103	1.4	102	-0.01
22 I	Perylene-d12	1.000	1.000	0.0	92	-0.02
23	Benzo(b)fluoranthene	1.209	1.282	-6.0	101	-0.01
24	Benzo(k)fluoranthene	1.146	1.167	-1.8	94	-0.01
25 S	Benzo(a)pyrene-d12	1.051	1.048	0.3	94	-0.01
26 C	Benzo(a)pyrene	1.155	1.156	-0.1	93	-0.02
27	Indeno(1,2,3-cd)pyrene	1.328	1.119	15.7	79	-0.02
28	Dibenzo(a,h)anthracene	1.110	0.961	13.4	81	-0.02
29	Benzo(g,h,i)perylene	1.106	0.895	19.1	76	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0