

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081415\
 Data File : BM002400.D
 Acq On : 15 Aug 2015 00:05
 Operator : UM/IZ
 Sample : SSTDICCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02037

Quant Time: Aug 17 01:56:26 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 17 01:53:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
3	Naphthalene	1.021	1.006	1.5	92	0.00
4	2-Methylnaphthalene	0.765	0.756	1.2	93	0.00
5	1-Methylnaphthalene	0.756	0.747	1.2	92	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
7 S	Acenaphthylene-d8	2.045	2.031	0.7	92	0.00
8	Acenaphthylene	2.140	2.116	1.1	92	0.00
9 C	Acenaphthene	1.410	1.395	1.1	93	0.00
10 S	Fluorene-d10	1.436	1.421	1.0	92	0.00
11	Fluorene	1.615	1.614	0.1	93	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
13	Phenanthrene	1.155	1.141	1.2	93	0.00
14 S	Anthracene-d10	0.965	0.955	1.0	93	0.00
15	Anthracene	1.170	1.157	1.1	92	0.00
16 C	Fluoranthene	1.308	1.267	3.1	92	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
18 S	Pyrene-d10	0.887	0.926	-4.4	93	0.00
19	Pyrene	1.172	1.229	-4.9	92	0.00
20	Benzo(a)anthracene	1.196	1.178	1.5	94	0.00
21	Chrysene	1.119	1.108	1.0	94	0.00
22 I	Perylene-d12	1.000	1.000	0.0	96	-0.01
23	Benzo(b)fluoranthene	1.209	1.174	2.9	96	0.00
24	Benzo(k)fluoranthene	1.146	1.141	0.4	95	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.035	1.5	96	-0.01
26 C	Benzo(a)pyrene	1.155	1.140	1.3	95	-0.01
27	Indeno(1,2,3-cd)pyrene	1.328	1.341	-1.0	98	-0.02
28	Dibenzo(a,h)anthracene	1.110	1.128	-1.6	98	-0.02
29	Benzo(g,h,i)perylene	1.106	1.118	-1.1	98	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0