

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081415\
 Data File : BM002427.D
 Acq On : 15 Aug 2015 16:35
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02039

Quant Time: Aug 17 02:58:02 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 17 02:54:04 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	143	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	129	0.00
3	Naphthalene	1.021	1.002	1.9	130	0.00
4	2-Methylnaphthalene	0.765	0.717	6.3	124	0.00
5	1-Methylnaphthalene	0.756	0.704	6.9	122	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
7 S	Acenaphthylene-d8	2.045	2.050	-0.2	117	0.00
8	Acenaphthylene	2.140	2.121	0.9	116	0.00
9 C	Acenaphthene	1.410	1.394	1.1	116	0.00
10 S	Fluorene-d10	1.436	1.395	2.9	114	0.00
11	Fluorene	1.615	1.557	3.6	113	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
13	Phenanthrene	1.155	1.132	2.0	108	0.00
14 S	Anthracene-d10	0.965	0.955	1.0	109	0.00
15	Anthracene	1.170	1.149	1.8	108	0.00
16 C	Fluoranthene	1.308	1.149	12.2	98	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	79	0.00
18 S	Pyrene-d10	0.887	1.135	-28.0#	95	0.00
19	Pyrene	1.172	1.497	-27.7#	94	0.00
20	Benzo(a)anthracene	1.196	1.179	1.4	79	0.00
21	Chrysene	1.119	1.095	2.1	78	0.00
22 I	Perylene-d12	1.000	1.000	0.0	63	0.00
23	Benzo(b)fluoranthene	1.209	1.248	-3.2	68	0.00
24	Benzo(k)fluoranthene	1.146	1.158	-1.0	64	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.043	0.8	64	0.00
26 C	Benzo(a)pyrene	1.155	1.139	1.4	63	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.291	2.8	63	0.00
28	Dibenzo(a,h)anthracene	1.110	1.092	1.6	63	0.00
29	Benzo(g,h,i)perylene	1.106	1.087	1.7	63	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0