

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081415\
 Data File : BM002443.D
 Acq On : 16 Aug 2015 02:18
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02040

Quant Time: Aug 17 07:26:42 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 17 05:12:59 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	149	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
3	Naphthalene	1.021	1.004	1.7	134	0.00
4	2-Methylnaphthalene	0.765	0.708	7.5	126	0.00
5	1-Methylnaphthalene	0.756	0.698	7.7	125	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
7 S	Acenaphthylene-d8	2.045	2.047	-0.1	118	0.00
8	Acenaphthylene	2.140	2.139	0.0	118	0.00
9 C	Acenaphthene	1.410	1.392	1.3	117	0.00
10 S	Fluorene-d10	1.436	1.366	4.9	112	0.00
11	Fluorene	1.615	1.544	4.4	113	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
13	Phenanthrene	1.155	1.133	1.9	104	0.00
14 S	Anthracene-d10	0.965	0.952	1.3	104	0.01
15	Anthracene	1.170	1.157	1.1	104	0.00
16 C	Fluoranthene	1.308	1.055	19.3	86	0.01
17 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
18 S	Pyrene-d10	0.887	1.150	-29.7#	82	0.00
19	Pyrene	1.172	1.513	-29.1#	81	0.00
20	Benzo(a)anthracene	1.196	1.183	1.1	68	0.00
21	Chrysene	1.119	1.104	1.3	67	0.00
22 I	Perylene-d12	1.000	1.000	0.0	62	0.01
23	Benzo(b)fluoranthene	1.209	1.186	1.9	62	0.02
24	Benzo(k)fluoranthene	1.146	1.142	0.3	61	0.02
25 S	Benzo(a)pyrene-d12	1.051	1.040	1.0	62	0.02
26 C	Benzo(a)pyrene	1.155	1.143	1.0	61	0.02
27	Indeno(1,2,3-cd)pyrene	1.328	1.344	-1.2	63	0.03
28	Dibenzo(a,h)anthracene	1.110	1.131	-1.9	63	0.02
29	Benzo(g,h,i)perylene	1.106	1.114	-0.7	63	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0