

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081422\
 Data File : BMCCC276.D
 Acq On : 14 Aug 2022 19:08
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 14 23:40:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 14 23:34:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.810	152	203664	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	887037	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	530439	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	1082614	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	1075858	20.000	ng/u1	0.00
88) Perylene-d12	23.709	264	1092057	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	46406	8.574	ng/uL	0.00
4) Pyridine-d5	3.640	84	295504	19.929	ng/u1	0.00
7) Phenol-d5	6.993	99	346270	20.274	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.145	67	232491	20.420	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	307152	21.666	ng/u1	0.00
15) 4-Methylphenol-d8	8.534	113	281480	20.344	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	148239	21.090	ng/u1	0.00
24) 2-Nitrophenol-d4	9.698	143	148161	21.635	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.239	165	266078	21.483	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	429812	20.499	ng/u1	0.00
46) Dimethylphthalate-d6	13.869	166	774066	21.423	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	962018	20.868	ng/u1	0.00
54) 4-Nitrophenol-d4	14.680	143	131653	17.744	ng/u1	0.00
60) Fluorene-d10	15.439	176	686217	21.017	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.568	200	112133	19.269	ng/u1	0.00
73) Anthracene-d10	17.292	188	1037042	21.188	ng/u1	0.00
81) Pyrene-d10	19.586	212	1195535	20.050	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.556	264	1175718	21.206	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.252	88	48496	8.646	ng/uL	93
5) Pyridine	3.657	79	316821	20.361	ng/u1	91
6) Benzaldehyde	6.951	77	116191	16.209	ng/u1	93
8) Phenol	7.022	94	379320	20.295	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.240	93	310241	20.647	ng/u1	96
12) 2-Chlorophenol	7.375	128	312810	21.510	ng/u1	99
13) 2-Methylphenol	8.269	108	281672	20.173	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.345	45	412611	20.803	ng/u1	98
16) Acetophenone	8.639	105	458660	20.647	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.622	70	227147	20.504	ng/u1	95
18) 4-Methylphenol	8.598	108	310944	20.759	ng/u1	97
19) Hexachloroethane	8.881	117	130879	21.812	ng/u1	91
22) Nitrobenzene	9.022	77	342382	21.286	ng/u1	99
23) Isophorone	9.545	82	610391	20.103	ng/u1	97
25) 2-Nitrophenol	9.728	139	165669	21.806	ng/u1	99
26) 2,4-Dimethylphenol	9.798	107	331248	21.337	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.034	93	401623	20.973	ng/u1	99
29) 2,4-Dichlorophenol	10.269	162	277023	21.549	ng/u1	97
30) Naphthalene	10.657	128	1010116	21.574	ng/u1	99
32) 4-Chloroaniline	10.781	127	427875	20.543	ng/u1	98
33) Hexachlorobutadiene	10.933	225	174087	21.936	ng/u1	99
34) Caprolactam	11.569	113	79105	18.619	ng/u1	97
35) 4-Chloro-3-methylphenol	11.922	107	283830	21.233	ng/u1	97
36) 2-Methylnaphthalene	12.275	142	648813	21.181	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081422\
 Data File : BMCCC276.D
 Acq On : 14 Aug 2022 19:08
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 11 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :


Quant Time: Aug 14 23:40:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 14 23:34:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	655596	21.233	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.639	216	312817	20.717	ng/ul	99
40) Hexachlorocyclopentadiene	12.610	237	170313	16.936	ng/ul	97
41) 2,4,6-Trichlorophenol	12.892	196	201944	20.899	ng/ul	100
42) 2,4,5-Trichlorophenol	12.969	196	213867	20.743	ng/ul	99
43) 1,1'-Biphenyl	13.286	154	859789	21.059	ng/ul	98
44) 2-Chloronaphthalene	13.327	162	657314	20.978	ng/ul	99
45) 2-Nitroaniline	13.545	65	174629	20.216	ng/ul	98
47) Dimethylphthalate	13.916	163	779298	21.442	ng/ul	99
48) 2,6-Dinitrotoluene	14.039	165	152049	21.072	ng/ul	92
50) Acenaphthylene	14.174	152	1080323	21.184	ng/ul	100
51) 3-Nitroaniline	14.374	138	136112	17.068	ng/ul	96
52) Acenaphthene	14.516	153	696399	21.049	ng/ul	98
53) 2,4-Dinitrophenol	14.586	184	68909	16.820	ng/ul	98
55) 4-Nitrophenol	14.692	109	103598	18.256	ng/ul	84
56) Dibenzofuran	14.851	168	985677	21.532	ng/ul	99
57) 2,4-Dinitrotoluene	14.827	165	219576	21.588	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.080	232	170674	21.069	ng/ul	95
59) Diethylphthalate	15.274	149	789276	21.470	ng/ul	99
61) Fluorene	15.498	166	787165	21.050	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.492	204	373335	20.837	ng/ul	98
63) 4-Nitroaniline	15.533	138	108119	14.727	ng/ul#	89
66) 4,6-Dinitro-2-methylph...	15.586	198	111902	19.217	ng/ul	94
67) N-Nitrosodiphenylamine	15.710	169	656375	20.623	ng/ul	99
68) 4-Bromophenyl-phenylether	16.386	248	217364	20.480	ng/ul	98
69) Hexachlorobenzene	16.492	284	262442	20.294	ng/ul	96
70) Atrazine	16.668	200	221372	20.053	ng/ul	97
71) Pentachlorophenol	16.851	266	129662	16.413	ng/ul	96
72) Phenanthrene	17.233	178	1224026	21.095	ng/ul	98
74) Anthracene	17.327	178	1248857	21.449	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.251	216	325918	19.359	ng/uL	97
76) Pentachlorobenzene	14.763	250	314233	20.591	ng/uL	99
77) Carbazole	17.604	167	1149356	21.373	ng/ul	99
78) Di-n-butylphthalate	18.162	149	1319221	21.743	ng/ul	99
80) Fluoranthene	19.251	202	1423457	19.981	ng/ul	98
82) Pyrene	19.615	202	1476386	20.106	ng/ul	96
83) Butylbenzylphthalate	20.515	149	585722	21.932	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.297	252	457647	23.052	ng/ul	99
85) Benzo(a)anthracene	21.356	228	1431117	21.134	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.292	149	846947	21.948	ng/ul	99
87) Chrysene	21.409	228	1393965	21.098	ng/ul	99
89) Di-n-octyl phthalate	22.192	149	1431557	20.299	ng/ul	100
90) Benzo(b)fluoranthene	22.997	252	1499439	20.574	ng/ul	98
91) Benzo(k)fluoranthene	23.044	252	1379135	19.941	ng/ul	98
93) Benzo(a)pyrene	23.603	252	1458198	21.006	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.121	276	1679243	22.694	ng/ul	94
95) Dibenzo(a,h)anthracene	26.138	278	1443962	22.746	ng/ul	96
96) Benzo(g,h,i)perylene	26.868	276	1425717	23.251	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Aug 14 23:40:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Aug 14 23:34:05 2022
Response via : Initial Calibration








#2
 1,4-Dioxane
 Concen: 8.646 ng/uL
 RT: 3.252 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BMCCC276.D
 Acq: 14 Aug 2022 19:08

Instrument :
 BNA_M
 ClientSampleId :







Tgt Ion: 88 Resp: 48496
 Ion Ratio Lower Upper
 88 100
 43 30.0 27.8 41.8
 58 71.5 61.8 92.8



#5
 Pyridine
 Concen: 20.361 ng/uL
 RT: 3.657 min Scan# 97
 Delta R.T. -0.000 min
 Lab File: BMCCC276.D
 Acq: 14 Aug 2022 19:08

Tgt Ion: 79 Resp: 316821
 Ion Ratio Lower Upper
 79 100
 52 75.1 67.8 101.6
 51 36.1 31.4 47.0









#6
Benzaldehyde
Concen: 16.209 ng/ul
RT: 6.951 min Scan# 657
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 116191
Ion Ratio Lower Upper
77 100
105 97.0 84.6 127.0
106 94.1 80.3 120.5



#8
Phenol
Concen: 20.295 ng/ul
RT: 7.022 min Scan# 669
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion: 94 Resp: 379320
Ion Ratio Lower Upper
94 100
65 27.5 21.3 31.9
66 34.9 27.4 41.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.647 ng/ul
 RT: 7.240 min Scan# 706
 Delta R.T. -0.000 min
 Lab File: BMCCC276.D
 Acq: 14 Aug 2022 19:08

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion: 93 Resp: 310241




Ion	Ratio	Lower	Upper
93	100		
63	75.2	64.0	96.0
95	32.4	25.8	38.6




#12
 2-Chlorophenol
 Concen: 21.510 ng/ul
 RT: 7.375 min Scan# 729
 Delta R.T. -0.000 min
 Lab File: BMCCC276.D
 Acq: 14 Aug 2022 19:08

Tgt Ion: 128 Resp: 312810

Ion	Ratio	Lower	Upper
128	100		
64	45.4	36.9	55.3
130	32.9	25.6	38.4









#13
2-Methylphenol
Concen: 20.173 ng/ul
RT: 8.269 min Scan# 881
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 281672
Ion Ratio Lower Upper
108 100
107 89.4 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.803 ng/ul
RT: 8.345 min Scan# 894
Delta R.T. 0.012 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion: 45 Resp: 412611
Ion Ratio Lower Upper
45 100
77 14.1 10.6 15.8
79 11.2 8.3 12.5









#16
Acetophenone
Concen: 20.647 ng/ul
RT: 8.639 min Scan# 944
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 458660
Ion Ratio Lower Upper
105 100
77 76.8 59.8 89.6
51 26.6 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 20.504 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion: 70 Resp: 227147
Ion Ratio Lower Upper
70 100
42 57.5 49.6 74.4
101 9.8 8.1 12.1
130 21.1 18.4 27.6









#18
4-Methylphenol
Concen: 20.759 ng/ul
RT: 8.598 min Scan# 91
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 310944
Ion Ratio Lower Upper
108 100
107 109.2 90.3 135.5



#19
Hexachloroethane
Concen: 21.812 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:117 Resp: 130879
Ion Ratio Lower Upper
117 100
201 92.9 81.8 122.6
199 57.2 51.6 77.4









#22
Nitrobenzene
Concen: 21.286 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 342382
Ion Ratio Lower Upper
77 100
123 46.0 37.4 56.0
65 12.7 10.8 16.2



#23
Isophorone
Concen: 20.103 ng/ul
RT: 9.545 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion: 82 Resp: 610391
Ion Ratio Lower Upper
82 100
95 7.0 5.5 8.3
138 17.4 15.3 22.9









#25
2-Nitrophenol
Concen: 21.806 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 165669
Ion Ratio Lower Upper
139 100
65 48.9 39.8 59.6
109 28.3 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 21.337 ng/ul
RT: 9.798 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:107 Resp: 331248
Ion Ratio Lower Upper
107 100
121 51.0 42.7 64.1
122 86.9 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 20.973 ng/ul

RT: 10.034 min Scan# 1181

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08

Instrument :

BNA_M

ClientSampleId :





Tgt Ion: 93 Resp: 401623

Ion Ratio Lower Upper

93 100

95 31.5 26.1 39.1

123 11.6 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 21.549 ng/ul

RT: 10.269 min Scan# 1221

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08




Tgt Ion: 162 Resp: 277023


Ion Ratio Lower Upper

162 100

164 66.1 51.8 77.6

98 34.6 29.5 44.3









#30
Naphthalene
Concen: 21.574 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 1010116
Ion Ratio Lower Upper
128 100
129 11.3 8.8 13.2
127 12.9 10.6 16.0



#32
4-Chloroaniline
Concen: 20.543 ng/ul
RT: 10.781 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:127 Resp: 427875
Ion Ratio Lower Upper
127 100
129 32.5 25.3 37.9









#33
Hexachlorobutadiene
Concen: 21.936 ng/ul
RT: 10.933 min Scan# 11
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:225 Resp: 174087
Ion Ratio Lower Upper
225 100
223 61.9 50.0 75.0
227 63.1 49.9 74.9



#34
Caprolactam
Concen: 18.619 ng/ul
RT: 11.569 min Scan# 1442
Delta R.T. -0.006 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:113 Resp: 79105
Ion Ratio Lower Upper
113 100
55 179.5 150.2 225.4
56 127.4 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 21.233 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:107 Resp: 283830
Ion Ratio Lower Upper
107 100
144 26.8 20.3 30.5
142 83.2 64.4 96.6



#36
2-Methylnaphthalene
Concen: 21.181 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:142 Resp: 648813
Ion Ratio Lower Upper
142 100
141 88.9 70.7 106.1









#37
1-Methylnaphthalene
Concen: 21.233 ng/ul
RT: 12.492 min Scan# 1599
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:142 Resp: 655596
Ion Ratio Lower Upper
142 100
141 91.9 75.2 112.8
116 3.7 2.6 3.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.717 ng/ul
RT: 12.639 min Scan# 1624
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:216 Resp: 312817
Ion Ratio Lower Upper
216 100
214 79.1 62.4 93.6
179 21.0 17.3 25.9
108 18.5 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 16.936 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:237 Resp: 170313
Ion Ratio Lower Upper
237 100
235 62.6 48.4 72.6
272 12.3 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 20.899 ng/ul
RT: 12.892 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:196 Resp: 201944
Ion Ratio Lower Upper
196 100
198 97.0 77.8 116.6
200 30.5 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 20.743 ng/ul
RT: 12.969 min Scan# 1
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:196 Resp: 213867
Ion Ratio Lower Upper
196 100
198 94.3 76.8 115.2
200 30.9 24.9 37.3



#43
1,1'-Biphenyl
Concen: 21.059 ng/ul
RT: 13.286 min Scan# 1734
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:154 Resp: 859789
Ion Ratio Lower Upper
154 100
153 39.3 32.4 48.6
76 13.2 10.1 15.1









#44
2-Chloronaphthalene
Concen: 20.978 ng/ul
RT: 13.327 min Scan# 1741
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:162 Resp: 657314
Ion Ratio Lower Upper
162 100
164 32.2 25.4 38.0
127 39.4 31.6 47.4



#45
2-Nitroaniline
Concen: 20.216 ng/ul
RT: 13.545 min Scan# 1778
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion: 65 Resp: 174629
Ion Ratio Lower Upper
65 100
92 69.0 54.5 81.7
138 110.4 91.2 136.8









#47
Dimethylphthalate
Concen: 21.442 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:163 Resp: 779298
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 21.072 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:165 Resp: 152049
Ion Ratio Lower Upper
165 100
89 60.5 42.5 63.7
121 23.2 17.2 25.8









#50
Acenaphthylene
Concen: 21.184 ng/ul
RT: 14.174 min Scan# 1885
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:152 Resp: 1080323
Ion Ratio Lower Upper
152 100
151 20.1 16.1 24.1
153 13.0 10.6 15.8



#51
3-Nitroaniline
Concen: 17.068 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:138 Resp: 136112
Ion Ratio Lower Upper
138 100
108 10.0 8.1 12.1
92 108.4 90.4 135.6









#52
Acenaphthene
Concen: 21.049 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:153 Resp: 696399
Ion Ratio Lower Upper
153 100
152 47.5 37.1 55.7
154 87.1 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 16.820 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:184 Resp: 68909
Ion Ratio Lower Upper
184 100
63 66.2 54.2 81.4
154 60.3 49.3 73.9









#55
4-Nitrophenol
Concen: 18.256 ng/ul
RT: 14.692 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 103598
Ion Ratio Lower Upper
109 100
139 136.6 124.9 187.3
65 131.0 122.2 183.4



#56
Dibenzofuran
Concen: 21.532 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:168 Resp: 985677
Ion Ratio Lower Upper
168 100
139 37.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 21.588 ng/ul
RT: 14.827 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:165 Resp: 219576
Ion Ratio Lower Upper
165 100
63 45.1 35.9 53.9
182 2.5 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.069 ng/ul
RT: 15.080 min Scan# 2039
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:232 Resp: 170674
Ion Ratio Lower Upper
232 100
131 50.8 36.7 55.1
130 2.4 1.7 2.5
166 31.4 24.4 36.6









#59
Diethylphthalate
Concen: 21.470 ng/ul
RT: 15.274 min Scan# 2072
Delta R.T. -0.006 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 789276
Ion Ratio Lower Upper
149 100
177 20.7 16.8 25.2
150 12.2 10.2 15.4



#61
Fluorene
Concen: 21.050 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:166 Resp: 787165
Ion Ratio Lower Upper
166 100
165 97.5 77.2 115.8
167 13.4 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 20.837 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:204 Resp: 373335
Ion Ratio Lower Upper
204 100
206 32.5 25.9 38.9
141 61.1 50.6 75.8



#63
4-Nitroaniline
Concen: 14.727 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:138 Resp: 108119
Ion Ratio Lower Upper
138 100
92 50.7 39.2 58.8
108 78.8 51.6 77.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 19.217 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:198 Resp: 111902
Ion Ratio Lower Upper
198 100
182 5.8 4.2 6.4
77 28.5 19.9 29.9



#67
N-Nitrosodiphenylamine
Concen: 20.623 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:169 Resp: 656375
Ion Ratio Lower Upper
169 100
168 67.9 53.8 80.6
167 37.4 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 20.480 ng/ul
RT: 16.386 min Scan# 217364
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:248 Resp: 217364
Ion Ratio Lower Upper
248 100
250 97.3 77.3 115.9
141 78.1 59.8 89.8



#69
Hexachlorobenzene
Concen: 20.294 ng/ul
RT: 16.492 min Scan# 2279
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:284 Resp: 262442
Ion Ratio Lower Upper
284 100
142 39.9 28.6 42.8
249 30.2 24.6 37.0





#70

Atrazine

Concen: 20.053 ng/ul

RT: 16.668 min Scan# 21

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:200 Resp: 221372

Ion Ratio Lower Upper

200 100

173 27.0 21.3 31.9

215 46.3 39.0 58.6



#71

Pentachlorophenol

Concen: 16.413 ng/ul

RT: 16.851 min Scan# 2340

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08




Tgt Ion:266 Resp: 129662


Ion Ratio Lower Upper

266 100

264 66.4 49.5 74.3


268 65.6 51.0 76.4








#72
Phenanthrene
Concen: 21.095 ng/u1
RT: 17.233 min Scan# 2405
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:178 Resp: 1224026
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 19.5 14.9 22.3



#74
Anthracene
Concen: 21.449 ng/u1
RT: 17.327 min Scan# 2421
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:178 Resp: 1248857
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 18.9 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.359 ng/uL

RT: 13.251 min Scan# 1

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:216 Resp: 325918

Ion Ratio Lower Upper

216 100

214 78.7 60.9 91.3

218 49.1 37.5 56.3



#76

Pentachlorobenzene

Concen: 20.591 ng/uL

RT: 14.763 min Scan# 1985

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08




Tgt Ion:250 Resp: 314233


Ion Ratio Lower Upper

250 100

248 63.7 50.4 75.6

252 64.4 51.1 76.7









#77
 Carbazole
 Concen: 21.373 ng/ul
 RT: 17.604 min Scan# 2468
 Delta R.T. -0.000 min
 Lab File: BMCCC276.D
 Acq: 14 Aug 2022 19:08




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:167 Resp: 1149356
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.4 26.0
 139 13.1 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 21.743 ng/ul
 RT: 18.162 min Scan# 2563
 Delta R.T. -0.006 min
 Lab File: BMCCC276.D
 Acq: 14 Aug 2022 19:08

Tgt Ion:149 Resp: 1319221
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 5.0 3.7 5.5









#80
Fluoranthene
Concen: 19.981 ng/ul
RT: 19.251 min Scan# 21
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:202 Resp: 1423457
Ion Ratio Lower Upper
202 100
101 11.9 10.1 15.1
100 9.1 7.8 11.6



#82
Pyrene
Concen: 20.106 ng/ul
RT: 19.615 min Scan# 2810
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:202 Resp: 1476386
Ion Ratio Lower Upper
202 100
101 13.4 11.9 17.9
100 11.0 10.2 15.4





#83

Butylbenzylphthalate

Concen: 21.932 ng/ul

RT: 20.515 min Scan# 2963

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:149 Resp: 585722

Ion Ratio Lower Upper

149 100

91 71.6 53.2 79.8

206 18.2 16.0 24.0



#84

3,3'-Dichlorobenzidine

Concen: 23.052 ng/ul

RT: 21.297 min Scan# 3096

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08




Tgt Ion:252 Resp: 457647


Ion Ratio Lower Upper

252 100

254 65.9 52.4 78.6

126 12.4 10.8 16.2





#85

Benzo(a)anthracene

Concen: 21.134 ng/ul

RT: 21.356 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:228 Resp: 1431117

Ion Ratio Lower Upper

228 100

229 19.7 15.6 23.4

226 27.0 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.948 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08




Tgt Ion:149 Resp: 846947


Ion Ratio Lower Upper

149 100

167 26.1 21.5 32.3

279 4.1 3.4 5.2









#87
Chrysene
Concen: 21.098 ng/ul
RT: 21.409 min Scan# 3115
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 1393965
Ion Ratio Lower Upper
228 100
226 29.9 23.7 35.5
229 19.7 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 20.299 ng/ul
RT: 22.192 min Scan# 3248
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08


Tgt Ion:149 Resp: 1431557








#90
Benzo(b)fluoranthene
Concen: 20.574 ng/ul
RT: 22.997 min Scan# 31
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:252 Resp: 1499439
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 11.1 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 19.941 ng/ul
RT: 23.044 min Scan# 3393
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:252 Resp: 1379135
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 11.2 10.6 15.8









#93
Benzo(a)pyrene
Concen: 21.006 ng/ul
RT: 23.603 min Scan# 3488
Delta R.T. -0.000 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 1458198
Ion Ratio Lower Upper
252 100
253 21.8 17.3 25.9
125 12.3 11.8 17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 22.694 ng/ul
RT: 26.121 min Scan# 3916
Delta R.T. -0.006 min
Lab File: BMCCC276.D
Acq: 14 Aug 2022 19:08

Tgt Ion:276 Resp: 1679243
Ion Ratio Lower Upper
276 100
138 25.7 25.6 38.4
277 25.0 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 22.746 ng/ul

RT: 26.138 min Scan# 3919

Delta R.T. -0.000 min


Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:278 Resp: 1443962

Ion Ratio Lower Upper

278 100

139 17.4 16.2 24.2

279 24.2 18.7 28.1



#96

Benzo(g,h,i)perylene

Concen: 23.251 ng/ul

RT: 26.868 min Scan# 4043

Delta R.T. -0.000 min

Lab File: BMCCC276.D

Acq: 14 Aug 2022 19:08

Tgt Ion:276 Resp: 1425717

Ion Ratio Lower Upper

276 100

138 23.7 21.3 31.9

277 24.1 18.6 28.0

