

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081424\
 Data File : BM047187.D
 Acq On : 14 Aug 2024 12:09
 Operator : RC/JU
 Sample : P3540-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0B39

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/14/2024
 Supervised By :mohammad ahmed 08/15/2024

Quant Time: Aug 14 12:40:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM080724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 16:46:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.372	152	5656	0.400	ng/ul	0.00
4) Naphthalene-d8	10.124	136	17205	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.026	164	8578	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.788	188	17932m	0.400	ng/ul	0.00
17) Chrysene-d12	21.015	240	12130	0.400	ng/ul	# 0.00
23) Perylene-d12	23.110	264	10897m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.005	96	36761	6.447	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.724	152	9264	0.361	ng/ul	0.00
18) Fluoranthene-d10	18.831	212	16597	0.474	ng/ul	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.038	88	15053	2.362	ng/ul#	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

