

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081424\
 Data File : BM047190.D
 Acq On : 14 Aug 2024 14:08
 Operator : RC/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4052

Quant Time: Aug 14 16:42:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM080724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 16:46:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.372	152	6507	0.400	ng/ul	0.00
4) Naphthalene-d8	10.124	136	18875	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.027	164	9416	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.792	188	18307	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.015	240	12937	0.400	ng/ul	# 0.00
23) Perylene-d12	23.113	264	13318	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.009	96	3224	0.491	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.729	152	10282	0.365	ng/ul	0.00
18) Fluoranthene-d10	18.836	212	15832	0.424	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.043	88	3869	0.528	ng/ul#	89
5) Naphthalene	10.173	128	19388	0.393	ng/ul	100
7) 2-Methylnaphthalene	11.806	142	11775	0.366	ng/ul	98
8) 1-Methylnaphthalene	12.026	142	12428	0.369	ng/ul	100
10) Acenaphthylene	13.744	152	16828	0.385	ng/ul	99
11) Acenaphthene	14.091	153	12226	0.411	ng/ul	98
12) Fluorene	15.091	166	13575	0.392	ng/ul	100
14) Pentachlorophenol	16.458	266	1645	0.282	ng/ul	95
15) Phenanthrene	16.834	178	20229	0.384	ng/ul	99
16) Anthracene	16.923	178	16946	0.360	ng/ul	100
19) Fluoranthene	18.868	202	21666	0.413	ng/ul#	95
20) Pyrene	19.231	202	22656	0.413	ng/ul	94
21) Benzo(a)anthracene	21.001	228	18600	0.391	ng/ul	99
22) Chrysene	21.053	228	20361	0.404	ng/ul	99
24) Benzo(b)fluoranthene	22.497	252	20326	0.384	ng/ul	96
25) Benzo(k)fluoranthene	22.538	252	22121	0.400	ng/ul#	96
26) Benzo(a)pyrene	23.026	252	16178	0.387	ng/ul#	94
27) Indeno(1,2,3-cd)pyrene	25.173	276	25187	0.397	ng/ul#	91
28) Dibenzo(a,h)anthracene	25.190	278	19313	0.391	ng/ul	97
29) Benzo(g,h,i)perylene	25.800	276	21279	0.404	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081424\
Data File : BM047190.D
Acq On : 14 Aug 2024 14:08
Operator : RC/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4052

Quant Time: Aug 14 16:42:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM080724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 13 16:46:53 2024
Response via : Initial Calibration

