

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081424\
 Data File : BM047183.D
 Acq On : 13 Aug 2024 20:23
 Operator : RC/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4199

Quant Time: Aug 14 00:21:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM080724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 13 16:46:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.372	152	5782	0.400	ng/ul	0.00
4) Naphthalene-d8	10.124	136	16710	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.027	164	8113	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.788	188	15457	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.013	240	11497	0.400	ng/ul	# 0.00
23) Perylene-d12	23.111	264	12229	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.009	96	2851	0.489	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.730	152	8958	0.359	ng/ul	0.00
18) Fluoranthene-d10	18.836	212	13392	0.403	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.043	88	3424	0.526	ng/ul	95
5) Naphthalene	10.173	128	17116	0.392	ng/ul	99
7) 2-Methylnaphthalene	11.801	142	10319	0.362	ng/ul	99
8) 1-Methylnaphthalene	12.026	142	10784	0.362	ng/ul	100
10) Acenaphthylene	13.740	152	14531	0.386	ng/ul	99
11) Acenaphthene	14.091	153	10574	0.413	ng/ul	99
12) Fluorene	15.091	166	11610	0.389	ng/ul	99
14) Pentachlorophenol	16.458	266	1442	0.293	ng/ul	96
15) Phenanthrene	16.830	178	17190	0.386	ng/ul	99
16) Anthracene	16.923	178	14507	0.365	ng/ul	99
19) Fluoranthene	18.864	202	18462	0.396	ng/ul#	94
20) Pyrene	19.231	202	19405	0.398	ng/ul	95
21) Benzo(a)anthracene	20.998	228	16290	0.386	ng/ul	99
22) Chrysene	21.048	228	18026	0.402	ng/ul	99
24) Benzo(b)fluoranthene	22.494	252	18712	0.385	ng/ul	96
25) Benzo(k)fluoranthene	22.535	252	19947	0.393	ng/ul#	95
26) Benzo(a)pyrene	23.023	252	14970	0.390	ng/ul#	93
27) Indeno(1,2,3-cd)pyrene	25.170	276	24010	0.412	ng/ul#	91
28) Dibenzo(a,h)anthracene	25.183	278	18581	0.410	ng/ul	97
29) Benzo(g,h,i)perylene	25.793	276	20370	0.422	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081424\
Data File : BM047183.D
Acq On : 13 Aug 2024 20:23
Operator : RC/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4199

Quant Time: Aug 14 00:21:26 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM080724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 13 16:46:53 2024
Response via : Initial Calibration

