

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081516\
 Data File : BM007119.D
 Acq On : 15 Aug 2016 15:54
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02040

Quant Time: Aug 16 01:51:54 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM081516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 15 15:51:10 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
3	Naphthalene	1.004	1.032	-2.8	107	0.00
4 SURR	2-Methylnaphthalene-d10	0.500	0.512	-2.4	107	0.00
5	2-Methylnaphthalene	0.676	0.692	-2.4	108	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
7	Acenaphthylene	2.166	2.192	-1.2	105	0.00
8 C	Acenaphthene	1.392	1.501	-7.8	111	0.00
9	Fluorene	1.578	1.684	-6.7	111	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
11	Pentachlorophenol	0.100	0.107	-7.0	120	0.00
12	Phenanthrene	1.136	1.218	-7.2	108	0.00
13	Anthracene	1.085	1.164	-7.3	109	0.00
14 SURR	Fluoranthene-d10	0.901	0.949	-5.3	106	0.00
15 C	Fluoranthene	1.320	1.398	-5.9	107	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
17	Pyrene	1.329	1.506	-13.3	107	0.00
18	Benzo(a)anthracene	1.264	1.404	-11.1	106	0.00
19	Chrysene	1.215	1.363	-12.2	106	0.00
20 I	Perylene-d12	1.000	1.000	0.0	104	0.00
21	Benzo(b)fluoranthene	1.482	1.628	-9.9	106	0.00
22	Benzo(k)fluoranthene	1.221	1.440	-17.9	108	0.00
23 C	Benzo(a)pyrene	1.272	1.426	-12.1	104	0.00
24	Indeno(1,2,3-cd)pyrene	1.584	1.790	-13.0	107	0.00
25	Dibenzo(a,h)anthracene	1.246	1.397	-12.1	105	0.00
26	Benzo(g,h,i)perylene	1.345	1.557	-15.8	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0