

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081522\
 Data File : BM036280.D
 Acq On : 15 Aug 2022 11:06
 Operator : CG/JU
 Sample : PB146648BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS648

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/16/2022
 Supervised By :Sohil Jodhani 08/16/2022

Quant Time: Aug 15 11:43:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 14 23:34:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	156978	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	647392	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	378130	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	774487	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	858495	20.000	ng/u1	0.00
88) Perylene-d12	23.709	264	910916	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	23805	5.706	ng/uL	0.00
4) Pyridine-d5	3.640	84	302727	26.489	ng/u1	0.00
7) Phenol-d5	6.992	99	399176	30.323	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.145	67	246126	28.046	ng/u1	0.00
11) 2-Chlorophenol-d4	7.339	132	331356	30.325	ng/u1	0.00
15) 4-Methylphenol-d8	8.534	113	299691	28.102	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	154668	30.150	ng/u1	0.00
24) 2-Nitrophenol-d4	9.698	143	156971	31.407	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.239	165	285717	31.607	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	323310	21.127	ng/u1	0.00
46) Dimethylphthalate-d6	13.868	166	782703	30.388	ng/u1	0.00
49) Acenaphthylene-d8	14.145	160	1005543	30.597	ng/u1	0.00
54) 4-Nitrophenol-d4	14.680	143	156961	29.675	ng/u1	0.00
60) Fluorene-d10	15.439	176	710289	30.517	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.568	200	123567	29.682	ng/u1	0.00
73) Anthracene-d10	17.292	188	1098526	31.373	ng/u1	0.00
81) Pyrene-d10	19.586	212	1332939	28.015	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.562	264	1479676	31.996	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	50345	11.645	ng/uL	97
5) Pyridine	3.657	79	347913	29.010	ng/u1	91
6) Benzaldehyde	6.951	77	253840	45.944	ng/u1	95
8) Phenol	7.016	94	438868	30.464	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.239	93	341736	29.506	ng/u1	97
12) 2-Chlorophenol	7.375	128	352445	31.443	ng/u1	99
13) 2-Methylphenol	8.269	108	315502	29.316	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.345	45	446365	29.198	ng/u1	97
16) Acetophenone	8.639	105	497160	29.036	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.622	70	242924	28.450	ng/u1	97
18) 4-Methylphenol	8.598	108	345455	29.922	ng/u1	97
19) Hexachloroethane	8.881	117	141998	30.703	ng/u1	91
22) Nitrobenzene	9.016	77	374286	31.883	ng/u1	99
23) Isophorone	9.545	82	655734	29.591	ng/u1	97
25) 2-Nitrophenol	9.728	139	180136	32.487	ng/u1	98
26) 2,4-Dimethylphenol	9.798	107	350626	30.946	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.028	93	425702	30.459	ng/u1	100
29) 2,4-Dichlorophenol	10.269	162	300425	32.020	ng/u1	98
30) Naphthalene	10.657	128	1083974	31.722	ng/u1	99
32) 4-Chloroaniline	10.780	127	370284	24.358	ng/u1	98
33) Hexachlorobutadiene	10.933	225	181600	31.353	ng/u1	100
34) Caprolactam	11.580	113	88840	28.651	ng/u1	88
35) 4-Chloro-3-methylphenol	11.916	107	309136	31.687	ng/u1	98
36) 2-Methylnaphthalene	12.274	142	686033	30.686	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081522\
 Data File : BM036280.D
 Acq On : 15 Aug 2022 11:06
 Operator : CG/JU
 Sample : PB146648BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS648


Quant Time: Aug 15 11:43:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM072522.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Aug 14 23:34:05 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/16/2022
 Supervised By :Sohil Jodhani 08/16/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	695407	30.859	ng/u1	97
39) 1,2,4,5-Tetrachloroben...	12.639	216	329415	30.604	ng/u1	99
40) Hexachlorocyclopentadiene	12.610	237	165478	23.083	ng/u1	95
41) 2,4,6-Trichlorophenol	12.886	196	207309	30.096	ng/u1	99
42) 2,4,5-Trichlorophenol	12.963	196	227806	30.995	ng/u1	100
43) 1,1'-Biphenyl	13.286	154	903452	31.042	ng/u1	99
44) 2-Chloronaphthalene	13.327	162	697749	31.238	ng/u1	99
45) 2-Nitroaniline	13.545	65	199557	32.407	ng/u1	98
47) Dimethylphthalate	13.916	163	819887	31.645	ng/u1	100
48) 2,6-Dinitrotoluene	14.039	165	162692	31.629	ng/u1	90
50) Acenaphthylene	14.168	152	1148388	31.589	ng/u1	100
51) 3-Nitroaniline	14.374	138	197232	34.695	ng/u1	95
52) Acenaphthene	14.515	153	750189	31.808	ng/u1	98
53) 2,4-Dinitrophenol	14.580	184	80854	27.684	ng/u1	97
55) 4-Nitrophenol	14.692	109	127344	31.479	ng/u1	87
56) Dibenzofuran	14.851	168	1055198	32.335	ng/u1	99
57) 2,4-Dinitrotoluene	14.827	165	240939	33.229	ng/u1	95
58) 2,3,4,6-Tetrachlorophenol	15.080	232	174184	30.163	ng/u1	93
59) Diethylphthalate	15.274	149	838783	32.007	ng/u1	99
61) Fluorene	15.498	166	856621	32.134	ng/u1	98
62) 4-Chlorophenyl-phenyle...	15.492	204	393913	30.841	ng/u1	98
63) 4-Nitroaniline	15.533	138	217514m	41.562	ng/u1	
66) 4,6-Dinitro-2-methylph...	15.586	198	128633	30.878	ng/u1#	94
67) N-Nitrosodiphenylamine	15.710	169	705165	30.970	ng/u1	97
68) 4-Bromophenyl-phenylether	16.386	248	226086	29.777	ng/u1	97
69) Hexachlorobenzene	16.492	284	282278	30.512	ng/u1	96
70) Atrazine	16.662	200	97413	12.335	ng/u1	98
71) Pentachlorophenol	16.845	266	141048	24.957	ng/u1	99
72) Phenanthrene	17.233	178	1356369	32.675	ng/u1	98
74) Anthracene	17.327	178	1373887	32.984	ng/u1	98
75) 1,2,3,4-Tetrachloroben...	13.245	216	328477	27.273	ng/uL	99
76) Pentachlorobenzene	14.763	250	331340	30.349	ng/uL	99
77) Carbazole	17.604	167	1320629	34.328	ng/u1	99
78) Di-n-butylphthalate	18.162	149	1441917	33.220	ng/u1	99
80) Fluoranthene	19.250	202	1647076	28.974	ng/u1	98
82) Pyrene	19.615	202	1724787	29.436	ng/u1	97
83) Butylbenzylphthalate	20.515	149	673805	31.618	ng/u1	94
84) 3,3'-Dichlorobenzidine	21.297	252	571352	36.065	ng/u1	99
85) Benzo(a)anthracene	21.362	228	1744247	32.280	ng/u1	100
86) Bis(2-ethylhexyl)phtha...	21.292	149	971362	31.545	ng/u1	100
87) Chrysene	21.415	228	1728299	32.781	ng/u1	99
89) Di-n-octyl phthalate	22.197	149	1695370	28.821	ng/u1	100
90) Benzo(b)fluoranthene	22.997	252	1905524	31.345	ng/u1	98
91) Benzo(k)fluoranthene	23.044	252	1822538	31.593	ng/u1	98
93) Benzo(a)pyrene	23.609	252	1884743	32.549	ng/u1	97
94) Indeno(1,2,3-cd)pyrene	26.127	276	2204348	35.715	ng/u1	95
95) Dibenzo(a,h)anthracene	26.144	278	1910495	36.080	ng/u1	96
96) Benzo(g,h,i)perylene	26.874	276	1885839	36.871	ng/u1	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.645 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion: 88 Resp: 50345
Ion Ratio Lower Upper
88 100
43 33.8 27.8 41.8
58 74.5 61.8 92.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#5
Pyridine
Concen: 29.010 ng/uL
RT: 3.657 min Scan# 97
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion: 79 Resp: 347913
Ion Ratio Lower Upper
79 100
52 74.1 67.8 101.6
51 37.1 31.4 47.0





#6
Benzaldehyde
Concen: 45.944 ng/ul
RT: 6.951 min Scan# 657
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion: 77 Resp: 253840
Ion Ratio Lower Upper
77 100
105 97.8 84.6 127.0
106 98.2 80.3 120.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#8
Phenol
Concen: 30.464 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion: 94 Resp: 438868
Ion Ratio Lower Upper
94 100
65 28.4 21.3 31.9
66 35.7 27.4 41.0





#10
Bis(2-Chloroethyl)ether
Concen: 29.506 ng/ul
RT: 7.239 min Scan# 705
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion: 93 Resp: 341736
Ion Ratio Lower Upper
93 100
63 77.0 64.0 96.0
95 32.8 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#12
2-Chlorophenol
Concen: 31.443 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:128 Resp: 352445
Ion Ratio Lower Upper
128 100
64 44.9 36.9 55.3
130 32.0 25.6 38.4





#13
2-Methylphenol
Concen: 29.316 ng/ul
RT: 8.269 min Scan# 881
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:108 Resp: 315502
Ion Ratio Lower Upper
108 100
107 88.2 75.0 112.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.198 ng/ul
RT: 8.345 min Scan# 894
Delta R.T. 0.012 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06



Tgt Ion: 45 Resp: 446365
Ion Ratio Lower Upper
45 100
77 14.3 10.6 15.8
79 11.3 8.3 12.5









#16
Acetophenone
Concen: 29.036 ng/ul
RT: 8.639 min Scan# 943
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06

Instrument :
BNA_M
ClientSampleId :
SLCS648

Tgt Ion:105 Resp: 497160
Ion Ratio Lower Upper
105 100
77 76.7 59.8 89.6
51 27.3 22.1 33.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#17
N-Nitroso-di-n-propylamine
Concen: 28.450 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion: 70 Resp: 242924
Ion Ratio Lower Upper
70 100
42 58.8 49.6 74.4
101 9.7 8.1 12.1
130 21.9 18.4 27.6





#18
4-Methylphenol
Concen: 29.922 ng/ul
RT: 8.598 min Scan# 931
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:108 Resp: 34545
Ion Ratio Lower Upper
108 100
107 109.4 90.3 135.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#19
Hexachloroethane
Concen: 30.703 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:117 Resp: 141998
Ion Ratio Lower Upper
117 100
201 92.3 81.8 122.6
199 58.3 51.6 77.4





#22
Nitrobenzene
Concen: 31.883 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion: 77 Resp: 374286
Ion Ratio Lower Upper
77 100
123 46.5 37.4 56.0
65 12.8 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#23
Isophorone
Concen: 29.591 ng/ul
RT: 9.545 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion: 82 Resp: 655734
Ion Ratio Lower Upper
82 100
95 6.3 5.5 8.3
138 17.6 15.3 22.9





#25
2-Nitrophenol
Concen: 32.487 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:139 Resp: 180136
Ion Ratio Lower Upper
139 100
65 50.9 39.8 59.6
109 29.8 22.2 33.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#26
2,4-Dimethylphenol
Concen: 30.946 ng/ul
RT: 9.798 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:107 Resp: 350626
Ion Ratio Lower Upper
107 100
121 51.6 42.7 64.1
122 88.7 71.5 107.3





#27
 Bis(2-Chloroethoxy)methane
 Concen: 30.459 ng/ul
 RT: 10.028 min Scan# 1180
 Delta R.T. -0.006 min
 Lab File: BM036280.D
 Acq: 15 Aug 2022 11:06




Instrument :
 BNA_M
 ClientSampleId :
 SLCS648



Tgt Ion: 93 Resp: 425702
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.1 39.1
 123 11.5 9.4 14.0




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
 Supervised By :Sohil Jodhani 08/16/2022



#29
 2,4-Dichlorophenol
 Concen: 32.020 ng/ul
 RT: 10.269 min Scan# 1221
 Delta R.T. -0.000 min
 Lab File: BM036280.D
 Acq: 15 Aug 2022 11:06


Tgt Ion:162 Resp: 300425
 Ion Ratio Lower Upper
 162 100
 164 65.5 51.8 77.6
 98 35.1 29.5 44.3





#30
Naphthalene
Concen: 31.722 ng/ul
RT: 10.657 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:128 Resp: 1083974
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 12.8 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#32
4-Chloroaniline
Concen: 24.358 ng/ul
RT: 10.780 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:127 Resp: 370284
Ion Ratio Lower Upper
127 100
129 32.5 25.3 37.9





#33
Hexachlorobutadiene
Concen: 31.353 ng/ul
RT: 10.933 min Scan# 1334
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:225 Resp: 181600
Ion Ratio Lower Upper
225 100
223 62.5 50.0 75.0
227 61.9 49.9 74.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#34
Caprolactam
Concen: 28.651 ng/ul
RT: 11.580 min Scan# 1444
Delta R.T. 0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:113 Resp: 88840
Ion Ratio Lower Upper
113 100
55 161.9 150.2 225.4
56 123.4 102.1 153.1





#35
4-Chloro-3-methylphenol
Concen: 31.687 ng/ul
RT: 11.916 min Scan# 1501
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:107 Resp: 309136
Ion Ratio Lower Upper
107 100
144 25.9 20.3 30.5
142 82.6 64.4 96.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#36
2-Methylnaphthalene
Concen: 30.686 ng/ul
RT: 12.274 min Scan# 1562
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06

Tgt Ion:142 Resp: 686033
Ion Ratio Lower Upper
142 100
141 90.0 70.7 106.1





#37

1-Methylnaphthalene

Concen: 30.859 ng/ul

RT: 12.492 min Scan# 1599

Delta R.T. -0.000 min

Lab File: BM036280.D


Acq: 15 Aug 2022 11:06

Instrument :

BNA_M

ClientSampleId :

SLCS648






Tgt Ion:	142	Resp:	69540
Ion Ratio	Lower	Upper	
142	100		
141	91.5	75.2	112.8
116	3.8	2.6	3.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/16/2022

Supervised By :Sohil Jodhani 08/16/2022



#39

1,2,4,5-Tetrachlorobenzene

Concen: 30.604 ng/ul




RT: 12.639 min Scan# 1624


Delta R.T. -0.000 min

Lab File: BM036280.D

Acq: 15 Aug 2022 11:06


Tgt Ion:	216	Resp:	329415
Ion Ratio	Lower	Upper	
216	100		
214	77.5	62.4	93.6
179	20.5	17.3	25.9
108	19.1	14.0	21.0





#40
Hexachlorocyclopentadiene
Concen: 23.083 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:237 Resp: 165478
Ion Ratio Lower Upper
237 100
235 64.8 48.4 72.6
272 13.9 11.6 17.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#41
2,4,6-Trichlorophenol
Concen: 30.096 ng/ul
RT: 12.886 min Scan# 1666
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:196 Resp: 207309
Ion Ratio Lower Upper
196 100
198 96.2 77.8 116.6
200 30.4 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 30.995 ng/ul
RT: 12.963 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:196 Resp: 227806
Ion Ratio Lower Upper
196 100
198 96.2 76.8 115.2
200 30.4 24.9 37.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#43
1,1'-Biphenyl
Concen: 31.042 ng/ul
RT: 13.286 min Scan# 1734
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:154 Resp: 903452
Ion Ratio Lower Upper
154 100
153 39.8 32.4 48.6
76 13.2 10.1 15.1





#44
2-Chloronaphthalene
Concen: 31.238 ng/ul
RT: 13.327 min Scan# 1741
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:162 Resp: 697749
Ion Ratio Lower Upper
162 100
164 31.9 25.4 38.0
127 38.7 31.6 47.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#45
2-Nitroaniline
Concen: 32.407 ng/ul
RT: 13.545 min Scan# 1778
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion: 65 Resp: 199557
Ion Ratio Lower Upper
65 100
92 68.8 54.5 81.7
138 111.4 91.2 136.8





#47
Dimethylphthalate
Concen: 31.645 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:163 Resp: 819887
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 9.7 7.9 11.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#48
2,6-Dinitrotoluene
Concen: 31.629 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:165 Resp: 162692
Ion Ratio Lower Upper
165 100
89 62.3 42.5 63.7
121 23.1 17.2 25.8





#50
Acenaphthylene
Concen: 31.589 ng/ul
RT: 14.168 min Scan# 1884
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument : BNA_M
ClientSampleId : SLCS648



Tgt Ion:152 Resp: 1148388
Ion Ratio Lower Upper
152 100
151 19.9 16.1 24.1
153 13.2 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#51
3-Nitroaniline
Concen: 34.695 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:138 Resp: 197232
Ion Ratio Lower Upper
138 100
108 9.0 8.1 12.1
92 107.1 90.4 135.6





#52
Acenaphthene
Concen: 31.808 ng/ul
RT: 14.515 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:153 Resp: 750189
Ion Ratio Lower Upper
153 100
152 47.7 37.1 55.7
154 87.7 71.3 106.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#53
2,4-Dinitrophenol
Concen: 27.684 ng/ul
RT: 14.580 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:184 Resp: 80854
Ion Ratio Lower Upper
184 100
63 70.5 54.2 81.4
154 64.2 49.3 73.9





#55
4-Nitrophenol
Concen: 31.479 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:109 Resp: 127344
Ion Ratio Lower Upper
109 100
139 140.6 124.9 187.3
65 134.9 122.2 183.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#56
Dibenzofuran
Concen: 32.335 ng/ul
RT: 14.851 min Scan# 2000
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:168 Resp: 1055198
Ion Ratio Lower Upper
168 100
139 38.3 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 33.229 ng/ul
RT: 14.827 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:165 Resp: 240939
Ion Ratio Lower Upper
165 100
63 48.4 35.9 53.9
182 2.7 2.5 3.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 30.163 ng/ul
RT: 15.080 min Scan# 2039
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:232 Resp: 174184
Ion Ratio Lower Upper
232 100
131 52.4 36.7 55.1
130 2.4 1.7 2.5
166 32.3 24.4 36.6





#59
Diethylphthalate
Concen: 32.007 ng/ul
RT: 15.274 min Scan# 2072
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:149 Resp: 838783
Ion Ratio Lower Upper
149 100
177 21.1 16.8 25.2
150 12.0 10.2 15.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#61
Fluorene
Concen: 32.134 ng/ul
RT: 15.498 min Scan# 2110
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:166 Resp: 856621
Ion Ratio Lower Upper
166 100
165 98.4 77.2 115.8
167 13.4 11.0 16.4





#62
4-Chlorophenyl-phenylether
Concen: 30.841 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:204 Resp: 393913
Ion Ratio Lower Upper
204 100
206 32.2 25.9 38.9
141 61.3 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#63
4-Nitroaniline
Concen: 41.562 ng/ul m
RT: 15.533 min Scan# 2116
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:138 Resp: 217514
Ion Ratio Lower Upper
138 100
92 50.9 39.2 58.8
108 77.0 51.6 77.4





#66
4,6-Dinitro-2-methylphenol
Concen: 30.878 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:198 Resp: 128633
Ion Ratio Lower Upper
198 100
182 4.2 4.2 6.4
77 28.4 19.9 29.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#67
N-Nitrosodiphenylamine
Concen: 30.970 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:169 Resp: 705165
Ion Ratio Lower Upper
169 100
168 68.6 53.8 80.6
167 38.3 28.7 43.1





#68
4-Bromophenyl-phenylether
Concen: 29.777 ng/ul
RT: 16.386 min Scan# 2261
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:248 Resp: 226086
Ion Ratio Lower Upper
248 100
250 99.5 77.3 115.9
141 77.9 59.8 89.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#69
Hexachlorobenzene
Concen: 30.512 ng/ul
RT: 16.492 min Scan# 2279
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:284 Resp: 282278
Ion Ratio Lower Upper
284 100
142 39.8 28.6 42.8
249 30.0 24.6 37.0





#70
 Atrazine
 Concen: 12.335 ng/ul
 RT: 16.662 min Scan# 2308
 Delta R.T. -0.006 min
 Lab File: BM036280.D
 Acq: 15 Aug 2022 11:06




Instrument :
 BNA_M
 ClientSampleId :
 SLCS648




Tgt Ion:200 Resp: 97413
 Ion Ratio Lower Upper
 200 100
 173 26.3 21.3 31.9
 215 47.1 39.0 58.6

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
 Supervised By :Sohil Jodhani 08/16/2022




#71
 Pentachlorophenol
 Concen: 24.957 ng/ul
 RT: 16.845 min Scan# 2339
 Delta R.T. -0.006 min
 Lab File: BM036280.D
 Acq: 15 Aug 2022 11:06




Tgt Ion:266 Resp: 141048
 Ion Ratio Lower Upper
 266 100
 264 62.1 49.5 74.3
 268 64.4 51.0 76.4





#72
Phenanthrene
Concen: 32.675 ng/u1
RT: 17.233 min Scan# 2405
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:178 Resp: 1356369
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 19.9 14.9 22.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#74
Anthracene
Concen: 32.984 ng/u1
RT: 17.327 min Scan# 2421
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:178 Resp: 1373887
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 19.2 14.4 21.6





#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.273 ng/uL
RT: 13.245 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:216 Resp: 328472
Ion Ratio Lower Upper
216 100
214 77.8 60.9 91.3
218 47.1 37.5 56.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#76
Pentachlorobenzene
Concen: 30.349 ng/uL
RT: 14.763 min Scan# 1985
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Tgt Ion:250 Resp: 331340
Ion Ratio Lower Upper
250 100
248 63.4 50.4 75.6
252 64.3 51.1 76.7





#77
 Carbazole
 Concen: 34.328 ng/ul
 RT: 17.604 min Scan# 2468
 Delta R.T. -0.000 min
 Lab File: BM036280.D
 Acq: 15 Aug 2022 11:06




Instrument :
 BNA_M
 ClientSampleId :
 SLCS648



Tgt Ion:167 Resp: 1320629
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 13.1 10.1 15.1




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
 Supervised By :Sohil Jodhani 08/16/2022



#78
 Di-n-butylphthalate
 Concen: 33.220 ng/ul
 RT: 18.162 min Scan# 2563
 Delta R.T. -0.006 min
 Lab File: BM036280.D
 Acq: 15 Aug 2022 11:06


Tgt Ion:149 Resp: 1441917
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 5.2 3.7 5.5





#80
Fluoranthene
Concen: 28.974 ng/ul
RT: 19.250 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648




Tgt Ion:202 Resp: 1647076
Ion Ratio Lower Upper
202 100
101 12.1 10.1 15.1
100 9.0 7.8 11.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022




#82
Pyrene
Concen: 29.436 ng/ul
RT: 19.615 min Scan# 2810
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06



Tgt Ion:202 Resp: 1724787
Ion Ratio Lower Upper
202 100
101 13.8 11.9 17.9
100 11.3 10.2 15.4





#83
Butylbenzylphthalate
Concen: 31.618 ng/ul
RT: 20.515 min Scan# 2963
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06

Instrument :

BNA_M





ClientSampleId :

SLCS648

Tgt Ion:149 Resp: 673805
Ion Ratio Lower Upper
149 100
91 71.6 53.2 79.8
206 17.4 16.0 24.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#84
3,3'-Dichlorobenzidine
Concen: 36.065 ng/ul
RT: 21.297 min Scan# 3096
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:252 Resp: 571352
Ion Ratio Lower Upper
252 100
254 65.0 52.4 78.6
126 12.6 10.8 16.2





#85
Benzo(a)anthracene
Concen: 32.280 ng/ul
RT: 21.362 min Scan# 3107
Delta R.T. 0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:228 Resp: 174424
Ion Ratio Lower Upper
228 100
229 20.0 15.6 23.4
226 27.1 21.7 32.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 31.545 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:149 Resp: 971362
Ion Ratio Lower Upper
149 100
167 26.8 21.5 32.3
279 3.8 3.4 5.2





#87
Chrysene
Concen: 32.781 ng/ul
RT: 21.415 min Scan# 3116
Delta R.T. 0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:228 Resp: 1728299
Ion Ratio Lower Upper
228 100
226 30.1 23.7 35.5
229 19.8 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#89
Di-n-octyl phthalate
Concen: 28.821 ng/ul
RT: 22.197 min Scan# 3249
Delta R.T. 0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06


Tgt Ion:149 Resp: 1695370





#90
Benzo(b)fluoranthene
Concen: 31.345 ng/ul
RT: 22.997 min Scan# 3385
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06




Instrument :
BNA_M
ClientSampleId :
SLCS648



Tgt Ion:252 Resp: 1905524
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 11.1 10.7 16.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#91
Benzo(k)fluoranthene
Concen: 31.593 ng/ul
RT: 23.044 min Scan# 3393
Delta R.T. -0.000 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06

Tgt Ion:252 Resp: 1822538
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 11.0 10.6 15.8





#93

Benzo(a)pyrene

Concen: 32.549 ng/ul

RT: 23.609 min Scan# 3488

Delta R.T. 0.006 min

Lab File: BM036280.D

Acq: 15 Aug 2022 11:06

Instrument :

BNA_M

ClientSampleId :

SLCS648





Tgt Ion:252 Resp: 188474

Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.3	25.9
125	11.9	11.8	17.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 35.715 ng/ul

RT: 26.127 min Scan# 3917




Delta R.T. -0.000 min


Lab File: BM036280.D

Acq: 15 Aug 2022 11:06

Tgt Ion:276 Resp: 2204348

Ion	Ratio	Lower	Upper
276	100		
138	27.1	25.6	38.4
277	24.9	20.1	30.1









#95
Dibenzo(a,h)anthracene
Concen: 36.080 ng/ul
RT: 26.144 min Scan# 3919
Delta R.T. 0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06

Instrument :
BNA_M
ClientSampleId :
SLCS648

Tgt Ion:278 Resp: 1910495
Ion Ratio Lower Upper
278 100
139 16.7 16.2 24.2
279 23.6 18.7 28.1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/16/2022
Supervised By :Sohil Jodhani 08/16/2022



#96
Benzo(g,h,i)perylene
Concen: 36.871 ng/ul
RT: 26.874 min Scan# 4044
Delta R.T. 0.006 min
Lab File: BM036280.D
Acq: 15 Aug 2022 11:06

Tgt Ion:276 Resp: 1885839
Ion Ratio Lower Upper
276 100
138 23.1 21.3 31.9
277 23.7 18.6 28.0

