

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036295.D
 Acq On : 16 Aug 2022 11:12
 Operator : CG/JU
 Sample : SSTD00560
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005012

Quant Time: Aug 16 14:07:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	243838	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	1075733	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	631801	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1141512	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	916728	20.000	ng/u1	0.00
88) Perylene-d12	23.703	264	922104	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	11817	1.824	ng/uL	0.00
4) Pyridine-d5	3.646	84	71698	4.039	ng/u1	0.00
7) Phenol-d5	6.987	99	87536	4.281	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	59089	4.335	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	71983	4.241	ng/u1	0.00
15) 4-Methylphenol-d8	8.534	113	65751	3.969	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	34230	4.016	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	33739	4.063	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.239	165	61819	4.116	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	94126	3.702	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	178253	4.142	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	209970	3.824	ng/u1	0.00
54) 4-Nitrophenol-d4	14.686	143	18656	2.111	ng/u1	0.00
60) Fluorene-d10	15.439	176	157845	4.059	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.574	200	14877	2.425	ng/u1	0.00
73) Anthracene-d10	17.286	188	223723	4.335	ng/u1	0.00
81) Pyrene-d10	19.580	212	216187	4.255	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.550	264	194854	4.162	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	11906	1.773	ng/uL#	93
5) Pyridine	3.663	79	76089	4.084	ng/u1	94
6) Benzaldehyde	6.951	77	40501	4.719	ng/u1	93
8) Phenol	7.016	94	90210	4.031	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.234	93	77966	4.334	ng/u1	97
12) 2-Chlorophenol	7.369	128	75683	4.347	ng/u1	98
13) 2-Methylphenol	8.263	108	66680	3.989	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.334	45	106521	4.486	ng/u1	98
16) Acetophenone	8.639	105	111734	4.201	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.616	70	57491	4.335	ng/u1	92
18) 4-Methylphenol	8.598	108	73652	4.107	ng/u1	96
19) Hexachloroethane	8.875	117	31812	4.428	ng/u1	89
22) Nitrobenzene	9.016	77	82149	4.211	ng/u1	98
23) Isophorone	9.539	82	153161	4.159	ng/u1	97
25) 2-Nitrophenol	9.728	139	38120	4.137	ng/u1	97
26) 2,4-Dimethylphenol	9.792	107	80843	4.294	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.028	93	102211	4.401	ng/u1	99
29) 2,4-Dichlorophenol	10.269	162	65151	4.179	ng/u1	98
30) Naphthalene	10.651	128	256909	4.525	ng/u1	99
32) 4-Chloroaniline	10.780	127	96986	3.840	ng/u1	97
33) Hexachlorobutadiene	10.928	225	42376	4.403	ng/u1	98
34) Caprolactam	11.575	113	17521	3.401	ng/u1	80
35) 4-Chloro-3-methylphenol	11.916	107	65219	4.023	ng/u1	97
36) 2-Methylnaphthalene	12.269	142	161571	4.349	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036295.D
 Acq On : 16 Aug 2022 11:12
 Operator : CG/JU
 Sample : SSTD00560
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD005012


Quant Time: Aug 16 14:07:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.486	142	161573	4.315	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.633	216	76907	4.276	ng/ul#	96
40) Hexachlorocyclopentadiene	12.610	237	30376	2.536	ng/ul	98
41) 2,4,6-Trichlorophenol	12.892	196	44904	3.902	ng/ul	97
42) 2,4,5-Trichlorophenol	12.969	196	47646	3.880	ng/ul	98
43) 1,1'-Biphenyl	13.280	154	211292	4.345	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	159192	4.266	ng/ul	98
45) 2-Nitroaniline	13.545	65	37370	3.632	ng/ul	92
47) Dimethylphthalate	13.910	163	183402	4.237	ng/ul	99
48) 2,6-Dinitrotoluene	14.039	165	28607	3.328	ng/ul#	78
50) Acenaphthylene	14.169	152	253487	4.173	ng/ul	99
51) 3-Nitroaniline	14.374	138	28027	2.951	ng/ul	96
52) Acenaphthene	14.510	153	173880	4.412	ng/ul	98
53) 2,4-Dinitrophenol	14.598	184	6750	1.383	ng/ul	87
55) 4-Nitrophenol	14.710	109	14483	2.143	ng/ul	83
56) Dibenzofuran	14.845	168	238230	4.369	ng/ul	96
57) 2,4-Dinitrotoluene	14.827	165	39482	3.259	ng/ul#	85
58) 2,3,4,6-Tetrachlorophenol	15.080	232	34984	3.626	ng/ul#	96
59) Diethylphthalate	15.268	149	176291	4.026	ng/ul	99
61) Fluorene	15.492	166	187567	4.211	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.492	204	88581	4.151	ng/ul	98
63) 4-Nitroaniline	15.533	138	17450	1.996	ng/ul#	82
66) 4,6-Dinitro-2-methylph...	15.592	198	16057	2.615	ng/ul#	64
67) N-Nitrosodiphenylamine	15.704	169	145338	4.331	ng/ul	97
68) 4-Bromophenyl-phenylether	16.386	248	47635	4.257	ng/ul	97
69) Hexachlorobenzene	16.486	284	58635	4.300	ng/ul	95
70) Atrazine	16.662	200	42359	3.639	ng/ul	99
71) Pentachlorophenol	16.857	266	17359	2.084	ng/ul	96
72) Phenanthrene	17.233	178	274701	4.490	ng/ul	99
74) Anthracene	17.321	178	264307	4.305	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	79058	4.454	ng/ul	98
76) Pentachlorobenzene	14.757	250	74407	4.624	ng/ul	99
77) Carbazole	17.604	167	228231	4.025	ng/ul	97
78) Di-n-butylphthalate	18.162	149	265216	4.146	ng/ul	100
80) Fluoranthene	19.251	202	268466	4.423	ng/ul	98
82) Pyrene	19.609	202	282924	4.522	ng/ul	97
83) Butylbenzylphthalate	20.515	149	94057	4.133	ng/ul	90
84) 3,3'-Dichlorobenzidine	21.303	252	73190	4.326	ng/ul	99
85) Benzo(a)anthracene	21.356	228	243561	4.221	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.292	149	120754	3.672	ng/ul	98
87) Chrysene	21.409	228	240003	4.263	ng/ul	99
89) Di-n-octyl phthalate	22.192	149	207245	3.480	ng/ul	100
90) Benzo(b)fluoranthene	22.991	252	239627	3.894	ng/ul	98
91) Benzo(k)fluoranthene	23.039	252	234554	4.017	ng/ul	100
93) Benzo(a)pyrene	23.597	252	236320	4.032	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.115	276	283307	4.534	ng/ul	96
95) Dibenzo(a,h)anthracene	26.132	278	242788	4.529	ng/ul	97
96) Benzo(g,h,i)perylene	26.868	276	242902	4.691	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 16 14:07:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 16 14:05:33 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.773 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion: 88 Resp: 11906
Ion Ratio Lower Upper
88 100
43 43.2 27.8 41.8#
58 74.3 61.8 92.8



#5
Pyridine
Concen: 4.084 ng/uL
RT: 3.663 min Scan# 98
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion: 79 Resp: 76089
Ion Ratio Lower Upper
79 100
52 78.4 67.8 101.6
51 37.6 31.4 47.0









#6
Benzaldehyde
Concen: 4.719 ng/ul
RT: 6.951 min Scan# 657
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion: 77 Resp: 40501
Ion Ratio Lower Upper
77 100
105 95.3 84.6 127.0
106 96.9 80.3 120.5



#8
Phenol
Concen: 4.031 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion: 94 Resp: 90210
Ion Ratio Lower Upper
94 100
65 29.6 21.3 31.9
66 35.7 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 4.334 ng/ul
RT: 7.234 min Scan# 705
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion: 93 Resp: 77966
Ion Ratio Lower Upper
93 100
63 76.0 64.0 96.0
95 32.3 25.8 38.6



#12
2-Chlorophenol
Concen: 4.347 ng/ul
RT: 7.369 min Scan# 728
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:128 Resp: 75683
Ion Ratio Lower Upper
128 100
64 48.1 36.9 55.3
130 31.2 25.6 38.4









#13
2-Methylphenol
Concen: 3.989 ng/ul
RT: 8.263 min Scan# 880
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:108 Resp: 66680
Ion Ratio Lower Upper
108 100
107 96.8 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.486 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion: 45 Resp: 106521
Ion Ratio Lower Upper
45 100
77 14.7 10.6 15.8
79 10.1 8.3 12.5









#16
Acetophenone
Concen: 4.201 ng/ul
RT: 8.639 min Scan# 943
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:105 Resp: 111734
Ion Ratio Lower Upper
105 100
77 78.5 59.8 89.6
51 26.9 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 4.335 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion: 70 Resp: 57491
Ion Ratio Lower Upper
70 100
42 55.8 49.6 74.4
101 9.2 8.1 12.1
130 18.8 18.4 27.6









#18
4-Methylphenol
Concen: 4.107 ng/ul
RT: 8.598 min Scan# 913
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:108 Resp: 73652
Ion Ratio Lower Upper
108 100
107 109.1 90.3 135.5



#19
Hexachloroethane
Concen: 4.428 ng/ul
RT: 8.875 min Scan# 984
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:117 Resp: 31812
Ion Ratio Lower Upper
117 100
201 90.4 81.8 122.6
199 56.8 51.6 77.4









#22
Nitrobenzene
Concen: 4.211 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion: 77 Resp: 82149
Ion Ratio Lower Upper
77 100
123 45.5 37.4 56.0
65 11.9 10.8 16.2



#23
Isophorone
Concen: 4.159 ng/ul
RT: 9.539 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion: 82 Resp: 153161
Ion Ratio Lower Upper
82 100
95 6.6 5.5 8.3
138 17.7 15.3 22.9









#25
2-Nitrophenol
Concen: 4.137 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:	139	Resp:	38120
Ion Ratio	Lower	Upper	
139	100		
65	51.1	39.8	59.6
109	25.6	22.2	33.4



#26
2,4-Dimethylphenol
Concen: 4.294 ng/ul
RT: 9.792 min Scan# 1140
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:	107	Resp:	80843
Ion Ratio	Lower	Upper	
107	100		
121	47.8	42.7	64.1
122	88.0	71.5	107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 4.401 ng/ul

RT: 10.028 min Scan# 1180

Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12

Instrument :

BNA_M

ClientSampleId :

SSTD005012





Tgt Ion: 93 Resp: 102211

Ion Ratio Lower Upper

93 100

95 33.0 26.1 39.1

123 11.4 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 4.179 ng/ul

RT: 10.269 min Scan# 1221

Delta R.T. 0.006 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12




Tgt Ion:162 Resp: 65151


Ion Ratio Lower Upper

162 100

164 62.7 51.8 77.6

98 35.9 29.5 44.3









#30
Naphthalene
Concen: 4.525 ng/ul
RT: 10.651 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:128 Resp: 256909
Ion Ratio Lower Upper
128 100
129 10.4 8.8 13.2
127 13.3 10.6 16.0



#32
4-Chloroaniline
Concen: 3.840 ng/ul
RT: 10.780 min Scan# 1308
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:127 Resp: 96986
Ion Ratio Lower Upper
127 100
129 33.4 25.3 37.9









#33
Hexachlorobutadiene
Concen: 4.403 ng/ul
RT: 10.928 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:225 Resp: 42376
Ion Ratio Lower Upper
225 100
223 63.9 50.0 75.0
227 60.5 49.9 74.9



#34
Caprolactam
Concen: 3.401 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:113 Resp: 17521
Ion Ratio Lower Upper
113 100
55 150.9 150.2 225.4
56 113.2 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 4.023 ng/ul
RT: 11.916 min Scan# 1501
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:107 Resp: 65219
Ion Ratio Lower Upper
107 100
144 26.3 20.3 30.5
142 83.9 64.4 96.6



#36
2-Methylnaphthalene
Concen: 4.349 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:142 Resp: 161571
Ion Ratio Lower Upper
142 100
141 87.6 70.7 106.1





#37

1-Methylnaphthalene

Concen: 4.315 ng/ul

RT: 12.486 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12

Instrument :

BNA_M

ClientSampleId :

SSTD005012





Tgt Ion:142 Resp: 161573

Ion Ratio Lower Upper

142 100

141 96.0 75.2 112.8

116 3.8 2.6 3.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.276 ng/ul

RT: 12.633 min Scan# 1623

Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12

Tgt Ion:216 Resp: 76907




Ion Ratio Lower Upper


216 100

214 75.2 62.4 93.6

179 20.5 17.3 25.9

108 21.5 14.0 21.0#









#40
Hexachlorocyclopentadiene
Concen: 2.536 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:237	Resp:	30376
Ion Ratio	Lower	Upper
237	100	
235	59.9	48.4 72.6
272	12.2	11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 3.902 ng/ul
RT: 12.892 min Scan# 1667
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:196	Resp:	44904
Ion Ratio	Lower	Upper
196	100	
198	94.2	77.8 116.6
200	30.6	25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 3.880 ng/ul
RT: 12.969 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:196 Resp: 47646
Ion Ratio Lower Upper
196 100
198 98.0 76.8 115.2
200 29.9 24.9 37.3



#43
1,1'-Biphenyl
Concen: 4.345 ng/ul
RT: 13.280 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:154 Resp: 211292
Ion Ratio Lower Upper
154 100
153 40.9 32.4 48.6
76 14.3 10.1 15.1









#44
2-Chloronaphthalene
Concen: 4.266 ng/ul
RT: 13.322 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:162 Resp: 159192
Ion Ratio Lower Upper
162 100
164 33.1 25.4 38.0
127 40.6 31.6 47.4



#45
2-Nitroaniline
Concen: 3.632 ng/ul
RT: 13.545 min Scan# 1778
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion: 65 Resp: 37370
Ion Ratio Lower Upper
65 100
92 72.8 54.5 81.7
138 104.3 91.2 136.8









#47
Dimethylphthalate
Concen: 4.237 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:163 Resp: 183402
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.2 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 3.328 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:165 Resp: 28607
Ion Ratio Lower Upper
165 100
89 71.1 42.5 63.7#
121 27.3 17.2 25.8#









#50
Acenaphthylene
Concen: 4.173 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:152 Resp: 253487
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.6 10.6 15.8



#51
3-Nitroaniline
Concen: 2.951 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:138 Resp: 28027
Ion Ratio Lower Upper
138 100
108 12.0 8.1 12.1
92 117.0 90.4 135.6









#52
Acenaphthene
Concen: 4.412 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:153 Resp: 173880
Ion Ratio Lower Upper
153 100
152 48.9 37.1 55.7
154 87.5 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 1.383 ng/ul
RT: 14.598 min Scan# 1957
Delta R.T. 0.012 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:184 Resp: 6750
Ion Ratio Lower Upper
184 100
63 80.7 54.2 81.4
154 69.0 49.3 73.9









#55
4-Nitrophenol
Concen: 2.143 ng/ul
RT: 14.710 min Scan# 1976
Delta R.T. 0.018 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:109 Resp: 14483
Ion Ratio Lower Upper
109 100
139 131.0 124.9 187.3
65 133.7 122.2 183.4



#56
Dibenzofuran
Concen: 4.369 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:168 Resp: 238230
Ion Ratio Lower Upper
168 100
139 40.0 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 3.259 ng/ul
RT: 14.827 min Scan# 1995
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:165 Resp: 39482
Ion Ratio Lower Upper
165 100
63 55.1 35.9 53.9#
182 2.7 2.5 3.7



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.626 ng/ul
RT: 15.080 min Scan# 2039
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:232 Resp: 34984
Ion Ratio Lower Upper
232 100
131 49.8 36.7 55.1
130 2.9 1.7 2.5#
166 31.4 24.4 36.6









#59
Diethylphthalate
Concen: 4.026 ng/ul
RT: 15.268 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:149 Resp: 176291
Ion Ratio Lower Upper
149 100
177 20.4 16.8 25.2
150 12.6 10.2 15.4



#61
Fluorene
Concen: 4.211 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:166 Resp: 187567
Ion Ratio Lower Upper
166 100
165 95.5 77.2 115.8
167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 4.151 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:204 Resp: 88581
Ion Ratio Lower Upper
204 100
206 32.2 25.9 38.9
141 61.2 50.6 75.8



#63
4-Nitroaniline
Concen: 1.996 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:138 Resp: 17450
Ion Ratio Lower Upper
138 100
92 58.4 39.2 58.8
108 81.0 51.6 77.4#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 2.615 ng/ul
 RT: 15.592 min Scan# 2126
 Delta R.T. 0.006 min
 Lab File: BM036295.D
 Acq: 16 Aug 2022 11:12




Instrument :
 BNA_M
 ClientSampleId :
 SST005012


Tgt Ion:198 Resp: 16057
 Ion Ratio Lower Upper
 198 100
 182 14.5 4.2 6.4#
 77 43.6 19.9 29.9#



#67
 N-Nitrosodiphenylamine
 Concen: 4.331 ng/ul
 RT: 15.704 min Scan# 2145
 Delta R.T. -0.006 min
 Lab File: BM036295.D
 Acq: 16 Aug 2022 11:12

Tgt Ion:169 Resp: 145338
 Ion Ratio Lower Upper
 169 100
 168 68.1 53.8 80.6
 167 38.9 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 4.257 ng/ul
RT: 16.386 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:248 Resp: 47635
Ion Ratio Lower Upper
248 100
250 95.9 77.3 115.9
141 69.4 59.8 89.8



#69
Hexachlorobenzene
Concen: 4.300 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:284 Resp: 58635
Ion Ratio Lower Upper
284 100
142 39.9 28.6 42.8
249 32.7 24.6 37.0





#70

Atrazine

Concen: 3.639 ng/ul

RT: 16.662 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12

Instrument :

BNA_M

ClientSampleId :

SSTD005012





Tgt Ion:200 Resp: 42359

Ion Ratio Lower Upper

200 100

173 25.8 21.3 31.9

215 49.5 39.0 58.6



#71

Pentachlorophenol

Concen: 2.084 ng/ul

RT: 16.857 min Scan# 2341

Delta R.T. 0.006 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12




Tgt Ion:266 Resp: 17359


Ion Ratio Lower Upper

266 100

264 60.1 49.5 74.3

268 67.4 51.0 76.4









#72
Phenanthrene
Concen: 4.490 ng/ul
RT: 17.233 min Scan# 2405
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:178 Resp: 274701
Ion Ratio Lower Upper
178 100
179 14.7 12.2 18.2
176 19.2 14.9 22.3



#74
Anthracene
Concen: 4.305 ng/ul
RT: 17.321 min Scan# 2420
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:178 Resp: 264307
Ion Ratio Lower Upper
178 100
179 14.9 12.2 18.2
176 18.9 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.454 ng/uL
RT: 13.245 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:	216	Resp:	79058
Ion Ratio	Lower	Upper	
216	100		
214	77.9	60.9	91.3
218	48.1	37.5	56.3



#76
Pentachlorobenzene
Concen: 4.624 ng/uL
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:	250	Resp:	74407
Ion Ratio	Lower	Upper	
250	100		
248	62.1	50.4	75.6
252	63.6	51.1	76.7





#77

Carbazole

Concen: 4.025 ng/ul

RT: 17.604 min Scan# 2468

Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12

Instrument :

BNA_M

ClientSampleId :

SSTD005012





Tgt Ion:167 Resp: 228231

Ion Ratio Lower Upper

167 100

166 23.0 17.4 26.0

139 13.5 10.1 15.1



#78

Di-n-butylphthalate

Concen: 4.146 ng/ul

RT: 18.162 min Scan# 2563

Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12




Tgt Ion:149 Resp: 265216


Ion Ratio Lower Upper

149 100

150 9.1 7.3 10.9

104 5.0 3.7 5.5









#80
Fluoranthene
Concen: 4.423 ng/ul
RT: 19.251 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:202 Resp: 268466
Ion Ratio Lower Upper
202 100
101 11.8 10.1 15.1
100 8.9 7.8 11.6



#82
Pyrene
Concen: 4.522 ng/ul
RT: 19.609 min Scan# 2809
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:202 Resp: 282924
Ion Ratio Lower Upper
202 100
101 13.7 11.9 17.9
100 11.7 10.2 15.4









#83
Butylbenzylphthalate
Concen: 4.133 ng/ul
RT: 20.515 min Scan# 2963
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:149 Resp: 94057
Ion Ratio Lower Upper
149 100
91 75.8 53.2 79.8
206 18.0 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 4.326 ng/ul
RT: 21.303 min Scan# 3097
Delta R.T. 0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:252 Resp: 73190
Ion Ratio Lower Upper
252 100
254 65.9 52.4 78.6
126 12.2 10.8 16.2









#85
Benzo(a)anthracene
Concen: 4.221 ng/ul
RT: 21.356 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:228 Resp: 243561
Ion Ratio Lower Upper
228 100
229 19.9 15.6 23.4
226 27.2 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 3.672 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:149 Resp: 120754
Ion Ratio Lower Upper
149 100
167 25.8 21.5 32.3
279 3.9 3.4 5.2





#87

Chrysene

Concen: 4.263 ng/ul

RT: 21.409 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12

Instrument :

BNA_M

ClientSampleId :

SSTD005012





Tgt Ion:228 Resp: 240003

Ion Ratio Lower Upper

228 100

226 29.9 23.7 35.5

229 19.2 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 3.480 ng/ul




RT: 22.192 min Scan# 3248


Delta R.T. 0.000 min

Lab File: BM036295.D

Acq: 16 Aug 2022 11:12

Tgt Ion:149 Resp: 207245









#90
Benzo(b)fluoranthene
Concen: 3.894 ng/ul
RT: 22.991 min Scan# 3384
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:252 Resp: 239627
Ion Ratio Lower Upper
252 100
253 22.6 17.5 26.3
125 12.7 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 4.017 ng/ul
RT: 23.039 min Scan# 3392
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:252 Resp: 234554
Ion Ratio Lower Upper
252 100
253 21.5 17.4 26.0
125 13.0 10.6 15.8









#93
Benzo(a)pyrene
Concen: 4.032 ng/ul
RT: 23.597 min Scan# 3488
Delta R.T. -0.006 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12




Instrument :
BNA_M
ClientSampleId :
SSTD005012


Tgt Ion:252 Resp: 236320
Ion Ratio Lower Upper
252 100
253 21.8 17.3 25.9
125 14.2 11.8 17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.534 ng/ul
RT: 26.115 min Scan# 3915
Delta R.T. 0.000 min
Lab File: BM036295.D
Acq: 16 Aug 2022 11:12

Tgt Ion:276 Resp: 283307
Ion Ratio Lower Upper
276 100
138 28.1 25.6 38.4
277 25.6 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 4.529 ng/ul
 RT: 26.132 min Scan# 3918
 Delta R.T. 0.000 min
 Lab File: BM036295.D
 Acq: 16 Aug 2022 11:12

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005012

Tgt Ion:278 Resp: 242788
 Ion Ratio Lower Upper
 278 100
 139 18.5 16.2 24.2
 279 24.9 18.7 28.1



#96
 Benzo(g,h,i)perylene
 Concen: 4.691 ng/ul
 RT: 26.868 min Scan# 4043
 Delta R.T. 0.006 min
 Lab File: BM036295.D
 Acq: 16 Aug 2022 11:12

Tgt Ion:276 Resp: 242902
 Ion Ratio Lower Upper
 276 100
 138 23.5 21.3 31.9
 277 23.2 18.6 28.0

