

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036296.D
 Acq On : 16 Aug 2022 11:49
 Operator : CG/JU
 Sample : SSTD01061
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010013

Quant Time: Aug 16 14:08:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	209533	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	888778	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	496756	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	921054	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	828613	20.000	ng/u1	0.00
88) Perylene-d12	23.703	264	865626	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	22479	4.037	ng/uL	0.00
4) Pyridine-d5	3.640	84	146754	9.620	ng/u1	0.00
7) Phenol-d5	6.987	99	169208	9.630	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.139	67	112011	9.562	ng/u1	0.00
11) 2-Chlorophenol-d4	7.339	132	138105	9.469	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	127756	8.975	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	64728	9.191	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	63412	9.242	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.239	165	118432	9.543	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	179581	8.548	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	311267	9.199	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	382261	8.854	ng/u1	0.00
54) 4-Nitrophenol-d4	14.680	143	43359	6.240	ng/u1	0.00
60) Fluorene-d10	15.439	176	281630	9.210	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.568	200	34550	6.978	ng/u1	0.00
73) Anthracene-d10	17.286	188	407034	9.775	ng/u1	0.00
81) Pyrene-d10	19.580	212	433065	9.430	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.550	264	419756	9.551	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	23493	4.071	ng/uL	97
5) Pyridine	3.663	79	150940	9.429	ng/u1	92
6) Benzaldehyde	6.951	77	79460	10.775	ng/u1	93
8) Phenol	7.016	94	178005	9.257	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.234	93	148703	9.619	ng/u1	99
12) 2-Chlorophenol	7.369	128	146872	9.816	ng/u1	98
13) 2-Methylphenol	8.263	108	129896	9.042	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.339	45	199623	9.783	ng/u1	96
16) Acetophenone	8.633	105	211770	9.266	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.616	70	104640	9.181	ng/u1	95
18) 4-Methylphenol	8.592	108	142410	9.241	ng/u1	100
19) Hexachloroethane	8.875	117	62666	10.151	ng/u1	89
22) Nitrobenzene	9.016	77	158679	9.846	ng/u1	97
23) Isophorone	9.539	82	276654	9.094	ng/u1	96
25) 2-Nitrophenol	9.728	139	72499	9.524	ng/u1	98
26) 2,4-Dimethylphenol	9.792	107	150337	9.665	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.028	93	184694	9.626	ng/u1	99
29) 2,4-Dichlorophenol	10.263	162	121917	9.465	ng/u1	98
30) Naphthalene	10.657	128	474760	10.120	ng/u1	99
32) 4-Chloroaniline	10.780	127	183364	8.786	ng/u1	98
33) Hexachlorobutadiene	10.933	225	79550	10.004	ng/u1	97
34) Caprolactam	11.574	113	30582	7.184	ng/u1	85
35) 4-Chloro-3-methylphenol	11.916	107	122413	9.140	ng/u1	94
36) 2-Methylnaphthalene	12.269	142	289936	9.446	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036296.D
 Acq On : 16 Aug 2022 11:49
 Operator : CG/JU
 Sample : SSTD01061
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD010013


Quant Time: Aug 16 14:08:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.486	142	297099	9.603	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.633	216	138818	9.817	ng/ul	98
40) Hexachlorocyclopentadiene	12.610	237	64091	6.805	ng/ul	98
41) 2,4,6-Trichlorophenol	12.886	196	82328	9.098	ng/ul	97
42) 2,4,5-Trichlorophenol	12.969	196	88342	9.149	ng/ul	98
43) 1,1'-Biphenyl	13.280	154	379660	9.930	ng/ul	99
44) 2-Chloronaphthalene	13.321	162	291417	9.931	ng/ul	99
45) 2-Nitroaniline	13.545	65	70458	8.710	ng/ul	96
47) Dimethylphthalate	13.910	163	322020	9.461	ng/ul	99
48) 2,6-Dinitrotoluene	14.039	165	55209	8.170	ng/ul#	84
50) Acenaphthylene	14.168	152	460832	9.649	ng/ul	99
51) 3-Nitroaniline	14.374	138	55345	7.411	ng/ul	100
52) Acenaphthene	14.510	153	303411	9.792	ng/ul	98
53) 2,4-Dinitrophenol	14.586	184	17605	4.588	ng/ul	91
55) 4-Nitrophenol	14.692	109	34826	6.553	ng/ul	89
56) Dibenzofuran	14.845	168	426457	9.948	ng/ul	98
57) 2,4-Dinitrotoluene	14.827	165	78589	8.250	ng/ul#	92
58) 2,3,4,6-Tetrachlorophenol	15.080	232	65948	8.693	ng/ul#	90
59) Diethylphthalate	15.274	149	312876	9.088	ng/ul	99
61) Fluorene	15.492	166	335165	9.571	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.492	204	155946	9.294	ng/ul	98
63) 4-Nitroaniline	15.668	138	111	0.016	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.586	198	34777	7.020	ng/ul#	78
67) N-Nitrosodiphenylamine	15.710	169	268808	9.927	ng/ul	99
68) 4-Bromophenyl-phenylether	16.386	248	84503	9.358	ng/ul	98
69) Hexachlorobenzene	16.486	284	102831	9.347	ng/ul	96
70) Atrazine	16.662	200	82082	8.740	ng/ul	97
71) Pentachlorophenol	16.851	266	43134	6.418	ng/ul	96
72) Phenanthrene	17.233	178	487760	9.881	ng/ul	98
74) Anthracene	17.321	178	489201	9.876	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	140601	9.816	ng/ul	94
76) Pentachlorobenzene	14.763	250	131093	10.097	ng/ul	99
77) Carbazole	17.604	167	433407	9.473	ng/ul	99
78) Di-n-butylphthalate	18.162	149	483345	9.364	ng/ul	99
80) Fluoranthene	19.250	202	523452	9.540	ng/ul	98
82) Pyrene	19.609	202	556050	9.832	ng/ul	98
83) Butylbenzylphthalate	20.515	149	192462	9.357	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.297	252	162840	10.650	ng/ul	94
85) Benzo(a)anthracene	21.356	228	500896	9.604	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.291	149	271315	9.129	ng/ul#	97
87) Chrysene	21.409	228	495725	9.742	ng/ul	100
89) Di-n-octyl phthalate	22.191	149	456816	8.172	ng/ul	100
90) Benzo(b)fluoranthene	22.991	252	518256	8.971	ng/ul	98
91) Benzo(k)fluoranthene	23.038	252	495191	9.033	ng/ul	98
93) Benzo(a)pyrene	23.597	252	513524	9.332	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.115	276	603864	10.296	ng/ul	95
95) Dibenzo(a,h)anthracene	26.132	278	523197	10.398	ng/ul	97
96) Benzo(g,h,i)perylene	26.862	276	520436	10.708	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 16 14:08:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 16 14:05:33 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 4.071 ng/uL
 RT: 3.246 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BM036296.D
 Acq: 16 Aug 2022 11:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010013


Tgt Ion: 88 Resp: 23493
 Ion Ratio Lower Upper
 88 100
 43 36.2 27.8 41.8
 58 80.1 61.8 92.8



#5
 Pyridine
 Concen: 9.429 ng/uL
 RT: 3.663 min Scan# 98
 Delta R.T. 0.006 min
 Lab File: BM036296.D
 Acq: 16 Aug 2022 11:49

Tgt Ion: 79 Resp: 150940
 Ion Ratio Lower Upper
 79 100
 52 75.0 67.8 101.6
 51 37.2 31.4 47.0









#6
Benzaldehyde
Concen: 10.775 ng/ul
RT: 6.951 min Scan# 657
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion: 77 Resp: 79460
Ion Ratio Lower Upper
77 100
105 97.3 84.6 127.0
106 94.0 80.3 120.5



#8
Phenol
Concen: 9.257 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion: 94 Resp: 178005
Ion Ratio Lower Upper
94 100
65 27.1 21.3 31.9
66 35.3 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 9.619 ng/ul
RT: 7.234 min Scan# 705
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion: 93 Resp: 148703
Ion Ratio Lower Upper
93 100
63 78.8 64.0 96.0
95 32.2 25.8 38.6



#12
2-Chlorophenol
Concen: 9.816 ng/ul
RT: 7.369 min Scan# 728
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:128 Resp: 146872
Ion Ratio Lower Upper
128 100
64 44.8 36.9 55.3
130 31.5 25.6 38.4





#13

2-Methylphenol

Concen: 9.042 ng/ul

RT: 8.263 min Scan# 880

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :





SSTD010013

Tgt Ion:108 Resp: 129896

Ion Ratio Lower Upper

108 100

107 91.2 75.0 112.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.783 ng/ul

RT: 8.339 min Scan# 893

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49




Tgt Ion: 45 Resp: 199623


Ion Ratio Lower Upper

45 100

77 15.3 10.6 15.8

79 11.4 8.3 12.5









#16
Acetophenone
Concen: 9.266 ng/ul
RT: 8.633 min Scan# 943
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:105 Resp: 211770
Ion Ratio Lower Upper
105 100
77 79.7 59.8 89.6
51 28.4 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 9.181 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. -0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion: 70 Resp: 104640
Ion Ratio Lower Upper
70 100
42 58.9 49.6 74.4
101 9.5 8.1 12.1
130 19.5 18.4 27.6









#18
4-Methylphenol
Concen: 9.241 ng/ul
RT: 8.592 min Scan# 911
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:108 Resp: 142410
Ion Ratio Lower Upper
108 100
107 112.5 90.3 135.5



#19
Hexachloroethane
Concen: 10.151 ng/ul
RT: 8.875 min Scan# 984
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:117 Resp: 62666
Ion Ratio Lower Upper
117 100
201 89.0 81.8 122.6
199 59.0 51.6 77.4









#22
Nitrobenzene
Concen: 9.846 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion: 77 Resp: 158679
Ion Ratio Lower Upper
77 100
123 44.8 37.4 56.0
65 12.3 10.8 16.2



#23
Isophorone
Concen: 9.094 ng/ul
RT: 9.539 min Scan# 1097
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion: 82 Resp: 276654
Ion Ratio Lower Upper
82 100
95 6.5 5.5 8.3
138 16.8 15.3 22.9









#25
2-Nitrophenol
Concen: 9.524 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:139 Resp: 72499
Ion Ratio Lower Upper
139 100
65 47.3 39.8 59.6
109 27.9 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 9.665 ng/ul
RT: 9.792 min Scan# 1140
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:107 Resp: 150337
Ion Ratio Lower Upper
107 100
121 52.0 42.7 64.1
122 89.6 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 9.626 ng/ul

RT: 10.028 min Scan# 1180

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion: 93 Resp: 184694

Ion Ratio Lower Upper

93 100

95 31.9 26.1 39.1

123 11.5 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 9.465 ng/ul

RT: 10.263 min Scan# 1220

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49




Tgt Ion:162 Resp: 121917


Ion Ratio Lower Upper

162 100

164 66.6 51.8 77.6

98 36.2 29.5 44.3









#30
Naphthalene
Concen: 10.120 ng/ul
RT: 10.657 min Scan# 1286
Delta R.T. 0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:128 Resp: 474760
Ion Ratio Lower Upper
128 100
129 10.6 8.8 13.2
127 12.9 10.6 16.0



#32
4-Chloroaniline
Concen: 8.786 ng/ul
RT: 10.780 min Scan# 1308
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:127 Resp: 183364
Ion Ratio Lower Upper
127 100
129 32.9 25.3 37.9









#33
Hexachlorobutadiene
Concen: 10.004 ng/ul
RT: 10.933 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:225 Resp: 79550
Ion Ratio Lower Upper
225 100
223 63.2 50.0 75.0
227 66.2 49.9 74.9



#34
Caprolactam
Concen: 7.184 ng/ul
RT: 11.574 min Scan# 1443
Delta R.T. 0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:113 Resp: 30582
Ion Ratio Lower Upper
113 100
55 163.9 150.2 225.4
56 112.7 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 9.140 ng/ul
RT: 11.916 min Scan# 1501
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:107 Resp: 122413
Ion Ratio Lower Upper
107 100
144 25.9 20.3 30.5
142 86.8 64.4 96.6



#36
2-Methylnaphthalene
Concen: 9.446 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:142 Resp: 289936
Ion Ratio Lower Upper
142 100
141 88.4 70.7 106.1





#37

1-Methylnaphthalene

Concen: 9.603 ng/ul

RT: 12.486 min Scan# 1598

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:142 Resp: 297099

Ion Ratio Lower Upper

142 100

141 90.5 75.2 112.8

116 4.0 2.6 3.8#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.817 ng/ul

RT: 12.633 min Scan# 1623

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Tgt Ion:216 Resp: 138818




Ion Ratio Lower Upper


216 100

214 76.1 62.4 93.6

179 21.5 17.3 25.9

108 19.7 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 6.805 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:237 Resp: 64091
Ion Ratio Lower Upper
237 100
235 61.5 48.4 72.6
272 13.3 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 9.098 ng/ul
RT: 12.886 min Scan# 1666
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:196 Resp: 82328
Ion Ratio Lower Upper
196 100
198 94.0 77.8 116.6
200 29.9 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 9.149 ng/ul
RT: 12.969 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:196 Resp: 88342
Ion Ratio Lower Upper
196 100
198 98.8 76.8 115.2
200 31.8 24.9 37.3



#43
1,1'-Biphenyl
Concen: 9.930 ng/ul
RT: 13.280 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:154 Resp: 379660
Ion Ratio Lower Upper
154 100
153 40.2 32.4 48.6
76 13.7 10.1 15.1









#44
2-Chloronaphthalene
Concen: 9.931 ng/ul
RT: 13.321 min Scan# 1740
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:162 Resp: 291417
Ion Ratio Lower Upper
162 100
164 32.8 25.4 38.0
127 40.2 31.6 47.4



#45
2-Nitroaniline
Concen: 8.710 ng/ul
RT: 13.545 min Scan# 1778
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion: 65 Resp: 70458
Ion Ratio Lower Upper
65 100
92 68.3 54.5 81.7
138 106.5 91.2 136.8









#47
Dimethylphthalate
Concen: 9.461 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:163 Resp: 322020
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 8.170 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:165 Resp: 55209
Ion Ratio Lower Upper
165 100
89 67.7 42.5 63.7#
121 24.5 17.2 25.8









#50
Acenaphthylene
Concen: 9.649 ng/ul
RT: 14.168 min Scan# 1884
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:152 Resp: 460832
Ion Ratio Lower Upper
152 100
151 20.6 16.1 24.1
153 13.3 10.6 15.8



#51
3-Nitroaniline
Concen: 7.411 ng/ul
RT: 14.374 min Scan# 1919
Delta R.T. 0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:138 Resp: 55345
Ion Ratio Lower Upper
138 100
108 10.7 8.1 12.1
92 112.8 90.4 135.6









#52
Acenaphthene
Concen: 9.792 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:153 Resp: 303411
Ion Ratio Lower Upper
153 100
152 48.3 37.1 55.7
154 87.4 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 4.588 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:184 Resp: 17605
Ion Ratio Lower Upper
184 100
63 79.6 54.2 81.4
154 63.8 49.3 73.9









#55
4-Nitrophenol
Concen: 6.553 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:109 Resp: 34826
Ion Ratio Lower Upper
109 100
139 139.7 124.9 187.3
65 140.9 122.2 183.4



#56
Dibenzofuran
Concen: 9.948 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:168 Resp: 426457
Ion Ratio Lower Upper
168 100
139 38.7 30.1 45.1





#57

2,4-Dinitrotoluene

Concen: 8.250 ng/ul

RT: 14.827 min Scan# 1995

Delta R.T. 0.006 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:165 Resp: 78589

Ion Ratio Lower Upper

165 100

63 50.6 35.9 53.9

182 2.3 2.5 3.7#



#58

2,3,4,6-Tetrachlorophenol

Concen: 8.693 ng/ul

RT: 15.080 min Scan# 2039

Delta R.T. 0.006 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Tgt Ion:232 Resp: 65948




Ion Ratio Lower Upper


232 100

131 54.2 36.7 55.1

130 3.1 1.7 2.5#

166 33.5 24.4 36.6









#59
Diethylphthalate
Concen: 9.088 ng/ul
RT: 15.274 min Scan# 2072
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:149 Resp: 312876
Ion Ratio Lower Upper
149 100
177 20.9 16.8 25.2
150 12.1 10.2 15.4



#61
Fluorene
Concen: 9.571 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:166 Resp: 335165
Ion Ratio Lower Upper
166 100
165 99.6 77.2 115.8
167 13.3 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 9.294 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:204 Resp: 155946
Ion Ratio Lower Upper
204 100
206 31.6 25.9 38.9
141 60.9 50.6 75.8



#63
4-Nitroaniline
Concen: 0.016 ng/ul
RT: 15.668 min Scan# 2139
Delta R.T. 0.135 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:138 Resp: 111
Ion Ratio Lower Upper
138 100
92 50.8 39.2 58.8
108 62.1 51.6 77.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 7.020 ng/ul
 RT: 15.586 min Scan# 2125
 Delta R.T. -0.000 min
 Lab File: BM036296.D
 Acq: 16 Aug 2022 11:49




Instrument :
 BNA_M
 ClientSampleId :
 SST010013


Tgt Ion:198 Resp: 34777
 Ion Ratio Lower Upper
 198 100
 182 7.7 4.2 6.4#
 77 37.2 19.9 29.9#



#67
 N-Nitrosodiphenylamine
 Concen: 9.927 ng/ul
 RT: 15.710 min Scan# 2146
 Delta R.T. -0.000 min
 Lab File: BM036296.D
 Acq: 16 Aug 2022 11:49

Tgt Ion:169 Resp: 268808
 Ion Ratio Lower Upper
 169 100
 168 67.8 53.8 80.6
 167 37.5 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 9.358 ng/ul
RT: 16.386 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:248 Resp: 84503
Ion Ratio Lower Upper
248 100
250 97.5 77.3 115.9
141 78.4 59.8 89.8



#69
Hexachlorobenzene
Concen: 9.347 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. -0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:284 Resp: 102831
Ion Ratio Lower Upper
284 100
142 38.7 28.6 42.8
249 29.8 24.6 37.0





#70

Atrazine

Concen: 8.740 ng/ul

RT: 16.662 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:200 Resp: 82082

Ion Ratio Lower Upper

200 100

173 26.6 21.3 31.9

215 46.0 39.0 58.6



#71

Pentachlorophenol

Concen: 6.418 ng/ul

RT: 16.851 min Scan# 2340

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49




Tgt Ion:266 Resp: 43134


Ion Ratio Lower Upper

266 100

264 60.8 49.5 74.3

268 58.7 51.0 76.4









#72
 Phenanthrene
 Concen: 9.881 ng/ul
 RT: 17.233 min Scan# 2405
 Delta R.T. -0.000 min
 Lab File: BM036296.D
 Acq: 16 Aug 2022 11:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010013


Tgt Ion:178 Resp: 487760
 Ion Ratio Lower Upper
 178 100
 179 15.6 12.2 18.2
 176 19.8 14.9 22.3



#74
 Anthracene
 Concen: 9.876 ng/ul
 RT: 17.321 min Scan# 2420
 Delta R.T. -0.000 min
 Lab File: BM036296.D
 Acq: 16 Aug 2022 11:49

Tgt Ion:178 Resp: 489201
 Ion Ratio Lower Upper
 178 100
 179 15.2 12.2 18.2
 176 19.0 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.816 ng/uL

RT: 13.245 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:216 Resp: 140601

Ion Ratio Lower Upper

216 100

214 82.5 60.9 91.3

218 49.0 37.5 56.3



#76

Pentachlorobenzene

Concen: 10.097 ng/uL

RT: 14.763 min Scan# 1985

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49




Tgt Ion:250 Resp: 131093


Ion Ratio Lower Upper

250 100

248 64.6 50.4 75.6

252 64.5 51.1 76.7





#77

Carbazole

Concen: 9.473 ng/ul

RT: 17.604 min Scan# 2468

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:167 Resp: 433407

Ion Ratio Lower Upper

167 100

166 21.6 17.4 26.0

139 13.7 10.1 15.1



#78

Di-n-butylphthalate

Concen: 9.364 ng/ul

RT: 18.162 min Scan# 2563

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49




Tgt Ion:149 Resp: 483345


Ion Ratio Lower Upper

149 100

150 9.0 7.3 10.9

104 5.4 3.7 5.5









#80
Fluoranthene
Concen: 9.540 ng/ul
RT: 19.250 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:202 Resp: 523452
Ion Ratio Lower Upper
202 100
101 11.9 10.1 15.1
100 8.9 7.8 11.6



#82
Pyrene
Concen: 9.832 ng/ul
RT: 19.609 min Scan# 2809
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:202 Resp: 556050
Ion Ratio Lower Upper
202 100
101 14.6 11.9 17.9
100 11.7 10.2 15.4





#83

Butylbenzylphthalate

Concen: 9.357 ng/ul

RT: 20.515 min Scan# 2963

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:149 Resp: 192462

Ion Ratio Lower Upper

149 100

91 69.8 53.2 79.8

206 18.6 16.0 24.0



#84

3,3'-Dichlorobenzidine

Concen: 10.650 ng/ul

RT: 21.297 min Scan# 3096

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49




Tgt Ion:252 Resp: 162840


Ion Ratio Lower Upper

252 100

254 60.0 52.4 78.6

126 11.8 10.8 16.2





#85

Benzo(a)anthracene

Concen: 9.604 ng/ul

RT: 21.356 min Scan# 3106

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:228 Resp: 500896

Ion Ratio Lower Upper

228 100

229 19.7 15.6 23.4

226 27.1 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.129 ng/ul

RT: 21.291 min Scan# 3095

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49




Tgt Ion:149 Resp: 271315


Ion Ratio Lower Upper

149 100

167 25.6 21.5 32.3

279 3.4 3.4 5.2#





#87

Chrysene

Concen: 9.742 ng/ul

RT: 21.409 min Scan# 3115

Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Instrument :

BNA_M

ClientSampleId :

SSTD010013





Tgt Ion:228 Resp: 495725

Ion Ratio Lower Upper

228 100

226 29.2 23.7 35.5

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 8.172 ng/ul




RT: 22.191 min Scan# 3248


Delta R.T. -0.000 min

Lab File: BM036296.D

Acq: 16 Aug 2022 11:49

Tgt Ion:149 Resp: 456816









#90
Benzo(b)fluoranthene
Concen: 8.971 ng/ul
RT: 22.991 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:252 Resp: 518256
Ion Ratio Lower Upper
252 100
253 21.3 17.5 26.3
125 12.1 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 9.033 ng/ul
RT: 23.038 min Scan# 3392
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:252 Resp: 495191
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 11.6 10.6 15.8









#93
Benzo(a)pyrene
Concen: 9.332 ng/ul
RT: 23.597 min Scan# 3488
Delta R.T. -0.006 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49




Instrument :
BNA_M
ClientSampleId :
SSTD010013


Tgt Ion:252 Resp: 513524
Ion Ratio Lower Upper
252 100
253 22.4 17.3 25.9
125 13.5 11.8 17.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 10.296 ng/ul
RT: 26.115 min Scan# 3915
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:276 Resp: 603864
Ion Ratio Lower Upper
276 100
138 27.9 25.6 38.4
277 24.1 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 10.398 ng/ul
RT: 26.132 min Scan# 3918
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Instrument :
BNA_M
ClientSampleId :
SSTD010013

Tgt Ion:278 Resp: 523197

Ion	Ratio	Lower	Upper
278	100		
139	16.9	16.2	24.2
279	23.4	18.7	28.1



#96
Benzo(g,h,i)perylene
Concen: 10.708 ng/ul
RT: 26.862 min Scan# 4042
Delta R.T. -0.000 min
Lab File: BM036296.D
Acq: 16 Aug 2022 11:49

Tgt Ion:276 Resp: 520436

Ion	Ratio	Lower	Upper
276	100		
138	24.1	21.3	31.9
277	23.6	18.6	28.0

