

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036297.D
 Acq On : 16 Aug 2022 12:26
 Operator : CG/JU
 Sample : SSTD02062
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020014

Quant Time: Aug 16 14:08:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	215491	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	909549	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	506559	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	959419	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	889556	20.000	ng/u1	0.00
88) Perylene-d12	23.703	264	913717	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	43379	7.575	ng/uL	0.00
4) Pyridine-d5	3.640	84	287153	18.303	ng/u1	0.00
7) Phenol-d5	6.987	99	338446	18.729	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	217343	18.041	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	276884	18.459	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	253268	17.300	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	131287	18.216	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	131953	18.792	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.233	165	240367	18.926	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	357571	16.631	ng/u1	0.00
46) Dimethylphthalate-d6	13.868	166	603346	17.485	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	768004	17.444	ng/u1	0.00
54) 4-Nitrophenol-d4	14.680	143	103260	14.573	ng/u1	0.00
60) Fluorene-d10	15.439	176	554497	17.783	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.568	200	79924	15.498	ng/u1	0.00
73) Anthracene-d10	17.286	188	810448	18.684	ng/u1	0.00
81) Pyrene-d10	19.580	212	869204	17.631	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.550	264	856715	18.468	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	46480	7.832	ng/uL	95
5) Pyridine	3.657	79	302829	18.394	ng/u1	89
6) Benzaldehyde	6.951	77	155589	20.514	ng/u1	94
8) Phenol	7.016	94	355959	18.000	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.234	93	289074	18.182	ng/u1	96
12) 2-Chlorophenol	7.369	128	290622	18.887	ng/u1	99
13) 2-Methylphenol	8.263	108	260899	17.660	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.339	45	383516	18.275	ng/u1	98
16) Acetophenone	8.634	105	413818	17.606	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.622	70	210516	17.960	ng/u1	94
18) 4-Methylphenol	8.592	108	280579	17.703	ng/u1	99
19) Hexachloroethane	8.875	117	120771	19.023	ng/u1	87
22) Nitrobenzene	9.016	77	315716	19.142	ng/u1	98
23) Isophorone	9.539	82	547482	17.585	ng/u1	97
25) 2-Nitrophenol	9.728	139	147719	18.962	ng/u1	98
26) 2,4-Dimethylphenol	9.792	107	298289	18.739	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.028	93	361538	18.412	ng/u1	99
29) 2,4-Dichlorophenol	10.263	162	244612	18.557	ng/u1	98
30) Naphthalene	10.651	128	917536	19.112	ng/u1	99
32) 4-Chloroaniline	10.780	127	365766	17.126	ng/u1	98
33) Hexachlorobutadiene	10.933	225	152150	18.697	ng/u1	96
34) Caprolactam	11.569	113	66161	15.187	ng/u1	93
35) 4-Chloro-3-methylphenol	11.916	107	245905	17.941	ng/u1	99
36) 2-Methylnaphthalene	12.269	142	568349	18.095	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036297.D
 Acq On : 16 Aug 2022 12:26
 Operator : CG/JU
 Sample : SSTD02062
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020014

Quant Time: Aug 16 14:08:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.486	142	574120	18.134	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.633	216	266731	18.498	ng/ul	99
40) Hexachlorocyclopentadiene	12.610	237	137906	14.360	ng/ul	97
41) 2,4,6-Trichlorophenol	12.886	196	167320	18.132	ng/ul	98
42) 2,4,5-Trichlorophenol	12.963	196	177299	18.007	ng/ul	98
43) 1,1'-Biphenyl	13.286	154	729905	18.720	ng/ul	100
44) 2-Chloronaphthalene	13.321	162	562670	18.804	ng/ul	99
45) 2-Nitroaniline	13.545	65	152929	18.538	ng/ul	97
47) Dimethylphthalate	13.916	163	623462	17.963	ng/ul	100
48) 2,6-Dinitrotoluene	14.039	165	118084	17.136	ng/ul	90
50) Acenaphthylene	14.168	152	911194	18.710	ng/ul	99
51) 3-Nitroaniline	14.368	138	117709	15.456	ng/ul	96
52) Acenaphthene	14.510	153	587184	18.584	ng/ul	98
53) 2,4-Dinitrophenol	14.586	184	46574	11.904	ng/ul	95
55) 4-Nitrophenol	14.692	109	81173	14.978	ng/ul	85
56) Dibenzofuran	14.845	168	817709	18.705	ng/ul	99
57) 2,4-Dinitrotoluene	14.821	165	169790	17.480	ng/ul#	95
58) 2,3,4,6-Tetrachlorophenol	15.074	232	135037	17.455	ng/ul#	93
59) Diethylphthalate	15.274	149	612937	17.459	ng/ul	99
61) Fluorene	15.492	166	649882	18.198	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.492	204	299284	17.491	ng/ul	98
63) 4-Nitroaniline	15.533	138	102844	14.669	ng/ul#	89
66) 4,6-Dinitro-2-methylph...	15.586	198	82849	16.054	ng/ul	93
67) N-Nitrosodiphenylamine	15.710	169	523771	18.570	ng/ul	98
68) 4-Bromophenyl-phenylether	16.386	248	165558	17.602	ng/ul	99
69) Hexachlorobenzene	16.492	284	201412	17.575	ng/ul	96
70) Atrazine	16.662	200	168740	17.248	ng/ul	97
71) Pentachlorophenol	16.851	266	94349	13.476	ng/ul	100
72) Phenanthrene	17.233	178	959519	18.660	ng/ul	98
74) Anthracene	17.321	178	961735	18.639	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	275156	18.442	ng/ul	99
76) Pentachlorobenzene	14.763	250	253575	18.749	ng/ul	100
77) Carbazole	17.604	167	883142	18.531	ng/ul	99
78) Di-n-butylphthalate	18.162	149	958399	17.824	ng/ul	99
80) Fluoranthene	19.250	202	1050591	17.836	ng/ul	97
82) Pyrene	19.609	202	1096174	18.055	ng/ul	99
83) Butylbenzylphthalate	20.515	149	400765	18.149	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.297	252	347664	21.179	ng/ul	98
85) Benzo(a)anthracene	21.356	228	1019927	18.216	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.292	149	556913	17.454	ng/ul	99
87) Chrysene	21.409	228	1001880	18.339	ng/ul	100
89) Di-n-octyl phthalate	22.191	149	949942	16.099	ng/ul	100
90) Benzo(b)fluoranthene	22.997	252	1049130	17.205	ng/ul	98
91) Benzo(k)fluoranthene	23.038	252	1023985	17.696	ng/ul	98
93) Benzo(a)pyrene	23.603	252	1048730	18.056	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.115	276	1226482	19.811	ng/ul	94
95) Dibenzo(a,h)anthracene	26.132	278	1057966	19.918	ng/ul	97
96) Benzo(g,h,i)perylene	26.862	276	1051587	20.497	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.832 ng/uL
 RT: 3.246 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020014


Tgt Ion: 88 Resp: 46480
 Ion Ratio Lower Upper
 88 100
 43 30.7 27.8 41.8
 58 73.6 61.8 92.8



#5
 Pyridine
 Concen: 18.394 ng/uL
 RT: 3.657 min Scan# 97
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26

Tgt Ion: 79 Resp: 302829
 Ion Ratio Lower Upper
 79 100
 52 72.3 67.8 101.6
 51 35.6 31.4 47.0









#6
Benzaldehyde
Concen: 20.514 ng/ul
RT: 6.951 min Scan# 657
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion: 77 Resp: 155589
Ion Ratio Lower Upper
77 100
105 97.7 84.6 127.0
106 95.3 80.3 120.5



#8
Phenol
Concen: 18.000 ng/ul
RT: 7.016 min Scan# 668
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion: 94 Resp: 355959
Ion Ratio Lower Upper
94 100
65 27.3 21.3 31.9
66 33.4 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 18.182 ng/ul
RT: 7.234 min Scan# 705
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Instrument :
BNA_M
ClientSampleId :
SSTD020014

Tgt Ion: 93 Resp: 289074




Ion	Ratio	Lower	Upper
93	100		
63	75.6	64.0	96.0
95	32.0	25.8	38.6




#12
2-Chlorophenol
Concen: 18.887 ng/ul
RT: 7.369 min Scan# 728
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:128 Resp: 290622

Ion	Ratio	Lower	Upper
128	100		
64	45.5	36.9	55.3
130	31.9	25.6	38.4









#13
2-Methylphenol
Concen: 17.660 ng/ul
RT: 8.263 min Scan# 880
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:108 Resp: 260899
Ion Ratio Lower Upper
108 100
107 87.6 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.275 ng/ul
RT: 8.339 min Scan# 893
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion: 45 Resp: 383516
Ion Ratio Lower Upper
45 100
77 14.4 10.6 15.8
79 10.6 8.3 12.5









#16
Acetophenone
Concen: 17.606 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:105 Resp: 413818
Ion Ratio Lower Upper
105 100
77 78.7 59.8 89.6
51 27.1 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 17.960 ng/ul
RT: 8.622 min Scan# 941
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion: 70 Resp: 210516
Ion Ratio Lower Upper
70 100
42 56.7 49.6 74.4
101 9.1 8.1 12.1
130 20.6 18.4 27.6









#18
4-Methylphenol
Concen: 17.703 ng/ul
RT: 8.592 min Scan# 936
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:108 Resp: 280579
Ion Ratio Lower Upper
108 100
107 114.1 90.3 135.5



#19
Hexachloroethane
Concen: 19.023 ng/ul
RT: 8.875 min Scan# 984
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:117 Resp: 120771
Ion Ratio Lower Upper
117 100
201 88.0 81.8 122.6
199 56.4 51.6 77.4









#22
Nitrobenzene
Concen: 19.142 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion: 77 Resp: 315716
Ion Ratio Lower Upper
77 100
123 45.1 37.4 56.0
65 12.8 10.8 16.2



#23
Isophorone
Concen: 17.585 ng/ul
RT: 9.539 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion: 82 Resp: 547482
Ion Ratio Lower Upper
82 100
95 6.5 5.5 8.3
138 17.2 15.3 22.9









#25
2-Nitrophenol
Concen: 18.962 ng/ul
RT: 9.728 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:139 Resp: 147719
Ion Ratio Lower Upper
139 100
65 50.8 39.8 59.6
109 28.6 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 18.739 ng/ul
RT: 9.792 min Scan# 1140
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:107 Resp: 298289
Ion Ratio Lower Upper
107 100
121 51.9 42.7 64.1
122 91.8 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.412 ng/ul
 RT: 10.028 min Scan# 1180
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020014


Tgt Ion: 93 Resp: 361538
 Ion Ratio Lower Upper
 93 100
 95 32.0 26.1 39.1
 123 12.0 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 18.557 ng/ul
 RT: 10.263 min Scan# 1220
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26

Tgt Ion: 162 Resp: 244612
 Ion Ratio Lower Upper
 162 100
 164 65.1 51.8 77.6
 98 33.9 29.5 44.3





#30

Naphthalene

Concen: 19.112 ng/ul

RT: 10.651 min Scan# 1286

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26

Instrument :

BNA_M

ClientSampleId :

SSTD020014





Tgt Ion:128 Resp: 917536

Ion Ratio Lower Upper

128 100

129 11.2 8.8 13.2

127 12.6 10.6 16.0



#32

4-Chloroaniline

Concen: 17.126 ng/ul

RT: 10.780 min Scan# 1308

Delta R.T. 0.000 min

Lab File: BM036297.D




Acq: 16 Aug 2022 12:26


Tgt Ion:127 Resp: 365766

Ion Ratio Lower Upper

127 100

129 32.8 25.3 37.9









#33
Hexachlorobutadiene
Concen: 18.697 ng/ul
RT: 10.933 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:225 Resp: 152150
Ion Ratio Lower Upper
225 100
223 63.5 50.0 75.0
227 67.3 49.9 74.9



#34
Caprolactam
Concen: 15.187 ng/ul
RT: 11.569 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:113 Resp: 66161
Ion Ratio Lower Upper
113 100
55 171.6 150.2 225.4
56 127.5 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 17.941 ng/ul
RT: 11.916 min Scan# 1501
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:107 Resp: 245905
Ion Ratio Lower Upper
107 100
144 26.0 20.3 30.5
142 81.7 64.4 96.6



#36
2-Methylnaphthalene
Concen: 18.095 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:142 Resp: 568349
Ion Ratio Lower Upper
142 100
141 89.7 70.7 106.1









#37
1-Methylnaphthalene
Concen: 18.134 ng/ul
RT: 12.486 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:142 Resp: 574120
Ion Ratio Lower Upper
142 100
141 92.3 75.2 112.8
116 4.0 2.6 3.8#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.498 ng/ul
RT: 12.633 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:216 Resp: 266731
Ion Ratio Lower Upper
216 100
214 77.2 62.4 93.6
179 21.7 17.3 25.9
108 19.1 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 14.360 ng/ul
RT: 12.610 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:237 Resp: 137906
Ion Ratio Lower Upper
237 100
235 62.8 48.4 72.6
272 13.8 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 18.132 ng/ul
RT: 12.886 min Scan# 1666
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:196 Resp: 167320
Ion Ratio Lower Upper
196 100
198 94.8 77.8 116.6
200 31.2 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 18.007 ng/u1
RT: 12.963 min Scan# 1679
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:196 Resp: 177299
Ion Ratio Lower Upper
196 100
198 97.6 76.8 115.2
200 31.8 24.9 37.3



#43
1,1'-Biphenyl
Concen: 18.720 ng/u1
RT: 13.286 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:154 Resp: 729905
Ion Ratio Lower Upper
154 100
153 40.1 32.4 48.6
76 12.6 10.1 15.1









#44
2-Chloronaphthalene
Concen: 18.804 ng/ul
RT: 13.321 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:162 Resp: 562670
Ion Ratio Lower Upper
162 100
164 32.5 25.4 38.0
127 40.1 31.6 47.4



#45
2-Nitroaniline
Concen: 18.538 ng/ul
RT: 13.545 min Scan# 1778
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion: 65 Resp: 152929
Ion Ratio Lower Upper
65 100
92 68.9 54.5 81.7
138 109.0 91.2 136.8









#47
Dimethylphthalate
Concen: 17.963 ng/ul
RT: 13.916 min Scan# 1841
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:163 Resp: 623462
Ion Ratio Lower Upper
163 100
194 3.8 3.2 4.8
164 9.9 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 17.136 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:165 Resp: 118084
Ion Ratio Lower Upper
165 100
89 61.5 42.5 63.7
121 23.3 17.2 25.8





#50

Acenaphthylene

Concen: 18.710 ng/ul

RT: 14.168 min Scan# 1884

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26

Instrument :

BNA_M

ClientSampleId :

SSTD020014





Tgt Ion:152 Resp: 911194

Ion Ratio Lower Upper

152 100

151 19.8 16.1 24.1

153 13.0 10.6 15.8



#51

3-Nitroaniline

Concen: 15.456 ng/ul

RT: 14.368 min Scan# 1918

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26




Tgt Ion:138 Resp: 117709


Ion Ratio Lower Upper

138 100

108 10.2 8.1 12.1

92 118.2 90.4 135.6









#52
Acenaphthene
Concen: 18.584 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:153 Resp: 587184
Ion Ratio Lower Upper
153 100
152 48.3 37.1 55.7
154 88.3 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 11.904 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:184 Resp: 46574
Ion Ratio Lower Upper
184 100
63 72.4 54.2 81.4
154 64.5 49.3 73.9









#55
4-Nitrophenol
Concen: 14.978 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:109 Resp: 81173
Ion Ratio Lower Upper
109 100
139 138.0 124.9 187.3
65 133.2 122.2 183.4



#56
Dibenzofuran
Concen: 18.705 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:168 Resp: 817709
Ion Ratio Lower Upper
168 100
139 37.9 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 17.480 ng/ul
RT: 14.821 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:165 Resp: 169790
Ion Ratio Lower Upper
165 100
63 48.2 35.9 53.9
182 2.0 2.5 3.7#



#58
2,3,4,6-Tetrachlorophenol
Concen: 17.455 ng/ul
RT: 15.074 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:232 Resp: 135037
Ion Ratio Lower Upper
232 100
131 51.4 36.7 55.1
130 2.8 1.7 2.5#
166 33.0 24.4 36.6









#59
Diethylphthalate
Concen: 17.459 ng/ul
RT: 15.274 min Scan# 2072
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:149 Resp: 612937
Ion Ratio Lower Upper
149 100
177 21.0 16.8 25.2
150 12.1 10.2 15.4



#61
Fluorene
Concen: 18.198 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:166 Resp: 649882
Ion Ratio Lower Upper
166 100
165 97.4 77.2 115.8
167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 17.491 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:204 Resp: 299284
Ion Ratio Lower Upper
204 100
206 32.2 25.9 38.9
141 61.0 50.6 75.8



#63
4-Nitroaniline
Concen: 14.669 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:138 Resp: 102844
Ion Ratio Lower Upper
138 100
92 51.2 39.2 58.8
108 78.0 51.6 77.4#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 16.054 ng/ul
 RT: 15.586 min Scan# 2125
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26




Instrument :
 BNA_M
 ClientSampleId :
 SST020014


Tgt Ion:198 Resp: 82849
 Ion Ratio Lower Upper
 198 100
 182 6.0 4.2 6.4
 77 28.9 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 18.570 ng/ul
 RT: 15.710 min Scan# 2146
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26

Tgt Ion:169 Resp: 523771
 Ion Ratio Lower Upper
 169 100
 168 68.4 53.8 80.6
 167 37.7 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 17.602 ng/ul
RT: 16.386 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:248 Resp: 165558
Ion Ratio Lower Upper
248 100
250 98.1 77.3 115.9
141 75.5 59.8 89.8



#69
Hexachlorobenzene
Concen: 17.575 ng/ul
RT: 16.492 min Scan# 2279
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:284 Resp: 201412
Ion Ratio Lower Upper
284 100
142 38.7 28.6 42.8
249 29.9 24.6 37.0





#70

Atrazine

Concen: 17.248 ng/ul

RT: 16.662 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26

Instrument :

BNA_M

ClientSampleId :

SSTD020014





Tgt Ion:200 Resp: 168740

Ion Ratio Lower Upper

200 100

173 27.3 21.3 31.9

215 46.6 39.0 58.6



#71

Pentachlorophenol

Concen: 13.476 ng/ul

RT: 16.851 min Scan# 2340

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26




Tgt Ion:266 Resp: 94349


Ion Ratio Lower Upper

266 100

264 61.7 49.5 74.3

268 63.6 51.0 76.4









#72
Phenanthrene
Concen: 18.660 ng/u1
RT: 17.233 min Scan# 2405
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:178 Resp: 959519
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.2
176 19.9 14.9 22.3



#74
Anthracene
Concen: 18.639 ng/u1
RT: 17.321 min Scan# 2420
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:178 Resp: 961735
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 19.0 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.442 ng/uL

RT: 13.245 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26

Instrument :

BNA_M

ClientSampleId :

SSTD020014





Tgt Ion:216 Resp: 275156

Ion Ratio Lower Upper

216 100

214 76.5 60.9 91.3

218 47.7 37.5 56.3



#76

Pentachlorobenzene

Concen: 18.749 ng/uL

RT: 14.763 min Scan# 1985

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26




Tgt Ion:250 Resp: 253575


Ion Ratio Lower Upper

250 100

248 63.2 50.4 75.6

252 64.1 51.1 76.7





#77

Carbazole

Concen: 18.531 ng/ul

RT: 17.604 min Scan# 2468

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26

Instrument :

BNA_M

ClientSampleId :

SSTD020014





Tgt Ion:167 Resp: 883142

Ion Ratio Lower Upper

167 100

166 21.3 17.4 26.0

139 13.2 10.1 15.1



#78

Di-n-butylphthalate

Concen: 17.824 ng/ul

RT: 18.162 min Scan# 2563

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26




Tgt Ion:149 Resp: 958399


Ion Ratio Lower Upper

149 100

150 8.9 7.3 10.9

104 5.0 3.7 5.5









#80
 Fluoranthene
 Concen: 17.836 ng/u1
 RT: 19.250 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26




Instrument :
 BNA_M
 ClientSampleId :
 SST020014


Tgt Ion:202 Resp: 1050591
 Ion Ratio Lower Upper
 202 100
 101 11.3 10.1 15.1
 100 8.9 7.8 11.6



#82
 Pyrene
 Concen: 18.055 ng/u1
 RT: 19.609 min Scan# 2809
 Delta R.T. 0.000 min
 Lab File: BM036297.D
 Acq: 16 Aug 2022 12:26

Tgt Ion:202 Resp: 1096174
 Ion Ratio Lower Upper
 202 100
 101 15.3 11.9 17.9
 100 12.1 10.2 15.4









#83
Butylbenzylphthalate
Concen: 18.149 ng/ul
RT: 20.515 min Scan# 2963
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:149 Resp: 400765
Ion Ratio Lower Upper
149 100
91 72.5 53.2 79.8
206 17.7 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 21.179 ng/ul
RT: 21.297 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:252 Resp: 347664
Ion Ratio Lower Upper
252 100
254 64.2 52.4 78.6
126 11.6 10.8 16.2





#85

Benzo(a)anthracene

Concen: 18.216 ng/ul

RT: 21.356 min Scan# 3106

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26

Instrument :

BNA_M

ClientSampleId :

SSTD020014





Tgt Ion:228 Resp: 1019927

Ion Ratio Lower Upper

228 100

229 19.4 15.6 23.4

226 27.0 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 17.454 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26




Tgt Ion:149 Resp: 556913


Ion Ratio Lower Upper

149 100

167 26.3 21.5 32.3

279 3.9 3.4 5.2









#87
Chrysene
Concen: 18.339 ng/ul
RT: 21.409 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:228 Resp: 1001880
Ion Ratio Lower Upper
228 100
226 29.5 23.7 35.5
229 19.5 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 16.099 ng/ul
RT: 22.191 min Scan# 3248
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:149 Resp: 949942









#90
Benzo(b)fluoranthene
Concen: 17.205 ng/ul
RT: 22.997 min Scan# 3385
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26




Instrument :
BNA_M
ClientSampleId :
SSTD020014


Tgt Ion:252 Resp: 1049130
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 11.2 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 17.696 ng/ul
RT: 23.038 min Scan# 3392
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:252 Resp: 1023985
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 11.4 10.6 15.8





#93

Benzo(a)pyrene

Concen: 18.056 ng/ul

RT: 23.603 min Scan# 3488

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26

Instrument :

BNA_M

ClientSampleId :

SSTD020014





Tgt Ion:252 Resp: 1048730

Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 12.7 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.811 ng/ul

RT: 26.115 min Scan# 3915

Delta R.T. 0.000 min

Lab File: BM036297.D

Acq: 16 Aug 2022 12:26




Tgt Ion:276 Resp: 1226482


Ion Ratio Lower Upper

276 100

138 25.8 25.6 38.4

277 24.7 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 19.918 ng/ul
RT: 26.132 min Scan# 3918
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Instrument :
BNA_M
ClientSampleId :
SSTD020014

Tgt Ion:278 Resp: 1057966
Ion Ratio Lower Upper
278 100
139 17.8 16.2 24.2
279 23.8 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 20.497 ng/ul
RT: 26.862 min Scan# 4042
Delta R.T. 0.000 min
Lab File: BM036297.D
Acq: 16 Aug 2022 12:26

Tgt Ion:276 Resp: 1051587
Ion Ratio Lower Upper
276 100
138 23.3 21.3 31.9
277 23.5 18.6 28.0

