

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036300.D
 Acq On : 16 Aug 2022 14:17
 Operator : CG/JU
 Sample : SSTD16065
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160017

Quant Time: Aug 16 14:49:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	216734	20.000	ng/u1	0.00
20) Naphthalene-d8	10.610	136	899604	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.451	164	498615	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	968237	20.000	ng/u1	0.00
79) Chrysene-d12	21.380	240	897562	20.000	ng/u1	0.00
88) Perylene-d12	23.709	264	959229	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	367532	63.811	ng/uL	0.00
4) Pyridine-d5	3.634	84	2546755	161.401	ng/u1	0.00
7) Phenol-d5	6.998	99	2992430	164.644	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.151	67	1817801	150.029	ng/u1	0.01
11) 2-Chlorophenol-d4	7.345	132	2443469	161.966	ng/u1	0.01
15) 4-Methylphenol-d8	8.545	113	2219211	150.720	ng/u1	0.02
21) Nitrobenzene-d5	8.981	128	1185314	166.277	ng/u1	0.00
24) 2-Nitrophenol-d4	9.704	143	1281486	184.516	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.245	165	2105118	167.589	ng/u1	0.01
31) 4-Chloroaniline-d4	10.763	131	3047890	143.332	ng/u1	0.00
46) Dimethylphthalate-d6	13.880	166	5200110	153.104	ng/u1	0.01
49) Acenaphthylene-d8	14.151	160	6583402	151.918	ng/u1	0.01
54) 4-Nitrophenol-d4	14.698	143	1102911	158.132	ng/u1	0.02
60) Fluorene-d10	15.445	176	4647672	151.430	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.586	200	951952	182.908	ng/u1	0.02
73) Anthracene-d10	17.298	188	6860959	156.734	ng/u1	0.01
81) Pyrene-d10	19.592	212	7420687	149.175	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.568	264	7759501	159.336	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	394306	66.061	ng/uL	94
5) Pyridine	3.652	79	2591118	156.484	ng/u1	92
6) Benzaldehyde	6.951	77	917430	120.268	ng/u1	94
8) Phenol	7.028	94	3072220	154.461	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.245	93	2392201	149.601	ng/u1	98
12) 2-Chlorophenol	7.375	128	2505251	161.879	ng/u1	100
13) 2-Methylphenol	8.269	108	2280981	153.510	ng/u1	95
14) 2,2'-oxybis(1-Chloropr...	8.345	45	3206066	151.898	ng/u1	97
16) Acetophenone	8.645	105	3406661	144.105	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.657	70	1720803	145.968	ng/u1	95
18) 4-Methylphenol	8.616	108	2386292	149.702	ng/u1	98
19) Hexachloroethane	8.881	117	1052581	164.843	ng/u1	93
22) Nitrobenzene	9.028	77	2639131	161.784	ng/u1	99
23) Isophorone	9.563	82	4721250	153.321	ng/u1	98
25) 2-Nitrophenol	9.733	139	1363153	176.916	ng/u1	98
26) 2,4-Dimethylphenol	9.804	107	2534097	160.954	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.039	93	2938429	151.302	ng/u1	100
29) 2,4-Dichlorophenol	10.269	162	2116653	162.352	ng/u1	99
30) Naphthalene	10.663	128	7253024	152.748	ng/u1	99
32) 4-Chloroaniline	10.786	127	3083460	145.972	ng/u1	98
33) Hexachlorobutadiene	10.933	225	1322798	164.352	ng/u1	99
34) Caprolactam	11.622	113	335869	77.950	ng/u1	93
35) 4-Chloro-3-methylphenol	11.927	107	2190429	161.576	ng/u1	99
36) 2-Methylnaphthalene	12.275	142	4691296	151.009	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036300.D
 Acq On : 16 Aug 2022 14:17
 Operator : CG/JU
 Sample : SSTD16065
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160017

Quant Time: Aug 16 14:49:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 14:05:33 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.492	142	4670893	149.164	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.645	216	2276379	160.383	ng/ul	99
40) Hexachlorocyclopentadiene	12.616	237	1561374	165.173	ng/ul	97
41) 2,4,6-Trichlorophenol	12.892	196	1557465	171.471	ng/ul	99
42) 2,4,5-Trichlorophenol	12.969	196	1662336	171.521	ng/ul	100
43) 1,1'-Biphenyl	13.292	154	5914215	154.104	ng/ul	99
44) 2-Chloronaphthalene	13.333	162	4619832	156.853	ng/ul	98
45) 2-Nitroaniline	13.557	65	1437615	177.045	ng/ul	98
47) Dimethylphthalate	13.927	163	5287095	154.757	ng/ul	99
48) 2,6-Dinitrotoluene	14.051	165	1190768	175.557	ng/ul	94
50) Acenaphthylene	14.180	152	7281441	151.892	ng/ul	98
51) 3-Nitroaniline	14.380	138	1165244	155.447	ng/ul	98
52) Acenaphthene	14.516	153	4745171	152.578	ng/ul	99
53) 2,4-Dinitrophenol	14.586	184	745626	193.611	ng/ul	95
55) 4-Nitrophenol	14.716	109	874886	164.010	ng/ul#	79
56) Dibenzofuran	14.857	168	6433896	149.517	ng/ul	96
57) 2,4-Dinitrotoluene	14.839	165	1644720	172.022	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.080	232	1284011	168.621	ng/ul#	91
59) Diethylphthalate	15.286	149	5348404	154.772	ng/ul	99
61) Fluorene	15.504	166	5147728	146.443	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.498	204	2444509	145.143	ng/ul	99
63) 4-Nitroaniline	15.551	138	915862	132.713	ng/ul#	90
66) 4,6-Dinitro-2-methylph...	15.598	198	949337	182.287	ng/ul#	94
67) N-Nitrosodiphenylamine	15.715	169	4428157	155.565	ng/ul	99
68) 4-Bromophenyl-phenylether	16.392	248	1511619	159.250	ng/ul	99
69) Hexachlorobenzene	16.498	284	1802295	155.832	ng/ul	99
70) Atrazine	16.680	200	1568035	158.820	ng/ul	99
71) Pentachlorophenol	16.851	266	1155340	163.517	ng/ul	99
72) Phenanthrene	17.239	178	7902311	152.276	ng/ul	97
74) Anthracene	17.333	178	7968284	153.021	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.251	216	2357654	156.583	ng/ul	97
76) Pentachlorobenzene	14.768	250	2145168	157.170	ng/ul	99
77) Carbazole	17.609	167	7411399	154.100	ng/ul	98
78) Di-n-butylphthalate	18.168	149	8684023	160.034	ng/ul	98
80) Fluoranthene	19.256	202	8967356	150.879	ng/ul	96
82) Pyrene	19.621	202	9008247	147.049	ng/ul	95
83) Butylbenzylphthalate	20.521	149	4084212	183.309	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.303	252	2830336	170.883	ng/ul	99
85) Benzo(a)anthracene	21.362	228	8710385	154.183	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.292	149	5496249	170.724	ng/ul	99
87) Chrysene	21.421	228	8387311	152.159	ng/ul	96
89) Di-n-octyl phthalate	22.197	149	9425714	152.163	ng/ul	100
90) Benzo(b)fluoranthene	23.003	252	9452735	147.660	ng/ul#	96
91) Benzo(k)fluoranthene	23.056	252	9066731	149.251	ng/ul#	95
93) Benzo(a)pyrene	23.627	252	9388142	153.966	ng/ul#	95
94) Indeno(1,2,3-cd)pyrene	26.156	276	11218300	172.605	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.174	278	9517415	170.683	ng/ul#	94
96) Benzo(g,h,i)perylene	26.909	276	9645536	179.087	ng/ul	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
Data File : BM036300.D
Acq On : 16 Aug 2022 14:17
Operator : CG/JU
Sample : SSTD16065
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD160017

Quant Time: Aug 16 14:49:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 16 14:05:33 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 66.061 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion: 88 Resp: 394306
Ion Ratio Lower Upper
88 100
43 29.0 27.8 41.8
58 73.4 61.8 92.8



#5
Pyridine
Concen: 156.484 ng/uL
RT: 3.652 min Scan# 96
Delta R.T. -0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion: 79 Resp: 2591118
Ion Ratio Lower Upper
79 100
52 76.7 67.8 101.6
51 36.1 31.4 47.0









#6
Benzaldehyde
Concen: 120.268 ng/ul
RT: 6.951 min Scan# 61
Delta R.T. 0.000 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion: 77 Resp: 917430
Ion Ratio Lower Upper
77 100
105 98.9 84.6 127.0
106 95.7 80.3 120.5



#8
Phenol
Concen: 154.461 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion: 94 Resp: 3072220
Ion Ratio Lower Upper
94 100
65 28.5 21.3 31.9
66 36.4 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 149.601 ng/ul
RT: 7.245 min Scan# 705
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Instrument :
BNA_M
ClientSampleId :
SSTD160017

Tgt Ion: 93 Resp: 2392201




Ion	Ratio	Lower	Upper
93	100		
63	77.4	64.0	96.0
95	32.9	25.8	38.6




#12
2-Chlorophenol
Concen: 161.879 ng/ul
RT: 7.375 min Scan# 729
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion: 128 Resp: 2505251

Ion	Ratio	Lower	Upper
128	100		
64	45.8	36.9	55.3
130	32.3	25.6	38.4









#13
2-Methylphenol
Concen: 153.510 ng/ul
RT: 8.269 min Scan# 880
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:108 Resp: 2280981
Ion Ratio Lower Upper
108 100
107 88.8 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 151.898 ng/ul
RT: 8.345 min Scan# 894
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion: 45 Resp: 3206066
Ion Ratio Lower Upper
45 100
77 14.5 10.6 15.8
79 11.1 8.3 12.5









#16
Acetophenone
Concen: 144.105 ng/ul
RT: 8.645 min Scan# 944
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:105 Resp: 3406661
Ion Ratio Lower Upper
105 100
77 77.1 59.8 89.6
51 27.0 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 145.968 ng/ul
RT: 8.657 min Scan# 947
Delta R.T. 0.035 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion: 70 Resp: 1720803
Ion Ratio Lower Upper
70 100
42 56.8 49.6 74.4
101 9.9 8.1 12.1
130 21.9 18.4 27.6









#18
4-Methylphenol
Concen: 149.702 ng/ul
RT: 8.616 min Scan# 94
Delta R.T. 0.024 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:108 Resp: 2386292
Ion Ratio Lower Upper
108 100
107 110.6 90.3 135.5



#19
Hexachloroethane
Concen: 164.843 ng/ul
RT: 8.881 min Scan# 985
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:117 Resp: 1052581
Ion Ratio Lower Upper
117 100
201 94.5 81.8 122.6
199 59.5 51.6 77.4









#22
Nitrobenzene
Concen: 161.784 ng/ul
RT: 9.028 min Scan# 1101
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion: 77 Resp: 2639131
Ion Ratio Lower Upper
77 100
123 47.0 37.4 56.0
65 13.3 10.8 16.2



#23
Isophorone
Concen: 153.321 ng/ul
RT: 9.563 min Scan# 1101
Delta R.T. 0.024 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion: 82 Resp: 4721250
Ion Ratio Lower Upper
82 100
95 7.0 5.5 8.3
138 17.7 15.3 22.9









#25
2-Nitrophenol
Concen: 176.916 ng/ul
RT: 9.733 min Scan# 1129
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:139 Resp: 1363153
Ion Ratio Lower Upper
139 100
65 49.2 39.8 59.6
109 29.8 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 160.954 ng/ul
RT: 9.804 min Scan# 1142
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:107 Resp: 2534097
Ion Ratio Lower Upper
107 100
121 50.6 42.7 64.1
122 86.6 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 151.302 ng/ul

RT: 10.039 min Scan# 1180

Delta R.T. 0.012 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Instrument :





BNA_M

ClientSampleId :

SSTD160017

Tgt Ion: 93 Resp: 2938429

Ion	Ratio	Lower	Upper
93	100		
95	32.9	26.1	39.1
123	11.6	9.4	14.0



#29

2,4-Dichlorophenol

Concen: 162.352 ng/ul

RT: 10.269 min Scan# 1221




Delta R.T. 0.006 min


Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Tgt Ion:162 Resp: 2116653

Ion	Ratio	Lower	Upper
162	100		
164	66.0	51.8	77.6
98	36.9	29.5	44.3









#30
Naphthalene
Concen: 152.748 ng/ul
RT: 10.663 min Scan# 1288
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:128 Resp: 7253024
Ion Ratio Lower Upper
128 100
129 11.3 8.8 13.2
127 13.5 10.6 16.0



#32
4-Chloroaniline
Concen: 145.972 ng/ul
RT: 10.786 min Scan# 1309
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:127 Resp: 3083460
Ion Ratio Lower Upper
127 100
129 32.4 25.3 37.9









#33
Hexachlorobutadiene
Concen: 164.352 ng/ul
RT: 10.933 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:225 Resp: 1322798
Ion Ratio Lower Upper
225 100
223 62.9 50.0 75.0
227 63.9 49.9 74.9



#34
Caprolactam
Concen: 77.950 ng/ul
RT: 11.622 min Scan# 1451
Delta R.T. 0.053 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:113 Resp: 335869
Ion Ratio Lower Upper
113 100
55 173.0 150.2 225.4
56 125.9 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 161.576 ng/ul
RT: 11.927 min Scan# 1503
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:107 Resp: 2190429
Ion Ratio Lower Upper
107 100
144 26.2 20.3 30.5
142 81.3 64.4 96.6



#36
2-Methylnaphthalene
Concen: 151.009 ng/ul
RT: 12.275 min Scan# 1562
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:142 Resp: 4691296
Ion Ratio Lower Upper
142 100
141 88.4 70.7 106.1





#37

1-Methylnaphthalene

Concen: 149.164 ng/ul

RT: 12.492 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Instrument :

BNA_M

ClientSampleId :

SSTD160017





Tgt Ion:142 Resp: 4670893

Ion Ratio Lower Upper

142 100

141 92.4 75.2 112.8

116 4.2 2.6 3.8#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 160.383 ng/ul

RT: 12.645 min Scan# 1625

Delta R.T. 0.012 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Tgt Ion:216 Resp: 2276379




Ion Ratio Lower Upper


216 100

214 78.9 62.4 93.6

179 21.5 17.3 25.9

108 17.4 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 165.173 ng/ul
RT: 12.616 min Scan# 1619
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:237 Resp: 1561374
Ion Ratio Lower Upper
237 100
235 62.8 48.4 72.6
272 13.0 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 171.471 ng/ul
RT: 12.892 min Scan# 1667
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:196 Resp: 1557465
Ion Ratio Lower Upper
196 100
198 95.7 77.8 116.6
200 30.6 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 171.521 ng/uI
RT: 12.969 min Scan# 1
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:196 Resp: 1662336
Ion Ratio Lower Upper
196 100
198 95.9 76.8 115.2
200 31.1 24.9 37.3



#43
1,1'-Biphenyl
Concen: 154.104 ng/uI
RT: 13.292 min Scan# 1735
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:154 Resp: 5914215
Ion Ratio Lower Upper
154 100
153 40.8 32.4 48.6
76 12.2 10.1 15.1









#44
2-Chloronaphthalene
Concen: 156.853 ng/ul
RT: 13.333 min Scan# 1742
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:162 Resp: 4619832
Ion Ratio Lower Upper
162 100
164 33.4 25.4 38.0
127 40.3 31.6 47.4



#45
2-Nitroaniline
Concen: 177.045 ng/ul
RT: 13.557 min Scan# 1780
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion: 65 Resp: 1437615
Ion Ratio Lower Upper
65 100
92 71.6 54.5 81.7
138 115.3 91.2 136.8









#47
Dimethylphthalate
Concen: 154.757 ng/ul
RT: 13.927 min Scan# 1843
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:163 Resp: 5287095
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 175.557 ng/ul
RT: 14.051 min Scan# 1864
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:165 Resp: 1190768
Ion Ratio Lower Upper
165 100
89 59.1 42.5 63.7
121 22.0 17.2 25.8





#50

Acenaphthylene

Concen: 151.892 ng/ul

RT: 14.180 min Scan# 1

Delta R.T. 0.012 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Instrument :

BNA_M

ClientSampleId :

SSTD160017





Tgt Ion:152 Resp: 7281441

Ion Ratio Lower Upper

152 100

151 21.1 16.1 24.1

153 13.8 10.6 15.8



#51

3-Nitroaniline

Concen: 155.447 ng/ul

RT: 14.380 min Scan# 1920

Delta R.T. 0.012 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17




Tgt Ion:138 Resp: 1165244


Ion Ratio Lower Upper

138 100

108 9.3 8.1 12.1

92 111.2 90.4 135.6









#52
Acenaphthene
Concen: 152.578 ng/ul
RT: 14.516 min Scan# 1943
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:153 Resp: 4745171
Ion Ratio Lower Upper
153 100
152 48.4 37.1 55.7
154 89.2 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 193.611 ng/ul
RT: 14.586 min Scan# 1955
Delta R.T. 0.000 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:184 Resp: 745626
Ion Ratio Lower Upper
184 100
63 71.4 54.2 81.4
154 66.6 49.3 73.9









#55
4-Nitrophenol
Concen: 164.010 ng/ul
RT: 14.716 min Scan# 1973
Delta R.T. 0.024 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:109 Resp: 874886
Ion Ratio Lower Upper
109 100
139 132.2 124.9 187.3
65 122.0 122.2 183.4#



#56
Dibenzofuran
Concen: 149.517 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:168 Resp: 6433896
Ion Ratio Lower Upper
168 100
139 40.1 30.1 45.1





#57

2,4-Dinitrotoluene

Concen: 172.022 ng/ul

RT: 14.839 min Scan# 1998

Delta R.T. 0.018 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Instrument :

BNA_M

ClientSampleId :

SSTD160017





Tgt Ion:165 Resp: 1644720

Ion Ratio Lower Upper

165 100

63 48.8 35.9 53.9

182 2.5 2.5 3.7



#58

2,3,4,6-Tetrachlorophenol

Concen: 168.621 ng/ul

RT: 15.080 min Scan# 2039

Delta R.T. 0.006 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Tgt Ion:232 Resp: 1284011




Ion Ratio Lower Upper


232 100

131 53.0 36.7 55.1

130 2.7 1.7 2.5#

166 33.7 24.4 36.6









#59
Diethylphthalate
Concen: 154.772 ng/ul
RT: 15.286 min Scan# 2074
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:149 Resp: 5348404
Ion Ratio Lower Upper
149 100
177 21.5 16.8 25.2
150 12.6 10.2 15.4



#61
Fluorene
Concen: 146.443 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:166 Resp: 5147728
Ion Ratio Lower Upper
166 100
165 97.2 77.2 115.8
167 13.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 145.143 ng/ul
RT: 15.498 min Scan# 2109
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:204 Resp: 2444509
Ion Ratio Lower Upper
204 100
206 33.3 25.9 38.9
141 63.1 50.6 75.8



#63
4-Nitroaniline
Concen: 132.713 ng/ul
RT: 15.551 min Scan# 2119
Delta R.T. 0.018 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:138 Resp: 915862
Ion Ratio Lower Upper
138 100
92 48.3 39.2 58.8
108 78.4 51.6 77.4#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 182.287 ng/ul
 RT: 15.598 min Scan# 2127
 Delta R.T. 0.012 min
 Lab File: BM036300.D
 Acq: 16 Aug 2022 14:17




Instrument :
 BNA_M
 ClientSampleId :
 SSTD160017


Tgt Ion:198 Resp: 949337
 Ion Ratio Lower Upper
 198 100
 182 3.7 4.2 6.4#
 77 27.9 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 155.565 ng/ul
 RT: 15.715 min Scan# 2147
 Delta R.T. 0.006 min
 Lab File: BM036300.D
 Acq: 16 Aug 2022 14:17

Tgt Ion:169 Resp: 4428157
 Ion Ratio Lower Upper
 169 100
 168 67.4 53.8 80.6
 167 37.2 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 159.250 ng/ul
RT: 16.392 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:248 Resp: 1511619
Ion Ratio Lower Upper
248 100
250 97.3 77.3 115.9
141 74.1 59.8 89.8



#69
Hexachlorobenzene
Concen: 155.832 ng/ul
RT: 16.498 min Scan# 2280
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:284 Resp: 1802295
Ion Ratio Lower Upper
284 100
142 36.7 28.6 42.8
249 30.4 24.6 37.0









#70
Atrazine
Concen: 158.820 ng/ul
RT: 16.680 min Scan# 21
Delta R.T. 0.018 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:200 Resp: 1568035
Ion Ratio Lower Upper
200 100
173 26.9 21.3 31.9
215 48.2 39.0 58.6



#71
Pentachlorophenol
Concen: 163.517 ng/ul
RT: 16.851 min Scan# 2340
Delta R.T. 0.000 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:266 Resp: 1155340
Ion Ratio Lower Upper
266 100
264 62.6 49.5 74.3
268 64.3 51.0 76.4









#72
Phenanthrene
Concen: 152.276 ng/ul
RT: 17.239 min Scan# 2405
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:178 Resp: 7902311
Ion Ratio Lower Upper
178 100
179 16.1 12.2 18.2
176 20.3 14.9 22.3



#74
Anthracene
Concen: 153.021 ng/ul
RT: 17.333 min Scan# 2422
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:178 Resp: 7968284
Ion Ratio Lower Upper
178 100
179 16.1 12.2 18.2
176 19.8 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 156.583 ng/uL

RT: 13.251 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Instrument :

BNA_M

ClientSampleId :

SSTD160017





Tgt Ion:216 Resp: 2357654

Ion Ratio Lower Upper

216 100

214 79.2 60.9 91.3

218 48.3 37.5 56.3



#76

Pentachlorobenzene

Concen: 157.170 ng/uL

RT: 14.768 min Scan# 1986

Delta R.T. 0.006 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17




Tgt Ion:250 Resp: 2145168


Ion Ratio Lower Upper

250 100

248 63.9 50.4 75.6

252 64.9 51.1 76.7





#77

Carbazole

Concen: 154.100 ng/ul

RT: 17.609 min Scan# 2469

Delta R.T. 0.006 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Instrument :

BNA_M

ClientSampleId :

SSTD160017





Tgt Ion:167 Resp: 7411399

Ion Ratio Lower Upper

167 100

166 22.0 17.4 26.0

139 14.1 10.1 15.1



#78

Di-n-butylphthalate

Concen: 160.034 ng/ul

RT: 18.168 min Scan# 2564

Delta R.T. 0.006 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17




Tgt Ion:149 Resp: 8684023


Ion Ratio Lower Upper

149 100

150 9.7 7.3 10.9

104 5.3 3.7 5.5









#80
 Fluoranthene
 Concen: 150.879 ng/ul
 RT: 19.256 min Scan# 21
 Delta R.T. 0.006 min
 Lab File: BM036300.D
 Acq: 16 Aug 2022 14:17




Instrument :
 BNA_M
 ClientSampleId :
 SSTD160017


Tgt Ion:202 Resp: 8967356
 Ion Ratio Lower Upper
 202 100
 101 11.0 10.1 15.1
 100 8.4 7.8 11.6



#82
 Pyrene
 Concen: 147.049 ng/ul
 RT: 19.621 min Scan# 2811
 Delta R.T. 0.012 min
 Lab File: BM036300.D
 Acq: 16 Aug 2022 14:17

Tgt Ion:202 Resp: 9008247
 Ion Ratio Lower Upper
 202 100
 101 13.1 11.9 17.9
 100 10.6 10.2 15.4









#83
Butylbenzylphthalate
Concen: 183.309 ng/ul
RT: 20.521 min Scan# 2964
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:149 Resp: 4084212
Ion Ratio Lower Upper
149 100
91 65.7 53.2 79.8
206 20.1 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 170.883 ng/ul
RT: 21.303 min Scan# 3097
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:252 Resp: 2830336
Ion Ratio Lower Upper
252 100
254 65.4 52.4 78.6
126 11.3 10.8 16.2









#85
Benzo(a)anthracene
Concen: 154.183 ng/ul
RT: 21.362 min Scan# 3107
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:228 Resp: 8710385
Ion Ratio Lower Upper
228 100
229 21.1 15.6 23.4
226 29.3 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 170.724 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. 0.000 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:149 Resp: 5496249
Ion Ratio Lower Upper
149 100
167 27.8 21.5 32.3
279 4.2 3.4 5.2









#87
Chrysene
Concen: 152.159 ng/ul
RT: 21.421 min Scan# 3117
Delta R.T. 0.012 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:228 Resp: 8387311
Ion Ratio Lower Upper
228 100
226 32.3 23.7 35.5
229 20.9 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 152.163 ng/ul
RT: 22.197 min Scan# 3249
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:149 Resp: 9425714









#90
Benzo(b)fluoranthene
Concen: 147.660 ng/ul
RT: 23.003 min Scan# 3386
Delta R.T. 0.006 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17




Instrument :
BNA_M
ClientSampleId :
SSTD160017


Tgt Ion:252 Resp: 9452735
Ion Ratio Lower Upper
252 100
253 23.0 17.5 26.3
125 10.4 10.7 16.1#



#91
Benzo(k)fluoranthene
Concen: 149.251 ng/ul
RT: 23.056 min Scan# 3395
Delta R.T. 0.018 min
Lab File: BM036300.D
Acq: 16 Aug 2022 14:17

Tgt Ion:252 Resp: 9066731
Ion Ratio Lower Upper
252 100
253 23.1 17.4 26.0
125 9.9 10.6 15.8#





#93

Benzo(a)pyrene

Concen: 153.966 ng/u1

RT: 23.627 min Scan# 34

Delta R.T. 0.024 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17

Instrument :

BNA_M

ClientSampleId :

SSTD160017





Tgt Ion:252 Resp: 9388142

Ion Ratio Lower Upper

252 100

253 22.4 17.3 25.9

125 10.9 11.8 17.8#



#94

Indeno(1,2,3-cd)pyrene

Concen: 172.605 ng/u1

RT: 26.156 min Scan# 3922

Delta R.T. 0.041 min

Lab File: BM036300.D

Acq: 16 Aug 2022 14:17




Tgt Ion:276 Resp:11218300


Ion Ratio Lower Upper

276 100

138 25.2 25.6 38.4#

277 25.7 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 170.683 ng/ul
 RT: 26.174 min Scan# 3918
 Delta R.T. 0.041 min
 Lab File: BM036300.D
 Acq: 16 Aug 2022 14:17

Instrument :
 BNA_M
 ClientSampleId :
 SSTD160017

Tgt Ion:278 Resp: 9517415
 Ion Ratio Lower Upper
 278 100
 139 15.9 16.2 24.2#
 279 24.5 18.7 28.1



#96
 Benzo(g,h,i)perylene
 Concen: 179.087 ng/ul
 RT: 26.909 min Scan# 4050
 Delta R.T. 0.047 min
 Lab File: BM036300.D
 Acq: 16 Aug 2022 14:17

Tgt Ion:276 Resp: 9645536
 Ion Ratio Lower Upper
 276 100
 138 21.9 21.3 31.9
 277 24.3 18.6 28.0

