

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
 Data File : BM036305.D
 Acq On : 16 Aug 2022 18:03
 Operator : CG/JU
 Sample : N4173-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EX7P4

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/17/2022
 Supervised By :Sohil Jodhani 08/18/2022

Quant Time: Aug 17 02:30:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 15:00:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	231144	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	963807	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	550781	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.192	188	1088559	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	1103430	20.000	ng/u1	0.00
88) Perylene-d12	23.703	264	1228724	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	21831	3.599	ng/uL	0.00
4) Pyridine-d5	3.663	84	27008m	1.615	ng/u1	0.02
7) Phenol-d5	6.987	99	387411	19.636	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	259365	20.884	ng/u1	0.00
11) 2-Chlorophenol-d4	7.340	132	321589	20.745	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	273656	18.576	ng/u1	0.00
21) Nitrobenzene-d5	8.975	128	155846	21.327	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	150002	20.237	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.233	165	263137	19.868	ng/u1	0.00
31) 4-Chloroaniline-d4	10.757	131	356604	17.368	ng/u1	0.00
46) Dimethylphthalate-d6	13.869	166	775174	22.246	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	953622	21.943	ng/u1	0.00
54) 4-Nitrophenol-d4	14.680	143	104710	15.859	ng/u1	0.00
60) Fluorene-d10	15.439	176	747599	23.743	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.568	200	84836	15.159	ng/u1	0.00
73) Anthracene-d10	17.286	188	1118746	23.113	ng/u1	0.00
81) Pyrene-d10	19.580	212	1346082	23.517	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.556	264	1461583	24.261	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081622\
Data File : BM036305.D
Acq On : 16 Aug 2022 18:03
Operator : CG/JU
Sample : N4173-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EX7P4

Quant Time: Aug 17 02:30:19 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 16 15:00:33 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/17/2022
Supervised By :Sohil Jodhani 08/18/2022

