

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081715\
 Data File : BM002461.D
 Acq On : 17 Aug 2015 21:09
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02042

Quant Time: Aug 18 03:01:23 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 18 01:54:16 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	163#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	144	0.00
3	Naphthalene	1.021	0.998	2.3	144	0.00
4	2-Methylnaphthalene	0.765	0.709	7.3	136	0.00
5	1-Methylnaphthalene	0.756	0.690	8.7	134	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
7 S	Acenaphthylene-d8	2.045	2.046	-0.0	124	0.00
8	Acenaphthylene	2.140	2.118	1.0	123	0.00
9 C	Acenaphthene	1.410	1.383	1.9	123	0.00
10 S	Fluorene-d10	1.436	1.345	6.3	117	0.00
11	Fluorene	1.615	1.511	6.4	116	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
13	Phenanthrene	1.155	1.133	1.9	101	0.00
14 S	Anthracene-d10	0.965	0.953	1.2	101	0.00
15	Anthracene	1.170	1.145	2.1	100	0.00
16 C	Fluoranthene	1.308	0.976	25.4#	77	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	64	0.00
18 S	Pyrene-d10	0.887	1.091	-23.0#	74	0.00
19	Pyrene	1.172	1.439	-22.8#	74	0.00
20	Benzo(a)anthracene	1.196	1.184	1.0	64	0.00
21	Chrysene	1.119	1.098	1.9	63	0.00
22 I	Perylene-d12	1.000	1.000	0.0	62	0.00
23	Benzo(b)fluoranthene	1.209	1.186	1.9	63	0.00
24	Benzo(k)fluoranthene	1.146	1.131	1.3	61	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.039	1.1	63	0.00
26 C	Benzo(a)pyrene	1.155	1.146	0.8	62	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.362	-2.6	65	0.00
28	Dibenzo(a,h)anthracene	1.110	1.146	-3.2	65	0.00
29	Benzo(g,h,i)perylene	1.106	1.141	-3.2	65	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1