

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081715\
 Data File : BM002475.D
 Acq On : 18 Aug 2015 05:45
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02043

Quant Time: Aug 18 06:57:40 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 18 03:02:54 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	136	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
3	Naphthalene	1.021	1.003	1.8	124	0.00
4	2-Methylnaphthalene	0.765	0.721	5.8	119	0.00
5	1-Methylnaphthalene	0.756	0.704	6.9	117	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	111	0.00
7 S	Acenaphthylene-d8	2.045	2.036	0.4	111	0.00
8	Acenaphthylene	2.140	2.111	1.4	110	0.00
9 C	Acenaphthene	1.410	1.380	2.1	111	0.00
10 S	Fluorene-d10	1.436	1.375	4.2	108	0.00
11	Fluorene	1.615	1.552	3.9	108	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
13	Phenanthrene	1.155	1.137	1.6	101	0.00
14 S	Anthracene-d10	0.965	0.952	1.3	100	0.00
15	Anthracene	1.170	1.152	1.5	100	0.00
16 C	Fluoranthene	1.308	1.090	16.7	86	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	74	0.00
18 S	Pyrene-d10	0.887	1.064	-20.0	84	0.00
19	Pyrene	1.172	1.397	-19.2	82	0.00
20	Benzo(a)anthracene	1.196	1.188	0.7	74	0.00
21	Chrysene	1.119	1.107	1.1	73	0.00
22 I	Perylene-d12	1.000	1.000	0.0	68	0.00
23	Benzo(b)fluoranthene	1.209	1.206	0.2	70	0.00
24	Benzo(k)fluoranthene	1.146	1.162	-1.4	69	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.035	1.5	68	0.00
26 C	Benzo(a)pyrene	1.155	1.145	0.9	68	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.300	2.1	68	0.00
28	Dibenzo(a,h)anthracene	1.110	1.098	1.1	68	0.00
29	Benzo(g,h,i)perylene	1.106	1.076	2.7	67	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0