


Data Path : Z:\HPCHEM1\BNA M\DATA\BM081716\
 Data File : BM007140.D
 Acq On : 16 Aug 2016 12:24
 Operator : UM/SJ
 Sample : DFTPP47
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP47

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081116.M
 Title : SVOA CALIBRATION
 Last Update : Wed Aug 17 04:01:43 2016



AutoFind: Scans 2555, 2556, 2557; Background Corrected with Scan 2550

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.6	86866	PASS
68	69	0.00	2	1.8	1742	PASS
69	198	0.00	100	36.5	94469	PASS
70	69	0.00	2	0.5	451	PASS
127	198	10	80	44.0	113850	PASS
197	198	0.00	2	0.4	1138	PASS
198	198	100	100	100.0	258730	PASS
199	198	5	9	7.0	18071	PASS
275	198	10	60	29.3	75698	PASS
365	198	1	100	3.9	10164	PASS
441	443	0.01	100	75.9	30536	PASS
442	198	50	100	79.8	206490	PASS
443	442	15	24	19.5	40258	PASS

m/z	Abundance
39.90	1148.0
41.00	382.0
44.00	2273.0
73.00	545.0
85.70	333.0
191.00	312.0
192.90	406.0
207.10	1573.0
208.10	314.0
281.00	500.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

m/z	Abundance
36.00	343.0
38.00	841.0
39.00	5187.0
40.00	1359.0
44.00	2306.0
49.10	900.0
50.10	18224.0
51.10	70120.0
52.10	3225.0
54.90	361.0
56.00	2162.0
57.00	5289.0
61.00	859.0
62.00	1300.0
63.00	3346.0
64.00	406.0
65.10	1713.0
68.00	1709.0
69.00	75368.0
69.90	315.0
73.00	1301.0
74.00	7775.0
75.00	14614.0
76.00	4326.0
77.10	86512.0
78.00	5723.0
79.00	7292.0
80.00	4896.0
81.00	6268.0
82.00	1800.0
83.00	1511.0
84.10	425.0
85.00	1313.0
86.00	2082.0
87.00	803.0
88.00	457.0
91.00	1404.0
92.00	1603.0
93.00	11806.0
94.00	874.0
96.10	542.0
96.90	375.0
98.00	9774.0
99.00	7999.0
99.90	679.0
101.00	3154.0
103.00	1354.0
103.90	2215.0
105.00	2453.0
106.00	508.0
107.00	28400.0
108.10	4082.0
109.10	987.0
110.00	48864.0
111.00	7553.0
112.00	1120.0
112.90	430.0
115.90	1509.0
117.00	28096.0
117.90	1894.0
119.90	494.0
122.00	1692.0
122.90	2246.0
124.00	1310.0
125.00	1167.0
127.00	84984.0
128.00	7664.0
129.00	35912.0
130.00	3493.0
130.90	493.0
132.00	326.0
132.90	320.0
134.10	1016.0
135.10	2917.0
135.90	1354.0
137.10	1711.0
137.80	316.0
139.00	395.0
139.90	602.0
141.00	5963.0
142.00	1292.0
143.00	1235.0
144.00	351.0
146.00	1140.0
147.00	3079.0
148.00	5266.0
149.00	1318.0
150.00	458.0
151.10	351.0
153.00	1528.0
153.90	1458.0
155.00	2894.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

156.00	4111.0
157.00	955.0
158.10	1064.0
159.10	533.0
160.10	1549.0
161.00	2391.0
161.90	465.0
163.80	336.0
164.90	1922.0
166.00	1329.0
167.00	10483.0
168.00	7498.0
169.00	1111.0
170.90	454.0
171.90	980.0
173.00	1275.0
174.00	2140.0
175.00	4363.0
176.00	1223.0
176.90	1559.0
177.90	618.0
179.00	7195.0
180.00	5082.0
181.00	2323.0
181.80	347.0
184.00	823.0
185.00	3790.0
186.00	25104.0
187.00	7123.0
188.00	1112.0
189.00	2212.0
191.00	824.0
192.00	2360.0
193.10	3015.0
193.90	561.0
195.10	401.0
196.00	5082.0
197.00	1117.0
198.00	179840.0
199.00	13534.0
200.00	1195.0
201.40	600.0
202.90	1255.0
204.00	7159.0
205.00	11832.0
206.00	49240.0
207.10	9015.0
208.00	2198.0
209.00	978.0
210.00	868.0
211.00	2452.0
211.80	545.0
215.00	801.0
216.00	1042.0
216.90	12775.0
217.90	2070.0
221.00	6915.0
222.00	1757.0
223.00	2775.0
224.00	27976.0
225.00	7842.0
226.00	993.0
227.00	13768.0
228.00	1798.0
229.00	2317.0
230.00	539.0
231.00	1070.0
234.00	954.0
235.00	917.0
236.00	692.0
237.00	1697.0
239.00	441.0
240.10	484.0
240.90	834.0
242.10	1401.0
243.00	1621.0
244.00	21496.0
245.00	3034.0
246.00	5083.0
247.10	631.0
249.00	853.0
250.90	435.0
253.00	587.0
253.90	530.0
255.00	113576.0
256.00	17352.0
257.00	1751.0
258.00	7703.0
258.90	1480.0
265.00	2601.0
266.00	630.0
270.90	419.0
273.00	3045.0
274.00	9059.0
275.00	46008.0
276.00	6090.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

277.00	4587.0
278.00	909.0
281.10	395.0
282.90	722.0
284.90	986.0
292.90	1079.0
294.00	308.0
294.90	433.0
296.00	14033.0
297.10	1821.0
303.00	1680.0
303.80	394.0
314.00	572.0
315.00	1904.0
316.00	879.0
320.80	428.0
323.00	4499.0
324.10	656.0
327.00	1216.0
328.00	339.0
333.00	603.0
334.00	2937.0
335.10	1092.0
341.10	583.0
345.90	983.0
352.00	1036.0
353.10	797.0
354.10	1843.0
354.90	395.0
365.00	5109.0
366.00	1121.0
372.10	2062.0
373.00	714.0
383.00	472.0
402.00	840.0
402.90	1031.0
404.10	304.0
420.90	774.0
421.90	819.0
423.00	5536.0
424.00	1811.0
441.00	16688.0
442.00	104272.0
443.10	22176.0
444.10	2165.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

m/z	Abundance
36.90	388.0
38.00	1106.0
39.10	7661.0
40.00	1398.0
44.00	1807.0
48.90	1020.0
50.10	26056.0
51.10	104440.0
52.00	5699.0
53.00	376.0
55.10	717.0
56.00	2854.0
57.00	6388.0
58.00	322.0
61.00	1399.0
62.00	1512.0
63.00	4622.0
63.90	632.0
65.00	2512.0
65.90	345.0
67.00	406.0
68.10	1953.0
69.00	113888.0
70.10	696.0
73.00	1151.0
74.00	12020.0
75.00	20752.0
76.10	6003.0
77.10	128728.0
78.10	9712.0
79.00	8595.0
80.00	6730.0
81.00	9319.0
82.00	2277.0
83.00	2179.0
84.00	517.0
85.00	1720.0
86.00	3292.0
87.00	1577.0
87.90	393.0
88.90	448.0
91.00	2126.0
92.00	2441.0
93.00	16166.0
94.00	1299.0
96.00	733.0
96.90	318.0
98.00	12431.0
99.00	11596.0
99.90	907.0
101.00	5554.0
101.90	410.0
102.90	1609.0
104.00	3361.0
105.00	2970.0
106.10	1132.0
107.00	43624.0
108.00	7184.0
109.00	1191.0
110.00	75176.0
111.00	10434.0
112.00	2080.0
112.90	642.0
116.00	2223.0
117.00	40464.0
118.00	2868.0
118.90	500.0
120.00	485.0
122.00	2876.0
123.00	4233.0
124.00	1871.0
125.10	1171.0
126.20	597.0
127.00	134656.0
128.00	9913.0
129.00	56880.0
129.90	4191.0
131.00	1271.0
132.00	579.0
132.90	454.0
133.90	2135.0
135.00	4826.0
135.90	1939.0
137.00	3230.0
137.90	604.0
140.00	840.0
141.00	6862.0
142.00	2334.0
143.00	1848.0
144.10	368.0
145.10	559.0
146.00	1514.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

147.00	4498.0
148.00	10807.0
149.00	2017.0
150.00	623.0
151.30	947.0
152.00	370.0
153.00	2926.0
153.90	1483.0
155.00	4150.0
156.00	6929.0
156.90	1335.0
158.00	1542.0
159.00	1380.0
160.00	2877.0
161.00	3435.0
162.00	642.0
163.00	381.0
165.00	3479.0
166.10	2234.0
167.00	16960.0
168.00	9565.0
169.00	1518.0
170.00	415.0
170.90	704.0
171.90	1477.0
172.90	1799.0
174.00	3050.0
175.00	6134.0
176.00	1324.0
177.00	2798.0
178.00	1160.0
179.00	11235.0
180.00	9006.0
181.00	3788.0
182.10	566.0
183.90	944.0
185.00	6107.0
186.00	45296.0
187.00	12468.0
188.00	1553.0
189.00	2646.0
189.90	774.0
191.00	1390.0
192.00	4416.0
193.00	4703.0
194.00	1172.0
194.90	314.0
196.00	8577.0
197.00	1163.0
198.00	306880.0
199.00	21544.0
200.00	1707.0
201.50	895.0
203.00	2557.0
204.00	12591.0
205.00	20920.0
206.00	84576.0
207.00	13085.0
208.00	3519.0
208.90	1206.0
209.90	1055.0
210.20	1094.0
211.10	3664.0
213.00	372.0
215.00	1593.0
216.00	2217.0
217.00	23744.0
218.00	2571.0
219.10	303.0
220.30	735.0
221.00	10991.0
222.00	2993.0
223.00	4781.0
224.00	51488.0
225.00	12570.0
226.00	1437.0
227.00	22248.0
228.00	3824.0
229.00	4512.0
230.10	674.0
231.00	1539.0
232.90	325.0
233.90	1771.0
235.00	1953.0
235.90	1153.0
237.00	2102.0
239.00	997.0
240.00	809.0
241.00	1038.0
242.00	2327.0
243.00	2644.0
244.10	35904.0
245.00	4923.0
246.00	8623.0
246.90	1719.0
248.10	359.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

248.90	1178.0
250.80	404.0
251.90	433.0
252.90	1013.0
254.10	1164.0
255.00	202880.0
256.00	29472.0
257.00	2500.0
258.00	13382.0
259.00	1403.0
265.00	5282.0
265.90	865.0
271.10	603.0
272.00	543.0
273.00	6153.0
274.00	16416.0
275.00	91680.0
276.00	11933.0
277.00	8125.0
278.10	1822.0
281.00	396.0
283.00	1046.0
284.00	750.0
285.00	1454.0
289.00	305.0
289.80	316.0
291.90	345.0
293.00	1750.0
294.00	355.0
294.90	562.0
296.00	28320.0
296.90	4270.0
297.90	359.0
302.00	574.0
303.10	2724.0
304.00	878.0
307.90	427.0
314.10	1020.0
315.00	3154.0
316.00	1634.0
321.10	726.0
321.90	422.0
323.00	7466.0
323.90	1147.0
327.00	1310.0
328.00	851.0
332.00	626.0
333.10	854.0
334.00	5202.0
335.10	1190.0
341.00	906.0
341.90	370.0
346.00	1343.0
347.00	489.0
352.00	2744.0
352.90	1399.0
354.00	2327.0
354.90	527.0
365.00	12138.0
366.00	1728.0
370.10	304.0
371.00	746.0
372.00	4309.0
373.10	756.0
373.90	300.0
383.10	1218.0
384.00	376.0
390.90	402.0
392.00	342.0
401.10	311.0
402.00	1311.0
403.00	2239.0
403.90	956.0
421.00	1742.0
422.00	1799.0
423.00	11580.0
424.00	2733.0
425.10	412.0
441.00	34752.0
442.00	231040.0
443.00	44032.0
444.00	3978.0
445.10	384.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

m/z	Abundance
36.90	448.0
38.00	826.0
39.00	6293.0
40.00	1376.0
41.00	314.0
44.00	2222.0
45.00	345.0
49.00	830.0
50.10	21064.0
51.10	86040.0
52.10	3967.0
55.00	698.0
56.00	2932.0
57.00	5646.0
58.00	363.0
60.90	1185.0
62.00	1297.0
63.00	3467.0
64.00	554.0
65.00	1914.0
67.90	1566.0
69.00	94152.0
70.10	657.0
73.10	1108.0
74.00	10488.0
75.00	17552.0
76.00	5378.0
77.10	112912.0
78.00	6965.0
79.00	8728.0
80.00	6015.0
81.00	8715.0
82.00	2296.0
83.00	1545.0
83.80	614.0
85.00	1564.0
86.00	2246.0
86.90	1209.0
87.90	385.0
88.90	443.0
91.00	2321.0
92.00	2158.0
93.00	14398.0
94.00	871.0
96.10	709.0
98.00	12446.0
99.00	9408.0
99.90	670.0
101.00	4776.0
101.90	315.0
102.90	1409.0
104.00	3254.0
105.00	3606.0
106.10	987.0
107.00	37192.0
108.00	5978.0
109.00	1175.0
110.00	64632.0
111.00	10622.0
112.00	1449.0
112.90	387.0
116.00	2184.0
117.00	36272.0
117.90	3182.0
119.00	372.0
119.90	602.0
122.00	2552.0
123.00	3959.0
124.10	1613.0
125.00	1784.0
126.20	496.0
127.00	121912.0
128.00	9614.0
129.00	47136.0
130.10	4149.0
131.00	787.0
132.10	393.0
132.80	448.0
133.90	1181.0
135.00	4048.0
136.00	2004.0
137.00	2326.0
140.00	749.0
141.00	7578.0
142.00	2463.0
142.90	1656.0
143.90	419.0
146.00	1394.0
147.00	3671.0
148.00	9903.0
149.00	1869.0
149.90	534.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

151.10	864.0
151.70	645.0
153.00	2806.0
154.00	1421.0
155.00	3852.0
156.00	5721.0
157.00	1374.0
157.90	1465.0
159.10	966.0
160.00	2009.0
161.00	4325.0
162.00	1164.0
164.00	720.0
165.00	2947.0
166.00	2125.0
167.00	16038.0
168.00	9533.0
169.00	1301.0
169.90	374.0
171.00	592.0
171.90	1608.0
173.00	1953.0
174.10	3278.0
175.00	6119.0
176.00	1515.0
176.90	3323.0
178.10	781.0
179.00	11364.0
180.00	7983.0
181.00	3485.0
182.10	421.0
184.00	1161.0
185.00	6002.0
186.00	40640.0
187.00	12827.0
188.10	1844.0
189.00	2983.0
190.00	740.0
191.00	1624.0
191.90	3750.0
193.00	4072.0
193.90	1087.0
194.90	397.0
196.00	7847.0
197.00	1135.0
198.00	289472.0
199.00	19136.0
200.00	1253.0
201.50	1235.0
202.10	300.0
202.90	2708.0
204.00	12818.0
205.10	20096.0
206.00	83088.0
207.00	11397.0
208.00	3400.0
208.90	1071.0
210.10	1681.0
211.00	3184.0
211.80	567.0
213.20	301.0
213.90	376.0
215.00	1073.0
216.00	1960.0
217.00	22936.0
218.00	2784.0
221.10	11203.0
221.90	2622.0
223.00	5324.0
224.00	50784.0
225.00	12642.0
226.00	1389.0
227.00	22432.0
228.00	3233.0
229.00	4735.0
230.00	925.0
231.00	1624.0
232.10	454.0
232.90	436.0
234.00	1838.0
235.00	1606.0
236.00	1149.0
237.00	1584.0
237.90	354.0
238.90	865.0
240.10	816.0
241.00	1065.0
242.00	2708.0
243.00	2514.0
244.00	37680.0
245.00	3969.0
246.00	8788.0
246.90	1635.0
248.00	644.0
248.90	1309.0
249.90	561.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47

252.90	1061.0
254.00	1650.0
255.00	212416.0
256.00	29672.0
257.00	2392.0
258.00	12591.0
259.00	2148.0
260.90	306.0
264.00	370.0
265.00	5504.0
266.00	803.0
270.00	429.0
271.00	422.0
272.00	574.0
273.00	5925.0
274.00	16936.0
275.00	89408.0
276.00	12800.0
277.00	9026.0
278.00	1484.0
279.00	461.0
281.10	657.0
281.90	501.0
282.90	954.0
283.90	617.0
285.00	1575.0
286.10	434.0
290.00	377.0
291.80	342.0
293.00	1514.0
293.90	657.0
295.00	331.0
296.00	28040.0
297.00	4221.0
297.90	350.0
301.00	319.0
302.00	312.0
303.00	3483.0
304.00	842.0
307.90	327.0
309.00	360.0
309.90	347.0
314.00	1392.0
315.00	2906.0
316.10	1643.0
321.10	701.0
322.00	426.0
323.00	8823.0
324.10	1480.0
327.00	1717.0
328.00	958.0
331.90	729.0
333.00	916.0
334.00	6617.0
335.00	1181.0
341.00	1116.0
346.00	2276.0
347.00	386.0
352.00	2849.0
353.00	1785.0
354.10	2749.0
354.90	560.0
365.00	13246.0
365.90	2046.0
371.00	577.0
372.00	4131.0
373.10	1132.0
383.00	943.0
383.80	432.0
390.00	731.0
390.80	356.0
402.00	1613.0
403.00	2823.0
404.00	1096.0
421.00	2231.0
421.90	2377.0
423.00	14579.0
424.10	3206.0
425.10	456.0
441.00	40168.0
442.00	284160.0
443.10	54568.0
444.10	4581.0

Instrument :
BNA_M
ClientSampleId :
DFTPP47