

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM081718\
 Data File : BM016452.D
 Acq On : 18 Aug 2018 00:08
 Operator : SJ/JU
 Sample : J4521-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 RT2286

Quant Time: Aug 18 03:08:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM081718.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 17 16:27:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	9269	20.00	ng	0.00
21) Naphthalene-d8	10.87	136	35616	20.00	ng	0.00
38) Acenaphthene-d10	14.70	164	24052	20.00	ng	0.00
63) Phenanthrene-d10	17.44	188	67724	20.00	ng	0.00
75) Chrysene-d12	21.58	240	98127	20.00	ng	0.00
86) Perylene-d12	23.89	264	112730	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.59	112	87684	158.98	ng	0.00
7) Phenol-d6	7.23	99	96143	129.69	ng	0.00
23) Nitrobenzene-d5	9.24	82	99528	116.38	ng	0.00
41) 2,4,6-Tribromophenol	16.19	330	77952	214.58	ng	0.00
44) 2-Fluorobiphenyl	13.32	172	244277	105.61	ng	0.00
78) Terphenyl-d14	20.03	244	790622	121.78	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM081718\
Data File : BM016452.D
Acq On : 18 Aug 2018 00:08
Operator : SJ/JU
Sample : J4521-03
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
RT2286

Quant Time: Aug 18 03:08:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\8270-BM081718.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 17 16:27:00 2018
Response via : Initial Calibration

