


Hexachlorocyclopentadiene

Response Ratio




$R = 3.21e-002 A^2 + 2.60e-001 A - 1.13e-001$
Coef of Det (r^2) = 0.994 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM081718.M
Calibration Table Last Updated: Fri Aug 17 16:27:00 2018

2,4-Dinitrophenol

Response Ratio



Resp Ratio = $2.16 \times 10^{-1} \times \text{Amt} - 8.00 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM081718.M
Calibration Table Last Updated: Fri Aug 17 16:27:00 2018