

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
 Data File : BMCCC323.D
 Acq On : 17 Aug 2022 19:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 18 02:25:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 15:00:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.798	152	269601	20.000	ng/u1	0.00
20) Naphthalene-d8	10.604	136	1173023	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	646515	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1171539	20.000	ng/u1	0.00
79) Chrysene-d12	21.368	240	1028129	20.000	ng/u1	0.00
88) Perylene-d12	23.697	264	1054211	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	50576	7.149	ng/uL	0.00
4) Pyridine-d5	3.634	84	338212	17.337	ng/u1	0.00
7) Phenol-d5	6.981	99	424605	18.451	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	272329	18.800	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	354052	19.581	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	327417	19.055	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	174274	19.595	ng/u1	0.00
24) 2-Nitrophenol-d4	9.692	143	179021	19.844	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.234	165	318229	19.742	ng/u1	0.00
31) 4-Chloroaniline-d4	10.751	131	454377	18.183	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	790671	19.331	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	998114	19.566	ng/u1	0.00
54) 4-Nitrophenol-d4	14.674	143	127326	16.429	ng/u1	0.00
60) Fluorene-d10	15.439	176	696413	18.843	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	102335	16.991	ng/u1	0.00
73) Anthracene-d10	17.286	188	998765	19.173	ng/u1	0.00
81) Pyrene-d10	19.580	212	1037420	19.452	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.550	264	1004225	19.429	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	52979	7.110	ng/uL	93
5) Pyridine	3.658	79	352082	17.471	ng/u1	88
6) Benzaldehyde	6.946	77	197094	20.850	ng/u1	95
8) Phenol	7.010	94	445251	18.609	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.234	93	361764	18.904	ng/u1	98
12) 2-Chlorophenol	7.369	128	369130	19.523	ng/u1	99
13) 2-Methylphenol	8.257	108	337697	19.147	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.340	45	492358	19.230	ng/u1	97
16) Acetophenone	8.634	105	542711	19.638	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.616	70	271175	19.714	ng/u1	96
18) 4-Methylphenol	8.593	108	360776	19.110	ng/u1	99
19) Hexachloroethane	8.875	117	155701	19.683	ng/u1	89
22) Nitrobenzene	9.010	77	401612	19.058	ng/u1	98
23) Isophorone	9.540	82	720901	19.264	ng/u1	97
25) 2-Nitrophenol	9.722	139	195891	19.490	ng/u1	97
26) 2,4-Dimethylphenol	9.787	107	390716	19.439	ng/u1	96
27) Bis(2-Chloroethoxy)met...	10.028	93	467510	19.190	ng/u1	99
29) 2,4-Dichlorophenol	10.257	162	322810	19.574	ng/u1	99
30) Naphthalene	10.651	128	1171237	19.024	ng/u1	99
32) 4-Chloroaniline	10.775	127	462636	18.202	ng/u1	98
33) Hexachlorobutadiene	10.928	225	206344	19.916	ng/u1	99
34) Caprolactam	11.569	113	85751	17.354	ng/u1	91
35) 4-Chloro-3-methylphenol	11.910	107	317487	19.040	ng/u1	99
36) 2-Methylnaphthalene	12.269	142	736527	19.067	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
 Data File : BMCCC323.D
 Acq On : 17 Aug 2022 19:50
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 16 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :


Quant Time: Aug 18 02:25:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 15:00:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.486	142	739217	19.002	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.633	216	352705	19.738	ng/ul	99
40) Hexachlorocyclopentadiene	12.604	237	180251	17.003	ng/ul	97
41) 2,4,6-Trichlorophenol	12.886	196	228485	20.320	ng/ul	99
42) 2,4,5-Trichlorophenol	12.957	196	237099	19.879	ng/ul	98
43) 1,1'-Biphenyl	13.280	154	946123	19.458	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	726963	19.483	ng/ul	99
45) 2-Nitroaniline	13.539	65	196278	19.426	ng/ul	96
47) Dimethylphthalate	13.910	163	810528	19.252	ng/ul	100
48) 2,6-Dinitrotoluene	14.033	165	158603	19.966	ng/ul	88
50) Acenaphthylene	14.169	152	1165032	19.366	ng/ul	99
51) 3-Nitroaniline	14.369	138	153998	17.727	ng/ul	99
52) Acenaphthene	14.510	153	747056	19.052	ng/ul	99
53) 2,4-Dinitrophenol	14.580	184	63012	14.555	ng/ul	97
55) 4-Nitrophenol	14.686	109	101148	16.518	ng/ul	83
56) Dibenzofuran	14.845	168	1027810	18.891	ng/ul	98
57) 2,4-Dinitrotoluene	14.822	165	219733	19.505	ng/ul#	96
58) 2,3,4,6-Tetrachlorophenol	15.074	232	177685	19.621	ng/ul#	94
59) Diethylphthalate	15.269	149	797353	19.212	ng/ul	99
61) Fluorene	15.492	166	805206	18.659	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.486	204	389037	19.299	ng/ul	97
63) 4-Nitroaniline	15.527	138	128036	15.800	ng/ul	90
66) 4,6-Dinitro-2-methylph...	15.580	198	104457	17.153	ng/ul	92
67) N-Nitrosodiphenylamine	15.704	169	656234	19.593	ng/ul	99
68) 4-Bromophenyl-phenylether	16.380	248	218507	20.143	ng/ul	95
69) Hexachlorobenzene	16.486	284	256782	19.533	ng/ul	94
70) Atrazine	16.663	200	215248	18.809	ng/ul	97
71) Pentachlorophenol	16.845	266	129968	17.933	ng/ul	97
72) Phenanthrene	17.227	178	1160812	18.848	ng/ul	98
74) Anthracene	17.321	178	1177772	19.075	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.245	216	359981	20.479	ng/ul	97
76) Pentachlorobenzene	14.757	250	325970	19.860	ng/ul	98
77) Carbazole	17.598	167	1052668	18.209	ng/ul	99
78) Di-n-butylphthalate	18.162	149	1231799	19.481	ng/ul	99
80) Fluoranthene	19.245	202	1268956	19.687	ng/ul	98
82) Pyrene	19.609	202	1309142	19.463	ng/ul	97
83) Butylbenzylphthalate	20.509	149	523516	20.805	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.292	252	408671	19.303	ng/ul	98
85) Benzo(a)anthracene	21.356	228	1204812	19.500	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.286	149	751783	21.638	ng/ul	99
87) Chrysene	21.403	228	1173930	19.337	ng/ul	99
89) Di-n-octyl phthalate	22.186	149	1253545	20.003	ng/ul	100
90) Benzo(b)fluoranthene	22.992	252	1276225	19.856	ng/ul	97
91) Benzo(k)fluoranthene	23.039	252	1148442	18.781	ng/ul	97
93) Benzo(a)pyrene	23.597	252	1226678	19.400	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.115	276	1428823	19.219	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.133	278	1238677	19.342	ng/ul	97
96) Benzo(g,h,i)perylene	26.856	276	1223178	19.240	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Aug 18 02:25:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 16 15:00:33 2022
Response via : Initial Calibration








#2
 1,4-Dioxane
 Concen: 7.110 ng/uL
 RT: 3.246 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Instrument :
 BNA_M
 ClientSampleId :



Tgt Ion: 88 Resp: 52979




Ion	Ratio	Lower	Upper
88	100		
43	28.7	27.8	41.8
58	73.1	61.8	92.8




#5
 Pyridine
 Concen: 17.471 ng/uL
 RT: 3.658 min Scan# 97
 Delta R.T. 0.000 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Tgt Ion: 79 Resp: 352082

Ion	Ratio	Lower	Upper
79	100		
52	71.1	67.8	101.6
51	35.0	31.4	47.0









#6
Benzaldehyde
Concen: 20.850 ng/ul
RT: 6.946 min Scan# 656
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 197094
Ion Ratio Lower Upper
77 100
105 99.8 84.6 127.0
106 95.8 80.3 120.5



#8
Phenol
Concen: 18.609 ng/ul
RT: 7.010 min Scan# 667
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion: 94 Resp: 445251
Ion Ratio Lower Upper
94 100
65 28.9 21.3 31.9
66 36.7 27.4 41.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 18.904 ng/ul
 RT: 7.234 min Scan# 705
 Delta R.T. 0.000 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion: 93 Resp: 361764




Ion	Ratio	Lower	Upper
93	100		
63	77.4	64.0	96.0
95	32.6	25.8	38.6




#12
 2-Chlorophenol
 Concen: 19.523 ng/ul
 RT: 7.369 min Scan# 728
 Delta R.T. 0.000 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Tgt Ion:128 Resp: 369130

Ion	Ratio	Lower	Upper
128	100		
64	45.2	36.9	55.3
130	31.5	25.6	38.4









#13
2-Methylphenol
Concen: 19.147 ng/ul
RT: 8.257 min Scan# 81
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 337697
Ion Ratio Lower Upper
108 100
107 88.1 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.230 ng/ul
RT: 8.340 min Scan# 893
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion: 45 Resp: 492358
Ion Ratio Lower Upper
45 100
77 14.6 10.6 15.8
79 11.7 8.3 12.5









#16
Acetophenone
Concen: 19.638 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 542711
Ion Ratio Lower Upper
105 100
77 75.9 59.8 89.6
51 26.9 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 19.714 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion: 70 Resp: 271175
Ion Ratio Lower Upper
70 100
42 59.1 49.6 74.4
101 9.9 8.1 12.1
130 21.1 18.4 27.6









#18
4-Methylphenol
Concen: 19.110 ng/ul
RT: 8.593 min Scan# 911
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 360776
Ion Ratio Lower Upper
108 100
107 111.9 90.3 135.5



#19
Hexachloroethane
Concen: 19.683 ng/ul
RT: 8.875 min Scan# 984
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:117 Resp: 155701
Ion Ratio Lower Upper
117 100
201 91.0 81.8 122.6
199 56.6 51.6 77.4









#22
Nitrobenzene
Concen: 19.058 ng/ul
RT: 9.010 min Scan# 1008
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 401612
Ion Ratio Lower Upper
77 100
123 45.4 37.4 56.0
65 13.4 10.8 16.2



#23
Isophorone
Concen: 19.264 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion: 82 Resp: 720901
Ion Ratio Lower Upper
82 100
95 6.5 5.5 8.3
138 17.5 15.3 22.9









#25
2-Nitrophenol
Concen: 19.490 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 195891
Ion Ratio Lower Upper
139 100
65 51.4 39.8 59.6
109 30.1 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 19.439 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:107 Resp: 390716
Ion Ratio Lower Upper
107 100
121 49.4 42.7 64.1
122 87.1 71.5 107.3





#27

Bis(2-Chloroethoxy)methane

Concen: 19.190 ng/ul

RT: 10.028 min Scan# 1180

Delta R.T. 0.000 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50

Instrument :

BNA_M

ClientSampleId :





Tgt Ion: 93 Resp: 467510

Ion Ratio Lower Upper

93 100

95 32.1 26.1 39.1

123 11.6 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 19.574 ng/ul

RT: 10.257 min Scan# 1219

Delta R.T. -0.006 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50




Tgt Ion:162 Resp: 322810


Ion Ratio Lower Upper

162 100

164 65.7 51.8 77.6

98 36.1 29.5 44.3









#30
Naphthalene
Concen: 19.024 ng/ul
RT: 10.651 min Scan# 11
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 1171237
Ion Ratio Lower Upper
128 100
129 10.5 8.8 13.2
127 12.6 10.6 16.0



#32
4-Chloroaniline
Concen: 18.202 ng/ul
RT: 10.775 min Scan# 1307
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:127 Resp: 462636
Ion Ratio Lower Upper
127 100
129 32.6 25.3 37.9









#33
Hexachlorobutadiene
Concen: 19.916 ng/ul
RT: 10.928 min Scan# 11
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:225 Resp: 206344
Ion Ratio Lower Upper
225 100
223 63.2 50.0 75.0
227 63.8 49.9 74.9



#34
Caprolactam
Concen: 17.354 ng/ul
RT: 11.569 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:113 Resp: 85751
Ion Ratio Lower Upper
113 100
55 167.7 150.2 225.4
56 125.7 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 19.040 ng/ul
RT: 11.910 min Scan# 1500
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:107 Resp: 317487
Ion Ratio Lower Upper
107 100
144 26.0 20.3 30.5
142 81.3 64.4 96.6



#36
2-Methylnaphthalene
Concen: 19.067 ng/ul
RT: 12.269 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:142 Resp: 736527
Ion Ratio Lower Upper
142 100
141 89.6 70.7 106.1









#37
1-Methylnaphthalene
Concen: 19.002 ng/ul
RT: 12.486 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:142 Resp: 739217
Ion Ratio Lower Upper
142 100
141 92.2 75.2 112.8
116 3.8 2.6 3.8



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.738 ng/ul
RT: 12.633 min Scan# 1623
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:216 Resp: 352705
Ion Ratio Lower Upper
216 100
214 78.3 62.4 93.6
179 20.8 17.3 25.9
108 18.5 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 17.003 ng/ul
RT: 12.604 min Scan# 1618
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:237 Resp: 180251
Ion Ratio Lower Upper
237 100
235 62.4 48.4 72.6
272 12.9 11.6 17.4



#41
2,4,6-Trichlorophenol
Concen: 20.320 ng/ul
RT: 12.886 min Scan# 1666
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:196 Resp: 228485
Ion Ratio Lower Upper
196 100
198 97.4 77.8 116.6
200 29.9 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 19.879 ng/ul
RT: 12.957 min Scan# 1
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:196 Resp: 237099
Ion Ratio Lower Upper
196 100
198 94.6 76.8 115.2
200 29.1 24.9 37.3



#43
1,1'-Biphenyl
Concen: 19.458 ng/ul
RT: 13.280 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:154 Resp: 946123
Ion Ratio Lower Upper
154 100
153 40.3 32.4 48.6
76 13.2 10.1 15.1









#44
2-Chloronaphthalene
Concen: 19.483 ng/ul
RT: 13.322 min Scan# 11
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:162 Resp: 726963
Ion Ratio Lower Upper
162 100
164 32.0 25.4 38.0
127 39.0 31.6 47.4



#45
2-Nitroaniline
Concen: 19.426 ng/ul
RT: 13.539 min Scan# 1777
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion: 65 Resp: 196278
Ion Ratio Lower Upper
65 100
92 69.3 54.5 81.7
138 107.7 91.2 136.8









#47
Dimethylphthalate
Concen: 19.252 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:163 Resp: 810528
Ion Ratio Lower Upper
163 100
194 3.7 3.2 4.8
164 9.8 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 19.966 ng/ul
RT: 14.033 min Scan# 1861
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:165 Resp: 158603
Ion Ratio Lower Upper
165 100
89 63.6 42.5 63.7
121 24.4 17.2 25.8





#50

Acenaphthylene

Concen: 19.366 ng/ul

RT: 14.169 min Scan# 1884

Delta R.T. 0.000 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:152 Resp: 1165032

Ion Ratio Lower Upper

152 100

151 19.7 16.1 24.1

153 12.9 10.6 15.8



#51

3-Nitroaniline

Concen: 17.727 ng/ul

RT: 14.369 min Scan# 1918

Delta R.T. 0.000 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50




Tgt Ion:138 Resp: 153998


Ion Ratio Lower Upper

138 100

108 9.0 8.1 12.1


92 111.6 90.4 135.6








#52
Acenaphthene
Concen: 19.052 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:153 Resp: 747056
Ion Ratio Lower Upper
153 100
152 48.0 37.1 55.7
154 88.5 71.3 106.9




#53
2,4-Dinitrophenol
Concen: 14.555 ng/ul
RT: 14.580 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50



Tgt Ion:184 Resp: 63012
Ion Ratio Lower Upper
184 100
63 70.2 54.2 81.4
154 63.7 49.3 73.9









#55
4-Nitrophenol
Concen: 16.518 ng/ul
RT: 14.686 min Scan# 1972
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 101148
Ion Ratio Lower Upper
109 100
139 133.9 124.9 187.3
65 130.0 122.2 183.4



#56
Dibenzofuran
Concen: 18.891 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:168 Resp: 1027810
Ion Ratio Lower Upper
168 100
139 38.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 19.505 ng/ul
RT: 14.822 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:165 Resp: 219733
Ion Ratio Lower Upper
165 100
63 47.4 35.9 53.9
182 2.3 2.5 3.7#



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.621 ng/ul
RT: 15.074 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:232 Resp: 177685
Ion Ratio Lower Upper
232 100
131 50.9 36.7 55.1
130 2.9 1.7 2.5#
166 32.6 24.4 36.6









#59
Diethylphthalate
Concen: 19.212 ng/ul
RT: 15.269 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 797353
Ion Ratio Lower Upper
149 100
177 20.7 16.8 25.2
150 12.4 10.2 15.4



#61
Fluorene
Concen: 18.659 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:166 Resp: 805206
Ion Ratio Lower Upper
166 100
165 98.8 77.2 115.8
167 13.9 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 19.299 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:204 Resp: 389037
Ion Ratio Lower Upper
204 100
206 32.7 25.9 38.9
141 60.2 50.6 75.8



#63
4-Nitroaniline
Concen: 15.800 ng/ul
RT: 15.527 min Scan# 2115
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:138 Resp: 128036
Ion Ratio Lower Upper
138 100
92 48.1 39.2 58.8
108 76.9 51.6 77.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.153 ng/ul
 RT: 15.580 min Scan# 2125
 Delta R.T. -0.006 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:198 Resp: 104457
 Ion Ratio Lower Upper
 198 100
 182 5.1 4.2 6.4
 77 29.4 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 19.593 ng/ul
 RT: 15.704 min Scan# 2145
 Delta R.T. -0.006 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Tgt Ion:169 Resp: 656234
 Ion Ratio Lower Upper
 169 100
 168 66.8 53.8 80.6
 167 36.3 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 20.143 ng/ul
RT: 16.380 min Scan# 21
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:248 Resp: 218507
Ion Ratio Lower Upper
248 100
250 100.4 77.3 115.9
141 79.4 59.8 89.8



#69
Hexachlorobenzene
Concen: 19.533 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:284 Resp: 256782
Ion Ratio Lower Upper
284 100
142 40.3 28.6 42.8
249 29.0 24.6 37.0





#70

Atrazine

Concen: 18.809 ng/ul

RT: 16.663 min Scan# 21

Delta R.T. 0.000 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:200 Resp: 215248

Ion Ratio Lower Upper

200 100

173 27.3 21.3 31.9

215 46.1 39.0 58.6



#71

Pentachlorophenol

Concen: 17.933 ng/ul

RT: 16.845 min Scan# 2339

Delta R.T. -0.006 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50




Tgt Ion:266 Resp: 129968


Ion Ratio Lower Upper

266 100

264 65.7 49.5 74.3

268 64.2 51.0 76.4





#72

Phenanthrene

Concen: 18.848 ng/ul

RT: 17.227 min Scan# 2404

Delta R.T. -0.006 min


Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:178 Resp: 1160812

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 19.7 14.9 22.3



#74

Anthracene

Concen: 19.075 ng/ul

RT: 17.321 min Scan# 2420

Delta R.T. 0.000 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50




Tgt Ion:178 Resp: 1177772


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 19.0 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.479 ng/uL
RT: 13.245 min Scan# 11
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:216 Resp: 359981
Ion Ratio Lower Upper
216 100
214 78.7 60.9 91.3
218 48.1 37.5 56.3



#76
Pentachlorobenzene
Concen: 19.860 ng/uL
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50


Tgt Ion:250 Resp: 325970
Ion Ratio Lower Upper
250 100
248 64.9 50.4 75.6
252 65.4 51.1 76.7








#77
 Carbazole
 Concen: 18.209 ng/ul
 RT: 17.598 min Scan# 2467
 Delta R.T. -0.006 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Instrument :
 BNA_M
 ClientSampleId :







Tgt Ion:167 Resp: 1052668
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.4 26.0
 139 13.4 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 19.481 ng/ul
 RT: 18.162 min Scan# 2563
 Delta R.T. 0.000 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Tgt Ion:149 Resp: 1231799
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.3 10.9
 104 4.9 3.7 5.5









#80
 Fluoranthene
 Concen: 19.687 ng/u1
 RT: 19.245 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:202 Resp: 1268956
 Ion Ratio Lower Upper
 202 100
 101 12.0 10.1 15.1
 100 9.0 7.8 11.6



#82
 Pyrene
 Concen: 19.463 ng/u1
 RT: 19.609 min Scan# 2809
 Delta R.T. 0.000 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Tgt Ion:202 Resp: 1309142
 Ion Ratio Lower Upper
 202 100
 101 13.8 11.9 17.9
 100 11.2 10.2 15.4









#83
 Butylbenzylphthalate
 Concen: 20.805 ng/ul
 RT: 20.509 min Scan# 2962
 Delta R.T. -0.006 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:149 Resp: 523516
 Ion Ratio Lower Upper
 149 100
 91 74.2 53.2 79.8
 206 17.9 16.0 24.0



#84
 3,3'-Dichlorobenzidine
 Concen: 19.303 ng/ul
 RT: 21.292 min Scan# 3095
 Delta R.T. -0.006 min
 Lab File: BMCCC323.D
 Acq: 17 Aug 2022 19:50

Tgt Ion:252 Resp: 408671
 Ion Ratio Lower Upper
 252 100
 254 63.6 52.4 78.6
 126 12.6 10.8 16.2









#85
Benzo(a)anthracene
Concen: 19.500 ng/ul
RT: 21.356 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 1204812
Ion Ratio Lower Upper
228 100
229 19.7 15.6 23.4
226 27.3 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.638 ng/ul
RT: 21.286 min Scan# 3094
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:149 Resp: 751783
Ion Ratio Lower Upper
149 100
167 26.3 21.5 32.3
279 3.7 3.4 5.2





#87

Chrysene

Concen: 19.337 ng/ul

RT: 21.403 min Scan# 3115

Delta R.T. -0.006 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:228 Resp: 1173930

Ion Ratio Lower Upper

228 100

226 29.8 23.7 35.5

229 19.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 20.003 ng/ul




RT: 22.186 min Scan# 3247


Delta R.T. -0.006 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50

Tgt Ion:149 Resp: 1253545









#90
Benzo(b)fluoranthene
Concen: 19.856 ng/ul
RT: 22.992 min Scan# 31
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 1276225
Ion Ratio Lower Upper
252 100
253 21.5 17.5 26.3
125 11.0 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 18.781 ng/ul
RT: 23.039 min Scan# 3392
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:252 Resp: 1148442
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 11.1 10.6 15.8





#93

Benzo(a)pyrene

Concen: 19.400 ng/ul

RT: 23.597 min Scan# 3488

Delta R.T. -0.006 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:252 Resp: 1226678

Ion Ratio Lower Upper

252 100

253 22.2 17.3 25.9

125 12.4 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.219 ng/ul

RT: 26.115 min Scan# 3915

Delta R.T. 0.000 min

Lab File: BMCCC323.D

Acq: 17 Aug 2022 19:50




Tgt Ion:276 Resp: 1428823


Ion Ratio Lower Upper

276 100

138 25.2 25.6 38.4#

277 24.9 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 19.342 ng/ul
RT: 26.133 min Scan# 3918
Delta R.T. 0.000 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Instrument :
BNA_M
ClientSampleId :

Tgt Ion:278 Resp: 1238677
Ion Ratio Lower Upper
278 100
139 17.0 16.2 24.2
279 23.5 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 19.240 ng/ul
RT: 26.856 min Scan# 4041
Delta R.T. -0.006 min
Lab File: BMCCC323.D
Acq: 17 Aug 2022 19:50

Tgt Ion:276 Resp: 1223178
Ion Ratio Lower Upper
276 100
138 23.5 21.3 31.9
277 23.4 18.6 28.0

