

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
 Data File : BM036338.D
 Acq On : 18 Aug 2022 04:48
 Operator : CG/JU
 Sample : N4176-08
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EX7Q3

Quant Time: Aug 18 05:21:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 15:00:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.798	152	270306	20.000	ng/u1	0.00
20) Naphthalene-d8	10.598	136	1161934	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.439	164	629390	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1155165	20.000	ng/u1	0.00
79) Chrysene-d12	21.368	240	1022877	20.000	ng/u1	0.00
88) Perylene-d12	23.697	264	1084608	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.210	96	24512	3.456	ng/uL	0.00
4) Pyridine-d5	3.640	84	112728	5.763	ng/u1	0.00
7) Phenol-d5	6.981	99	444587	19.269	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	291621	20.080	ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	370675	20.447	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	353081	20.495	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	186326	21.150	ng/u1	0.00
24) 2-Nitrophenol-d4	9.686	143	188405	21.084	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	315275	19.746	ng/u1	0.00
31) 4-Chloroaniline-d4	10.745	131	482783	19.504	ng/u1	-0.01
46) Dimethylphthalate-d6	13.863	166	935487	23.494	ng/u1	0.00
49) Acenaphthylene-d8	14.133	160	1097693	22.103	ng/u1	0.00
54) 4-Nitrophenol-d4	14.674	143	114372	15.159	ng/u1	0.00
60) Fluorene-d10	15.433	176	816972	22.706	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.568	200	66237	11.153	ng/u1	0.00
73) Anthracene-d10	17.280	188	1343836	26.163	ng/u1	0.00
81) Pyrene-d10	19.574	212	1490716	28.095	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.544	264	1484243	27.911	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
Data File : BM036338.D
Acq On : 18 Aug 2022 04:48
Operator : CG/JU
Sample : N4176-08
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EX7Q3

Quant Time: Aug 18 05:21:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 16 15:00:33 2022
Response via : Initial Calibration

