

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
 Data File : BM036329.D
 Acq On : 17 Aug 2022 23:26
 Operator : CG/JU
 Sample : N4176-12MS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EX7Q6MS

Quant Time: Aug 18 02:22:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 15:00:33 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 08/18/2022
 Supervised By :Sohil Jodhani 08/19/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	257191	20.000	ng/u1	0.00
20) Naphthalene-d8	10.599	136	1112863	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.445	164	646421	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1271692	20.000	ng/u1	0.00
79) Chrysene-d12	21.368	240	1221386	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	1280161	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	30912	4.581	ng/uL	0.00
4) Pyridine-d5	3.634	84	319044	17.144	ng/u1	0.00
7) Phenol-d5	6.981	99	602642	27.451	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.140	67	369393	26.732	ng/u1	0.00
11) 2-Chlorophenol-d4	7.334	132	493269	28.597	ng/u1	0.00
15) 4-Methylphenol-d8	8.528	113	440563	26.877	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	245419	29.086	ng/u1	0.00
24) 2-Nitrophenol-d4	9.687	143	259729	30.347	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	459660	30.058	ng/u1	0.00
31) 4-Chloroaniline-d4	10.751	131	594504	25.077	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	1279816	31.295	ng/u1	0.00
49) Acenaphthylene-d8	14.139	160	1539029	30.173	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	252414	32.574	ng/u1	0.00
60) Fluorene-d10	15.433	176	1150105	31.122	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	187293	28.647	ng/u1	0.00
73) Anthracene-d10	17.286	188	1823914	32.256	ng/u1	0.00
81) Pyrene-d10	19.580	212	2108244	33.275	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.551	264	2083555	33.196	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.246	88	74847	10.530	ng/uL	96
5) Pyridine	3.658	79	380857	19.811	ng/u1	90
6) Benzaldehyde	6.946	77	304073	33.720	ng/u1	94
8) Phenol	7.011	94	689087	30.190	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.234	93	542843	29.735	ng/u1	96
12) 2-Chlorophenol	7.363	128	559611	31.025	ng/u1	99
13) 2-Methylphenol	8.257	108	484616	28.803	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.340	45	729031	29.848	ng/u1	96
16) Acetophenone	8.634	105	833149	31.602	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.616	70	406645	30.990	ng/u1	97
18) 4-Methylphenol	8.593	108	532978	29.593	ng/u1	96
19) Hexachloroethane	8.869	117	231223	30.641	ng/u1	88
22) Nitrobenzene	9.010	77	617944	30.909	ng/u1	99
23) Isophorone	9.540	82	1125809	31.711	ng/u1	98
25) 2-Nitrophenol	9.722	139	309283	32.436	ng/u1	97
26) 2,4-Dimethylphenol	9.787	107	310386	16.277	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.022	93	695525	30.092	ng/u1	99
29) 2,4-Dichlorophenol	10.257	162	502923	32.144	ng/u1	98
30) Naphthalene	10.651	128	1781418	30.499	ng/u1	99
32) 4-Chloroaniline	10.775	127	538137	22.317	ng/u1	99
33) Hexachlorobutadiene	10.928	225	308658	31.402	ng/u1	99
34) Caprolactam	11.575	113	159191	33.958	ng/u1	93
35) 4-Chloro-3-methylphenol	11.910	107	538405	34.034	ng/u1	97
36) 2-Methylnaphthalene	12.263	142	1159188	31.631	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
 Data File : BM036329.D
 Acq On : 17 Aug 2022 23:26
 Operator : CG/JU
 Sample : N4176-12MS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :

BNA_M

ClientSampleId :

EX7Q6MS

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/18/2022

Supervised By :Sohil Jodhani 08/19/2022

Quant Time: Aug 18 02:22:31 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M


Quant Title : SVOA CALIBRATION

QLast Update : Tue Aug 16 15:00:33 2022

Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.487	142	1161749	31.478	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.634	216	560284	31.359	ng/ul	99
40) Hexachlorocyclopentadiene	12.604	237	202880	19.140	ng/ul	97
41) 2,4,6-Trichlorophenol	12.881	196	384973	34.241	ng/ul	99
42) 2,4,5-Trichlorophenol	12.957	196	423743	35.533	ng/ul	100
43) 1,1'-Biphenyl	13.281	154	1517833	31.220	ng/ul	99
44) 2-Chloronaphthalene	13.322	162	1174516	31.482	ng/ul	99
45) 2-Nitroaniline	13.539	65	359406	35.577	ng/ul	97
47) Dimethylphthalate	13.910	163	1437001	34.137	ng/ul	100
48) 2,6-Dinitrotoluene	14.034	165	302261	38.056	ng/ul	90
50) Acenaphthylene	14.169	152	1931223	32.107	ng/ul	99
51) 3-Nitroaniline	14.369	138	335664	38.644	ng/ul	96
52) Acenaphthene	14.510	153	1247968	31.832	ng/ul	99
53) 2,4-Dinitrophenol	14.575	184	148848	34.388	ng/ul	96
55) 4-Nitrophenol	14.686	109	235247	38.423	ng/ul	83
56) Dibenzofuran	14.845	168	1765790	32.460	ng/ul	98
57) 2,4-Dinitrotoluene	14.822	165	449520	39.907	ng/ul#	98
58) 2,3,4,6-Tetrachlorophenol	15.075	232	336645	37.179	ng/ul#	94
59) Diethylphthalate	15.275	149	1516657	36.550	ng/ul	99
61) Fluorene	15.492	166	1439353	33.359	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.486	204	681702	33.822	ng/ul	98
63) 4-Nitroaniline	15.533	138	372289m	45.947	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.586	198	222270	33.625	ng/ul#	97
67) N-Nitrosodiphenylamine	15.704	169	1234078	33.944	ng/ul	99
68) 4-Bromophenyl-phenylether	16.380	248	412818	35.058	ng/ul	97
69) Hexachlorobenzene	16.486	284	503710	35.299	ng/ul	96
70) Atrazine	16.663	200	230031	18.517	ng/ul	97
71) Pentachlorophenol	16.839	266	264302	33.596	ng/ul	99
72) Phenanthrene	17.227	178	2359137	35.288	ng/ul	98
74) Anthracene	17.322	178	2338726	34.894	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.240	216	559698	29.333	ng/ul	96
76) Pentachlorobenzene	14.757	250	565731	31.753	ng/ul	99
77) Carbazole	17.598	167	2254163	35.921	ng/ul	99
78) Di-n-butylphthalate	18.163	149	2742853	39.961	ng/ul	99
80) Fluoranthene	19.245	202	2870964	37.494	ng/ul	98
82) Pyrene	19.610	202	2922458	36.574	ng/ul	97
83) Butylbenzylphthalate	20.510	149	1255053	41.984	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.292	252	714838	28.422	ng/ul	98
85) Benzo(a)anthracene	21.357	228	2748551	37.447	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.286	149	1815585	43.989	ng/ul	99
87) Chrysene	21.410	228	2705504	37.514	ng/ul	99
89) Di-n-octyl phthalate	22.186	149	3050671	40.087	ng/ul	100
90) Benzo(b)fluoranthene	22.992	252	2959281	37.915	ng/ul	97
91) Benzo(k)fluoranthene	23.039	252	2727442	36.730	ng/ul	97
93) Benzo(a)pyrene	23.598	252	2825958	36.804	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.127	276	3327408	36.858	ng/ul	95
95) Dibenzo(a,h)anthracene	26.139	278	2843009	36.558	ng/ul	96
96) Benzo(g,h,i)perylene	26.862	276	2509442	32.505	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.530 ng/uL
RT: 3.246 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion: 88 Resp: 7484
Ion Ratio Lower Upper
88 100
43 31.7 27.8 41.8
58 73.9 61.8 92.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#5
Pyridine
Concen: 19.811 ng/uL
RT: 3.658 min Scan# 97
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Tgt Ion: 79 Resp: 380857
Ion Ratio Lower Upper
79 100
52 73.7 67.8 101.6
51 35.5 31.4 47.0









#6
Benzaldehyde
Concen: 33.720 ng/ul
RT: 6.946 min Scan# 656
Delta R.T. -0.006 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Instrument :
BNA_M
ClientSampleId :
EX7Q6MS

Tgt Ion: 77 Resp: 30407
Ion Ratio Lower Upper
77 100
105 99.6 84.6 127.0
106 94.5 80.3 120.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#8
Phenol
Concen: 30.190 ng/ul
RT: 7.011 min Scan# 667
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion: 94 Resp: 689087
Ion Ratio Lower Upper
94 100
65 27.7 21.3 31.9
66 35.7 27.4 41.0





#10
Bis(2-Chloroethyl)ether
Concen: 29.735 ng/ul
RT: 7.234 min Scan# 705
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion: 93 Resp: 54284
Ion Ratio Lower Upper
93 100
63 75.4 64.0 96.0
95 32.4 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#12
2-Chlorophenol
Concen: 31.025 ng/ul
RT: 7.363 min Scan# 727
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:128 Resp: 559611
Ion Ratio Lower Upper
128 100
64 45.3 36.9 55.3
130 31.7 25.6 38.4





#13
2-Methylphenol
Concen: 28.803 ng/ul
RT: 8.257 min Scan# 879
Delta R.T. -0.006 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion: 108 Resp: 484610
Ion Ratio Lower Upper
108 100
107 88.1 75.0 112.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.848 ng/ul
RT: 8.340 min Scan# 893
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26



Tgt Ion: 45 Resp: 729031
Ion Ratio Lower Upper
45 100
77 15.3 10.6 15.8
79 11.4 8.3 12.5









#16
Acetophenone
Concen: 31.602 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Instrument :
BNA_M
ClientSampleId :
EX7Q6MS

Tgt Ion: 105 Resp: 833149
Ion Ratio Lower Upper
105 100
77 78.2 59.8 89.6
51 26.5 22.1 33.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#17
N-Nitroso-di-n-propylamine
Concen: 30.990 ng/ul
RT: 8.616 min Scan# 940
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion: 70 Resp: 406645
Ion Ratio Lower Upper
70 100
42 59.5 49.6 74.4
101 9.7 8.1 12.1
130 21.7 18.4 27.6





#18
4-Methylphenol
Concen: 29.593 ng/ul
RT: 8.593 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:108 Resp: 532978
Ion Ratio Lower Upper
108 100
107 109.1 90.3 135.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#19
Hexachloroethane
Concen: 30.641 ng/ul
RT: 8.869 min Scan# 983
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:117 Resp: 231223
Ion Ratio Lower Upper
117 100
201 90.3 81.8 122.6
199 55.5 51.6 77.4





#22
Nitrobenzene
Concen: 30.909 ng/ul
RT: 9.010 min Scan# 1007
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion: 77 Resp: 617944
Ion Ratio Lower Upper
77 100
123 45.8 37.4 56.0
65 13.3 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#23
Isophorone
Concen: 31.711 ng/ul
RT: 9.540 min Scan# 1097
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Tgt Ion: 82 Resp: 1125809
Ion Ratio Lower Upper
82 100
95 6.7 5.5 8.3
138 17.8 15.3 22.9





#25
2-Nitrophenol
Concen: 32.436 ng/ul
RT: 9.722 min Scan# 1128
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:139 Resp: 30928
Ion Ratio Lower Upper
139 100
65 48.3 39.8 59.6
109 29.6 22.2 33.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#26
2,4-Dimethylphenol
Concen: 16.277 ng/ul
RT: 9.787 min Scan# 1139
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:107 Resp: 310386
Ion Ratio Lower Upper
107 100
121 50.3 42.7 64.1
122 88.1 71.5 107.3





#27
 Bis(2-Chloroethoxy)methane
 Concen: 30.092 ng/ul
 RT: 10.022 min Scan# 1180
 Delta R.T. -0.005 min
 Lab File: BM036329.D
 Acq: 17 Aug 2022 23:26




Instrument :
 BNA_M
 ClientSampleId :
 EX7Q6MS



Tgt Ion: 93 Resp: 695529
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.1 39.1
 123 11.8 9.4 14.0




Manual Integrations
 APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
 Supervised By :Sohil Jodhani 08/19/2022



#29
 2,4-Dichlorophenol
 Concen: 32.144 ng/ul
 RT: 10.257 min Scan# 1219
 Delta R.T. -0.005 min
 Lab File: BM036329.D
 Acq: 17 Aug 2022 23:26


Tgt Ion:162 Resp: 502923
 Ion Ratio Lower Upper
 162 100
 164 66.1 51.8 77.6
 98 34.9 29.5 44.3





#30
Naphthalene
Concen: 30.499 ng/ul
RT: 10.651 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:128 Resp: 1781418
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.0 10.6 16.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#32
4-Chloroaniline
Concen: 22.317 ng/ul
RT: 10.775 min Scan# 1307
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:127 Resp: 538137
Ion Ratio Lower Upper
127 100
129 31.9 25.3 37.9





#33
Hexachlorobutadiene
Concen: 31.402 ng/ul
RT: 10.928 min Scan# 11
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion:225 Resp: 308658
Ion Ratio Lower Upper
225 100
223 62.1 50.0 75.0
227 63.9 49.9 74.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#34
Caprolactam
Concen: 33.958 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. 0.006 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Tgt Ion:113 Resp: 159191
Ion Ratio Lower Upper
113 100
55 173.1 150.2 225.4
56 124.2 102.1 153.1





#35
4-Chloro-3-methylphenol
Concen: 34.034 ng/ul
RT: 11.910 min Scan# 1500
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:107 Resp: 53840
Ion Ratio Lower Upper
107 100
144 26.6 20.3 30.5
142 82.8 64.4 96.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#36
2-Methylnaphthalene
Concen: 31.631 ng/ul
RT: 12.263 min Scan# 1560
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Tgt Ion:142 Resp: 1159188
Ion Ratio Lower Upper
142 100
141 89.3 70.7 106.1





#37

1-Methylnaphthalene

Concen: 31.478 ng/ul

RT: 12.487 min Scan# 1598

Delta R.T. 0.000 min

Lab File: BM036329.D

Acq: 17 Aug 2022 23:26

Instrument :

BNA_M

ClientSampleId :

EX7Q6MS

Tgt Ion:142 Resp: 1161749

Ion Ratio Lower Upper

142 100

141 93.3 75.2 112.8





116 3.9 2.6 3.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/18/2022

Supervised By :Sohil Jodhani 08/19/2022



#39

1,2,4,5-Tetrachlorobenzene

Concen: 31.359 ng/ul

RT: 12.634 min Scan# 1623

Delta R.T. 0.000 min

Lab File: BM036329.D

Acq: 17 Aug 2022 23:26

Tgt Ion:216 Resp: 560284




Ion Ratio Lower Upper


216 100

214 78.2 62.4 93.6

179 21.9 17.3 25.9


108 18.5 14.0 21.0





#40
Hexachlorocyclopentadiene
Concen: 19.140 ng/ul
RT: 12.604 min Scan# 1618
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion:237 Resp: 202880
Ion Ratio Lower Upper
237 100
235 62.8 48.4 72.6
272 13.6 11.6 17.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#41
2,4,6-Trichlorophenol
Concen: 34.241 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Tgt Ion:196 Resp: 384973
Ion Ratio Lower Upper
196 100
198 96.1 77.8 116.6
200 30.2 25.0 37.6





#42
2,4,5-Trichlorophenol
Concen: 35.533 ng/u1
RT: 12.957 min Scan# 1678
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:196 Resp: 42374
Ion Ratio Lower Upper
196 100
198 96.3 76.8 115.2
200 30.9 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#43
1,1'-Biphenyl
Concen: 31.220 ng/u1
RT: 13.281 min Scan# 1733
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:154 Resp: 1517833
Ion Ratio Lower Upper
154 100
153 39.9 32.4 48.6
76 12.9 10.1 15.1





#44
2-Chloronaphthalene
Concen: 31.482 ng/ul
RT: 13.322 min Scan# 1740
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:162 Resp: 1174510
Ion Ratio Lower Upper
162 100
164 32.9 25.4 38.0
127 39.8 31.6 47.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#45
2-Nitroaniline
Concen: 35.577 ng/ul
RT: 13.539 min Scan# 1777
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion: 65 Resp: 359406
Ion Ratio Lower Upper
65 100
92 68.4 54.5 81.7
138 109.7 91.2 136.8





#47
Dimethylphthalate
Concen: 34.137 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:163 Resp: 143700
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 9.9 7.9 11.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#48
2,6-Dinitrotoluene
Concen: 38.056 ng/ul
RT: 14.034 min Scan# 1861
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:165 Resp: 302261
Ion Ratio Lower Upper
165 100
89 62.2 42.5 63.7
121 23.0 17.2 25.8





#50
Acenaphthylene
Concen: 32.107 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:152 Resp: 193122
Ion Ratio Lower Upper
152 100
151 19.7 16.1 24.1
153 13.1 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#51
3-Nitroaniline
Concen: 38.644 ng/ul
RT: 14.369 min Scan# 1918
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:138 Resp: 335664
Ion Ratio Lower Upper
138 100
108 9.5 8.1 12.1
92 108.6 90.4 135.6





#52
Acenaphthene
Concen: 31.832 ng/ul
RT: 14.510 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion:153 Resp: 1247968
Ion Ratio Lower Upper
153 100
152 48.2 37.1 55.7
154 88.9 71.3 106.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#53
2,4-Dinitrophenol
Concen: 34.388 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. -0.011 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Tgt Ion:184 Resp: 148848
Ion Ratio Lower Upper
184 100
63 68.5 54.2 81.4
154 68.0 49.3 73.9





#55
4-Nitrophenol
Concen: 38.423 ng/ul
RT: 14.686 min Scan# 1972
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:109 Resp: 23524
Ion Ratio Lower Upper
109 100
139 135.4 124.9 187.3
65 128.3 122.2 183.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#56
Dibenzofuran
Concen: 32.460 ng/ul
RT: 14.845 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:168 Resp: 1765790
Ion Ratio Lower Upper
168 100
139 38.6 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 39.907 ng/ul
RT: 14.822 min Scan# 1995
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:165 Resp: 449520
Ion Ratio Lower Upper
165 100
63 46.2 35.9 53.9
182 2.4 2.5 3.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.179 ng/ul
RT: 15.075 min Scan# 2038
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:232 Resp: 336645
Ion Ratio Lower Upper
232 100
131 51.8 36.7 55.1
130 2.5 1.7 2.5#
166 32.0 24.4 36.6





#59
Diethylphthalate
Concen: 36.550 ng/ul
RT: 15.275 min Scan# 2072
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:149 Resp: 151665
Ion Ratio Lower Upper
149 100
177 21.5 16.8 25.2
150 12.3 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#61
Fluorene
Concen: 33.359 ng/ul
RT: 15.492 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:166 Resp: 1439353
Ion Ratio Lower Upper
166 100
165 97.2 77.2 115.8
167 13.8 11.0 16.4





#62
4-Chlorophenyl-phenylether
Concen: 33.822 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:204 Resp: 68170
Ion Ratio Lower Upper
204 100
206 32.0 25.9 38.9
141 61.6 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#63
4-Nitroaniline
Concen: 45.947 ng/ul m
RT: 15.533 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:138 Resp: 372289
Ion Ratio Lower Upper
138 100
92 50.3 39.2 58.8
108 78.0 51.6 77.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 33.625 ng/ul
RT: 15.586 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:198 Resp: 222270
Ion Ratio Lower Upper
198 100
182 3.7 4.2 6.4
77 26.4 19.9 29.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#67
N-Nitrosodiphenylamine
Concen: 33.944 ng/ul
RT: 15.704 min Scan# 2145
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Tgt Ion:169 Resp: 1234078
Ion Ratio Lower Upper
169 100
168 67.4 53.8 80.6
167 37.3 28.7 43.1






#68
4-Bromophenyl-phenylether
Concen: 35.058 ng/ul
RT: 16.380 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Instrument :

BNA_M

ClientSampleId :

EX7Q6MS






Tgt Ion:248 Resp: 412818
Ion Ratio Lower Upper
248 100
250 98.4 77.3 115.9
141 78.3 59.8 89.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#69
Hexachlorobenzene
Concen: 35.299 ng/ul
RT: 16.486 min Scan# 2278
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:284 Resp: 503710
Ion Ratio Lower Upper
284 100
142 39.5 28.6 42.8
249 30.7 24.6 37.0





#70
Atrazine
Concen: 18.517 ng/ul
RT: 16.663 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:200 Resp: 23003
Ion Ratio Lower Upper
200 100
173 25.8 21.3 31.9
215 46.4 39.0 58.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#71
Pentachlorophenol
Concen: 33.596 ng/ul
RT: 16.839 min Scan# 2338
Delta R.T. -0.011 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:266 Resp: 264302
Ion Ratio Lower Upper
266 100
264 61.1 49.5 74.3
268 64.1 51.0 76.4





#72
Phenanthrene
Concen: 35.288 ng/u1
RT: 17.227 min Scan# 2404
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:178 Resp: 235913
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.2
176 20.1 14.9 22.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#74
Anthracene
Concen: 34.894 ng/u1
RT: 17.322 min Scan# 2420
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:178 Resp: 2338726
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.2
176 19.1 14.4 21.6





#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.333 ng/uL
RT: 13.240 min Scan# 1727
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion:216 Resp: 559698
Ion Ratio Lower Upper
216 100
214 80.8 60.9 91.3
218 47.3 37.5 56.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#76
Pentachlorobenzene
Concen: 31.753 ng/uL
RT: 14.757 min Scan# 1984
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Tgt Ion:250 Resp: 565731
Ion Ratio Lower Upper
250 100
248 62.4 50.4 75.6
252 62.8 51.1 76.7





#77
 Carbazole
 Concen: 35.921 ng/ul
 RT: 17.598 min Scan# 2467
 Delta R.T. -0.006 min
 Lab File: BM036329.D
 Acq: 17 Aug 2022 23:26




Instrument :
 BNA_M
 ClientSampleId :
 EX7Q6MS




Tgt Ion:167 Resp: 225416
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.4 26.0
 139 13.5 10.1 15.1

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
 Supervised By :Sohil Jodhani 08/19/2022




#78
 Di-n-butylphthalate
 Concen: 39.961 ng/ul
 RT: 18.163 min Scan# 2563
 Delta R.T. 0.000 min
 Lab File: BM036329.D
 Acq: 17 Aug 2022 23:26




Tgt Ion:149 Resp: 2742853
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.3 10.9
 104 4.9 3.7 5.5





#80
Fluoranthene
Concen: 37.494 ng/ul
RT: 19.245 min Scan# 21
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:202 Resp: 2870964
Ion Ratio Lower Upper
202 100
101 11.9 10.1 15.1
100 9.1 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#82
Pyrene
Concen: 36.574 ng/ul
RT: 19.610 min Scan# 2809
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:202 Resp: 2922458
Ion Ratio Lower Upper
202 100
101 14.0 11.9 17.9
100 11.5 10.2 15.4





#83
Butylbenzylphthalate
Concen: 41.984 ng/ul
RT: 20.510 min Scan# 2962
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion:149 Resp: 125505
Ion Ratio Lower Upper
149 100
91 71.6 53.2 79.8
206 17.7 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#84
3,3'-Dichlorobenzidine
Concen: 28.422 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Tgt Ion:252 Resp: 714838
Ion Ratio Lower Upper
252 100
254 63.8 52.4 78.6
126 12.7 10.8 16.2





#85
Benzo(a)anthracene
Concen: 37.447 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:228 Resp: 274855
Ion Ratio Lower Upper
228 100
229 19.9 15.6 23.4
226 27.0 21.7 32.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 43.989 ng/ul
RT: 21.286 min Scan# 3094
Delta R.T. -0.006 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:149 Resp: 1815585
Ion Ratio Lower Upper
149 100
167 26.6 21.5 32.3
279 3.9 3.4 5.2





#87
Chrysene
Concen: 37.514 ng/ul
RT: 21.410 min Scan# 3114
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS



Tgt Ion:228 Resp: 2705504
Ion Ratio Lower Upper
228 100
226 30.4 23.7 35.5
229 19.9 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#89
Di-n-octyl phthalate
Concen: 40.087 ng/ul
RT: 22.186 min Scan# 3247
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26


Tgt Ion:149 Resp: 3050671





#90
Benzo(b)fluoranthene
Concen: 37.915 ng/ul
RT: 22.992 min Scan# 3384
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion:252 Resp: 295928
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 11.0 10.7 16.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#91
Benzo(k)fluoranthene
Concen: 36.730 ng/ul
RT: 23.039 min Scan# 3392
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Tgt Ion:252 Resp: 2727442
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 11.0 10.6 15.8





#93
Benzo(a)pyrene
Concen: 36.804 ng/ul
RT: 23.598 min Scan# 3487
Delta R.T. -0.005 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26




Instrument :
BNA_M
ClientSampleId :
EX7Q6MS




Tgt Ion:252 Resp: 2825958
Ion Ratio Lower Upper
252 100
253 22.0 17.3 25.9
125 12.0 11.8 17.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 36.858 ng/ul
RT: 26.127 min Scan# 3917
Delta R.T. 0.012 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26



Tgt Ion:276 Resp: 3327408
Ion Ratio Lower Upper
276 100
138 26.9 25.6 38.4
277 24.9 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 36.558 ng/ul
RT: 26.139 min Scan# 3918
Delta R.T. 0.006 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Instrument :
BNA_M
ClientSampleId :
EX7Q6MS

Tgt Ion:278 Resp: 2843009
Ion Ratio Lower Upper
278 100
139 16.6 16.2 24.2
279 23.8 18.7 28.1

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/18/2022
Supervised By :Sohil Jodhani 08/19/2022



#96
Benzo(g,h,i)perylene
Concen: 32.505 ng/ul
RT: 26.862 min Scan# 4042
Delta R.T. 0.000 min
Lab File: BM036329.D
Acq: 17 Aug 2022 23:26

Tgt Ion:276 Resp: 2509442
Ion Ratio Lower Upper
276 100
138 22.9 21.3 31.9
277 24.1 18.6 28.0

