

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
 Data File : BM036337.D
 Acq On : 18 Aug 2022 04:12
 Operator : CG/JU
 Sample : N4176-05
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EX7Q0

Quant Time: Aug 18 05:10:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 16 15:00:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	312095	20.000	ng/u1	0.00
20) Naphthalene-d8	10.599	136	1334412	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.439	164	731817	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.186	188	1363130	20.000	ng/u1	0.00
79) Chrysene-d12	21.368	240	1239091	20.000	ng/u1	0.00
88) Perylene-d12	23.698	264	1308332	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.211	96	17272	2.109	ng/uL	0.00
4) Pyridine-d5	3.640	84	118439	5.245	ng/u1	0.00
7) Phenol-d5	6.981	99	317823	11.930	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	207860	12.396	ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	272428	13.015	ng/u1	0.00
15) 4-Methylphenol-d8	8.522	113	247839	12.460	ng/u1	0.00
21) Nitrobenzene-d5	8.969	128	134305	13.275	ng/u1	0.00
24) 2-Nitrophenol-d4	9.687	143	136523	13.303	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.228	165	223958	12.214	ng/u1	0.00
31) 4-Chloroaniline-d4	10.751	131	334674	11.773	ng/u1	0.00
46) Dimethylphthalate-d6	13.863	166	677320	14.630	ng/u1	0.00
49) Acenaphthylene-d8	14.134	160	795986	13.785	ng/u1	0.00
54) 4-Nitrophenol-d4	14.681	143	79062	9.012	ng/u1	0.00
60) Fluorene-d10	15.434	176	592920	14.172	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.569	200	45964	6.559	ng/u1	0.00
73) Anthracene-d10	17.280	188	894327	14.755	ng/u1	0.00
81) Pyrene-d10	19.574	212	1055314	16.418	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.545	264	1061036	16.541	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM081722\
Data File : BM036337.D
Acq On : 18 Aug 2022 04:12
Operator : CG/JU
Sample : N4176-05
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EX7Q0

Quant Time: Aug 18 05:10:37 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 16 15:00:33 2022
Response via : Initial Calibration

