

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081815\
 Data File : BM002477.D
 Acq On : 18 Aug 2015 08:55
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02044

Quant Time: Aug 18 17:55:33 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 18 17:52:34 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	172#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	149	0.00
3	Naphthalene	1.021	1.000	2.1	149	0.00
4	2-Methylnaphthalene	0.765	0.700	8.5	139	0.00
5	1-Methylnaphthalene	0.756	0.677	10.4	135	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
7 S	Acenaphthylene-d8	2.045	2.060	-0.7	123	0.00
8	Acenaphthylene	2.140	2.133	0.3	122	0.00
9 C	Acenaphthene	1.410	1.395	1.1	122	0.00
10 S	Fluorene-d10	1.436	1.353	5.8	116	0.00
11	Fluorene	1.615	1.516	6.1	115	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
13	Phenanthrene	1.155	1.128	2.3	97	0.00
14 S	Anthracene-d10	0.965	0.942	2.4	96	0.00
15	Anthracene	1.170	1.135	3.0	96	0.00
16 C	Fluoranthene	1.308	0.990	24.3	76	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
18 S	Pyrene-d10	0.887	1.040	-17.2	74	0.00
19	Pyrene	1.172	1.375	-17.3	73	0.00
20	Benzo(a)anthracene	1.196	1.178	1.5	67	0.00
21	Chrysene	1.119	1.106	1.2	66	0.00
22 I	Perylene-d12	1.000	1.000	0.0	66	0.00
23	Benzo(b)fluoranthene	1.209	1.184	2.1	67	0.00
24	Benzo(k)fluoranthene	1.146	1.134	1.0	65	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.039	1.1	66	0.00
26 C	Benzo(a)pyrene	1.155	1.151	0.3	66	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.356	-2.1	68	0.00
28	Dibenzo(a,h)anthracene	1.110	1.147	-3.3	69	0.00
29	Benzo(g,h,i)perylene	1.106	1.137	-2.8	69	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0