

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081815\
 Data File : BM002484.D
 Acq On : 18 Aug 2015 14:35
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02045

Quant Time: Aug 18 18:27:25 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 18 17:52:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
3	Naphthalene	1.021	0.999	2.2	123	0.00
4	2-Methylnaphthalene	0.765	0.711	7.1	116	0.00
5	1-Methylnaphthalene	0.756	0.703	7.0	116	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
7 S	Acenaphthylene-d8	2.045	2.027	0.9	109	0.00
8	Acenaphthylene	2.140	2.115	1.2	109	0.00
9 C	Acenaphthene	1.410	1.384	1.8	109	0.00
10 S	Fluorene-d10	1.436	1.386	3.5	107	0.00
11	Fluorene	1.615	1.562	3.3	107	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
13	Phenanthrene	1.155	1.130	2.2	100	0.00
14 S	Anthracene-d10	0.965	0.951	1.5	101	0.00
15	Anthracene	1.170	1.142	2.4	99	0.00
16 C	Fluoranthene	1.308	1.068	18.3	85	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	68	0.00
18 S	Pyrene-d10	0.887	1.140	-28.5#	82	0.00
19	Pyrene	1.172	1.496	-27.6#	81	0.00
20	Benzo(a)anthracene	1.196	1.197	-0.1	69	0.00
21	Chrysene	1.119	1.107	1.1	67	0.00
22 I	Perylene-d12	1.000	1.000	0.0	63	0.00
23	Benzo(b)fluoranthene	1.209	1.168	3.4	63	0.00
24	Benzo(k)fluoranthene	1.146	1.168	-1.9	64	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.046	0.5	64	0.00
26 C	Benzo(a)pyrene	1.155	1.145	0.9	63	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.357	-2.2	65	0.00
28	Dibenzo(a,h)anthracene	1.110	1.139	-2.6	65	0.00
29	Benzo(g,h,i)perylene	1.106	1.145	-3.5	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0