

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081815\
 Data File : BM002486.D
 Acq On : 18 Aug 2015 16:57
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02046

Quant Time: Aug 18 18:31:52 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 18 18:25:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
3	Naphthalene	1.021	0.997	2.4	123	0.00
4	2-Methylnaphthalene	0.765	0.715	6.5	118	0.00
5	1-Methylnaphthalene	0.756	0.698	7.7	116	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
7 S	Acenaphthylene-d8	2.045	2.038	0.3	111	0.00
8	Acenaphthylene	2.140	2.119	1.0	110	0.00
9 C	Acenaphthene	1.410	1.380	2.1	110	0.00
10 S	Fluorene-d10	1.436	1.384	3.6	108	0.00
11	Fluorene	1.615	1.565	3.1	108	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
13	Phenanthrene	1.155	1.124	2.7	102	0.00
14 S	Anthracene-d10	0.965	0.948	1.8	102	0.00
15	Anthracene	1.170	1.150	1.7	102	0.00
16 C	Fluoranthene	1.308	1.083	17.2	88	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
18 S	Pyrene-d10	0.887	1.130	-27.4#	85	0.00
19	Pyrene	1.172	1.485	-26.7#	83	0.00
20	Benzo(a)anthracene	1.196	1.190	0.5	71	0.00
21	Chrysene	1.119	1.094	2.2	69	0.00
22 I	Perylene-d12	1.000	1.000	0.0	64	0.00
23	Benzo(b)fluoranthene	1.209	1.227	-1.5	67	0.00
24	Benzo(k)fluoranthene	1.146	1.116	2.6	62	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.036	1.4	64	0.00
26 C	Benzo(a)pyrene	1.155	1.141	1.2	64	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.341	-1.0	66	0.00
28	Dibenzo(a,h)anthracene	1.110	1.135	-2.3	66	0.00
29	Benzo(g,h,i)perylene	1.106	1.121	-1.4	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0