


Data Path : Z:\HPCHEM1\BNA M\DATA\BM081816\
Data File : BM007175.D
Acq On : 17 Aug 2016 12:04
Operator : UM/SJ
Sample : DFTPP48
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP48

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM071416.M
Title : SVOA CALIBRATION
Last Update : Wed Jul 20 04:22:37 2016



AutoFind: Scans 2555, 2556, 2557; Background Corrected with Scan 2549

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.2	75496	PASS
68	69	0.00	2	1.8	1507	PASS
69	198	0.00	100	35.1	84949	PASS
70	69	0.00	2	0.6	526	PASS
127	198	10	80	43.2	104344	PASS
197	198	0.00	2	0.3	756	PASS
198	198	100	100	100.0	241706	PASS
199	198	5	9	6.9	16694	PASS
275	198	10	60	30.3	73309	PASS
365	198	1	100	4.4	10621	PASS
441	443	0.01	100	77.7	31626	PASS
442	198	50	100	87.7	212053	PASS
443	442	15	24	19.2	40682	PASS

Scan 2549 (17.580 min): BM007175.D
DFTPP48

m/z	Abundance
39.90	829.0
44.00	1717.0
48.90	349.0
57.00	384.0
73.10	344.0
96.00	367.0
207.00	2568.0
208.20	412.0
209.00	413.0
280.90	316.0

Instrument :

BNA_M

ClientSampleId :

DFTPP48

m/z	Abundance
36.90	339.0
38.00	1143.0
39.10	6409.0
39.90	1047.0
40.90	333.0
44.00	2200.0
44.90	380.0
49.10	1043.0
50.10	23456.0
51.10	88704.0
52.10	4739.0
55.10	542.0
56.00	2947.0
57.00	6223.0
57.80	307.0
61.10	1542.0
62.00	1169.0
63.00	3926.0
63.90	600.0
65.00	1469.0
68.00	2054.0
69.00	96936.0
70.10	665.0
73.10	1026.0
74.00	11158.0
75.00	15974.0
76.10	6211.0
77.10	113472.0
78.00	7785.0
79.00	9248.0
80.00	5005.0
81.00	8623.0
82.00	2369.0
83.10	1911.0
83.90	657.0
84.90	1443.0
86.00	2591.0
87.00	1020.0
88.00	483.0
91.00	1711.0
92.00	1934.0
93.00	14948.0
94.00	761.0
96.00	498.0
97.20	377.0
98.00	11597.0
99.00	9451.0
99.90	663.0
101.00	4588.0
103.00	1719.0
104.00	3139.0
105.00	2820.0
106.00	858.0
107.00	36752.0
108.00	6223.0
109.10	1135.0
110.00	61856.0
111.00	10174.0
111.90	1239.0
112.90	350.0
116.00	1691.0
117.00	39112.0
118.00	3223.0
119.00	505.0
119.90	447.0
122.00	2562.0
123.00	4416.0
124.00	1932.0
125.00	1620.0
126.10	667.0
127.00	114336.0
128.00	8845.0
129.00	47136.0
130.00	3934.0
130.90	710.0
132.00	509.0
133.00	491.0
134.00	1438.0
135.00	4536.0
135.90	1469.0
137.00	2482.0
137.80	509.0
140.00	680.0
141.00	6381.0
142.00	2025.0
143.00	1892.0
144.00	324.0
144.90	323.0
145.90	1781.0
147.00	4031.0
148.00	8124.0
149.00	1548.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

149.90	406.0
151.10	1009.0
152.10	334.0
153.10	2226.0
153.90	1297.0
155.00	3728.0
156.00	5535.0
157.00	1301.0
158.00	1400.0
159.00	1257.0
160.00	2545.0
161.10	3450.0
161.90	929.0
162.90	412.0
164.90	2373.0
166.00	2252.0
167.00	14080.0
168.00	8672.0
168.90	1267.0
170.10	602.0
171.00	755.0
172.00	1416.0
173.00	2089.0
174.00	2641.0
175.00	5294.0
176.00	1755.0
177.00	2635.0
178.00	956.0
179.00	10027.0
180.00	7274.0
181.00	3726.0
182.10	503.0
182.70	433.0
183.90	863.0
185.00	5140.0
186.00	40128.0
187.00	10793.0
188.00	1831.0
188.90	2985.0
190.00	431.0
191.00	1570.0
192.00	2932.0
193.00	3807.0
194.00	902.0
196.00	7187.0
197.10	1092.0
198.00	252864.0
199.00	18544.0
200.00	1262.0
201.60	1020.0
203.00	2394.0
204.00	10341.0
205.00	16464.0
206.00	75936.0
207.00	11897.0
208.00	3590.0
209.00	1489.0
210.00	1396.0
211.10	2886.0
212.00	850.0
214.90	682.0
216.00	1505.0
217.00	19584.0
218.00	2758.0
220.20	522.0
221.00	7937.0
221.90	2725.0
222.90	3454.0
224.00	38496.0
225.00	9134.0
226.00	956.0
227.00	18872.0
228.00	2422.0
229.00	4176.0
230.00	605.0
231.00	1335.0
234.00	1282.0
235.00	1968.0
236.00	1184.0
237.00	1257.0
237.90	462.0
238.90	678.0
240.00	503.0
241.00	1130.0
242.00	1792.0
243.00	2091.0
244.00	31344.0
245.10	3931.0
246.00	6602.0
246.90	1487.0
249.10	1150.0
249.90	320.0
253.00	385.0
254.10	1188.0
255.00	167424.0
256.00	25200.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

257.00	1900.0
258.00	9664.0
259.00	1834.0
263.90	306.0
265.00	3597.0
266.00	883.0
271.00	530.0
273.00	4942.0
274.00	12391.0
275.00	69760.0
276.00	9627.0
277.00	6552.0
278.00	1084.0
281.00	436.0
282.00	307.0
283.00	887.0
284.00	340.0
284.90	1203.0
288.90	331.0
293.00	1481.0
295.00	325.0
296.00	22296.0
296.90	3831.0
297.90	320.0
302.00	438.0
303.00	2392.0
304.10	564.0
308.00	424.0
308.90	314.0
310.00	391.0
312.90	300.0
314.00	1292.0
315.00	2388.0
315.90	1287.0
316.90	350.0
321.00	729.0
322.00	459.0
323.00	6414.0
324.00	908.0
326.90	1218.0
327.90	798.0
332.00	455.0
333.00	531.0
334.00	4833.0
335.00	949.0
341.00	902.0
346.00	1724.0
352.00	2483.0
353.10	1552.0
354.10	2024.0
355.00	549.0
365.00	9599.0
366.00	1759.0
371.00	621.0
372.00	2957.0
373.00	806.0
382.90	520.0
402.00	988.0
403.00	1759.0
403.90	688.0
421.00	1565.0
422.00	1332.0
423.10	8975.0
424.10	2324.0
441.00	26464.0
442.00	176384.0
443.00	34136.0
444.00	2746.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

m/z	Abundance
36.00	307.0
38.10	885.0
39.00	6350.0
40.00	916.0
41.00	317.0
44.00	1970.0
49.00	1196.0
50.10	23432.0
51.10	86680.0
52.10	4595.0
53.00	346.0
55.00	628.0
56.00	3076.0
57.00	6389.0
61.00	1022.0
61.90	1250.0
63.00	3276.0
64.00	676.0
65.00	1786.0
68.00	1442.0
69.00	100112.0
70.00	488.0
73.00	1028.0
74.00	11584.0
75.00	16792.0
76.10	5413.0
77.10	112376.0
78.10	7668.0
79.10	8375.0
80.00	6807.0
81.00	8400.0
82.00	2246.0
83.00	1953.0
83.90	586.0
84.90	1748.0
86.00	2829.0
87.00	978.0
90.90	2484.0
92.00	2162.0
93.00	14872.0
93.90	683.0
94.80	300.0
96.10	573.0
98.00	12484.0
99.00	9880.0
100.00	1067.0
101.00	4107.0
102.10	357.0
103.00	1767.0
104.00	3539.0
105.00	3404.0
105.90	848.0
107.00	38984.0
108.00	5732.0
108.90	883.0
110.00	64208.0
111.00	10196.0
112.10	1421.0
113.00	404.0
116.00	1867.0
117.00	39080.0
118.00	2393.0
120.00	570.0
122.00	2360.0
123.00	4120.0
124.00	1564.0
125.10	1664.0
125.90	391.0
127.00	123784.0
128.00	9008.0
129.00	47872.0
130.00	4050.0
131.10	929.0
132.10	411.0
133.00	475.0
134.00	1745.0
135.00	4007.0
136.00	2085.0
137.00	2171.0
137.90	422.0
139.00	515.0
139.90	543.0
141.00	6744.0
142.00	2468.0
143.00	1596.0
143.90	325.0
144.80	323.0
146.00	1254.0
147.00	3762.0
148.00	8879.0
149.00	1574.0
150.00	515.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

150.80	857.0
151.90	607.0
153.00	2215.0
154.10	1705.0
155.10	4164.0
156.00	5891.0
156.90	1322.0
158.00	1253.0
159.00	1129.0
159.90	2486.0
161.00	3514.0
162.00	1365.0
162.90	409.0
163.90	354.0
165.00	2695.0
166.00	2166.0
167.00	14793.0
168.00	8636.0
168.90	1326.0
170.00	565.0
170.90	513.0
171.90	1249.0
173.00	1432.0
174.00	3645.0
175.10	6024.0
176.00	1600.0
177.00	3265.0
178.00	913.0
178.90	11826.0
180.00	7992.0
181.00	4089.0
182.00	490.0
183.00	345.0
184.00	1126.0
185.00	5824.0
186.00	43632.0
187.00	11352.0
188.00	1311.0
189.00	3468.0
190.00	730.0
191.00	1621.0
192.00	4433.0
193.00	5110.0
194.00	1057.0
194.80	512.0
196.00	7305.0
196.70	1761.0
198.00	286272.0
198.90	18352.0
200.00	1660.0
201.30	1155.0
202.90	2733.0
204.00	11250.0
205.00	19872.0
206.00	81352.0
207.00	12988.0
208.00	3702.0
209.00	1251.0
209.90	1333.0
210.90	3557.0
213.10	336.0
214.90	931.0
216.00	1794.0
217.00	23368.0
218.00	2881.0
221.00	10772.0
221.90	2816.0
223.00	4715.0
224.00	48336.0
225.10	11097.0
226.00	1748.0
227.00	23008.0
228.00	3134.0
229.00	4395.0
230.00	417.0
231.00	1426.0
232.20	313.0
232.80	371.0
234.00	1244.0
235.00	1607.0
236.00	1109.0
237.00	1912.0
239.00	1147.0
240.00	436.0
240.90	1101.0
241.90	2384.0
243.00	2534.0
244.00	36432.0
245.00	4287.0
246.00	9454.0
247.00	1445.0
248.00	403.0
248.90	1240.0
249.90	303.0
251.00	507.0
252.00	455.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

253.00	946.0
254.20	1212.0
255.00	198912.0
256.00	27968.0
257.00	2306.0
258.00	11678.0
259.00	2163.0
260.00	326.0
265.00	4843.0
266.00	825.0
270.90	671.0
272.00	1098.0
273.00	6344.0
274.00	14245.0
275.00	87248.0
276.00	12030.0
277.00	7697.0
278.00	1472.0
279.00	315.0
281.10	680.0
282.90	1155.0
284.00	596.0
284.90	1060.0
286.00	303.0
290.00	411.0
292.10	338.0
293.00	2124.0
293.90	479.0
296.00	28232.0
297.00	3815.0
298.00	330.0
302.10	477.0
302.90	2492.0
304.20	736.0
308.10	430.0
314.00	1347.0
315.00	3243.0
316.00	1656.0
316.90	413.0
320.90	879.0
322.00	367.0
323.00	7161.0
324.00	1628.0
326.90	1335.0
327.80	583.0
331.90	575.0
333.00	1027.0
334.00	4979.0
335.00	1314.0
341.00	1189.0
341.90	332.0
345.90	2182.0
346.80	596.0
352.10	2506.0
353.00	1727.0
354.00	2579.0
355.10	473.0
359.00	313.0
365.00	12862.0
365.90	1478.0
371.00	631.0
372.10	4494.0
373.00	936.0
383.00	1205.0
384.00	478.0
390.10	639.0
391.10	379.0
402.00	1945.0
403.00	2518.0
404.10	660.0
421.00	1997.0
422.10	1562.0
423.10	12735.0
424.00	3274.0
424.90	312.0
441.10	37080.0
442.00	251072.0
443.00	46712.0
444.00	4656.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

m/z	Abundance
38.00	517.0
39.10	3577.0
40.00	973.0
44.00	1999.0
49.10	810.0
50.00	13448.0
51.00	51104.0
52.10	2587.0
55.10	414.0
55.90	1707.0
57.00	3980.0
61.00	663.0
62.10	992.0
63.00	1696.0
65.00	1025.0
67.10	300.0
68.00	1025.0
69.00	57800.0
70.00	426.0
73.10	1030.0
74.10	6264.0
75.00	9898.0
76.10	3828.0
77.10	66704.0
78.00	3986.0
79.00	5027.0
80.00	3834.0
81.00	5477.0
82.00	1371.0
83.00	1072.0
84.00	410.0
85.00	1105.0
86.00	1534.0
87.10	427.0
91.00	1394.0
92.10	1303.0
93.00	8871.0
94.00	589.0
96.10	519.0
98.00	8078.0
99.00	5602.0
100.10	517.0
101.00	2783.0
103.00	949.0
104.00	2112.0
105.00	1832.0
106.00	624.0
107.00	23552.0
108.00	3728.0
108.90	603.0
110.00	38192.0
111.00	6329.0
111.90	646.0
113.00	338.0
116.00	1352.0
117.00	23096.0
118.00	1637.0
122.00	1684.0
123.10	2655.0
123.90	986.0
125.00	1013.0
127.00	74912.0
128.00	6038.0
129.00	30192.0
130.00	2425.0
131.00	645.0
133.20	476.0
134.00	991.0
135.00	2837.0
135.90	1484.0
136.90	1320.0
140.10	523.0
141.00	4437.0
142.00	1363.0
143.00	1117.0
145.90	915.0
147.00	2772.0
147.90	5781.0
149.00	1396.0
150.00	308.0
151.00	822.0
152.90	1525.0
153.90	969.0
155.00	2423.0
156.00	3560.0
157.00	747.0
158.00	675.0
158.90	720.0
160.00	1595.0
160.90	2386.0
162.00	811.0
165.00	1829.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

166.00	1142.0
167.00	10444.0
168.00	5700.0
169.00	802.0
170.80	416.0
172.00	1105.0
173.00	975.0
174.00	2444.0
175.10	3791.0
176.00	1028.0
177.00	1786.0
178.00	696.0
179.00	7485.0
180.00	5086.0
181.00	2314.0
181.90	366.0
183.90	492.0
185.00	3350.0
186.00	27464.0
187.00	8219.0
188.10	1278.0
188.90	1921.0
189.80	357.0
191.00	1308.0
192.00	3214.0
193.00	3132.0
194.00	504.0
195.90	4820.0
196.90	1178.0
198.00	185984.0
199.00	13188.0
200.00	1220.0
200.80	315.0
201.40	636.0
202.90	1440.0
204.00	7903.0
205.00	12435.0
206.00	54296.0
207.00	8820.0
208.00	2224.0
209.10	757.0
210.10	1003.0
211.10	1904.0
212.20	483.0
215.00	779.0
215.90	1199.0
217.00	16145.0
218.00	1893.0
221.00	6894.0
221.90	1605.0
223.00	3569.0
224.00	31440.0
225.00	8266.0
226.00	808.0
227.00	14603.0
227.90	2699.0
229.00	2846.0
231.00	1117.0
234.00	939.0
234.90	1391.0
235.80	592.0
236.90	797.0
239.10	455.0
239.90	344.0
240.90	467.0
242.00	1945.0
243.00	1327.0
244.00	24944.0
245.00	3008.0
246.00	5501.0
246.90	983.0
249.00	1203.0
253.00	552.0
254.00	640.0
255.00	141888.0
256.00	19640.0
257.00	1585.0
258.00	8932.0
259.00	1250.0
265.00	3665.0
265.90	510.0
272.00	449.0
273.00	4438.0
274.00	10781.0
275.00	62920.0
276.00	8554.0
277.00	5948.0
278.00	919.0
281.00	729.0
283.00	764.0
283.90	650.0
285.00	1563.0
293.00	1037.0
293.90	391.0
295.00	320.0
296.00	20416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48

297.00	2647.0
303.00	2454.0
304.00	706.0
310.00	373.0
314.00	1187.0
314.90	2276.0
315.90	1522.0
321.10	478.0
323.00	5137.0
324.00	1087.0
327.00	1317.0
327.90	725.0
331.90	624.0
333.00	642.0
334.00	3723.0
335.00	1032.0
341.20	749.0
345.90	1163.0
346.90	403.0
352.00	1818.0
353.10	965.0
354.10	2451.0
355.00	516.0
365.00	9404.0
365.90	1488.0
371.00	634.0
372.00	3050.0
373.10	503.0
383.00	746.0
390.00	387.0
391.00	479.0
391.90	304.0
402.00	1449.0
403.00	2305.0
404.00	619.0
421.10	1730.0
422.10	1148.0
423.00	10562.0
424.00	2493.0
425.00	389.0
441.00	31336.0
442.00	208704.0
443.00	41200.0
444.00	3506.0

Instrument :
BNA_M
ClientSampleId :
DFTPP48