

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081817\  
 Data File : BM011296.D  
 Acq On : 18 Aug 2017 20:16  
 Operator : SJ/JU  
 Sample : PB101645TB  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 PB101645TB

Quant Time: Aug 18 23:40:44 2017  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Aug 16 02:25:04 2017  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.27  | 152  | 88978    | 20.00  | ng    | -0.02    |
| 21) Naphthalene-d8          | 10.03 | 136  | 292453   | 20.00  | ng    | -0.02    |
| 38) Acenaphthene-d10        | 13.94 | 164  | 189633   | 20.00  | ng    | -0.02    |
| 63) Phenanthrene-d10        | 16.70 | 188  | 458770   | 20.00  | ng    | -0.01    |
| 75) Chrysene-d12            | 20.93 | 240  | 582017   | 20.00  | ng    | -0.02    |
| 86) Perylene-d12            | 23.00 | 264  | 437418   | 20.00  | ng    | -0.02    |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 4.91  | 112  | 722365   | 137.24 | ng    | -0.02    |
| 7) Phenol-d6                | 6.47  | 99   | 993514   | 132.85 | ng    | -0.02    |
| 23) Nitrobenzene-d5         | 8.42  | 82   | 909465   | 116.72 | ng    | -0.02    |
| 41) 2,4,6-Tribromophenol    | 15.44 | 330  | 351964   | 115.80 | ng    | -0.02    |
| 44) 2-Fluorobiphenyl        | 12.55 | 172  | 1397224  | 92.41  | ng    | -0.02    |
| 78) Terphenyl-d14           | 19.37 | 244  | 2271448  | 77.31  | ng    | -0.02    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081817\  
Data File : BM011296.D  
Acq On : 18 Aug 2017 20:16  
Operator : SJ/JU  
Sample : PB101645TB  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
PB101645TB

Quant Time: Aug 18 23:40:44 2017  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 16 02:25:04 2017  
Response via : Initial Calibration

