

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081915\
 Data File : BM002509.D
 Acq On : 20 Aug 2015 02:09
 Operator : UM/IZ
 Sample : MDL-BLK-S-ML
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 MDL-BLK-S-ML

Quant Time: Aug 20 05:39:53 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 20 05:32:43 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.82	152	117755	20.00	ng/ul	0.00
18) Naphthalene-d8	11.69	136	573901	20.00	ng/ul	0.00
36) Acenaphthene-d10	15.41	164	333003	20.00	ng/ul	0.00
62) Phenanthrene-d10	18.12	188	735060	20.00	ng/ul	0.00
78) Chrysene-d12	22.20	240	809720	20.00	ng/ul	0.00
86) Perylene-d12	24.85	264	821633	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.85	96	15237	5.98	ng/uL	0.00
5) Phenol-d5	7.93	99	330841	31.06	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	8.11	67	190507	30.29	ng/ul	0.00
9) 2-Chlorophenol-d4	8.33	132	296028	33.63	ng/ul	0.00
13) 4-Methylphenol-d8	9.52	113	297003	32.55	ng/ul	0.00
19) Nitrobenzene-d5	10.02	128	159302	37.02	ng/ul	0.00
22) 2-Nitrophenol-d4	10.75	143	171870	45.16	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	11.29	165	277321	33.21	ng/ul	0.00
29) 4-Chloroaniline-d4	11.81	131	450302	40.64	ng/ul	0.00
44) Dimethylphthalate-d6	14.79	166	964217	35.17	ng/ul	0.00
47) Acenaphthylene-d8	15.11	160	1195279	34.65	ng/ul	0.00
52) 4-Nitrophenol-d4	15.57	143	179861	36.35	ng/ul	0.00
58) Fluorene-d10	16.38	176	804609	33.54	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	16.49	200	133022	46.05	ng/ul	0.00
71) Anthracene-d10	18.22	188	1260874	35.35	ng/ul	0.00
79) Pyrene-d10	20.44	212	1335849	34.02	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.69	264	1571337	36.02	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM081915\
Data File : BM002509.D
Acq On : 20 Aug 2015 02:09
Operator : UM/IZ
Sample : MDL-BLK-S-ML
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
MDL-BLK-S-ML

Quant Time: Aug 20 05:39:53 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 20 05:32:43 2015
Response via : Initial Calibration

