

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081915\
 Data File : BM002517.D
 Acq On : 20 Aug 2015 07:00
 Operator : UM/IZ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02050

Quant Time: Aug 20 07:38:07 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 20 05:32:43 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
2	1,4-Dioxane	0.488	0.408	16.4	57	0.00
3 S	1,4-Dioxane-d8	0.433	0.394	9.0	61	0.00
4	Benzaldehyde	1.067	1.053	1.3	58	0.00
5 S	Phenol-d5	1.809	1.692	6.5	61	0.00
6	Phenol	1.854	1.711	7.7	60	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.068	0.919	14.0	56	0.00
8	Bis(2-Chloroethyl)ether	1.427	1.303	8.7	59	0.00
9 S	2-Chlorophenol-d4	1.495	1.483	0.8	65	0.00
10	2-Chlorophenol	1.525	1.471	3.5	64	0.00
11	2-Methylphenol	1.467	1.371	6.5	61	0.00
12	2,2'-oxybis(1-Chloropropane	2.545	1.835	27.9#	46	0.00
13 S	4-Methylphenol-d8	1.550	1.490	3.9	62	0.00
14	Acetophenone	2.309	2.201	4.7	60	0.00
15 P	N-Nitroso-di-n-propylamine	1.240	1.052	15.2	54	0.00
16	4-Methylphenol	1.606	1.516	5.6	61	0.00
17	Hexachloroethane	0.602	0.533	11.5	59	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	66	0.00
19 S	Nitrobenzene-d5	0.150	0.153	-2.0	68	0.00
20	Nitrobenzene	0.348	0.302	13.2	57	0.00
21	Isophorone	0.694	0.612	11.8	57	0.00
22 S	2-Nitrophenol-d4	0.133	0.162	-21.8#	81	0.00
23 C	2-Nitrophenol	0.151	0.179	-18.5	77	0.00
24	2,4-Dimethylphenol	0.370	0.337	8.9	59	0.00
25	Bis(2-Chloroethoxy)methane	0.429	0.380	11.4	57	0.00
26 S	2,4-Dichlorophenol-d3	0.291	0.294	-1.0	66	0.00
27 C	2,4-Dichlorophenol	0.280	0.284	-1.4	66	0.00
28	Naphthalene	1.053	1.006	4.5	62	0.00
29 S	4-Chloroaniline-d4	0.386	0.429	-11.1	62	0.00
30	4-Chloroaniline	0.383	0.439	-14.6	64	0.00
31 C	Hexachlorobutadiene	0.162	0.154	4.9	62	0.00
32	Caprolactam	0.115	0.107	7.0	59	0.00
33 C	4-Chloro-3-methylphenol	0.350	0.326	6.9	60	0.00
34	2-Methylnaphthalene	0.752	0.743	1.2	64	0.00
35	1-Methylnaphthalene	0.741	0.734	0.9	64	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	66	0.00
37	1,2,4,5-Tetrachlorobenzene	0.574	0.568	1.0	64	0.00
38	Hexachlorocyclopentadiene	0.340	0.233	31.5#	45	0.00
39 C	2,4,6-Trichlorophenol	0.358	0.378	-5.6	68	0.00
40	2,4,5-Trichlorophenol	0.401	0.424	-5.7	69	0.00
41	1,1'-Biphenyl	1.691	1.649	2.5	63	0.00
42	2-Chloronaphthalene	1.285	1.255	2.3	63	0.00
43	2-Nitroaniline	0.390	0.349	10.5	59	0.00
44 S	Dimethylphthalate-d6	1.646	1.580	4.0	62	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081915\
 Data File : BM002517.D
 Acq On : 20 Aug 2015 07:00
 Operator : UM/IZ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02050

Quant Time: Aug 20 07:38:07 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 20 05:32:43 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.680	1.592	5.2	61	0.00
46	2,6-Dinitrotoluene	0.315	0.343	-8.9	72	0.00
47 S	Acenaphthylene-d8	2.072	2.034	1.8	63	0.00
48	Acenaphthylene	2.151	2.127	1.1	63	0.00
49	3-Nitroaniline	0.340	0.383	-12.6	68	0.00
50 C	Acenaphthene	1.447	1.390	3.9	62	0.00
51	2,4-Dinitrophenol	0.109	0.128	-17.4	95	0.00
52 S	4-Nitrophenol-d4	0.297	0.302	-1.7	68	0.00
53	4-Nitrophenol	0.218	0.190	12.8	57	0.00
54	Dibenzofuran	1.931	1.898	1.7	63	0.00
55	2,4-Dinitrotoluene	0.446	0.500	-12.1	72	0.00
56	2,3,4,6-Tetrachlorophenol	0.305	0.312	-2.3	66	0.00
57	Diethylphthalate	1.796	1.656	7.8	59	0.00
58 S	Fluorene-d10	1.441	1.415	1.8	63	0.00
59	Fluorene	1.629	1.598	1.9	63	0.00
60	4-Chlorophenyl-phenylether	0.747	0.725	2.9	62	0.00
61	4-Nitroaniline	0.394	0.416	-5.6	69	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	67	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.079	0.102	-29.1#	102	0.00
64	4,6-Dinitro-2-methylphenol	0.088	0.107	-21.6#	92	0.00
65	N-Nitrosodiphenylamine	0.656	0.626	4.6	63	0.00
66	4-Bromophenyl-phenylether	0.209	0.198	5.3	62	0.00
67	Hexachlorobenzene	0.238	0.226	5.0	62	0.00
68	Atrazine	0.228	0.222	2.6	61	0.00
69 C	Pentachlorophenol	0.117	0.076	35.0#	45	0.00
70	Phenanthrene	1.151	1.137	1.2	64	0.00
71 S	Anthracene-d10	0.970	0.958	1.2	64	0.00
72	Anthracene	1.189	1.165	2.0	64	0.00
73	1,2,3,4-Tetrachlorobenzene	0.272	0.252	7.4	61	0.00
74	Pentachlorobenzene	0.263	0.260	1.1	65	0.00
75	Carbazole	1.076	1.078	-0.2	63	0.00
76	Di-n-butylphthalate	1.443	1.334	7.6	60	0.00
77 C	Fluoranthene	1.218	1.278	-4.9	65	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	73	0.00
79 S	Pyrene-d10	0.970	0.914	5.8	66	0.00
80	Pyrene	1.293	1.213	6.2	66	0.00
81	Butylbenzylphthalate	0.639	0.572	10.5	64	0.00
82	3,3'-Dichlorobenzidine	0.384	0.436	-13.5	73	0.00
83	Benzo(a)anthracene	1.221	1.181	3.3	68	0.00
84	Bis(2-ethylhexyl)phthalate	0.924	0.847	8.3	65	0.00
85	Chrysene	1.157	1.110	4.1	68	0.00
86 I	Perylene-d12	1.000	1.000	0.0	76	0.00
87	Di-n-octyl phthalate	1.527	1.420	7.0	67	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081915\
Data File : BM002517.D
Acq On : 20 Aug 2015 07:00
Operator : UM/IZ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02050

Quant Time: Aug 20 07:38:07 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM080715.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 20 05:32:43 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.243	1.160	6.7	70	0.00
89	Benzo(k)fluoranthene	1.194	1.146	4.0	70	0.00
90 S	Benzo(a)pyrene-d12	1.062	1.038	2.3	73	0.00
91 C	Benzo(a)pyrene	1.189	1.139	4.2	71	0.00
92	Indeno(1,2,3-cd)pyrene	1.363	1.353	0.7	72	-0.01
93	Dibenzo(a,h)anthracene	1.155	1.132	2.0	71	0.00
94	Benzo(g,h,i)perylene	1.150	1.130	1.7	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1