

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081916\
 Data File : BM007241.D
 Acq On : 19 Aug 2016 17:29
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Aug 19 18:00:01 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 19 09:10:25 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.476	0.464	2.5	108	0.00
3 S	1,4-Dioxane-d8	0.458	0.438	4.4	102	0.00
4	Benzaldehyde	0.974	1.025	-5.2	106	0.00
5 S	Phenol-d5	1.620	1.579	2.5	107	0.00
6	Phenol	1.707	1.638	4.0	105	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.929	0.902	2.9	104	0.00
8	Bis(2-Chloroethyl)ether	1.259	1.234	2.0	105	0.00
9 S	2-Chlorophenol-d4	1.290	1.306	-1.2	110	0.00
10	2-Chlorophenol	1.320	1.305	1.1	106	0.00
11	2-Methylphenol	1.283	1.245	3.0	108	0.00
12	2,2'-oxybis(1-Chloropropane	1.443	1.368	5.2	104	0.00
13 S	4-Methylphenol-d8	1.333	1.265	5.1	106	0.00
14	Acetophenone	2.053	1.967	4.2	104	0.00
15 P	N-Nitroso-di-n-propylamine	1.068	0.990	7.3	101	0.00
16	4-Methylphenol	1.401	1.316	6.1	105	0.00
17	Hexachloroethane	0.551	0.525	4.7	104	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
19 S	Nitrobenzene-d5	0.145	0.155	-6.9	110	0.00
20	Nitrobenzene	0.374	0.383	-2.4	107	0.00
21	Isophorone	0.700	0.668	4.6	100	0.00
22 S	2-Nitrophenol-d4	0.155	0.165	-6.5	111	0.00
23 C	2-Nitrophenol	0.170	0.178	-4.7	107	0.00
24	2,4-Dimethylphenol	0.389	0.394	-1.3	106	0.00
25	Bis(2-Chloroethoxy)methane	0.420	0.405	3.6	101	0.00
26 S	2,4-Dichlorophenol-d3	0.325	0.329	-1.2	106	0.00
27 C	2,4-Dichlorophenol	0.324	0.327	-0.9	106	0.00
28	Naphthalene	0.991	1.004	-1.3	106	0.00
29 S	4-Chloroaniline-d4	0.387	0.396	-2.3	103	0.00
30	4-Chloroaniline	0.397	0.405	-2.0	102	0.00
31 C	Hexachlorobutadiene	0.239	0.245	-2.5	107	0.00
32	Caprolactam	0.110	0.096	12.7	95	0.01
33 C	4-Chloro-3-methylphenol	0.354	0.338	4.5	100	0.00
34	2-Methylnaphthalene	0.764	0.740	3.1	102	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
36	1,2,4,5-Tetrachlorobenzene	0.707	0.760	-7.5	101	0.00
37	Hexachlorocyclopentadiene	0.450	0.280	37.8#	59	0.00
38 C	2,4,6-Trichlorophenol	0.454	0.480	-5.7	100	0.00
39	2,4,5-Trichlorophenol	0.475	0.503	-5.9	99	0.00
40	1,1'-Biphenyl	1.605	1.674	-4.3	99	0.00
41	2-Chloronaphthalene	1.259	1.321	-4.9	100	0.00
42	2-Nitroaniline	0.361	0.373	-3.3	98	0.00
43 S	Dimethylphthalate-d6	1.653	1.600	3.2	93	0.00
44	Dimethylphthalate	1.652	1.586	4.0	92	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081916\
 Data File : BM007241.D
 Acq On : 19 Aug 2016 17:29
 Operator : UM/SJ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Aug 19 18:00:01 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081116.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 19 09:10:25 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.321	0.327	-1.9	98	0.00
46 S	Acenaphthylene-d8	1.985	2.032	-2.4	97	0.00
47	Acenaphthylene	2.043	2.100	-2.8	97	0.00
48	3-Nitroaniline	0.317	0.336	-6.0	96	0.00
49 C	Acenaphthene	1.327	1.335	-0.6	96	0.00
50	2,4-Dinitrophenol	0.190	0.106	44.2#	63	0.00
51 S	4-Nitrophenol-d4	0.296	0.266	10.1	87	0.00
52	4-Nitrophenol	0.301	0.276	8.3	89	0.00
53	Dibenzofuran	1.969	1.969	0.0	96	0.00
54	2,4-Dinitrotoluene	0.474	0.463	2.3	93	0.00
55	2,3,4,6-Tetrachlorophenol	0.459	0.459	0.0	95	0.00
56	Diethylphthalate	1.742	1.604	7.9	91	0.00
57 S	Fluorene-d10	1.445	1.423	1.5	95	0.00
58	Fluorene	1.588	1.579	0.6	94	0.00
59	4-Chlorophenyl-phenylether	0.838	0.827	1.3	94	0.00
60	4-Nitroaniline	0.372	0.363	2.4	91	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.111	0.081	27.0#	71	0.00
63	4,6-Dinitro-2-methylphenol	0.120	0.088	26.7#	70	0.00
64	N-Nitrosodiphenylamine	0.566	0.591	-4.4	92	0.00
65	4-Bromophenyl-phenylether	0.227	0.239	-5.3	95	0.00
66	Hexachlorobenzene	0.256	0.258	-0.8	91	0.00
67	Atrazine	0.228	0.229	-0.4	88	0.00
68 C	Pentachlorophenol	0.158	0.148	6.3	87	0.00
69	Phenanthrene	1.076	1.085	-0.8	91	0.00
70 S	Anthracene-d10	0.927	0.948	-2.3	91	0.00
71	Anthracene	1.104	1.117	-1.2	90	0.00
72	Carbazole	0.951	0.978	-2.8	88	0.00
73	Di-n-butylphthalate	1.163	1.187	-2.1	90	0.00
74 C	Fluoranthene	1.305	1.361	-4.3	88	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
76 S	Pyrene-d10	0.844	0.902	-6.9	88	0.00
77	Pyrene	1.106	1.158	-4.7	86	0.00
78	Butylbenzylphthalate	0.430	0.467	-8.6	88	0.00
79	3,3'-Dichlorobenzidine	0.386	0.378	2.1	69	0.00
80	Benzo(a)anthracene	1.182	1.198	-1.4	83	0.00
81	Bis(2-ethylhexyl)phthalate	0.636	0.673	-5.8	85	0.00
82	Chrysene	1.080	1.080	0.0	81	0.00
83 I	Perylene-d12	1.000	1.000	0.0	80	0.00
84	Di-n-octyl phthalate	1.032	1.143	-10.8	84	0.00
85	Benzo(b)fluoranthene	1.205	1.170	2.9	75	0.00
86	Benzo(k)fluoranthene	1.131	1.167	-3.2	81	0.00
87 S	Benzo(a)pyrene-d12	0.915	0.934	-2.1	80	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM081916\
Data File : BM007241.D
Acq On : 19 Aug 2016 17:29
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02055

Quant Time: Aug 19 18:00:01 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081116.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 19 09:10:25 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.142	1.152	-0.9	79	0.00
89	Indeno(1,2,3-cd)pyrene	1.400	1.382	1.3	77	0.02
90	Dibenzo(a,h)anthracene	1.175	1.160	1.3	77	0.01
91	Benzo(g,h,i)perylene	1.183	1.153	2.5	78	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0