

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081917\
 Data File : BM011320.D
 Acq On : 19 Aug 2017 16:15
 Operator : SJ/JU
 Sample : I4849-05
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 CAPTAINSCOVE-WC-002

Quant Time: Aug 21 01:35:12 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 02:25:04 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.27	152	118204	20.00	ng	-0.02
21) Naphthalene-d8	10.02	136	410656	20.00	ng	-0.02
38) Acenaphthene-d10	13.94	164	251155	20.00	ng	-0.02
63) Phenanthrene-d10	16.70	188	719519	20.00	ng	-0.02
75) Chrysene-d12	20.93	240	941833	20.00	ng	-0.02
86) Perylene-d12	23.00	264	425157	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	4.91	112	777526	111.19	ng	-0.02
7) Phenol-d6	6.47	99	976866	98.33	ng	-0.02
23) Nitrobenzene-d5	8.42	82	1080537	98.75	ng	-0.02
41) 2,4,6-Tribromophenol	15.44	330	496047	123.23	ng	-0.02
44) 2-Fluorobiphenyl	12.54	172	1673170	83.55	ng	-0.02
78) Terphenyl-d14	19.37	244	3338522	70.22	ng	-0.02
Target Compounds						
15) Benzyl Alcohol	7.57	79	23082	2.42	ng	Qvalue # 40


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM081917\
Data File : BM011320.D
Acq On : 19 Aug 2017 16:15
Operator : SJ/JU
Sample : I4849-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
CAPTAINSCOVE-WC-002

Quant Time: Aug 21 01:35:12 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-BM072617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 16 02:25:04 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.27 min Scan# 762
Delta R.T. -0.02 min
Lab File: BM011320.D
Acq: 19 Aug 2017 16:15

Instrument :
BNA_M
ClientSampleId :
CAPTAINSCOVE-WC-002

Tot Ion:152 Resp: 118204
Ion Ratio Lower Upper
152 100
150 160.9 119.5 179.3
115 75.1 43.0 64.4#




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 111.19 ng
RT: 4.91 min Scan# 361
Delta R.T. -0.02 min
Lab File: BM011320.D
Acq: 19 Aug 2017 16:15

Tgt Ion:112 Resp: 777526
Ion Ratio Lower Upper
112 100
64 86.2 38.6 57.8#
63 55.7 19.3 28.9#



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BM0
Ion 63.00 (62.70 to 63.70): BM0





#7

Phenol-d6

Concen: 98.33 ng

RT: 6.47 min Scan# 627

Delta R.T. -0.02 min

Lab File: BM011320.D

Acq: 19 Aug 2017 16:15

Instrument :

BNA_M

ClientSampleId :

CAPTAINSCOVE-WC-002


Tot Ion: 99 Resp: 976866

Ion Ratio Lower Upper

99 100

42 29.0 11.0 16.4#




71 68.4 23.4 35.2#



Abundance Ion 99.00 (98.70 to 99.70): BM011320.D (-613) (-)

Ion 42.00 (41.70 to 42.70): BM011320.D (-613) (-)

Ion 71.00 (70.70 to 71.70): BM011320.D (-613) (-)



#15

Benzyl Alcohol

Concen: 2.42 ng

RT: 7.57 min Scan# 814

Delta R.T. 0.04 min

Lab File: BM011320.D

Acq: 19 Aug 2017 16:15


Tgt Ion: 79 Resp: 23082

Ion Ratio Lower Upper

79 100

108 46.8 65.0 97.4#



77 135.8 53.0 79.6#




Abundance Ion 79.00 (78.70 to 79.70): BM011320.D (-791) (-)

Ion 108.00 (107.70 to 108.70): BM011320.D (-791) (-)

Ion 77.00 (76.70 to 77.70): BM011320.D (-791) (-)









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.02 min Scan# 1230
Delta R.T. -0.02 min
Lab File: BM011320.D
Acq: 19 Aug 2017 16:15




Instrument :
BNA_M
ClientSampleId :
CAPTAINSCOVE-WC-002


Tot Ion:136	Resp:	410656
Ion Ratio	Lower	Upper
136	100	
137	11.5	8.7 13.1
54	10.5	4.6 6.8#
68	7.8	3.7 5.5#



#23
Nitrobenzene-d5
Concen: 98.75 ng
RT: 8.42 min Scan# 957
Delta R.T. -0.02 min
Lab File: BM011320.D
Acq: 19 Aug 2017 16:15

Tgt Ion: 82	Resp:	1080537
Ion Ratio	Lower	Upper
82	100	
128	20.8	32.8 49.2#
54	49.9	40.6 60.8





#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 13.94 min Scan# 1896

Delta R.T. -0.02 min

Lab File: BM011320.D

Acq: 19 Aug 2017 16:15

Instrument :

BNA_M

ClientSampleId :

CAPTAINSCOVE-WC-002


Tot Ion:164 Resp: 251155

Ion Ratio Lower Upper

164 100

162 98.8 82.4 123.6




160 48.2 35.8 53.6



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#41

2,4,6-Tribromophenol

Concen: 123.23 ng

RT: 15.44 min Scan# 2152

Delta R.T. -0.02 min

Lab File: BM011320.D

Acq: 19 Aug 2017 16:15


Tgt Ion:330 Resp: 496047

Ion Ratio Lower Upper

330 100

332 97.1 0.0 0.0#



141 48.0 0.0 0.0#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#44
 2-Fluorobiphenyl
 Concen: 83.55 ng
 RT: 12.54 min Scan# 1658
 Delta R.T. -0.02 min
 Lab File: BM011320.D
 Acq: 19 Aug 2017 16:15

Instrument :
 BNA_M
 ClientSampleId :
 CAPTAINSCOVE-WC-002

Tot Ion:172 Resp: 1673170




Ion	Ratio	Lower	Upper
172	100		
171	34.2	28.5	42.7
170	23.2	18.8	28.2




#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 16.70 min Scan# 2365
 Delta R.T. -0.02 min
 Lab File: BM011320.D
 Acq: 19 Aug 2017 16:15

Tgt Ion:188 Resp: 719519

Ion	Ratio	Lower	Upper
188	100		
94	7.5	8.7	13.1#
80	10.2	9.3	13.9





#75

Chrysene-d12

Concen: 20.00 ng

RT: 20.93 min Scan# 3084

Delta R.T. -0.02 min

Lab File: BM011320.D

Acq: 19 Aug 2017 16:15

Instrument :

BNA_M

ClientSampleId :

CAPTAINSCOVE-WC-002


Tot Ion:240 Resp: 941833

Ion Ratio Lower Upper

240 100

120 3.8 8.2 12.2#




236 24.8 20.0 30.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 70.22 ng

RT: 19.37 min Scan# 2819

Delta R.T. -0.02 min

Lab File: BM011320.D

Acq: 19 Aug 2017 16:15


Tgt Ion:244 Resp: 3338522

Ion Ratio Lower Upper

244 100

212 9.3 4.9 7.3#



122 6.2 6.2 9.4#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
 Pervlene-d12
 Concen: 20.00 ng
 RT: 23.00 min Scan# 3436
 Delta R.T. -0.02 min
 Lab File: BM011320.D
 Acq: 19 Aug 2017 16:15

Instrument :
 BNA_M
 ClientSampleId :
 CAPTAINSCOVE-WC-002

Tot Ion	Ratio	Lower	Upper
264	100		
260	26.5	18.6	27.8
265	18.8	17.3	25.9

