

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM082115\
 Data File : BM002547.D
 Acq On : 22 Aug 2015 07:08
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02053

Quant Time: Aug 24 00:45:29 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 21 23:16:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.82	152	158021	20.00	ng/ul	0.00
2) Naphthalene-d8	11.69	136	740007	20.00	ng/ul	0.00
6) Acenaphthene-d10	15.40	164	395183	20.00	ng/ul	0.00
12) Phenanthrene-d10	18.12	188	816541	20.00	ng/ul	0.00
17) Chrysene-d12	22.20	240	663325	20.00	ng/ul	0.00
22) Perylene-d12	24.84	264	520851	20.00	ng/ul	0.00

System Monitoring Compounds

7) Acenaphthylene-d8	15.11	160	805868	19.95	ng/ul	0.00
10) Fluorene-d10	16.37	176	542682	19.12	ng/ul	0.00
14) Anthracene-d10	18.21	188	779197	19.78	ng/ul	0.00
18) Pyrene-d10	20.43	212	743366	25.28	ng/ul	0.00
25) Benzo(a)pyrene-d12	24.67	264	543277	19.86	ng/ul	0.00

Target Compounds

						Qvalue
3) Naphthalene	11.74	128	745615	19.74	ng/ul	99
4) 2-Methylnaphthalene	13.29	142	537206	18.99	ng/ul	99
5) 1-Methylnaphthalene	13.50	142	528684	18.89	ng/ul	99
8) Acenaphthylene	15.14	152	843614	19.95	ng/ul	100
9) Acenaphthene	15.47	153	548068	19.67	ng/ul	100
11) Fluorene	16.43	166	617264	19.34	ng/ul	99
13) Phenanthrene	18.16	178	932707	19.78	ng/ul	99
15) Anthracene	18.24	178	946616	19.81	ng/ul	100
16) Fluoranthene	20.10	202	974738	18.26	ng/ul	97
19) Pyrene	20.46	202	986945	25.40	ng/ul	99
20) Benzo(a)anthracene	22.17	228	790673	19.93	ng/ul	100
21) Chrysene	22.23	228	730692	19.69	ng/ul	100
23) Benzo(b)fluoranthene	24.02	252	638300	20.27	ng/ul	99
24) Benzo(k)fluoranthene	24.07	252	628492	21.06	ng/ul	99
26) Benzo(a)pyrene	24.72	252	593935	19.75	ng/ul	98
27) Indeno(1,2,3-cd)pyrene	27.63	276	646429	18.69	ng/ul	98
28) Dibenzo(a,h)anthracene	27.63	278	545989	18.89	ng/ul	100
29) Benzo(g,h,i)perylene	28.50	276	539790	18.75	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM082115\
Data File : BM002547.D
Acq On : 22 Aug 2015 07:08
Operator : UM/IZ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD02053

Quant Time: Aug 24 00:45:29 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Aug 21 23:16:58 2015
Response via : Initial Calibration

