

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM082115\
 Data File : BM002519.D
 Acq On : 21 Aug 2015 11:50
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02051

Quant Time: Aug 21 22:56:10 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 21 22:55:48 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
3	Naphthalene	1.021	1.011	1.0	122	0.00
4	2-Methylnaphthalene	0.765	0.751	1.8	121	0.00
5	1-Methylnaphthalene	0.756	0.745	1.5	121	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
7 S	Acenaphthylene-d8	2.045	2.043	0.1	120	0.00
8	Acenaphthylene	2.140	2.131	0.4	120	0.00
9 C	Acenaphthene	1.410	1.393	1.2	120	0.00
10 S	Fluorene-d10	1.436	1.414	1.5	119	0.00
11	Fluorene	1.615	1.609	0.4	120	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
13	Phenanthrene	1.155	1.140	1.3	120	0.00
14 S	Anthracene-d10	0.965	0.954	1.1	120	0.00
15	Anthracene	1.170	1.161	0.8	120	0.00
16 C	Fluoranthene	1.308	1.269	3.0	120	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
18 S	Pyrene-d10	0.887	0.918	-3.5	120	0.00
19	Pyrene	1.172	1.219	-4.0	120	0.00
20	Benzo(a)anthracene	1.196	1.178	1.5	123	0.00
21	Chrysene	1.119	1.112	0.6	123	0.00
22 I	Perylene-d12	1.000	1.000	0.0	126	0.00
23	Benzo(b)fluoranthene	1.209	1.211	-0.2	131	0.00
24	Benzo(k)fluoranthene	1.146	1.101	3.9	121	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.033	1.7	126	0.00
26 C	Benzo(a)pyrene	1.155	1.136	1.6	125	0.01
27	Indeno(1,2,3-cd)pyrene	1.328	1.311	1.3	127	0.01
28	Dibenzo(a,h)anthracene	1.110	1.104	0.5	127	0.01
29	Benzo(g,h,i)perylene	1.106	1.095	1.0	127	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0