

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM082115\
 Data File : BM002547.D
 Acq On : 22 Aug 2015 07:08
 Operator : UM/IZ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Aug 24 00:45:29 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 21 23:16:58 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	182#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	170#	0.00
3	Naphthalene	1.021	1.008	1.3	172#	0.00
4	2-Methylnaphthalene	0.765	0.726	5.1	165#	0.00
5	1-Methylnaphthalene	0.756	0.714	5.6	164#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	156#	0.00
7 S	Acenaphthylene-d8	2.045	2.039	0.3	157#	0.00
8	Acenaphthylene	2.140	2.135	0.2	157#	0.00
9 C	Acenaphthene	1.410	1.387	1.6	157#	0.00
10 S	Fluorene-d10	1.436	1.373	4.4	152#	0.00
11	Fluorene	1.615	1.562	3.3	153#	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	148	0.00
13	Phenanthrene	1.155	1.142	1.1	148	0.00
14 S	Anthracene-d10	0.965	0.954	1.1	148	0.00
15	Anthracene	1.170	1.159	0.9	148	0.00
16 C	Fluoranthene	1.308	1.194	8.7	139	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
18 S	Pyrene-d10	0.887	1.121	-26.4#	135	0.00
19	Pyrene	1.172	1.488	-27.0#	135	0.00
20	Benzo(a)anthracene	1.196	1.192	0.3	115	0.00
21	Chrysene	1.119	1.102	1.5	112	0.00
22 I	Perylene-d12	1.000	1.000	0.0	88	0.00
23	Benzo(b)fluoranthene	1.209	1.225	-1.3	92	0.00
24	Benzo(k)fluoranthene	1.146	1.207	-5.3	93	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.043	0.8	89	0.00
26 C	Benzo(a)pyrene	1.155	1.140	1.3	88	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.241	6.6	84	0.00
28	Dibenzo(a,h)anthracene	1.110	1.048	5.6	84	0.00
29	Benzo(g,h,i)perylene	1.106	1.036	6.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0