

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082222\
 Data File : BM036378.D
 Acq On : 22 Aug 2022 22:57
 Operator : CG/JU
 Sample : PB147176BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 23 01:07:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 23 00:55:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.792	152	267048	20.000	ng/u1	0.00
20) Naphthalene-d8	10.592	136	1156350	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.433	164	672224	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.180	188	1302505	20.000	ng/u1	0.00
79) Chrysene-d12	21.362	240	1157067	20.000	ng/u1	0.00
88) Perylene-d12	23.685	264	1182924	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.204	96	41021	5.854	ng/uL	0.00
4) Pyridine-d5	3.628	84	559255	28.942	ng/u1	0.00
7) Phenol-d5	6.969	99	677811	29.735	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	449963	31.360	ng/u1	0.00
11) 2-Chlorophenol-d4	7.322	132	539584	30.128	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	513454	30.168	ng/u1	0.00
21) Nitrobenzene-d5	8.963	128	266300	30.374	ng/u1	0.00
24) 2-Nitrophenol-d4	9.680	143	268062	30.143	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	462385	29.099	ng/u1	0.00
31) 4-Chloroaniline-d4	10.745	131	741186	30.088	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	1393311	32.762	ng/u1	0.00
49) Acenaphthylene-d8	14.133	160	1674766	31.574	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	205011	25.441	ng/u1	0.00
60) Fluorene-d10	15.427	176	1264024	32.892	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.562	200	141912	21.192	ng/u1	0.00
73) Anthracene-d10	17.280	188	1909945	32.978	ng/u1	0.00
81) Pyrene-d10	19.574	212	2041686	34.016	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.538	264	1871523	32.269	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082222\
Data File : BM036378.D
Acq On : 22 Aug 2022 22:57
Operator : CG/JU
Sample : PB147176BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Aug 23 01:07:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 23 00:55:01 2022
Response via : Initial Calibration

