

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082222\
 Data File : BM036380.D
 Acq On : 23 Aug 2022 00:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 23 01:09:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 23 00:55:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.792	152	309107	20.000	ng/u1	0.00
20) Naphthalene-d8	10.592	136	1389466	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.433	164	767024	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.180	188	1335352	20.000	ng/u1	0.00
79) Chrysene-d12	21.362	240	1131615	20.000	ng/u1	0.00
88) Perylene-d12	23.686	264	1122139	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	58573	7.222	ng/uL	0.00
4) Pyridine-d5	3.628	84	419503	18.756	ng/u1	0.00
7) Phenol-d5	6.969	99	508235	19.262	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	340955	20.530	ng/u1	0.00
11) 2-Chlorophenol-d4	7.322	132	403127	19.446	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	395457	20.073	ng/u1	0.00
21) Nitrobenzene-d5	8.963	128	203942	19.359	ng/u1	0.00
24) 2-Nitrophenol-d4	9.681	143	206215	19.298	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	368220	19.285	ng/u1	0.00
31) 4-Chloroaniline-d4	10.739	131	549125	18.552	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	952300	19.625	ng/u1	0.00
49) Acenaphthylene-d8	14.127	160	1188197	19.632	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	143370	15.593	ng/u1	0.00
60) Fluorene-d10	15.427	176	825820	18.833	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	112934	16.450	ng/u1	0.00
73) Anthracene-d10	17.274	188	1134686	19.110	ng/u1	0.00
81) Pyrene-d10	19.568	212	1153576	19.652	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.533	264	1051973	19.121	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	62363	7.300	ng/uL	97
5) Pyridine	3.652	79	435387	18.844	ng/u1	91
6) Benzaldehyde	6.940	77	245104	22.615	ng/u1	94
8) Phenol	6.998	94	535634	19.526	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.222	93	441938	20.142	ng/u1	100
12) 2-Chlorophenol	7.357	128	422823	19.505	ng/u1	97
13) 2-Methylphenol	8.245	108	402148	19.887	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.334	45	633284	21.573	ng/u1	99
16) Acetophenone	8.628	105	662555	20.910	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.610	70	344991	21.875	ng/u1	97
18) 4-Methylphenol	8.581	108	438159	20.242	ng/u1	98
19) Hexachloroethane	8.863	117	175891	19.394	ng/u1	85
22) Nitrobenzene	9.004	77	503382	20.167	ng/u1	97
23) Isophorone	9.534	82	907323	20.469	ng/u1	96
25) 2-Nitrophenol	9.716	139	227909	19.144	ng/u1	98
26) 2,4-Dimethylphenol	9.781	107	469895	19.736	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.016	93	584709	20.262	ng/u1	99
29) 2,4-Dichlorophenol	10.245	162	379707	19.438	ng/u1	99
30) Naphthalene	10.639	128	1408543	19.315	ng/u1	99
32) 4-Chloroaniline	10.769	127	560003	18.600	ng/u1	98
33) Hexachlorobutadiene	10.916	225	228447	18.615	ng/u1	98
34) Caprolactam	11.557	113	100778	17.218	ng/u1	97
35) 4-Chloro-3-methylphenol	11.898	107	381992	19.340	ng/u1	99
36) 2-Methylnaphthalene	12.257	142	902450	19.723	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082222\
 Data File : BM036380.D
 Acq On : 23 Aug 2022 00:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 23 01:09:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 23 00:55:01 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.475	142	911247	19.775	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.627	216	415573	19.602	ng/ul	99
40) Hexachlorocyclopentadiene	12.598	237	260092	20.679	ng/ul	98
41) 2,4,6-Trichlorophenol	12.875	196	259407	19.445	ng/ul	99
42) 2,4,5-Trichlorophenol	12.951	196	268968	19.008	ng/ul	99
43) 1,1'-Biphenyl	13.269	154	1158829	20.088	ng/ul	99
44) 2-Chloronaphthalene	13.316	162	868063	19.609	ng/ul	99
45) 2-Nitroaniline	13.533	65	227663	18.992	ng/ul	96
47) Dimethylphthalate	13.904	163	978591	19.592	ng/ul	100
48) 2,6-Dinitrotoluene	14.027	165	182540	19.369	ng/ul#	84
50) Acenaphthylene	14.157	152	1409825	19.753	ng/ul	100
51) 3-Nitroaniline	14.363	138	177625	17.234	ng/ul	98
52) Acenaphthene	14.498	153	904927	19.453	ng/ul	97
53) 2,4-Dinitrophenol	14.574	184	81708	15.909	ng/ul	93
55) 4-Nitrophenol	14.674	109	113950	15.685	ng/ul	88
56) Dibenzofuran	14.833	168	1230778	19.068	ng/ul	98
57) 2,4-Dinitrotoluene	14.816	165	249372	18.658	ng/ul#	91
58) 2,3,4,6-Tetrachlorophenol	15.063	232	196542	18.293	ng/ul#	90
59) Diethylphthalate	15.263	149	961220	19.522	ng/ul	100
61) Fluorene	15.480	166	977532	19.093	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.480	204	456363	19.082	ng/ul	99
63) 4-Nitroaniline	15.521	138	140034	14.565	ng/ul#	88
66) 4,6-Dinitro-2-methylph...	15.574	198	117139	16.876	ng/ul	92
67) N-Nitrosodiphenylamine	15.698	169	782842	20.506	ng/ul	98
68) 4-Bromophenyl-phenylether	16.374	248	245290	19.838	ng/ul	98
69) Hexachlorobenzene	16.480	284	285869	19.078	ng/ul	96
70) Atrazine	16.651	200	241064	18.481	ng/ul	97
71) Pentachlorophenol	16.833	266	155273	18.796	ng/ul	97
72) Phenanthrene	17.221	178	1356611	19.325	ng/ul	99
74) Anthracene	17.310	178	1364199	19.384	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.233	216	420696	20.997	ng/uL	99
76) Pentachlorobenzene	14.751	250	380328	20.329	ng/uL	100
77) Carbazole	17.592	167	1211909	18.391	ng/ul	100
78) Di-n-butylphthalate	18.151	149	1409469	19.556	ng/ul	99
80) Fluoranthene	19.239	202	1416800	19.971	ng/ul	99
82) Pyrene	19.598	202	1475317	19.928	ng/ul	99
83) Butylbenzylphthalate	20.503	149	541307	19.544	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.286	252	415148	17.815	ng/ul	97
85) Benzo(a)anthracene	21.345	228	1321445	19.432	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.274	149	771352	20.171	ng/ul#	98
87) Chrysene	21.397	228	1268895	18.990	ng/ul	99
89) Di-n-octyl phthalate	22.174	149	1266907	18.992	ng/ul	100
90) Benzo(b)fluoranthene	22.974	252	1295258	18.932	ng/ul	99
91) Benzo(k)fluoranthene	23.027	252	1280007	19.665	ng/ul	99
93) Benzo(a)pyrene	23.580	252	1305603	19.398	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.091	276	1463810	18.498	ng/ul	96
95) Dibenzo(a,h)anthracene	26.109	278	1269766	18.627	ng/ul	99
96) Benzo(g,h,i)perylene	26.838	276	1236325	18.269	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082222\
 Data File : BM036380.D
 Acq On : 23 Aug 2022 00:10
 Operator : CG/JU
 Sample : SSTCCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :


Quant Time: Aug 23 01:09:08 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 23 00:55:01 2022
 Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.300 ng/uL
RT: 3.240 min Scan# 20
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Instrument :
BNA_M
ClientSampleId :







Tgt Ion: 88 Resp: 62363
Ion Ratio Lower Upper
88 100
43 31.0 27.8 41.8
58 76.3 61.8 92.8



#5
Pyridine
Concen: 18.844 ng/uL
RT: 3.652 min Scan# 96
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion: 79 Resp: 435387
Ion Ratio Lower Upper
79 100
52 74.5 67.8 101.6
51 35.7 31.4 47.0









#6
Benzaldehyde
Concen: 22.615 ng/ul
RT: 6.940 min Scan# 655
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 245104
Ion Ratio Lower Upper
77 100
105 98.8 84.6 127.0
106 95.7 80.3 120.5



#8
Phenol
Concen: 19.526 ng/ul
RT: 6.998 min Scan# 665
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion: 94 Resp: 535634
Ion Ratio Lower Upper
94 100
65 28.2 21.3 31.9
66 35.5 27.4 41.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.142 ng/ul
 RT: 7.222 min Scan# 704
 Delta R.T. -0.006 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion: 93 Resp: 441938
 Ion Ratio Lower Upper
 93 100
 63 79.8 64.0 96.0
 95 32.8 25.8 38.6



#12
 2-Chlorophenol
 Concen: 19.505 ng/ul
 RT: 7.357 min Scan# 726
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10

Tgt Ion: 128 Resp: 422823
 Ion Ratio Lower Upper
 128 100
 64 48.4 36.9 55.3
 130 33.1 25.6 38.4









#13
2-Methylphenol
Concen: 19.887 ng/ul
RT: 8.245 min Scan# 811
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 402148
Ion Ratio Lower Upper
108 100
107 88.4 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.573 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion: 45 Resp: 633284
Ion Ratio Lower Upper
45 100
77 13.7 10.6 15.8
79 11.1 8.3 12.5









#16
Acetophenone
Concen: 20.910 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 662555
Ion Ratio Lower Upper
105 100
77 77.9 59.8 89.6
51 27.4 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 21.875 ng/ul
RT: 8.610 min Scan# 939
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion: 70 Resp: 344991
Ion Ratio Lower Upper
70 100
42 60.3 49.6 74.4
101 9.3 8.1 12.1
130 20.9 18.4 27.6









#18
4-Methylphenol
Concen: 20.242 ng/ul
RT: 8.581 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 438159
Ion Ratio Lower Upper
108 100
107 110.9 90.3 135.5



#19
Hexachloroethane
Concen: 19.394 ng/ul
RT: 8.863 min Scan# 982
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:117 Resp: 175891
Ion Ratio Lower Upper
117 100
201 86.2 81.8 122.6
199 54.6 51.6 77.4









#22
Nitrobenzene
Concen: 20.167 ng/ul
RT: 9.004 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 503382
Ion Ratio Lower Upper
77 100
123 44.9 37.4 56.0
65 12.7 10.8 16.2



#23
Isophorone
Concen: 20.469 ng/ul
RT: 9.534 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion: 82 Resp: 907323
Ion Ratio Lower Upper
82 100
95 6.5 5.5 8.3
138 17.0 15.3 22.9









#25
2-Nitrophenol
Concen: 19.144 ng/ul
RT: 9.716 min Scan# 1127
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 227909
Ion Ratio Lower Upper
139 100
65 50.1 39.8 59.6
109 29.5 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 19.736 ng/ul
RT: 9.781 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:107 Resp: 469895
Ion Ratio Lower Upper
107 100
121 51.3 42.7 64.1
122 87.7 71.5 107.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.262 ng/ul
 RT: 10.016 min Scan# 1178
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion: 93 Resp: 584709
 Ion Ratio Lower Upper
 93 100
 95 31.7 26.1 39.1
 123 11.6 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 19.438 ng/ul
 RT: 10.245 min Scan# 1217
 Delta R.T. -0.006 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10

Tgt Ion: 162 Resp: 379707
 Ion Ratio Lower Upper
 162 100
 164 63.6 51.8 77.6
 98 36.9 29.5 44.3









#30
Naphthalene
Concen: 19.315 ng/ul
RT: 10.639 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 1408543
Ion Ratio Lower Upper
128 100
129 11.2 8.8 13.2
127 13.0 10.6 16.0



#32
4-Chloroaniline
Concen: 18.600 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:127 Resp: 560003
Ion Ratio Lower Upper
127 100
129 32.5 25.3 37.9









#33
Hexachlorobutadiene
Concen: 18.615 ng/ul
RT: 10.916 min Scan# 1331
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:225 Resp: 228447
Ion Ratio Lower Upper
225 100
223 64.3 50.0 75.0
227 64.4 49.9 74.9



#34
Caprolactam
Concen: 17.218 ng/ul
RT: 11.557 min Scan# 1440
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:113 Resp: 100778
Ion Ratio Lower Upper
113 100
55 188.7 150.2 225.4
56 135.6 102.1 153.1









#35
 4-Chloro-3-methylphenol
 Concen: 19.340 ng/ul
 RT: 11.898 min Scan# 1498
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:107 Resp: 381992
 Ion Ratio Lower Upper
 107 100
 144 26.2 20.3 30.5
 142 81.5 64.4 96.6



#36
 2-Methylnaphthalene
 Concen: 19.723 ng/ul
 RT: 12.257 min Scan# 1559
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10

Tgt Ion:142 Resp: 902450
 Ion Ratio Lower Upper
 142 100
 141 87.9 70.7 106.1





#37

1-Methylnaphthalene

Concen: 19.775 ng/ul

RT: 12.475 min Scan# 1596

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:142 Resp: 911247

Ion Ratio Lower Upper

142 100

141 92.3 75.2 112.8

116 3.9 2.6 3.8#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.602 ng/ul

RT: 12.627 min Scan# 1622

Delta R.T. 0.006 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Tgt Ion:216 Resp: 415573




Ion Ratio Lower Upper


216 100

214 78.1 62.4 93.6

179 20.9 17.3 25.9


108 18.6 14.0 21.0








#40
Hexachlorocyclopentadiene
Concen: 20.679 ng/ul
RT: 12.598 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:237 Resp: 260092
Ion Ratio Lower Upper
237 100
235 62.5 48.4 72.6
272 14.3 11.6 17.4




#41
2,4,6-Trichlorophenol
Concen: 19.445 ng/ul
RT: 12.875 min Scan# 1664
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10



Tgt Ion:196 Resp: 259407
Ion Ratio Lower Upper
196 100
198 96.6 77.8 116.6
200 31.1 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 19.008 ng/uI

RT: 12.951 min Scan# 1

Delta R.T. 0.000 min


Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:196 Resp: 268968

Ion Ratio Lower Upper

196 100

198 97.8 76.8 115.2

200 31.0 24.9 37.3



#43

1,1'-Biphenyl

Concen: 20.088 ng/uI

RT: 13.269 min Scan# 1731

Delta R.T. -0.006 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:154 Resp: 1158829


Ion Ratio Lower Upper

154 100

153 40.3 32.4 48.6


76 13.9 10.1 15.1








#44
2-Chloronaphthalene
Concen: 19.609 ng/ul
RT: 13.316 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:162 Resp: 868063
Ion Ratio Lower Upper
162 100
164 32.6 25.4 38.0
127 39.8 31.6 47.4



#45
2-Nitroaniline
Concen: 18.992 ng/ul
RT: 13.533 min Scan# 1776
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion: 65 Resp: 227663
Ion Ratio Lower Upper
65 100
92 69.3 54.5 81.7
138 108.5 91.2 136.8





#47

Dimethylphthalate

Concen: 19.592 ng/ul

RT: 13.904 min Scan# 1839

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:163 Resp: 978591

Ion Ratio Lower Upper

163 100

194 3.9 3.2 4.8

164 10.0 7.9 11.9



#48

2,6-Dinitrotoluene

Concen: 19.369 ng/ul

RT: 14.027 min Scan# 1860

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:165 Resp: 182540


Ion Ratio Lower Upper

165 100

89 67.7 42.5 63.7#

121 24.4 17.2 25.8





#50

Acenaphthylene

Concen: 19.753 ng/ul

RT: 14.157 min Scan# 1882

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:152 Resp: 1409825

Ion Ratio Lower Upper

152 100

151 20.1 16.1 24.1

153 13.2 10.6 15.8



#51

3-Nitroaniline

Concen: 17.234 ng/ul

RT: 14.363 min Scan# 1917

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:138 Resp: 177625


Ion Ratio Lower Upper

138 100

108 9.8 8.1 12.1

92 114.7 90.4 135.6









#52
Acenaphthene
Concen: 19.453 ng/ul
RT: 14.498 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:153 Resp: 904927
Ion Ratio Lower Upper
153 100
152 48.6 37.1 55.7
154 86.8 71.3 106.9



#53
2,4-Dinitrophenol
Concen: 15.909 ng/ul
RT: 14.574 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:184 Resp: 81708
Ion Ratio Lower Upper
184 100
63 75.6 54.2 81.4
154 64.8 49.3 73.9









#55
4-Nitrophenol
Concen: 15.685 ng/ul
RT: 14.674 min Scan# 1971
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 113950
Ion Ratio Lower Upper
109 100
139 140.1 124.9 187.3
65 138.8 122.2 183.4



#56
Dibenzofuran
Concen: 19.068 ng/ul
RT: 14.833 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:168 Resp: 1230778
Ion Ratio Lower Upper
168 100
139 38.8 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 18.658 ng/ul
RT: 14.816 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:165 Resp: 249372
Ion Ratio Lower Upper
165 100
63 51.0 35.9 53.9
182 2.2 2.5 3.7#



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.293 ng/ul
RT: 15.063 min Scan# 2036
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:232 Resp: 196542
Ion Ratio Lower Upper
232 100
131 54.2 36.7 55.1
130 3.0 1.7 2.5#
166 34.4 24.4 36.6





#59

Diethylphthalate

Concen: 19.522 ng/ul

RT: 15.263 min Scan# 2070

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:149 Resp: 961220

Ion Ratio Lower Upper

149 100

177 21.0 16.8 25.2

150 12.3 10.2 15.4



#61

Fluorene

Concen: 19.093 ng/ul

RT: 15.480 min Scan# 2107

Delta R.T. -0.006 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:166 Resp: 977532


Ion Ratio Lower Upper

166 100

165 99.2 77.2 115.8

167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 19.082 ng/ul
RT: 15.480 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:204 Resp: 456363
Ion Ratio Lower Upper
204 100
206 32.2 25.9 38.9
141 61.6 50.6 75.8



#63
4-Nitroaniline
Concen: 14.565 ng/ul
RT: 15.521 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:138 Resp: 140034
Ion Ratio Lower Upper
138 100
92 51.5 39.2 58.8
108 78.8 51.6 77.4#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 16.876 ng/ul
 RT: 15.574 min Scan# 2123
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:198 Resp: 117139
 Ion Ratio Lower Upper
 198 100
 182 5.2 4.2 6.4
 77 29.9 19.9 29.9



#67
 N-Nitrosodiphenylamine
 Concen: 20.506 ng/ul
 RT: 15.698 min Scan# 2144
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10

Tgt Ion:169 Resp: 782842
 Ion Ratio Lower Upper
 169 100
 168 68.5 53.8 80.6
 167 37.8 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 19.838 ng/ul
RT: 16.374 min Scan# 2259
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:248 Resp: 245290
Ion Ratio Lower Upper
248 100
250 97.5 77.3 115.9
141 77.5 59.8 89.8



#69
Hexachlorobenzene
Concen: 19.078 ng/ul
RT: 16.480 min Scan# 2277
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:284 Resp: 285869
Ion Ratio Lower Upper
284 100
142 40.3 28.6 42.8
249 30.6 24.6 37.0





#70

Atrazine

Concen: 18.481 ng/ul

RT: 16.651 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:200 Resp: 241064

Ion Ratio Lower Upper

200 100

173 27.8 21.3 31.9

215 47.0 39.0 58.6



#71

Pentachlorophenol

Concen: 18.796 ng/ul

RT: 16.833 min Scan# 2337

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:266 Resp: 155273


Ion Ratio Lower Upper

266 100

264 60.5 49.5 74.3

268 61.0 51.0 76.4





#72

Phenanthrene

Concen: 19.325 ng/ul

RT: 17.221 min Scan# 2403

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:178 Resp: 1356611

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.2

176 19.6 14.9 22.3



#74

Anthracene

Concen: 19.384 ng/ul

RT: 17.310 min Scan# 2418

Delta R.T. -0.006 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:178 Resp: 1364199


Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.2

176 19.2 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.997 ng/uL

RT: 13.233 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:216 Resp: 420696

Ion Ratio Lower Upper

216 100

214 77.2 60.9 91.3

218 47.1 37.5 56.3



#76

Pentachlorobenzene

Concen: 20.329 ng/uL

RT: 14.751 min Scan# 1983

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:250 Resp: 380328


Ion Ratio Lower Upper

250 100

248 63.5 50.4 75.6

252 64.1 51.1 76.7









#77
 Carbazole
 Concen: 18.391 ng/ul
 RT: 17.592 min Scan# 2466
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:167 Resp: 1211909
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 12.9 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 19.556 ng/ul
 RT: 18.151 min Scan# 2561
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10

Tgt Ion:149 Resp: 1409469
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 5.1 3.7 5.5









#80
Fluoranthene
Concen: 19.971 ng/ul
RT: 19.239 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:202 Resp: 1416800
Ion Ratio Lower Upper
202 100
101 12.0 10.1 15.1
100 9.2 7.8 11.6



#82
Pyrene
Concen: 19.928 ng/ul
RT: 19.598 min Scan# 2807
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:202 Resp: 1475317
Ion Ratio Lower Upper
202 100
101 15.1 11.9 17.9
100 12.5 10.2 15.4





#83

Butylbenzylphthalate

Concen: 19.544 ng/ul

RT: 20.503 min Scan# 2961

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:149 Resp: 541307

Ion Ratio Lower Upper

149 100

91 74.6 53.2 79.8

206 17.8 16.0 24.0



#84

3,3'-Dichlorobenzidine

Concen: 17.815 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. 0.000 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:252 Resp: 415148


Ion Ratio Lower Upper

252 100

254 63.3 52.4 78.6

126 12.4 10.8 16.2









#85
Benzo(a)anthracene
Concen: 19.432 ng/ul
RT: 21.345 min Scan# 3105
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 1321445
Ion Ratio Lower Upper
228 100
229 19.8 15.6 23.4
226 27.3 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 20.171 ng/ul
RT: 21.274 min Scan# 3092
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10


Tgt Ion:149 Resp: 771352
Ion Ratio Lower Upper
149 100
167 25.7 21.5 32.3
279 3.3 3.4 5.2#








#87
Chrysene
Concen: 18.990 ng/ul
RT: 21.397 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Instrument :
BNA_M
ClientSampleId :







Tgt Ion:228 Resp: 1268895
Ion Ratio Lower Upper
228 100
226 30.5 23.7 35.5
229 19.8 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 18.992 ng/ul
RT: 22.174 min Scan# 3245
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:149 Resp: 1266907









#90
Benzo(b)fluoranthene
Concen: 18.932 ng/ul
RT: 22.974 min Scan# 3381
Delta R.T. -0.006 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 1295258
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 12.3 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 19.665 ng/ul
RT: 23.027 min Scan# 3390
Delta R.T. 0.000 min
Lab File: BM036380.D
Acq: 23 Aug 2022 00:10

Tgt Ion:252 Resp: 1280007
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 12.1 10.6 15.8





#93

Benzo(a)pyrene

Concen: 19.398 ng/ul

RT: 23.580 min Scan# 3485

Delta R.T. -0.006 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:252 Resp: 1305603

Ion Ratio Lower Upper

252 100

253 21.8 17.3 25.9

125 13.7 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.498 ng/ul

RT: 26.091 min Scan# 3911

Delta R.T. -0.012 min

Lab File: BM036380.D

Acq: 23 Aug 2022 00:10




Tgt Ion:276 Resp: 1463810


Ion Ratio Lower Upper

276 100

138 28.4 25.6 38.4

277 25.1 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 18.627 ng/ul
 RT: 26.109 min Scan# 3915
 Delta R.T. -0.006 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:278 Resp: 1269766
 Ion Ratio Lower Upper
 278 100
 139 18.7 16.2 24.2
 279 23.4 18.7 28.1



#96
 Benzo(g,h,i)perylene
 Concen: 18.269 ng/ul
 RT: 26.838 min Scan# 4038
 Delta R.T. 0.000 min
 Lab File: BM036380.D
 Acq: 23 Aug 2022 00:10

Tgt Ion:276 Resp: 1236325
 Ion Ratio Lower Upper
 276 100
 138 25.4 21.3 31.9
 277 24.2 18.6 28.0

