

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082222\
 Data File : BMCCC377.D
 Acq On : 22 Aug 2022 22:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Aug 23 01:09:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 23 00:55:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.792	152	300331	20.000	ng/u1	0.00
20) Naphthalene-d8	10.592	136	1390309	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.433	164	805020	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.180	188	1395022	20.000	ng/u1	0.00
79) Chrysene-d12	21.362	240	1104012	20.000	ng/u1	0.00
88) Perylene-d12	23.686	264	1083779	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.205	96	58242	7.391	ng/uL	0.00
4) Pyridine-d5	3.628	84	427858	19.688	ng/u1	0.00
7) Phenol-d5	6.969	99	503320	19.634	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.134	67	340141	21.079	ng/u1	0.00
11) 2-Chlorophenol-d4	7.322	132	399568	19.837	ng/u1	0.00
15) 4-Methylphenol-d8	8.516	113	395253	20.649	ng/u1	0.00
21) Nitrobenzene-d5	8.963	128	204136	19.366	ng/u1	0.00
24) 2-Nitrophenol-d4	9.681	143	202517	18.940	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.222	165	375567	19.658	ng/u1	0.00
31) 4-Chloroaniline-d4	10.745	131	562317	18.986	ng/u1	0.00
46) Dimethylphthalate-d6	13.857	166	1005076	19.735	ng/u1	0.00
49) Acenaphthylene-d8	14.127	160	1233061	19.412	ng/u1	0.00
54) 4-Nitrophenol-d4	14.663	143	148057	15.343	ng/u1	0.00
60) Fluorene-d10	15.427	176	875810	19.031	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	117438	16.375	ng/u1	0.00
73) Anthracene-d10	17.280	188	1180436	19.030	ng/u1	0.00
81) Pyrene-d10	19.574	212	1158675	20.232	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.539	264	1018994	19.177	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.240	88	63571	7.659	ng/uL	93
5) Pyridine	3.652	79	438668	19.540	ng/u1	90
6) Benzaldehyde	6.940	77	245549	23.318	ng/u1	93
8) Phenol	6.998	94	534970	20.071	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.222	93	437705	20.532	ng/u1	99
12) 2-Chlorophenol	7.357	128	418216	19.856	ng/u1	98
13) 2-Methylphenol	8.245	108	397437	20.229	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.334	45	635707	22.288	ng/u1	99
16) Acetophenone	8.628	105	657117	21.344	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.610	70	347551	22.682	ng/u1	99
18) 4-Methylphenol	8.581	108	436582	20.759	ng/u1	98
19) Hexachloroethane	8.863	117	173610	19.702	ng/u1	87
22) Nitrobenzene	9.004	77	500059	20.021	ng/u1	97
23) Isophorone	9.534	82	925387	20.864	ng/u1	96
25) 2-Nitrophenol	9.716	139	228440	19.177	ng/u1	98
26) 2,4-Dimethylphenol	9.781	107	474693	19.926	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.016	93	596175	20.647	ng/u1	99
29) 2,4-Dichlorophenol	10.245	162	380579	19.470	ng/u1	100
30) Naphthalene	10.639	128	1423602	19.509	ng/u1	100
32) 4-Chloroaniline	10.769	127	570599	18.941	ng/u1	100
33) Hexachlorobutadiene	10.916	225	227000	18.486	ng/u1	99
34) Caprolactam	11.557	113	106458	18.178	ng/u1	96
35) 4-Chloro-3-methylphenol	11.898	107	395730	20.023	ng/u1	99
36) 2-Methylnaphthalene	12.257	142	923808	20.178	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082222\
 Data File : BMCCC377.D
 Acq On : 22 Aug 2022 22:21
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 18 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :


Quant Time: Aug 23 01:09:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 23 00:55:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.475	142	929688	20.163	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.627	216	426844	19.183	ng/ul	98
40) Hexachlorocyclopentadiene	12.598	237	267146	20.238	ng/ul	99
41) 2,4,6-Trichlorophenol	12.874	196	266832	19.058	ng/ul	98
42) 2,4,5-Trichlorophenol	12.951	196	281278	18.939	ng/ul	99
43) 1,1'-Biphenyl	13.274	154	1203759	19.882	ng/ul	99
44) 2-Chloronaphthalene	13.316	162	897863	19.325	ng/ul	99
45) 2-Nitroaniline	13.533	65	247032	19.635	ng/ul	94
47) Dimethylphthalate	13.904	163	1032119	19.688	ng/ul	99
48) 2,6-Dinitrotoluene	14.027	165	191181	19.328	ng/ul#	82
50) Acenaphthylene	14.157	152	1469469	19.617	ng/ul	99
51) 3-Nitroaniline	14.363	138	185599	17.158	ng/ul	99
52) Acenaphthene	14.498	153	955050	19.561	ng/ul	99
53) 2,4-Dinitrophenol	14.574	184	88425	16.404	ng/ul	92
55) 4-Nitrophenol	14.674	109	118624	15.558	ng/ul	85
56) Dibenzofuran	14.833	168	1301549	19.212	ng/ul	97
57) 2,4-Dinitrotoluene	14.816	165	263151	18.759	ng/ul#	90
58) 2,3,4,6-Tetrachlorophenol	15.063	232	204901	18.171	ng/ul#	88
59) Diethylphthalate	15.263	149	1025009	19.835	ng/ul	99
61) Fluorene	15.486	166	1027375	19.120	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.480	204	481620	19.188	ng/ul	99
63) 4-Nitroaniline	15.521	138	141793	14.052	ng/ul#	87
66) 4,6-Dinitro-2-methylph...	15.574	198	119629	16.498	ng/ul#	87
67) N-Nitrosodiphenylamine	15.698	169	830976	20.836	ng/ul	98
68) 4-Bromophenyl-phenylether	16.374	248	259687	20.104	ng/ul	97
69) Hexachlorobenzene	16.480	284	299419	19.128	ng/ul	95
70) Atrazine	16.657	200	258552	18.973	ng/ul	98
71) Pentachlorophenol	16.833	266	160070	18.548	ng/ul	98
72) Phenanthrene	17.221	178	1421651	19.385	ng/ul	98
74) Anthracene	17.310	178	1436686	19.540	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.233	216	433983	20.733	ng/ul	97
76) Pentachlorobenzene	14.751	250	398814	20.406	ng/ul	98
77) Carbazole	17.592	167	1244068	18.072	ng/ul	99
78) Di-n-butylphthalate	18.151	149	1511758	20.078	ng/ul	99
80) Fluoranthene	19.239	202	1427894	20.631	ng/ul	99
82) Pyrene	19.603	202	1473126	20.396	ng/ul	99
83) Butylbenzylphthalate	20.503	149	548037	20.282	ng/ul	88
84) 3,3'-Dichlorobenzidine	21.286	252	399610	17.577	ng/ul	99
85) Benzo(a)anthracene	21.345	228	1288506	19.421	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.280	149	784999	21.041	ng/ul#	96
87) Chrysene	21.397	228	1258811	19.310	ng/ul	99
89) Di-n-octyl phthalate	22.180	149	1252803	19.445	ng/ul	100
90) Benzo(b)fluoranthene	22.980	252	1285987	19.462	ng/ul	99
91) Benzo(k)fluoranthene	23.027	252	1216163	19.346	ng/ul	99
93) Benzo(a)pyrene	23.586	252	1248933	19.213	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.097	276	1410279	18.452	ng/ul	98
95) Dibenzo(a,h)anthracene	26.109	278	1213146	18.426	ng/ul	98
96) Benzo(g,h,i)perylene	26.838	276	1195824	18.296	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 23 01:09:58 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 23 00:55:01 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.659 ng/uL
RT: 3.240 min Scan# 20
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 88 Resp: 63571
Ion Ratio Lower Upper
88 100
43 30.9 27.8 41.8
58 71.2 61.8 92.8



#5
Pyridine
Concen: 19.540 ng/uL
RT: 3.652 min Scan# 96
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion: 79 Resp: 438668
Ion Ratio Lower Upper
79 100
52 74.0 67.8 101.6
51 36.3 31.4 47.0









#6
Benzaldehyde
Concen: 23.318 ng/ul
RT: 6.940 min Scan# 655
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 245549
Ion Ratio Lower Upper
77 100
105 96.6 84.6 127.0
106 94.9 80.3 120.5



#8
Phenol
Concen: 20.071 ng/ul
RT: 6.998 min Scan# 665
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion: 94 Resp: 534970
Ion Ratio Lower Upper
94 100
65 28.7 21.3 31.9
66 35.0 27.4 41.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.532 ng/ul
RT: 7.222 min Scan# 704
Delta R.T. -0.006 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 93 Resp: 437705
Ion Ratio Lower Upper
93 100
63 78.6 64.0 96.0
95 31.9 25.8 38.6



#12
2-Chlorophenol
Concen: 19.856 ng/ul
RT: 7.357 min Scan# 726
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:128 Resp: 418216
Ion Ratio Lower Upper
128 100
64 47.6 36.9 55.3
130 31.6 25.6 38.4









#13
2-Methylphenol
Concen: 20.229 ng/ul
RT: 8.245 min Scan# 811
Delta R.T. -0.006 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 397437
Ion Ratio Lower Upper
108 100
107 88.1 75.0 112.4



#14
2,2'-oxybis(1-Chloropropane)
Concen: 22.288 ng/ul
RT: 8.334 min Scan# 892
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion: 45 Resp: 635707
Ion Ratio Lower Upper
45 100
77 13.5 10.6 15.8
79 10.9 8.3 12.5









#16
Acetophenone
Concen: 21.344 ng/ul
RT: 8.628 min Scan# 942
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 657117
Ion Ratio Lower Upper
105 100
77 77.8 59.8 89.6
51 27.0 22.1 33.1



#17
N-Nitroso-di-n-propylamine
Concen: 22.682 ng/ul
RT: 8.610 min Scan# 939
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion: 70 Resp: 347551
Ion Ratio Lower Upper
70 100
42 61.8 49.6 74.4
101 9.8 8.1 12.1
130 20.8 18.4 27.6









#18
4-Methylphenol
Concen: 20.759 ng/ul
RT: 8.581 min Scan# 91
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 436582
Ion Ratio Lower Upper
108 100
107 111.3 90.3 135.5



#19
Hexachloroethane
Concen: 19.702 ng/ul
RT: 8.863 min Scan# 982
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:117 Resp: 173610
Ion Ratio Lower Upper
117 100
201 89.3 81.8 122.6
199 53.2 51.6 77.4









#22
Nitrobenzene
Concen: 20.021 ng/ul
RT: 9.004 min Scan# 1006
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 500059
Ion Ratio Lower Upper
77 100
123 44.2 37.4 56.0
65 13.4 10.8 16.2



#23
Isophorone
Concen: 20.864 ng/ul
RT: 9.534 min Scan# 1096
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion: 82 Resp: 925387
Ion Ratio Lower Upper
82 100
95 6.6 5.5 8.3
138 17.0 15.3 22.9









#25
2-Nitrophenol
Concen: 19.177 ng/ul
RT: 9.716 min Scan# 1127
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 228440
Ion Ratio Lower Upper
139 100
65 51.4 39.8 59.6
109 28.6 22.2 33.4



#26
2,4-Dimethylphenol
Concen: 19.926 ng/ul
RT: 9.781 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21


Tgt Ion:107 Resp: 474693
Ion Ratio Lower Upper
107 100
121 51.3 42.7 64.1
122 87.5 71.5 107.3








#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.647 ng/ul
 RT: 10.016 min Scan# 1178
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Instrument :
 BNA_M
 ClientSampleId :







Tgt Ion: 93 Resp: 596175
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.1 39.1
 123 11.4 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 19.470 ng/ul
 RT: 10.245 min Scan# 1217
 Delta R.T. -0.006 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Tgt Ion: 162 Resp: 380579
 Ion Ratio Lower Upper
 162 100
 164 64.8 51.8 77.6
 98 37.2 29.5 44.3









#30
Naphthalene
Concen: 19.509 ng/ul
RT: 10.639 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 1423602
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.0 10.6 16.0



#32
4-Chloroaniline
Concen: 18.941 ng/ul
RT: 10.769 min Scan# 1306
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:127 Resp: 570599
Ion Ratio Lower Upper
127 100
129 31.5 25.3 37.9









#33
Hexachlorobutadiene
Concen: 18.486 ng/ul
RT: 10.916 min Scan# 11
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:225 Resp: 227000
Ion Ratio Lower Upper
225 100
223 62.9 50.0 75.0
227 64.1 49.9 74.9



#34
Caprolactam
Concen: 18.178 ng/ul
RT: 11.557 min Scan# 1440
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:113 Resp: 106458
Ion Ratio Lower Upper
113 100
55 182.8 150.2 225.4
56 133.0 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 20.023 ng/ul
RT: 11.898 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:107 Resp: 395730
Ion Ratio Lower Upper
107 100
144 26.7 20.3 30.5
142 81.3 64.4 96.6



#36
2-Methylnaphthalene
Concen: 20.178 ng/ul
RT: 12.257 min Scan# 1559
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:142 Resp: 923808
Ion Ratio Lower Upper
142 100
141 90.1 70.7 106.1





#37

1-Methylnaphthalene

Concen: 20.163 ng/ul

RT: 12.475 min Scan# 1

Delta R.T. 0.000 min


Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:142 Resp: 929688

Ion Ratio Lower Upper

142 100

141 91.0 75.2 112.8

116 3.8 2.6 3.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.183 ng/ul

RT: 12.627 min Scan# 1622

Delta R.T. 0.006 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Tgt Ion:216 Resp: 426844




Ion Ratio Lower Upper


216 100

214 76.2 62.4 93.6

179 21.4 17.3 25.9


108 19.6 14.0 21.0








#40
Hexachlorocyclopentadiene
Concen: 20.238 ng/ul
RT: 12.598 min Scan# 1
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:237 Resp: 267146
Ion Ratio Lower Upper
237 100
235 61.2 48.4 72.6
272 13.6 11.6 17.4




#41
2,4,6-Trichlorophenol
Concen: 19.058 ng/ul
RT: 12.874 min Scan# 1664
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21



Tgt Ion:196 Resp: 266832
Ion Ratio Lower Upper
196 100
198 95.2 77.8 116.6
200 31.2 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 18.939 ng/ul

RT: 12.951 min Scan# 1

Delta R.T. 0.000 min


Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:196 Resp: 281278

Ion Ratio Lower Upper

196 100

198 94.7 76.8 115.2

200 31.4 24.9 37.3



#43

1,1'-Biphenyl

Concen: 19.882 ng/ul

RT: 13.274 min Scan# 1732

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21




Tgt Ion:154 Resp: 1203759


Ion Ratio Lower Upper

154 100

153 39.9 32.4 48.6

76 13.5 10.1 15.1





#44

2-Chloronaphthalene

Concen: 19.325 ng/ul

RT: 13.316 min Scan# 1

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:162 Resp: 897863

Ion Ratio Lower Upper

162 100

164 32.7 25.4 38.0

127 39.3 31.6 47.4



#45

2-Nitroaniline

Concen: 19.635 ng/ul

RT: 13.533 min Scan# 1776

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21




Tgt Ion: 65 Resp: 247032


Ion Ratio Lower Upper

65 100

92 68.3 54.5 81.7

138 103.0 91.2 136.8









#47
Dimethylphthalate
Concen: 19.688 ng/ul
RT: 13.904 min Scan# 1839
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:163 Resp: 1032119
Ion Ratio Lower Upper
163 100
194 3.8 3.2 4.8
164 10.4 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 19.328 ng/ul
RT: 14.027 min Scan# 1860
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:165 Resp: 191181
Ion Ratio Lower Upper
165 100
89 68.6 42.5 63.7#
121 25.1 17.2 25.8









#50
Acenaphthylene
Concen: 19.617 ng/ul
RT: 14.157 min Scan# 1882
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:152 Resp: 1469469
Ion Ratio Lower Upper
152 100
151 19.9 16.1 24.1
153 12.8 10.6 15.8



#51
3-Nitroaniline
Concen: 17.158 ng/ul
RT: 14.363 min Scan# 1917
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:138 Resp: 185599
Ion Ratio Lower Upper
138 100
108 10.7 8.1 12.1
92 114.2 90.4 135.6





#52

Acenaphthene

Concen: 19.561 ng/ul

RT: 14.498 min Scan# 1940

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:153 Resp: 955050

Ion Ratio Lower Upper

153 100

152 48.8 37.1 55.7

154 89.1 71.3 106.9



#53

2,4-Dinitrophenol

Concen: 16.404 ng/ul

RT: 14.574 min Scan# 1953

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21




Tgt Ion:184 Resp: 88425


Ion Ratio Lower Upper

184 100

63 76.2 54.2 81.4

154 66.2 49.3 73.9









#55
4-Nitrophenol
Concen: 15.558 ng/ul
RT: 14.674 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 118624
Ion Ratio Lower Upper
109 100
139 133.3 124.9 187.3
65 138.1 122.2 183.4



#56
Dibenzofuran
Concen: 19.212 ng/ul
RT: 14.833 min Scan# 1997
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:168 Resp: 1301549
Ion Ratio Lower Upper
168 100
139 39.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 18.759 ng/ul
RT: 14.816 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:165 Resp: 263151
Ion Ratio Lower Upper
165 100
63 51.8 35.9 53.9
182 2.4 2.5 3.7#



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.171 ng/ul
RT: 15.063 min Scan# 2036
Delta R.T. -0.006 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:232 Resp: 204901
Ion Ratio Lower Upper
232 100
131 56.1 36.7 55.1#
130 3.0 1.7 2.5#
166 34.3 24.4 36.6









#59
Diethylphthalate
Concen: 19.835 ng/ul
RT: 15.263 min Scan# 2070
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 1025009
Ion Ratio Lower Upper
149 100
177 21.1 16.8 25.2
150 12.4 10.2 15.4



#61
Fluorene
Concen: 19.120 ng/ul
RT: 15.486 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:166 Resp: 1027375
Ion Ratio Lower Upper
166 100
165 98.8 77.2 115.8
167 13.3 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 19.188 ng/ul
RT: 15.480 min Scan# 2107
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:204 Resp: 481620
Ion Ratio Lower Upper
204 100
206 32.3 25.9 38.9
141 62.1 50.6 75.8



#63
4-Nitroaniline
Concen: 14.052 ng/ul
RT: 15.521 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:138 Resp: 141793
Ion Ratio Lower Upper
138 100
92 53.5 39.2 58.8
108 78.4 51.6 77.4#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 16.498 ng/ul
 RT: 15.574 min Scan# 2123
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:198 Resp: 119629
 Ion Ratio Lower Upper
 198 100
 182 5.5 4.2 6.4
 77 32.7 19.9 29.9#



#67
 N-Nitrosodiphenylamine
 Concen: 20.836 ng/ul
 RT: 15.698 min Scan# 2144
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Tgt Ion:169 Resp: 830976
 Ion Ratio Lower Upper
 169 100
 168 68.0 53.8 80.6
 167 37.6 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 20.104 ng/ul
RT: 16.374 min Scan# 21
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:248 Resp: 259687
Ion Ratio Lower Upper
248 100
250 97.8 77.3 115.9
141 79.3 59.8 89.8



#69
Hexachlorobenzene
Concen: 19.128 ng/ul
RT: 16.480 min Scan# 2277
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:284 Resp: 299419
Ion Ratio Lower Upper
284 100
142 41.3 28.6 42.8
249 30.5 24.6 37.0





#70

Atrazine

Concen: 18.973 ng/ul

RT: 16.657 min Scan# 21

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:200 Resp: 258552

Ion Ratio Lower Upper

200 100

173 26.1 21.3 31.9

215 46.5 39.0 58.6



#71

Pentachlorophenol

Concen: 18.548 ng/ul

RT: 16.833 min Scan# 2337

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21




Tgt Ion:266 Resp: 160070


Ion Ratio Lower Upper

266 100

264 64.0 49.5 74.3

268 65.0 51.0 76.4









#72
Phenanthrene
Concen: 19.385 ng/ul
RT: 17.221 min Scan# 2403
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:178 Resp: 1421651
Ion Ratio Lower Upper
178 100
179 15.6 12.2 18.2
176 19.6 14.9 22.3



#74
Anthracene
Concen: 19.540 ng/ul
RT: 17.310 min Scan# 2418
Delta R.T. -0.006 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:178 Resp: 1436686
Ion Ratio Lower Upper
178 100
179 15.8 12.2 18.2
176 19.1 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.733 ng/uL

RT: 13.233 min Scan# 1

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:216 Resp: 433983

Ion Ratio Lower Upper

216 100

214 79.5 60.9 91.3

218 47.1 37.5 56.3



#76

Pentachlorobenzene

Concen: 20.406 ng/uL

RT: 14.751 min Scan# 1983

Delta R.T. 0.000 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21




Tgt Ion:250 Resp: 398814


Ion Ratio Lower Upper

250 100

248 62.1 50.4 75.6


252 61.2 51.1 76.7








#77
 Carbazole
 Concen: 18.072 ng/ul
 RT: 17.592 min Scan# 2466
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Instrument :
 BNA_M
 ClientSampleId :







Tgt Ion:167 Resp: 1244068
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 13.4 10.1 15.1



#78
 Di-n-butylphthalate
 Concen: 20.078 ng/ul
 RT: 18.151 min Scan# 2561
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Tgt Ion:149 Resp: 1511758
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 5.3 3.7 5.5









#80
 Fluoranthene
 Concen: 20.631 ng/u1
 RT: 19.239 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:202 Resp: 1427894
 Ion Ratio Lower Upper
 202 100
 101 12.9 10.1 15.1
 100 9.9 7.8 11.6



#82
 Pyrene
 Concen: 20.396 ng/u1
 RT: 19.603 min Scan# 2808
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Tgt Ion:202 Resp: 1473126
 Ion Ratio Lower Upper
 202 100
 101 14.7 11.9 17.9
 100 11.9 10.2 15.4









#83
Butylbenzylphthalate
Concen: 20.282 ng/ul
RT: 20.503 min Scan# 2961
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 548037
Ion Ratio Lower Upper
149 100
91 77.6 53.2 79.8
206 17.4 16.0 24.0



#84
3,3'-Dichlorobenzidine
Concen: 17.577 ng/ul
RT: 21.286 min Scan# 3094
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:252 Resp: 399610
Ion Ratio Lower Upper
252 100
254 64.6 52.4 78.6
126 13.6 10.8 16.2









#85
Benzo(a)anthracene
Concen: 19.421 ng/ul
RT: 21.345 min Scan# 3105
Delta R.T. -0.006 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 1288506
Ion Ratio Lower Upper
228 100
229 19.6 15.6 23.4
226 27.5 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.041 ng/ul
RT: 21.280 min Scan# 3093
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:149 Resp: 784999
Ion Ratio Lower Upper
149 100
167 25.0 21.5 32.3
279 3.4 3.4 5.2#









#87
Chrysene
Concen: 19.310 ng/ul
RT: 21.397 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 1258811
Ion Ratio Lower Upper
228 100
226 30.4 23.7 35.5
229 19.7 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 19.445 ng/ul
RT: 22.180 min Scan# 3246
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:149 Resp: 1252803









#90
Benzo(b)fluoranthene
Concen: 19.462 ng/ul
RT: 22.980 min Scan# 3382
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 1285987
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 12.6 10.7 16.1



#91
Benzo(k)fluoranthene
Concen: 19.346 ng/ul
RT: 23.027 min Scan# 3390
Delta R.T. 0.000 min
Lab File: BMCCC377.D
Acq: 22 Aug 2022 22:21

Tgt Ion:252 Resp: 1216163
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 12.2 10.6 15.8





#93

Benzo(a)pyrene

Concen: 19.213 ng/ul

RT: 23.586 min Scan# 3485

Delta R.T. 0.000 min


Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:252 Resp: 1248933

Ion Ratio Lower Upper

252 100

253 21.3 17.3 25.9

125 13.9 11.8 17.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.452 ng/ul

RT: 26.097 min Scan# 3912

Delta R.T. -0.006 min

Lab File: BMCCC377.D

Acq: 22 Aug 2022 22:21




Tgt Ion:276 Resp: 1410279


Ion Ratio Lower Upper

276 100

138 29.9 25.6 38.4

277 24.8 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 18.426 ng/ul
 RT: 26.109 min Scan# 3915
 Delta R.T. -0.006 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:278 Resp: 1213146
 Ion Ratio Lower Upper
 278 100
 139 19.0 16.2 24.2
 279 23.7 18.7 28.1



#96
 Benzo(g,h,i)perylene
 Concen: 18.296 ng/ul
 RT: 26.838 min Scan# 4038
 Delta R.T. 0.000 min
 Lab File: BMCCC377.D
 Acq: 22 Aug 2022 22:21

Tgt Ion:276 Resp: 1195824
 Ion Ratio Lower Upper
 276 100
 138 25.4 21.3 31.9
 277 23.5 18.6 28.0

