

Data Path : Z:\HPCHEM1\BNA_M\Data\BM082415\
 Data File : BM002549.D
 Acq On : 24 Aug 2015 09:32
 Operator : UM/IZ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02054

Quant Time: Aug 24 17:43:15 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 24 17:38:40 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	160#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	153#	0.00
3	Naphthalene	1.021	1.008	1.3	155#	0.00
4	2-Methylnaphthalene	0.765	0.733	4.2	150#	0.00
5	1-Methylnaphthalene	0.756	0.724	4.2	149	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	146	0.00
7 S	Acenaphthylene-d8	2.045	2.039	0.3	147	0.00
8	Acenaphthylene	2.140	2.138	0.1	148	0.00
9 C	Acenaphthene	1.410	1.396	1.0	148	0.00
10 S	Fluorene-d10	1.436	1.404	2.2	145	0.00
11	Fluorene	1.615	1.596	1.2	146	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	146	0.00
13	Phenanthrene	1.155	1.150	0.4	148	0.00
14 S	Anthracene-d10	0.965	0.963	0.2	147	0.00
15	Anthracene	1.170	1.172	-0.2	148	0.00
16 C	Fluoranthene	1.308	1.286	1.7	148	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	142	0.00
18 S	Pyrene-d10	0.887	0.979	-10.4	148	0.00
19	Pyrene	1.172	1.298	-10.8	147	0.00
20	Benzo(a)anthracene	1.196	1.180	1.3	142	0.00
21	Chrysene	1.119	1.115	0.4	142	0.00
22 I	Perylene-d12	1.000	1.000	0.0	115	0.00
23	Benzo(b)fluoranthene	1.209	1.257	-4.0	124	0.00
24	Benzo(k)fluoranthene	1.146	1.214	-5.9	122	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.054	-0.3	118	0.00
26 C	Benzo(a)pyrene	1.155	1.157	-0.2	117	0.00
27	Indeno(1,2,3-cd)pyrene	1.328	1.195	10.0	105	0.00
28	Dibenzo(a,h)anthracene	1.110	1.023	7.8	108	0.00
29	Benzo(g,h,i)perylene	1.106	0.990	10.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0