

Data Path : Z:\HPCHEM1\BNA_M\Data\BM082415\
 Data File : BM002551.D
 Acq On : 24 Aug 2015 14:25
 Operator : UM/IZ
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Aug 24 17:39:16 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM081415MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Aug 21 23:16:58 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	205#	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	195#	0.00
3	Naphthalene	1.021	1.005	1.6	196#	0.00
4	2-Methylnaphthalene	0.765	0.732	4.3	191#	0.00
5	1-Methylnaphthalene	0.756	0.721	4.6	189#	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	182#	0.00
7 S	Acenaphthylene-d8	2.045	2.042	0.1	184#	0.00
8	Acenaphthylene	2.140	2.139	0.0	184#	0.00
9 C	Acenaphthene	1.410	1.395	1.1	184#	0.00
10 S	Fluorene-d10	1.436	1.392	3.1	179#	0.00
11	Fluorene	1.615	1.588	1.7	181#	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	177#	0.00
13	Phenanthrene	1.155	1.139	1.4	177#	0.00
14 S	Anthracene-d10	0.965	0.955	1.0	177#	0.00
15	Anthracene	1.170	1.163	0.6	177#	0.00
16 C	Fluoranthene	1.308	1.222	6.6	170#	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	148	0.00
18 S	Pyrene-d10	0.887	1.068	-20.4#	168#	0.00
19	Pyrene	1.172	1.417	-20.9#	167#	0.00
20	Benzo(a)anthracene	1.196	1.190	0.5	149	0.00
21	Chrysene	1.119	1.107	1.1	147	0.00
22 I	Perylene-d12	1.000	1.000	0.0	117	0.00
23	Benzo(b)fluoranthene	1.209	1.238	-2.4	123	0.00
24	Benzo(k)fluoranthene	1.146	1.199	-4.6	122	0.00
25 S	Benzo(a)pyrene-d12	1.051	1.040	1.0	118	0.00
26 C	Benzo(a)pyrene	1.155	1.142	1.1	116	0.01
27	Indeno(1,2,3-cd)pyrene	1.328	1.209	9.0	108	0.00
28	Dibenzo(a,h)anthracene	1.110	1.029	7.3	109	0.00
29	Benzo(g,h,i)perylene	1.106	1.003	9.3	108	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0