

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
 Data File : BM031904.D
 Acq On : 26 Aug 2021 01:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Aug 26 03:24:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 14:16:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.522	152	41385	20.000	ng	0.00
21) Naphthalene-d8	10.280	136	185992	20.000	ng	0.00
39) Acenaphthene-d10	14.162	164	132728	20.000	ng	0.00
64) Phenanthrene-d10	16.915	188	293044	20.000	ng	0.00
76) Chrysene-d12	21.121	240	289149	20.000	ng	# 0.00
86) Perylene-d12	23.256	264	260750	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.157	112	191474	76.489	ng	0.00
7) Phenol-d6	6.722	99	268633	75.844	ng	0.00
23) Nitrobenzene-d5	8.675	82	318272	74.122	ng	0.00
42) 2,4,6-Tribromophenol	15.662	330	120850	77.027	ng	0.00
45) 2-Fluorobiphenyl	12.774	172	722290	71.796	ng	0.00
79) Terphenyl-d14	19.562	244	1327524	66.273	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.169	88	36203	40.363	ng	# 84
3) Pyridine	3.551	79	102884	40.380	ng	# 82
4) n-Nitrosodimethylamine	3.469	42	67049	38.740	ng	# 49
6) Aniline	6.875	93	141346	37.567	ng	99
8) 2-Chlorophenol	7.092	128	109194	37.466	ng	97
9) Benzaldehyde	6.686	77	82260	41.655	ng	99
10) Phenol	6.745	94	132382	37.794	ng	97
11) bis(2-Chloroethyl)ether	6.969	93	96358	37.712	ng	97
12) 1,3-Dichlorobenzene	7.404	146	118548	36.829	ng	99
13) 1,4-Dichlorobenzene	7.557	146	123499	37.445	ng	99
14) 1,2-Dichlorobenzene	7.863	146	118399	36.868	ng	99
15) Benzyl Alcohol	7.775	79	117846	39.003	ng	98
16) 2,2'-oxybis(1-Chloropr...	8.051	45	126672	36.953	ng	96
17) 2-Methylphenol	7.969	107	92852	37.156	ng	96
18) Hexachloroethane	8.569	117	52048	38.095	ng	98
19) n-Nitroso-di-n-propyla...	8.333	70	100700	36.902	ng	98
20) 3+4-Methylphenols	8.298	107	130708	38.019	ng	97
22) Acetophenone	8.345	105	185089	36.609	ng	98
24) Nitrobenzene	8.716	77	159514	36.690	ng	98
25) Isophorone	9.239	82	270536	36.482	ng	97
26) 2-Nitrophenol	9.416	139	65651	37.720	ng	93
27) 2,4-Dimethylphenol	9.480	122	95344	36.399	ng	94
28) bis(2-Chloroethoxy)met...	9.716	93	131916	36.419	ng	99
29) 2,4-Dichlorophenol	9.939	162	115105	37.416	ng	98
30) 1,2,4-Trichlorobenzene	10.145	180	122920	36.420	ng	99
31) Naphthalene	10.327	128	380831	36.641	ng	99
32) Benzoic acid	9.675	122	82956	37.171	ng	99
33) 4-Chloroaniline	10.457	127	154936	37.288	ng	98
34) Hexachlorobutadiene	10.604	225	79481	35.605	ng	96
35) Caprolactam	11.280	113	37278	39.762	ng	95
36) 4-Chloro-3-methylphenol	11.592	107	138514	37.969	ng	98
37) 2-Methylnaphthalene	11.945	142	282684	36.529	ng	99
38) 1-Methylnaphthalene	12.169	142	271208	37.051	ng	96
40) 1,2,4,5-Tetrachloroben...	12.321	216	137651	35.400	ng	99
41) Hexachlorocyclopentadiene	12.292	237	67019	35.348	ng	99
43) 2,4,6-Trichlorophenol	12.580	196	101564	36.376	ng	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
 Data File : BM031904.D
 Acq On : 26 Aug 2021 01:37
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 22 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040EC


Quant Time: Aug 26 03:24:28 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 14:16:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.651	196	111661	37.008	ng	98
46) 1,1'-Biphenyl	12.986	154	367009	35.728	ng	98
47) 2-Chloronaphthalene	13.021	162	290707	35.945	ng	99
48) 2-Nitroaniline	13.251	65	107225	38.100	ng	95
49) Acenaphthylene	13.880	152	474307	36.457	ng	99
50) Dimethylphthalate	13.639	163	379655	36.083	ng	100
51) 2,6-Dinitrotoluene	13.763	165	86809	37.464	ng	88
52) Acenaphthene	14.227	154	287596	36.309	ng	98
53) 3-Nitroaniline	14.092	138	88211	39.137	ng	97
54) 2,4-Dinitrophenol	14.304	184	54516	41.543	ng	89
55) Dibenzofuran	14.562	168	487339	36.536	ng	99
56) 4-Nitrophenol	14.415	139	75753	42.539	ng	93
57) 2,4-Dinitrotoluene	14.557	165	129697	39.299	ng	95
58) Fluorene	15.215	166	409975	37.484	ng	98
59) 2,3,4,6-Tetrachlorophenol	14.798	232	100476	38.144	ng	97
60) Diethylphthalate	15.009	149	430003	37.446	ng	99
61) 4-Chlorophenyl-phenyle...	15.221	204	200358	36.904	ng	94
62) 4-Nitroaniline	15.268	138	91054	41.834	ng	93
63) Azobenzene	15.515	77	474506	37.476	ng	98
65) 4,6-Dinitro-2-methylph...	15.321	198	78834	39.057	ng	93
66) n-Nitrosodiphenylamine	15.439	169	343913	35.212	ng	96
67) 4-Bromophenyl-phenylether	16.115	248	113236	34.596	ng	97
68) Hexachlorobenzene	16.215	284	122182	35.318	ng	98
69) Atrazine	16.409	200	123979	37.656	ng	97
70) Pentachlorophenol	16.568	266	85305	39.261	ng	98
71) Phenanthrene	16.956	178	632668	36.669	ng	99
72) Anthracene	17.051	178	626228	37.107	ng	100
73) Carbazole	17.333	167	620380	38.926	ng	99
74) Di-n-butylphthalate	17.898	149	766670	38.321	ng	100
75) Fluoranthene	18.980	202	772354	40.990	ng	98
77) Benzidine	19.186	184	302564	40.002	ng	98
78) Pyrene	19.345	202	783984	32.420	ng	100
80) Butylbenzylphthalate	20.268	149	367160	35.304	ng	97
81) Benzo(a)anthracene	21.103	228	772391	37.137	ng	99
82) 3,3'-Dichlorobenzidine	21.044	252	244419	40.870	ng	98
83) Chrysene	21.156	228	747475	37.277	ng	99
84) Bis(2-ethylhexyl)phtha...	21.050	149	572526	38.263	ng	# 98
85) Di-n-octyl phthalate	21.903	149	914860	40.574	ng	97
87) Indeno(1,2,3-cd)pyrene	25.379	276	666551	36.943	ng	97
88) Benzo(b)fluoranthene	22.627	252	724417	36.695	ng	98
89) Benzo(k)fluoranthene	22.668	252	705648	37.707	ng	99
90) Benzo(a)pyrene	23.168	252	642503	37.575	ng	99
91) Dibenzo(a,h)anthracene	25.385	278	587071	37.568	ng	99
92) Benzo(g,h,i)perylene	26.015	276	528989	35.800	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 26 03:24:28 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 25 14:16:52 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.522 min Scan# 754
 Delta R.T. -0.000 min
 Lab File: BM031904.D
 Acq: 26 Aug 2021 01:37




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:152 Resp: 41385
 Ion Ratio Lower Upper
 152 100
 150 153.6 122.4 183.6
 115 62.4 53.8 80.8



#2
 1,4-Dioxane
 Concen: 40.363 ng
 RT: 3.169 min Scan# 14
 Delta R.T. -0.006 min
 Lab File: BM031904.D
 Acq: 26 Aug 2021 01:37

Tgt Ion: 88 Resp: 36203
 Ion Ratio Lower Upper
 88 100
 58 72.2 50.4 75.6
 43 40.9 22.2 33.2#









#3
Pyridine
Concen: 40.380 ng
RT: 3.551 min Scan# 79
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 79 Resp: 102884
Ion Ratio Lower Upper
79 100
52 72.7 48.6 73.0
51 47.7 27.4 41.2#



#4
n-Nitrosodimethylamine
Concen: 38.740 ng
RT: 3.469 min Scan# 65
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion: 42 Resp: 67049
Ion Ratio Lower Upper
42 100
74 85.2 99.5 149.3#
44 4.6 57.3 85.9#









#5
2-Fluorophenol
Concen: 76.489 ng
RT: 5.157 min Scan# 352
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	112	Resp:	191474
Ion Ratio	Lower	Upper	
112	100		
64	62.6	53.7	80.5
63	41.1	33.9	50.9



#6
Aniline
Concen: 37.567 ng
RT: 6.875 min Scan# 644
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	93	Resp:	141346
Ion Ratio	Lower	Upper	
93	100		
66	43.1	35.3	52.9
65	23.3	18.7	28.1









#7
Phenol-d6
Concen: 75.844 ng
RT: 6.722 min Scan# 618
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	99	Resp:	268633
Ion	Ratio	Lower	Upper
99	100		
42	28.4	23.0	34.6
71	48.5	37.8	56.8



#8
2-Chlorophenol
Concen: 37.466 ng
RT: 7.092 min Scan# 681
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	128	Resp:	109194
Ion	Ratio	Lower	Upper
128	100		
130	33.7	13.0	53.0
64	50.7	33.4	73.4









#9
Benzaldehyde
Concen: 41.655 ng
RT: 6.686 min Scan# 612
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 77 Resp: 82260
Ion Ratio Lower Upper
77 100
105 90.0 68.9 108.9
106 82.5 61.2 101.2



#10
Phenol
Concen: 37.794 ng
RT: 6.745 min Scan# 622
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion: 94 Resp: 132382
Ion Ratio Lower Upper
94 100
65 36.9 16.9 56.9
66 53.6 30.0 70.0









#11
bis(2-Chloroethyl)ether
Concen: 37.712 ng
RT: 6.969 min Scan# 660
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 93 Resp: 96358
Ion Ratio Lower Upper
93 100
63 80.2 63.7 103.7
95 32.5 12.4 52.4



#12
1,3-Dichlorobenzene
Concen: 36.829 ng
RT: 7.404 min Scan# 734
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion: 146 Resp: 118548
Ion Ratio Lower Upper
146 100
148 66.5 52.3 78.5
75 43.2 34.2 51.2









#13
1,4-Dichlorobenzene
Concen: 37.445 ng
RT: 7.557 min Scan# 760
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:146 Resp: 123499
Ion Ratio Lower Upper
146 100
148 64.0 51.1 76.7
111 43.2 35.3 52.9



#14
1,2-Dichlorobenzene
Concen: 36.868 ng
RT: 7.863 min Scan# 812
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:146 Resp: 118399
Ion Ratio Lower Upper
146 100
148 64.0 51.7 77.5
111 46.0 38.2 57.2









#15
Benzyl Alcohol
Concen: 39.003 ng
RT: 7.775 min Scan# 797
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 79 Resp: 117846
Ion Ratio Lower Upper
79 100
108 65.0 53.6 80.4
77 70.2 57.7 86.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 36.953 ng
RT: 8.051 min Scan# 844
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion: 45 Resp: 126672
Ion Ratio Lower Upper
45 100
77 17.4 0.0 36.1
79 12.5 0.0 30.5









#17
2-Methylphenol
Concen: 37.156 ng
RT: 7.969 min Scan# 830
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	107	Resp:	92852
Ion	Ratio	Lower	Upper
107	100		
108	111.2	87.9	131.9
77	63.7	48.2	72.2
79	61.3	44.7	67.1



#18
Hexachloroethane
Concen: 38.095 ng
RT: 8.569 min Scan# 932
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	117	Resp:	52048
Ion	Ratio	Lower	Upper
117	100		
119	94.3	77.7	116.5
201	88.9	70.2	105.2









#19

n-Nitroso-di-n-propylamine
Concen: 36.902 ng
RT: 8.333 min Scan# 892
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC




Tgt Ion:	70	Resp:	100700
Ion	Ratio	Lower	Upper
70	100		
42	68.9	56.3	84.5
101	9.0	7.8	11.6
130	19.7	14.2	21.4




#20

3+4-Methylphenols
Concen: 38.019 ng
RT: 8.298 min Scan# 886
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	107	Resp:	130708
Ion	Ratio	Lower	Upper
107	100		
108	82.6	67.6	107.6
77	42.2	21.9	61.9
79	35.1	15.3	55.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.280 min Scan# 1223
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	136	Resp:	185992
Ion Ratio	Lower	Upper	
136	100		
137	11.0	8.7	13.1
54	8.1	7.4	11.0
68	4.7	3.8	5.8



#22
Acetophenone
Concen: 36.609 ng
RT: 8.345 min Scan# 894
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	105	Resp:	185089
Ion Ratio	Lower	Upper	
105	100		
71	4.7	3.8	5.6
51	34.9	29.5	44.3
120	20.8	16.3	24.5





#23

Nitrobenzene-d5

Concen: 74.122 ng

RT: 8.675 min Scan# 950

Delta R.T. -0.006 min

Lab File: BM031904.D

Acq: 26 Aug 2021 01:37





Instrument :

BNA_M

ClientSampleId :

SSTDCCC040EC

Tgt Ion:	82	Resp:	318272
Ion	Ratio	Lower	Upper
82	100		
128	36.1	29.0	43.6
54	56.3	45.8	68.8



#24

Nitrobenzene

Concen: 36.690 ng




RT: 8.716 min Scan# 957


Delta R.T. -0.006 min

Lab File: BM031904.D

Acq: 26 Aug 2021 01:37

Tgt Ion:	77	Resp:	159514
Ion	Ratio	Lower	Upper
77	100		
123	39.0	30.1	45.1
65	14.5	11.7	17.5









#25
Isophorone
Concen: 36.482 ng
RT: 9.239 min Scan# 1046
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 82 Resp: 270536
Ion Ratio Lower Upper
82 100
95 8.8 6.4 9.6
138 17.0 12.4 18.6



#26
2-Nitrophenol
Concen: 37.720 ng
RT: 9.416 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:139 Resp: 65651
Ion Ratio Lower Upper
139 100
109 46.0 34.8 52.2
65 68.0 59.9 89.9









#27
2,4-Dimethylphenol
Concen: 36.399 ng
RT: 9.480 min Scan# 1087
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:122 Resp: 95344
Ion Ratio Lower Upper
122 100
107 132.6 100.6 150.8
121 63.0 47.3 70.9



#28
bis(2-Chloroethoxy)methane
Concen: 36.419 ng
RT: 9.716 min Scan# 1127
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion: 93 Resp: 131916
Ion Ratio Lower Upper
93 100
95 32.6 25.8 38.8
123 10.7 8.8 13.2









#29
2,4-Dichlorophenol
Concen: 37.416 ng
RT: 9.939 min Scan# 1165
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:162 Resp: 115105
Ion Ratio Lower Upper
162 100
164 65.9 47.4 87.4
98 38.1 18.9 58.9



#30
1,2,4-Trichlorobenzene
Concen: 36.420 ng
RT: 10.145 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37


Tgt Ion:180 Resp: 122920
Ion Ratio Lower Upper
180 100
182 95.8 75.5 113.3
145 32.9 26.2 39.2








#31
Naphthalene
Concen: 36.641 ng
RT: 10.327 min Scan# 1231
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC







Tgt Ion:128 Resp: 380831
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 12.6 10.4 15.6



#32
Benzoic acid
Concen: 37.171 ng
RT: 9.675 min Scan# 1120
Delta R.T. 0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:122 Resp: 82956
Ion Ratio Lower Upper
122 100
105 140.2 120.3 160.3
77 118.3 100.0 140.0









#33
4-Chloroaniline
Concen: 37.288 ng
RT: 10.457 min Scan# 1253
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	127	Resp:	154936
Ion	Ratio	Lower	Upper
127	100		
129	33.5	25.6	38.4
65	37.0	30.4	45.6
92	20.6	16.2	24.4



#34
Hexachlorobutadiene
Concen: 35.605 ng
RT: 10.604 min Scan# 1278
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	225	Resp:	79481
Ion	Ratio	Lower	Upper
225	100		
223	60.4	52.6	78.8
227	62.6	51.2	76.8









#35
Caprolactam
Concen: 39.762 ng
RT: 11.280 min Scan# 1393
Delta R.T. 0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:113 Resp: 37278
Ion Ratio Lower Upper
113 100
55 179.2 170.5 210.5
56 144.7 125.2 165.2



#36
4-Chloro-3-methylphenol
Concen: 37.969 ng
RT: 11.592 min Scan# 1446
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:107 Resp: 138514
Ion Ratio Lower Upper
107 100
144 25.8 19.3 28.9
142 76.9 60.5 90.7









#37
2-Methylnaphthalene
Concen: 36.529 ng
RT: 11.945 min Scan# 1506
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:142 Resp: 282684
Ion Ratio Lower Upper
142 100
141 87.9 71.8 107.6
115 45.7 36.3 54.5



#38
1-Methylnaphthalene
Concen: 37.051 ng
RT: 12.169 min Scan# 1544
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:142 Resp: 271208
Ion Ratio Lower Upper
142 100
141 90.8 75.6 113.4
116 5.0 4.0 6.0









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.162 min Scan# 1883
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:164 Resp: 132728
Ion Ratio Lower Upper
164 100
162 98.1 80.6 121.0
160 43.0 35.1 52.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 35.400 ng
RT: 12.321 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:216 Resp: 137651
Ion Ratio Lower Upper
216 100
214 77.9 62.5 93.7
179 23.1 18.7 28.1
108 24.9 21.0 31.4









#41
Hexachlorocyclopentadiene
Concen: 35.348 ng
RT: 12.292 min Scan# 1565
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:237 Resp: 67019
Ion Ratio Lower Upper
237 100
235 63.3 44.4 84.4
272 11.7 0.0 31.4



#42
2,4,6-Tribromophenol
Concen: 77.027 ng
RT: 15.662 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:330 Resp: 120850
Ion Ratio Lower Upper
330 100
332 96.3 75.7 113.5
141 58.9 50.7 76.1









#43
2,4,6-Trichlorophenol
Concen: 36.376 ng
RT: 12.580 min Scan# 1614
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:196 Resp: 101564
Ion Ratio Lower Upper
196 100
198 96.2 73.6 110.4
200 32.0 22.1 33.1



#44
2,4,5-Trichlorophenol
Concen: 37.008 ng
RT: 12.651 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:196 Resp: 111661
Ion Ratio Lower Upper
196 100
198 94.9 77.0 115.4
97 64.3 53.9 80.9
132 28.8 23.8 35.8









#45
2-Fluorobiphenyl
Concen: 71.796 ng
RT: 12.774 min Scan# 1647
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:172 Resp: 722290
Ion Ratio Lower Upper
172 100
171 35.2 28.1 42.1
170 23.6 18.6 27.8



#46
1,1'-Biphenyl
Concen: 35.728 ng
RT: 12.986 min Scan# 1683
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:154 Resp: 367009
Ion Ratio Lower Upper
154 100
153 40.2 20.3 60.3
76 11.3 0.0 33.5









#47
2-Chloronaphthalene
Concen: 35.945 ng
RT: 13.021 min Scan# 1689
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	162	Resp:	290707
Ion Ratio	Lower	Upper	
162	100		
127	41.9	33.2	49.8
164	32.8	25.7	38.5



#48
2-Nitroaniline
Concen: 38.100 ng
RT: 13.251 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	65	Resp:	107225
Ion Ratio	Lower	Upper	
65	100		
92	63.0	51.2	76.8
138	91.3	66.8	100.2









#49
Acenaphthylene
Concen: 36.457 ng
RT: 13.880 min Scan# 1835
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:152 Resp: 474307
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.1 10.6 15.8



#50
Dimethylphthalate
Concen: 36.083 ng
RT: 13.639 min Scan# 1794
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:163 Resp: 379655
Ion Ratio Lower Upper
163 100
194 3.9 2.9 4.3
164 9.9 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 37.464 ng
RT: 13.763 min Scan# 1815
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion	Ratio	Lower	Upper
165	100		
63	77.7	72.4	108.6
89	76.3	69.3	103.9



#52
Acenaphthene
Concen: 36.309 ng
RT: 14.227 min Scan# 1894
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion	Ratio	Lower	Upper
154	100		
153	114.0	92.4	138.6
152	52.7	43.7	65.5









#53
3-Nitroaniline
Concen: 39.137 ng
RT: 14.092 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:138 Resp: 88211
Ion Ratio Lower Upper
138 100
108 11.0 8.3 12.5
92 128.0 105.2 157.8



#54
2,4-Dinitrophenol
Concen: 41.543 ng
RT: 14.304 min Scan# 1907
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:184 Resp: 54516
Ion Ratio Lower Upper
184 100
63 98.4 93.8 140.6
154 81.1 64.0 96.0









#55
Dibenzenofuran
Concen: 36.536 ng
RT: 14.562 min Scan# 1951
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:168 Resp: 487339
Ion Ratio Lower Upper
168 100
139 44.3 35.0 52.4
169 13.4 10.8 16.2



#56
4-Nitrophenol
Concen: 42.539 ng
RT: 14.415 min Scan# 1926
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:139 Resp: 75753
Ion Ratio Lower Upper
139 100
109 120.7 92.3 132.3
65 123.5 110.5 150.5









#57
2,4-Dinitrotoluene
Concen: 39.299 ng
RT: 14.557 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:165 Resp: 129697
Ion Ratio Lower Upper
165 100
63 82.4 70.3 105.5
89 100.0 83.1 124.7
182 3.0 2.2 3.4



#58
Fluorene
Concen: 37.484 ng
RT: 15.215 min Scan# 2062
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:166 Resp: 409975
Ion Ratio Lower Upper
166 100
165 97.0 79.3 118.9
167 13.5 10.6 16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 38.144 ng
RT: 14.798 min Scan# 1991
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	232	Resp:	100476
Ion Ratio	Lower	Upper	
232	100		
131	59.6	50.6	75.8
130	3.8	3.0	4.4
166	35.6	28.8	43.2



#60
Diethylphthalate
Concen: 37.446 ng
RT: 15.009 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	149	Resp:	430003
Ion Ratio	Lower	Upper	
149	100		
177	19.8	15.7	23.5
150	12.6	9.6	14.4









#61
4-Chlorophenyl-phenylether
Concen: 36.904 ng
RT: 15.221 min Scan# 2063
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	204	Resp:	200358
Ion Ratio	Lower	Upper	
204	100		
206	33.0	26.7	40.1
141	63.3	55.8	83.8



#62
4-Nitroaniline
Concen: 41.834 ng
RT: 15.268 min Scan# 2071
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:	138	Resp:	91054
Ion Ratio	Lower	Upper	
138	100		
92	59.9	40.6	80.6
108	135.3	103.8	143.8









#63
Azobenzene
Concen: 37.476 ng
RT: 15.515 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 77	Resp: 474506
Ion Ratio	Lower Upper
77 100	
182 21.3	0.0 39.3
105 16.1	0.0 36.1
51 38.1	19.1 59.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.915 min Scan# 2351
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:188	Resp: 293044
Ion Ratio	Lower Upper
188 100	
94 7.5	6.3 9.5
80 8.9	7.3 10.9









#65
4,6-Dinitro-2-methylphenol
Concen: 39.057 ng
RT: 15.321 min Scan# 2080
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:198 Resp: 78834
Ion Ratio Lower Upper
198 100
51 64.2 49.7 89.7
105 50.2 34.9 74.9



#66
n-Nitrosodiphenylamine
Concen: 35.212 ng
RT: 15.439 min Scan# 2100
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:169 Resp: 343913
Ion Ratio Lower Upper
169 100
168 65.0 55.0 82.4
167 35.4 29.8 44.8









#67
4-Bromophenyl-phenylether
Concen: 34.596 ng
RT: 16.115 min Scan# 2215
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:248 Resp: 113236
Ion Ratio Lower Upper
248 100
250 101.2 77.4 116.2
141 93.2 74.1 111.1



#68
Hexachlorobenzene
Concen: 35.318 ng
RT: 16.215 min Scan# 2232
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:284 Resp: 122182
Ion Ratio Lower Upper
284 100
142 52.4 43.0 64.6
249 34.0 27.6 41.4









#69
Atrazine
Concen: 37.656 ng
RT: 16.409 min Scan# 2265
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:200 Resp: 123979
Ion Ratio Lower Upper
200 100
173 24.5 5.8 45.8
215 45.2 23.6 63.6



#70
Pentachlorophenol
Concen: 39.261 ng
RT: 16.568 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:266 Resp: 85305
Ion Ratio Lower Upper
266 100
268 64.7 50.3 75.5
264 65.8 52.1 78.1









#71
Phenanthrene
Concen: 36.669 ng
RT: 16.956 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:178	Resp: 632668
Ion Ratio	Lower Upper
178	100
176	19.1 15.6 23.4
179	15.3 12.4 18.6



#72
Anthracene
Concen: 37.107 ng
RT: 17.051 min Scan# 2374
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:178	Resp: 626228
Ion Ratio	Lower Upper
178	100
176	18.6 14.8 22.2
179	15.5 12.3 18.5









#73
 Carbazole
 Concen: 38.926 ng
 RT: 17.333 min Scan# 2422
 Delta R.T. -0.000 min
 Lab File: BM031904.D
 Acq: 26 Aug 2021 01:37




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:167	Resp: 620380		
Ion Ratio	Lower	Upper	
167	100		
166	21.0	17.4	26.0
139	14.3	11.9	17.9



#74
 Di-n-butylphthalate
 Concen: 38.321 ng
 RT: 17.898 min Scan# 2518
 Delta R.T. -0.006 min
 Lab File: BM031904.D
 Acq: 26 Aug 2021 01:37

Tgt Ion:149	Resp: 766670
Ion Ratio	Lower Upper
149 100	
150 9.1	7.3 10.9
104 7.2	5.6 8.4









#75
Fluoranthene
Concen: 40.990 ng
RT: 18.980 min Scan# 2702
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:202 Resp: 772354
Ion Ratio Lower Upper
202 100
101 9.3 0.0 29.9
203 17.6 0.0 36.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.121 min Scan# 3066
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:240 Resp: 289149
Ion Ratio Lower Upper
240 100
120 7.7 7.8 11.6#
236 26.0 20.9 31.3









#77
Benzidine
Concen: 40.002 ng
RT: 19.186 min Scan# 2737
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:184 Resp: 302564
Ion Ratio Lower Upper
184 100
185 14.2 10.6 16.0
183 12.1 9.0 13.4



#78
Pyrene
Concen: 32.420 ng
RT: 19.345 min Scan# 2764
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:202 Resp: 783984
Ion Ratio Lower Upper
202 100
200 21.1 16.6 25.0
203 17.2 13.7 20.5









#79
 Terphenyl-d14
 Concen: 66.273 ng
 RT: 19.562 min Scan# 2801
 Delta R.T. -0.006 min
 Lab File: BM031904.D
 Acq: 26 Aug 2021 01:37




Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:244 Resp: 1327524
 Ion Ratio Lower Upper
 244 100
 212 8.1 6.5 9.7
 122 10.9 8.7 13.1



#80
 Butylbenzylphthalate
 Concen: 35.304 ng
 RT: 20.268 min Scan# 2921
 Delta R.T. -0.006 min
 Lab File: BM031904.D
 Acq: 26 Aug 2021 01:37

Tgt Ion:149 Resp: 367160
 Ion Ratio Lower Upper
 149 100
 91 83.2 64.0 96.0
 206 18.5 14.4 21.6









#81
Benzo(a)anthracene
Concen: 37.137 ng
RT: 21.103 min Scan# 3063
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:228 Resp: 772391
Ion Ratio Lower Upper
228 100
226 26.7 21.4 32.2
229 19.9 15.3 22.9



#82
3,3'-Dichlorobenzidine
Concen: 40.870 ng
RT: 21.044 min Scan# 3053
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:252 Resp: 244419
Ion Ratio Lower Upper
252 100
254 62.7 51.6 77.4
126 11.6 9.7 14.5









#83
Chrysene
Concen: 37.277 ng
RT: 21.156 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:228 Resp: 747475
Ion Ratio Lower Upper
228 100
226 29.5 24.0 36.0
229 19.9 16.0 24.0



#84
Bis(2-ethylhexyl)phthalate
Concen: 38.263 ng
RT: 21.050 min Scan# 3054
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:149 Resp: 572526
Ion Ratio Lower Upper
149 100
167 25.7 19.7 29.5
279 4.0 2.6 4.0#









#85
Di-n-octyl phthalate
Concen: 40.574 ng
RT: 21.903 min Scan# 3199
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:149 Resp: 914860
Ion Ratio Lower Upper
149 100
167 1.5 1.2 1.8
43 11.5 10.4 15.6



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.256 min Scan# 3429
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:264 Resp: 260750
Ion Ratio Lower Upper
264 100
260 23.7 19.3 28.9
265 21.8 18.3 27.5









#87
Indeno(1,2,3-cd)pyrene
Concen: 36.943 ng
RT: 25.379 min Scan# 3790
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:276 Resp: 666551
Ion Ratio Lower Upper
276 100
138 20.9 18.4 27.6
277 25.6 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 36.695 ng
RT: 22.627 min Scan# 3322
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:252 Resp: 724417
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 8.4 7.5 11.3









#89
Benzo(k)fluoranthene
Concen: 37.707 ng
RT: 22.668 min Scan# 3329
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37




Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC


Tgt Ion:252 Resp: 705648
Ion Ratio Lower Upper
252 100
253 21.5 17.3 25.9
125 8.7 7.5 11.3



#90
Benzo(a)pyrene
Concen: 37.575 ng
RT: 23.168 min Scan# 3414
Delta R.T. -0.000 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:252 Resp: 642503
Ion Ratio Lower Upper
252 100
253 21.2 17.4 26.2
125 9.4 8.1 12.1









#91
Dibenzo(a,h)anthracene
Concen: 37.568 ng
RT: 25.385 min Scan# 3791
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Instrument :
BNA_M
ClientSampleId :
SSTDCCC040EC

Tgt Ion:278 Resp: 587071
Ion Ratio Lower Upper
278 100
139 12.8 10.9 16.3
279 23.3 18.2 27.2



#92
Benzo(g,h,i)perylene
Concen: 35.800 ng
RT: 26.015 min Scan# 3898
Delta R.T. -0.006 min
Lab File: BM031904.D
Acq: 26 Aug 2021 01:37

Tgt Ion:276 Resp: 528989
Ion Ratio Lower Upper
276 100
277 23.4 19.4 29.0
138 18.5 14.6 22.0

