

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
 Data File : BM031892.D
 Acq On : 25 Aug 2021 14:51
 Operator : CG/JU
 Sample : PB138680BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB138680BL

Quant Time: Aug 25 15:29:45 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 14:16:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.522	152	32887	20.000	ng	0.00
21) Naphthalene-d8	10.280	136	131073	20.000	ng	0.00
39) Acenaphthene-d10	14.163	164	80340	20.000	ng	0.00
64) Phenanthrene-d10	16.915	188	154217	20.000	ng	0.00
76) Chrysene-d12	21.121	240	97039	20.000	ng	0.00
86) Perylene-d12	23.268	264	84881	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.163	112	301121	151.372	ng	0.00
7) Phenol-d6	6.722	99	373269	132.618	ng	0.00
23) Nitrobenzene-d5	8.675	82	266904	88.203	ng	0.00
42) 2,4,6-Tribromophenol	15.662	330	117597	123.830	ng	0.00
45) 2-Fluorobiphenyl	12.774	172	541251	88.883	ng	0.00
79) Terphenyl-d14	19.568	244	640264	95.242	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
Data File : BM031892.D
Acq On : 25 Aug 2021 14:51
Operator : CG/JU
Sample : PB138680BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB138680BL

Quant Time: Aug 25 15:29:45 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 25 14:16:52 2021
Response via : Initial Calibration

