

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
 Data File : BM031886.D
 Acq On : 25 Aug 2021 10:42
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020

Quant Time: Aug 25 12:06:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 12:03:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.522	152	32133	20.000	ng	0.00
21) Naphthalene-d8	10.280	136	146415	20.000	ng	0.00
39) Acenaphthene-d10	14.163	164	103647	20.000	ng	0.00
64) Phenanthrene-d10	16.915	188	220666	20.000	ng	0.00
76) Chrysene-d12	21.121	240	175105	20.000	ng	0.00
86) Perylene-d12	23.262	264	138143	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.157	112	80011	36.630	ng	0.00
7) Phenol-d6	6.716	99	111069	37.858	ng	0.00
23) Nitrobenzene-d5	8.675	82	134569	37.426	ng	0.00
42) 2,4,6-Tribromophenol	15.662	330	48010	42.037	ng	0.00
45) 2-Fluorobiphenyl	12.774	172	306797	38.085	ng	0.00
79) Terphenyl-d14	19.562	244	480523	41.285	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.175	88	15505	21.857	ng	# 86
3) Pyridine	3.557	79	42154	19.716	ng	# 85
4) n-Nitrosodimethylamine	3.475	42	27549	29.522	ng	# 48
6) Aniline	6.869	93	60478	19.369	ng	97
8) 2-Chlorophenol	7.092	128	46056	19.664	ng	95
9) Benzaldehyde	6.692	77	39181	20.204	ng	96
10) Phenol	6.745	94	54891	18.728	ng	95
11) bis(2-Chloroethyl)ether	6.969	93	41143	18.555	ng	98
12) 1,3-Dichlorobenzene	7.410	146	51124	20.172	ng	97
13) 1,4-Dichlorobenzene	7.557	146	52395	20.357	ng	99
14) 1,2-Dichlorobenzene	7.863	146	50552	20.167	ng	97
15) Benzyl Alcohol	7.775	79	47415	19.665	ng	97
16) 2,2'-oxybis(1-Chloropr...	8.045	45	56036	16.911	ng	95
17) 2-Methylphenol	7.969	107	40731	20.642	ng	95
18) Hexachloroethane	8.569	117	22196	20.253	ng	94
19) n-Nitroso-di-n-propyla...	8.328	70	43559	19.723	ng	99
20) 3+4-Methylphenols	8.298	107	55736	20.255	ng	97
22) Acetophenone	8.345	105	78520	18.391	ng	# 99
24) Nitrobenzene	8.716	77	68661	18.643	ng	97
25) Isophorone	9.239	82	117201	18.167	ng	98
26) 2-Nitrophenol	9.416	139	27095	19.493	ng	96
27) 2,4-Dimethylphenol	9.480	122	41057	19.189	ng	95
28) bis(2-Chloroethoxy)met...	9.722	93	57423	18.312	ng	99
29) 2,4-Dichlorophenol	9.939	162	48612	20.048	ng	97
30) 1,2,4-Trichlorobenzene	10.151	180	52879	20.539	ng	99
31) Naphthalene	10.333	128	162262	19.326	ng	99
32) Benzoic acid	9.633	122	36647	20.207	ng	85
33) 4-Chloroaniline	10.457	127	65662	19.697	ng	98
34) Hexachlorobutadiene	10.604	225	35168	21.856	ng	99
35) Caprolactam	11.257	113	15174	19.776	ng	# 81
36) 4-Chloro-3-methylphenol	11.586	107	58926	20.468	ng	99
37) 2-Methylnaphthalene	11.951	142	121390	20.056	ng	99
38) 1-Methylnaphthalene	12.174	142	114663	20.123	ng	99
40) 1,2,4,5-Tetrachloroben...	12.327	216	59429	19.951	ng	99
41) Hexachlorocyclopentadiene	12.298	237	26956	18.660	ng	98
43) 2,4,6-Trichlorophenol	12.580	196	43746	19.893	ng	93

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
 Data File : BM031886.D
 Acq On : 25 Aug 2021 10:42
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020


Quant Time: Aug 25 12:06:13 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 12:03:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.657	196	46082	19.645	ng	95
46) 1,1'-Biphenyl	12.986	154	156655	18.738	ng	99
47) 2-Chloronaphthalene	13.021	162	124087	19.113	ng	99
48) 2-Nitroaniline	13.251	65	43437	19.226	ng	95
49) Acenaphthylene	13.880	152	200469	19.034	ng	100
50) Dimethylphthalate	13.639	163	165283	19.631	ng	99
51) 2,6-Dinitrotoluene	13.757	165	36211	19.713	ng	97
52) Acenaphthene	14.227	154	121716	19.211	ng	98
53) 3-Nitroaniline	14.092	138	35638	19.819	ng	99
54) 2,4-Dinitrophenol	14.310	184	20075	19.571	ng	91
55) Dibenzofuran	14.563	168	203867	19.435	ng	99
56) 4-Nitrophenol	14.415	139	27997	18.906	ng	89
57) 2,4-Dinitrotoluene	14.557	165	50775	19.857	ng	98
58) Fluorene	15.215	166	166465	19.372	ng	98
59) 2,3,4,6-Tetrachlorophenol	14.798	232	41636	20.732	ng	97
60) Diethylphthalate	15.010	149	179487	19.578	ng	98
61) 4-Chlorophenyl-phenyle...	15.221	204	83182	20.199	ng	98
62) 4-Nitroaniline	15.262	138	34404	20.113	ng	96
63) Azobenzene	15.515	77	200166	18.938	ng	99
65) 4,6-Dinitro-2-methylph...	15.321	198	30835	20.062	ng	92
66) n-Nitrosodiphenylamine	15.439	169	142872	19.175	ng	97
67) 4-Bromophenyl-phenylether	16.115	248	47522	19.827	ng	98
68) Hexachlorobenzene	16.221	284	51414	20.271	ng	92
69) Atrazine	16.404	200	49811	19.998	ng	99
70) Pentachlorophenol	16.568	266	32191	19.404	ng	96
71) Phenanthrene	16.956	178	256567	19.305	ng	98
72) Anthracene	17.051	178	253317	19.375	ng	99
73) Carbazole	17.333	167	239094	19.012	ng	100
74) Di-n-butylphthalate	17.903	149	304093	18.151	ng	100
75) Fluoranthene	18.986	202	288236	19.511	ng	97
77) Benzidine	19.192	184	83921	17.877	ng	97
78) Pyrene	19.350	202	292207	20.826	ng	99
80) Butylbenzylphthalate	20.274	149	129788	18.764	ng	97
81) Benzo(a)anthracene	21.103	228	252867	19.697	ng	99
82) 3,3'-Dichlorobenzidine	21.050	252	73066	19.227	ng	98
83) Chrysene	21.156	228	241041	19.510	ng	100
84) Bis(2-ethylhexyl)phtha...	21.050	149	179375	16.665	ng	99
85) Di-n-octyl phthalate	21.909	149	262813	15.261	ng	98
87) Indeno(1,2,3-cd)pyrene	25.374	276	183363	17.098	ng	97
88) Benzo(b)fluoranthene	22.627	252	211381	20.775	ng	98
89) Benzo(k)fluoranthene	22.668	252	193369	19.748	ng	98
90) Benzo(a)pyrene	23.168	252	179816	19.684	ng	99
91) Dibenzo(a,h)anthracene	25.385	278	157422	16.814	ng	97
92) Benzo(g,h,i)perylene	26.015	276	148559	17.047	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 25 12:06:13 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 25 12:03:54 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.522 min Scan# 754
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion:152 Resp: 32133
Ion Ratio Lower Upper
152 100
150 154.5 122.4 183.6
115 66.3 53.8 80.8




#2
1,4-Dioxane
Concen: 21.857 ng
RT: 3.175 min Scan# 15
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 88 Resp: 15505
Ion Ratio Lower Upper
88 100
58 69.7 50.4 75.6
43 42.1 22.2 33.2#



Abundance Scan 79 (3.551 min): BM031887.D\data.ms (-74)



#3

Pyridine

Concen: 19.716 ng

RT: 3.557 min Scan# 80

Delta R.T. 0.006 min

Lab File: BM031886.D

Acq: 25 Aug 2021 10:42

Tgt Ion: 79 Resp: 42154

Ion Ratio Lower Upper

79 100

52 70.0 48.6 73.0

51 46.5 27.4 41.2#


Instrument :

BNA_M


ClientSampleId :

SSTDIC020


Abundance Scan 80 (3.557 min): BM031886.D\data.ms




Abundance



Abundance Scan 80 (3.557 min): BM031886.D\data.ms (-62)



Abundance Scan 66 (3.475 min): BM031887.D\data.ms (-61)



#4

n-Nitrosodimethylamine

Concen: 29.522 ng

RT: 3.475 min Scan# 66

Delta R.T. -0.000 min

Lab File: BM031886.D

Acq: 25 Aug 2021 10:42

Tgt Ion: 42 Resp: 27549


Ion Ratio Lower Upper

42 100


74 83.0 99.5 149.3#

44 5.0 57.3 85.9#


Abundance Scan 66 (3.475 min): BM031886.D\data.ms




Abundance



Abundance Scan 66 (3.475 min): BM031886.D\data.ms (-49)









#5
2-Fluorophenol
Concen: 36.630 ng
RT: 5.157 min Scan# 352
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:112 Resp: 80011
Ion Ratio Lower Upper
112 100
64 63.5 53.7 80.5
63 41.4 33.9 50.9



#6
Aniline
Concen: 19.369 ng
RT: 6.869 min Scan# 643
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 93 Resp: 60478
Ion Ratio Lower Upper
93 100
66 41.2 35.3 52.9
65 23.9 18.7 28.1









#7
Phenol-d6
Concen: 37.858 ng
RT: 6.716 min Scan# 617
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	99	Resp:	111069
Ion	Ratio	Lower	Upper
99	100		
42	27.9	23.0	34.6
71	50.5	37.8	56.8



#8
2-Chlorophenol
Concen: 19.664 ng
RT: 7.092 min Scan# 681
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	128	Resp:	46056
Ion	Ratio	Lower	Upper
128	100		
130	32.8	13.0	53.0
64	47.6	33.4	73.4









#9
Benzaldehyde
Concen: 20.204 ng
RT: 6.692 min Scan# 613
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 77 Resp: 39181
Ion Ratio Lower Upper
77 100
105 87.9 68.9 108.9
106 88.1 61.2 101.2



#10
Phenol
Concen: 18.728 ng
RT: 6.745 min Scan# 622
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 94 Resp: 54891
Ion Ratio Lower Upper
94 100
65 36.9 16.9 56.9
66 55.7 30.0 70.0









#11
bis(2-Chloroethyl)ether
Concen: 18.555 ng
RT: 6.969 min Scan# 660
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 93 Resp: 41143
Ion Ratio Lower Upper
93 100
63 81.5 63.7 103.7
95 32.3 12.4 52.4



#12
1,3-Dichlorobenzene
Concen: 20.172 ng
RT: 7.410 min Scan# 735
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:146 Resp: 51124
Ion Ratio Lower Upper
146 100
148 67.9 52.3 78.5
75 44.3 34.2 51.2









#13
1,4-Dichlorobenzene
Concen: 20.357 ng
RT: 7.557 min Scan# 760
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:146 Resp: 52395
Ion Ratio Lower Upper
146 100
148 65.3 51.1 76.7
111 43.6 35.3 52.9



#14
1,2-Dichlorobenzene
Concen: 20.167 ng
RT: 7.863 min Scan# 812
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:146 Resp: 50552
Ion Ratio Lower Upper
146 100
148 62.4 51.7 77.5
111 45.1 38.2 57.2









#15
Benzyl Alcohol
Concen: 19.665 ng
RT: 7.775 min Scan# 797
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 79 Resp: 47415
Ion Ratio Lower Upper
79 100
108 64.9 53.6 80.4
77 68.8 57.7 86.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 16.911 ng
RT: 8.045 min Scan# 843
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 45 Resp: 56036
Ion Ratio Lower Upper
45 100
77 18.6 0.0 36.1
79 11.9 0.0 30.5









#17
2-Methylphenol
Concen: 20.642 ng
RT: 7.969 min Scan# 830
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	107	Resp:	40731
Ion	Ratio	Lower	Upper
107	100		
108	102.3	87.9	131.9
77	62.5	48.2	72.2
79	58.9	44.7	67.1



#18
Hexachloroethane
Concen: 20.253 ng
RT: 8.569 min Scan# 932
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	117	Resp:	22196
Ion	Ratio	Lower	Upper
117	100		
119	92.1	77.7	116.5
201	81.7	70.2	105.2









#19
n-Nitroso-di-n-propylamine
Concen: 19.723 ng
RT: 8.328 min Scan# 891
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion: 70 Resp: 43559




Ion	Ratio	Lower	Upper
70	100		
42	70.3	56.3	84.5
101	9.1	7.8	11.6
130	20.6	14.2	21.4




#20
3+4-Methylphenols
Concen: 20.255 ng
RT: 8.298 min Scan# 886
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 107 Resp: 55736

Ion	Ratio	Lower	Upper
107	100		
108	83.5	67.6	107.6
77	43.0	21.9	61.9
79	36.4	15.3	55.3









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.280 min Scan# 1223
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	136	Resp:	146415
Ion Ratio	Lower	Upper	
136	100		
137	11.2	8.7	13.1
54	7.7	7.4	11.0
68	4.4	3.8	5.8



#22
Acetophenone
Concen: 18.391 ng
RT: 8.345 min Scan# 894
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	105	Resp:	78520
Ion Ratio	Lower	Upper	
105	100		
71	2.6	3.8	5.6#
51	36.9	29.5	44.3
120	20.2	16.3	24.5









#23
Nitrobenzene-d5
Concen: 37.426 ng
RT: 8.675 min Scan# 950
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 134569
Ion Ratio Lower Upper
82 100
128 38.7 29.0 43.6
54 58.7 45.8 68.8



#24
Nitrobenzene
Concen: 18.643 ng
RT: 8.716 min Scan# 957
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 77 Resp: 68661
Ion Ratio Lower Upper
77 100
123 39.8 30.1 45.1
65 14.9 11.7 17.5









#25
Isophorone
Concen: 18.167 ng
RT: 9.239 min Scan# 1046
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion: 82 Resp: 117201
Ion Ratio Lower Upper
82 100
95 8.7 6.4 9.6
138 16.4 12.4 18.6



#26
2-Nitrophenol
Concen: 19.493 ng
RT: 9.416 min Scan# 1076
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:139 Resp: 27095
Ion Ratio Lower Upper
139 100
109 48.8 34.8 52.2
65 73.4 59.9 89.9









#27
2,4-Dimethylphenol
Concen: 19.189 ng
RT: 9.480 min Scan# 1087
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:122 Resp: 41057
Ion Ratio Lower Upper
122 100
107 132.8 100.6 150.8
121 61.1 47.3 70.9



#28
bis(2-Chloroethoxy)methane
Concen: 18.312 ng
RT: 9.722 min Scan# 1128
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 93 Resp: 57423
Ion Ratio Lower Upper
93 100
95 31.5 25.8 38.8
123 11.4 8.8 13.2









#29
2,4-Dichlorophenol
Concen: 20.048 ng
RT: 9.939 min Scan# 1165
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	162	Resp:	48612
Ion Ratio	Lower	Upper	
162	100		
164	64.5	47.4	87.4
98	39.7	18.9	58.9



#30
1,2,4-Trichlorobenzene
Concen: 20.539 ng
RT: 10.151 min Scan# 1201
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	180	Resp:	52879
Ion Ratio	Lower	Upper	
180	100		
182	93.6	75.5	113.3
145	30.8	26.2	39.2









#31
Naphthalene
Concen: 19.326 ng
RT: 10.333 min Scan# 1232
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:128 Resp: 162262
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.5 10.4 15.6



#32
Benzoic acid
Concen: 20.207 ng
RT: 9.633 min Scan# 1113
Delta R.T. -0.035 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:122 Resp: 36647
Ion Ratio Lower Upper
122 100
105 123.2 120.3 160.3
77 101.7 100.0 140.0









#33
4-Chloroaniline
Concen: 19.697 ng
RT: 10.457 min Scan# 1253
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	127	Resp:	65662
Ion	Ratio	Lower	Upper
127	100		
129	33.3	25.6	38.4
65	37.4	30.4	45.6
92	22.0	16.2	24.4



#34
Hexachlorobutadiene
Concen: 21.856 ng
RT: 10.604 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	225	Resp:	35168
Ion	Ratio	Lower	Upper
225	100		
223	66.2	52.6	78.8
227	65.3	51.2	76.8









#35
Caprolactam
Concen: 19.776 ng
RT: 11.257 min Scan# 1389
Delta R.T. -0.018 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:113 Resp: 15174
Ion Ratio Lower Upper
113 100
55 157.0 170.5 210.5#
56 126.5 125.2 165.2



#36
4-Chloro-3-methylphenol
Concen: 20.468 ng
RT: 11.586 min Scan# 1445
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:107 Resp: 58926
Ion Ratio Lower Upper
107 100
144 23.9 19.3 28.9
142 76.6 60.5 90.7









#37
2-Methylnaphthalene
Concen: 20.056 ng
RT: 11.951 min Scan# 1507
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:142 Resp: 121390
Ion Ratio Lower Upper
142 100
141 88.3 71.8 107.6
115 45.0 36.3 54.5



#38
1-Methylnaphthalene
Concen: 20.123 ng
RT: 12.174 min Scan# 1545
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:142 Resp: 114663
Ion Ratio Lower Upper
142 100
141 93.2 75.6 113.4
116 5.1 4.0 6.0









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.163 min Scan# 1883
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:164 Resp: 103647
Ion Ratio Lower Upper
164 100
162 101.2 80.6 121.0
160 45.8 35.1 52.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 19.951 ng
RT: 12.327 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:216 Resp: 59429
Ion Ratio Lower Upper
216 100
214 77.3 62.5 93.7
179 23.4 18.7 28.1
108 25.1 21.0 31.4









#41
Hexachlorocyclopentadiene
Concen: 18.660 ng
RT: 12.298 min Scan# 1566
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	237	Resp:	26956
Ion Ratio	Lower	Upper	
237	100		
235	62.9	44.4	84.4
272	11.7	0.0	31.4



#42
2,4,6-Tribromophenol
Concen: 42.037 ng
RT: 15.662 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	330	Resp:	48010
Ion Ratio	Lower	Upper	
330	100		
332	97.8	75.7	113.5
141	61.4	50.7	76.1









#43
2,4,6-Trichlorophenol
Concen: 19.893 ng
RT: 12.580 min Scan# 1614
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:196 Resp: 43746
Ion Ratio Lower Upper
196 100
198 97.8 73.6 110.4
200 32.9 22.1 33.1



#44
2,4,5-Trichlorophenol
Concen: 19.645 ng
RT: 12.657 min Scan# 1627
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42


Tgt Ion:196 Resp: 46082
Ion Ratio Lower Upper
196 100
198 93.2 77.0 115.4
97 61.6 53.9 80.9
132 28.0 23.8 35.8








#45
2-Fluorobiphenyl
Concen: 38.085 ng
RT: 12.774 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42


Instrument :
BNA_M
ClientSampleId :
SSTDIC020





Tgt Ion:172	Resp:	306797	
Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.1	42.1
170	23.5	18.6	27.8




#46
1,1'-Biphenyl
Concen: 18.738 ng
RT: 12.986 min Scan# 1683
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42



Tgt Ion:154	Resp:	156655	
Ion	Ratio	Lower	Upper
154	100		
153	40.8	20.3	60.3
76	12.2	0.0	33.5









#47
2-Chloronaphthalene
Concen: 19.113 ng
RT: 13.021 min Scan# 1689
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampled :
SSTDIC020

Tgt Ion:162 Resp: 124087




Ion	Ratio	Lower	Upper
162	100		
127	42.3	33.2	49.8
164	32.8	25.7	38.5




#48
2-Nitroaniline
Concen: 19.226 ng
RT: 13.251 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 65 Resp: 43437


Ion	Ratio	Lower	Upper
65	100		
92	64.6	51.2	76.8
138	90.4	66.8	100.2








#49
Acenaphthylene
Concen: 19.034 ng
RT: 13.880 min Scan# 1835
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020




Tgt Ion:152 Resp: 200469

Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.7	23.5
153	13.4	10.6	15.8






#50
Dimethylphthalate
Concen: 19.631 ng
RT: 13.639 min Scan# 1794
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42



Tgt Ion:163 Resp: 165283

Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.9	4.3
164	9.6	8.0	12.0









#51
2,6-Dinitrotoluene
Concen: 19.713 ng
RT: 13.757 min Scan# 1814
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:165 Resp: 36211
Ion Ratio Lower Upper
165 100
63 89.3 72.4 108.6
89 81.5 69.3 103.9



#52
Acenaphthene
Concen: 19.211 ng
RT: 14.227 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:154 Resp: 121716
Ion Ratio Lower Upper
154 100
153 112.9 92.4 138.6
152 52.7 43.7 65.5









#53
3-Nitroaniline
Concen: 19.819 ng
RT: 14.092 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:138 Resp: 35638
Ion Ratio Lower Upper
138 100
108 10.3 8.3 12.5
92 130.5 105.2 157.8



#54
2,4-Dinitrophenol
Concen: 19.571 ng
RT: 14.310 min Scan# 1908
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:184 Resp: 20075
Ion Ratio Lower Upper
184 100
63 101.2 93.8 140.6
154 81.1 64.0 96.0









#55
Dibenzenofuran
Concen: 19.435 ng
RT: 14.563 min Scan# 1951
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:168 Resp: 203867
Ion Ratio Lower Upper
168 100
139 44.1 35.0 52.4
169 13.3 10.8 16.2



#56
4-Nitrophenol
Concen: 18.906 ng
RT: 14.415 min Scan# 1926
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:139 Resp: 27997
Ion Ratio Lower Upper
139 100
109 130.1 92.3 132.3
65 137.2 110.5 150.5









#57
2,4-Dinitrotoluene
Concen: 19.857 ng
RT: 14.557 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	165	Resp:	50775
Ion	Ratio	Lower	Upper
165	100		
63	85.0	70.3	105.5
89	104.4	83.1	124.7
182	2.9	2.2	3.4



#58
Fluorene
Concen: 19.372 ng
RT: 15.215 min Scan# 2062
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	166	Resp:	166465
Ion	Ratio	Lower	Upper
166	100		
165	100.6	79.3	118.9
167	13.8	10.6	16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 20.732 ng
RT: 14.798 min Scan# 1991
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion:232 Resp: 41636




Ion	Ratio	Lower	Upper
232	100		
131	60.5	50.6	75.8
130	3.7	3.0	4.4
166	34.4	28.8	43.2




#60
Diethylphthalate
Concen: 19.578 ng
RT: 15.010 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:149 Resp: 179487

Ion	Ratio	Lower	Upper
149	100		
177	20.8	15.7	23.5
150	12.7	9.6	14.4









#61
4-Chlorophenyl-phenylether
Concen: 20.199 ng
RT: 15.221 min Scan# 2063
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:204 Resp: 83182
Ion Ratio Lower Upper
204 100
206 33.1 26.7 40.1
141 67.5 55.8 83.8



#62
4-Nitroaniline
Concen: 20.113 ng
RT: 15.262 min Scan# 2070
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:138 Resp: 34404
Ion Ratio Lower Upper
138 100
92 60.1 40.6 80.6
108 130.9 103.8 143.8









#63
Azobenzene
Concen: 18.938 ng
RT: 15.515 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion: 77 Resp: 200166




Ion	Ratio	Lower	Upper
77	100		
182	21.1	0.0	39.3
105	15.5	0.0	36.1
51	39.3	19.1	59.1




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 16.915 min Scan# 2351
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion: 188 Resp: 220666

Ion	Ratio	Lower	Upper
188	100		
94	7.5	6.3	9.5
80	8.4	7.3	10.9









#65
4,6-Dinitro-2-methylphenol
Concen: 20.062 ng
RT: 15.321 min Scan# 2080
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:198	Resp:	30835
Ion Ratio	Lower	Upper
198	100	
51	63.4	49.7 89.7
105	48.9	34.9 74.9



#66
n-Nitrosodiphenylamine
Concen: 19.175 ng
RT: 15.439 min Scan# 2100
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:169	Resp:	142872
Ion Ratio	Lower	Upper
169	100	
168	65.6	55.0 82.4
167	36.3	29.8 44.8









#67
4-Bromophenyl-phenylether
Concen: 19.827 ng
RT: 16.115 min Scan# 2215
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:248 Resp: 47522
Ion Ratio Lower Upper
248 100
250 98.6 77.4 116.2
141 94.8 74.1 111.1



#68
Hexachlorobenzene
Concen: 20.271 ng
RT: 16.221 min Scan# 2233
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:284 Resp: 51414
Ion Ratio Lower Upper
284 100
142 45.6 43.0 64.6
249 32.8 27.6 41.4









#69
Atrazine
Concen: 19.998 ng
RT: 16.404 min Scan# 2264
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	200	Resp:	49811
Ion Ratio	Lower	Upper	
200	100		
173	26.8	5.8	45.8
215	44.0	23.6	63.6



#70
Pentachlorophenol
Concen: 19.404 ng
RT: 16.568 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	266	Resp:	32191
Ion Ratio	Lower	Upper	
266	100		
268	61.0	50.3	75.5
264	60.1	52.1	78.1









#71
Phenanthrene
Concen: 19.305 ng
RT: 16.956 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:178 Resp: 256567
Ion Ratio Lower Upper
178 100
176 18.5 15.6 23.4
179 15.0 12.4 18.6



#72
Anthracene
Concen: 19.375 ng
RT: 17.051 min Scan# 2374
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:178 Resp: 253317
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 14.4 12.3 18.5









#73
 Carbazole
 Concen: 19.012 ng
 RT: 17.333 min Scan# 2422
 Delta R.T. -0.000 min
 Lab File: BM031886.D
 Acq: 25 Aug 2021 10:42

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020

Tgt Ion:167 Resp: 239094




Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.4	26.0
139	14.5	11.9	17.9




#74
 Di-n-butylphthalate
 Concen: 18.151 ng
 RT: 17.903 min Scan# 2519
 Delta R.T. -0.000 min
 Lab File: BM031886.D
 Acq: 25 Aug 2021 10:42

Tgt Ion:149 Resp: 304093

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.3	10.9
104	6.8	5.6	8.4









#75
Fluoranthene
Concen: 19.511 ng
RT: 18.986 min Scan# 2703
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion:202 Resp: 288236




Ion	Ratio	Lower	Upper
202	100		
101	8.0	0.0	29.9
203	17.4	0.0	36.8




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.121 min Scan# 3066
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:240 Resp: 175105

Ion	Ratio	Lower	Upper
240	100		
120	8.4	7.8	11.6
236	25.6	20.9	31.3









#77
Benzidine
Concen: 17.877 ng
RT: 19.192 min Scan# 2738
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:184 Resp: 83921
Ion Ratio Lower Upper
184 100
185 14.2 10.6 16.0
183 12.5 9.0 13.4



#78
Pyrene
Concen: 20.826 ng
RT: 19.350 min Scan# 2765
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:202 Resp: 292207
Ion Ratio Lower Upper
202 100
200 21.0 16.6 25.0
203 16.8 13.7 20.5









#79
 Terphenyl-d14
 Concen: 41.285 ng
 RT: 19.562 min Scan# 2801
 Delta R.T. -0.006 min
 Lab File: BM031886.D
 Acq: 25 Aug 2021 10:42




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC020


Tgt Ion:244 Resp: 480523
 Ion Ratio Lower Upper
 244 100
 212 8.4 6.5 9.7
 122 10.2 8.7 13.1



#80
 Butylbenzylphthalate
 Concen: 18.764 ng
 RT: 20.274 min Scan# 2922
 Delta R.T. -0.000 min
 Lab File: BM031886.D
 Acq: 25 Aug 2021 10:42

Tgt Ion:149 Resp: 129788
 Ion Ratio Lower Upper
 149 100
 91 77.8 64.0 96.0
 206 19.5 14.4 21.6









#81
Benzo(a)anthracene
Concen: 19.697 ng
RT: 21.103 min Scan# 3063
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion:228 Resp: 252867




Ion	Ratio	Lower	Upper
228	100		
226	26.1	21.4	32.2
229	18.6	15.3	22.9




#82
3,3'-Dichlorobenzidine
Concen: 19.227 ng
RT: 21.050 min Scan# 3054
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:252 Resp: 73066

Ion	Ratio	Lower	Upper
252	100		
254	66.2	51.6	77.4
126	11.1	9.7	14.5









#83
Chrysene
Concen: 19.510 ng
RT: 21.156 min Scan# 3072
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion:228 Resp: 241041




Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.0	36.0
229	20.2	16.0	24.0




#84
Bis(2-ethylhexyl)phthalate
Concen: 16.665 ng
RT: 21.050 min Scan# 3054
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:149 Resp: 179375

Ion	Ratio	Lower	Upper
149	100		
167	23.9	19.7	29.5
279	3.5	2.6	4.0









#85
Di-n-octyl phthalate
Concen: 15.261 ng
RT: 21.909 min Scan# 3200
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:149 Resp: 262813
Ion Ratio Lower Upper
149 100
167 1.6 1.2 1.8
43 12.3 10.4 15.6



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.262 min Scan# 3430
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:264 Resp: 138143
Ion Ratio Lower Upper
264 100
260 25.0 19.3 28.9
265 21.9 18.3 27.5









#87
Indeno(1,2,3-cd)pyrene
Concen: 17.098 ng
RT: 25.374 min Scan# 3789
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:	276	Resp:	183363
Ion Ratio	Lower	Upper	
276	100		
138	20.1	18.4	27.6
277	25.1	20.1	30.1



#88
Benzo(b)fluoranthene
Concen: 20.775 ng
RT: 22.627 min Scan# 3322
Delta R.T. 0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:	252	Resp:	211381
Ion Ratio	Lower	Upper	
252	100		
253	22.7	18.0	27.0
125	7.8	7.5	11.3









#89
Benzo(k)fluoranthene
Concen: 19.748 ng
RT: 22.668 min Scan# 3329
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42




Instrument :
BNA_M
ClientSampleId :
SSTDIC020


Tgt Ion:252 Resp: 193369
Ion Ratio Lower Upper
252 100
253 21.0 17.3 25.9
125 8.3 7.5 11.3



#90
Benzo(a)pyrene
Concen: 19.684 ng
RT: 23.168 min Scan# 3414
Delta R.T. -0.000 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:252 Resp: 179816
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.2
125 9.0 8.1 12.1









#91
Dibenzo(a,h)anthracene
Concen: 16.814 ng
RT: 25.385 min Scan# 3791
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Instrument :
BNA_M
ClientSampleId :
SSTDIC020

Tgt Ion:278 Resp: 157422
Ion Ratio Lower Upper
278 100
139 12.9 10.9 16.3
279 24.2 18.2 27.2



#92
Benzo(g,h,i)perylene
Concen: 17.047 ng
RT: 26.015 min Scan# 3898
Delta R.T. -0.006 min
Lab File: BM031886.D
Acq: 25 Aug 2021 10:42

Tgt Ion:276 Resp: 148559
Ion Ratio Lower Upper
276 100
277 23.7 19.4 29.0
138 17.6 14.6 22.0

