



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
 Data File : BM031883.D
 Acq On : 25 Aug 2021 08:52
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Aug 26 12:15:55 2021






AutoFind: Scans 2427, 2428, 2429; Background Corrected with Scan 2421

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	59.4	30912	PASS
68	69	0.00	2	1.8	517	PASS
69	198	0.00	100	56.7	29477	PASS
70	69	0.00	2	0.4	113	PASS
127	198	10	80	57.3	29784	PASS
197	198	0.00	2	0.6	300	PASS
198	198	100	100	100.0	52000	PASS
199	198	5	9	7.2	3742	PASS
275	198	10	60	28.4	14781	PASS
365	198	1	100	5.3	2761	PASS
441	198	0.01	100	9.3	4826	PASS
442	442	50	100	100.0	29169	PASS
443	442	15	24	19.7	5733	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
Data File : BM031883.D
Acq On : 25 Aug 2021 08:52
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 25 10:49:22 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 25 10:44:01 2021
Response via : Initial Calibration



TIC: BM031883.D\data.ms

(70) Pentachlorophenol (C)

16.568min (-0.012) 0.00 ng




response 53141

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.90	65.80
264.00	65.10	60.74
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM082521\
Data File : BM031883.D
Acq On : 25 Aug 2021 08:52
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Aug 25 10:49:22 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM082521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 25 10:44:01 2021
Response via : Initial Calibration



TIC: BM031883.D\data.ms

(77) Benzidine

19.192min (-0.006) 0.00 ng

response 324687

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.30	13.96
183.00	11.20	11.81
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/25/2021	BNA_M	<u>BM031883.D</u>
Compound Name	Response	Retention Time
DDT	191150	20.42
DDD	1572	19.98
DDE	306	19.46
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
1878	193028	0.97

Instrument :
BNA_M
ClientSampleId :
DFTPP