

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM082616\
 Data File : BM007295.D
 Acq On : 26 Aug 2016 15:28
 Operator : UM/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.434

Quant Time: Aug 26 23:44:51 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM081516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 16 01:54:24 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	55	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	60	0.00
3	Naphthalene	1.004	1.033	-2.9	60	0.00
4 SURR	2-Methylnaphthalene-d10	0.500	0.539	-7.8	63	0.00
5	2-Methylnaphthalene	0.676	0.736	-8.9	64	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
7	Acenaphthylene	2.166	1.920	11.4	66	0.00
8 C	Acenaphthene	1.392	1.441	-3.5	77	0.00
9	Fluorene	1.578	1.793	-13.6	85	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
11	Pentachlorophenol	0.100	0.087	13.0	90	0.00
12	Phenanthrene	1.136	1.188	-4.6	97	0.00
13	Anthracene	1.085	1.085	0.0	94	0.00
14 SURR	Fluoranthene-d10	0.901	1.018	-13.0	105	0.00
15 C	Fluoranthene	1.320	1.512	-14.5	107	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
17	Pyrene	1.329	1.390	-4.6	110	0.00
18	Benzo(a)anthracene	1.264	1.263	0.1	107	0.00
19	Chrysene	1.215	1.367	-12.5	118	0.00
20 I	Perylene-d12	1.000	1.000	0.0	119	0.00
21	Benzo(b)fluoranthene	1.482	1.548	-4.5	116	0.00
22	Benzo(k)fluoranthene	1.221	1.410	-15.5	122	0.00
23 C	Benzo(a)pyrene	1.272	1.432	-12.6	121	0.00
24	Indeno(1,2,3-cd)pyrene	1.584	1.699	-7.3	117	0.00
25	Dibenzo(a,h)anthracene	1.246	1.337	-7.3	115	0.00
26	Benzo(g,h,i)perylene	1.345	1.457	-8.3	118	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0