

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM082616\
 Data File : BM007304.D
 Acq On : 26 Aug 2016 22:05
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.435

Quant Time: Aug 27 00:18:15 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM081516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 26 23:46:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	55	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	60	0.00
3	Naphthalene	1.004	1.025	-2.1	60	0.00
4 SURR	2-Methylnaphthalene-d10	0.500	0.536	-7.2	63	0.00
5	2-Methylnaphthalene	0.676	0.729	-7.8	64	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
7	Acenaphthylene	2.166	1.907	12.0	66	0.00
8 C	Acenaphthene	1.392	1.456	-4.6	77	0.00
9	Fluorene	1.578	1.773	-12.4	84	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
11	Pentachlorophenol	0.100	0.082	18.0	86	0.00
12	Phenanthrene	1.136	1.176	-3.5	97	0.00
13	Anthracene	1.085	1.051	3.1	92	0.00
14 SURR	Fluoranthene-d10	0.901	1.006	-11.7	105	0.00
15 C	Fluoranthene	1.320	1.503	-13.9	108	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
17	Pyrene	1.329	1.390	-4.6	110	0.00
18	Benzo(a)anthracene	1.264	1.202	4.9	101	0.00
19	Chrysene	1.215	1.415	-16.5	122	0.00
20 I	Perylene-d12	1.000	1.000	0.0	117	-0.01
21	Benzo(b)fluoranthene	1.482	1.674	-13.0	123	-0.01
22	Benzo(k)fluoranthene	1.221	1.370	-12.2	116	0.00
23 C	Benzo(a)pyrene	1.272	1.450	-14.0	120	-0.01
24	Indeno(1,2,3-cd)pyrene	1.584	1.692	-6.8	114	0.00
25	Dibenzo(a,h)anthracene	1.246	1.331	-6.8	113	-0.01
26	Benzo(g,h,i)perylene	1.345	1.461	-8.6	117	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0